



# NATIONAL PHYSICAL LABORATORY

Teddington Middlesex UK TW11 0LW Telephone +44 20 8977 3222



## Test Report



This test report is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This test report may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

### OXIDES OF NITROGEN (AS NO<sub>2</sub>) AND CARBON MONOXIDE COMPLIANCE TESTING AT SOUTH HOOK LNG TERMINAL ON BEHALF OF INTERTEK LIMITED

Permit Number: **XP3538LD**

Operator Name: **South Hook LNG (on behalf of Intertek Limited)**

Installation Name: **South Hook LNG Terminal**

Dates of Monitoring Visit: **9th - 11th March 2020**

Contract Reference: **E08040221**

Client Contact: **Adrian Walsh**

Client Organisation: **Intertek Limited**

Address: **Unit 14 - Waterston Trading Estate  
Main Road, Waterston  
Milford Haven  
SA73 3SL**

Monitoring Organisation: **National Physical Laboratory**

Address: **Hampton Road  
Teddington  
Middlesex  
TW11 0LW**

Date of Report: **9th April 2020**

Report Author: **Matthew Ellison**

---

Reference: XP3538LD/INTERTEK/SHLNG/MAR2020/SCV/PPC/Q1/V1

Report Approver: **Kevin Blakley**  
MCERTS Registration: **MM-03-317**  
Level & TEs Held: **Level 2, TE1, TE2, TE3 & TE4**  
Signature:

NPL Authorised Signatory  
Name: Mr R Robinson (for NPLML)  
Signature:

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**CONTENTS PAGE**

	<b>PAGE</b>
<b>Part One: Executive Summary</b>	
1.1 Monitoring Objectives	3
1.2.1 SCV 1H (Phase One) Monitoring Results	4
1.2.2 SCV 2G (Phase Two) Monitoring Results	5
1.3 Operating Information	6
1.4 Monitoring Deviations	7
1.5 Conclusions	7
1.6 References	7
<b>Part Two: Supporting Information</b>	
<b>Appendix One</b>	
2.1.1 Emissions Testing Personnel Details	9
2.1.2 Emissions Testing Procedures	9
2.1.3 Equipment Checklist Reference	10
2.1.4 Data Capture Location Reference	10
<b>Appendix Two</b>	
2.2.1 Stack Diagram	12
2.2.2 Flow Criteria Measurements	14
2.2.3 One Minute Averaged Gaseous Emissions Data	19
2.2.4 Gaseous Emissions Graphical Data	22
2.2.5 Gas Calibration Log	25
2.2.6 Uncertainty Calculations	28
2.2.7 Calculations Used in Reporting Results	39

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**1.1 Monitoring Objectives**

NPL were awarded a contract by Intertek Limited to carry out emissions compliance testing at South Hook LNG plant near Milford Haven, Pembrokeshire. The scope of work includes carrying out emissions monitoring on the SCV flue stacks phases one and two.

There are a total of fifteen SCV units on the South Hook site that require monitoring, eight on phase one and seven on phase two. Two units were measured for oxides of nitrogen and carbon monoxide.

In addition, oxygen measurements were taken to allow a correction to reference conditions. Water vapour and flow measurements were also taken to determine the moisture content and velocity of the flue gas.

NPL carried out the monitoring visit between the 9th and 11th March 2020. The report documents the results obtained.

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**1.2.1 SCV H (Phase One) Monitoring Results**

Client: Intertek  
Site: South Hook LNG  
Emission Point: SCV H (Phase One)

Field	Units	Oxides of Nitrogen (as NO <sub>2</sub> )	Carbon Monoxide	Oxygen	Moisture	Stack Flow
Emission Limit Value	mg/m <sup>3</sup> , Reference Conditions	107	N/A	N/A	N/A	N/A
Periodic Monitoring Result	Reference Conditions	52.3	432	12.2	1.5	5.1
Uncertainty (95% Confidence Level)	Reference Conditions	5.8	38.0	0.4	N/A	1.1
	Units	mg/m <sup>3</sup>	mg/m <sup>3</sup>	%Vol/Vol	%Vol/Vol	m <sup>3</sup> /s
Reference Conditions			273K, 101.3 kPa, 3% Oxygen on a dry gas basis			
Date	dd/mm/yyyy			11/03/2020		
Sample Period	From hh:mm	To hh:mm		12:00		
				13:00		
Monitoring Method	BS EN 14792:2017	BS EN 15058:2017	BS EN 14789:2017	BS EN 14790:2017	BS EN ISO 16911-1:2013	
Accreditation	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	
Process Status	Load (Tonnes/Hour)			115		
Process Status	Burner Demand (%)			31		

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**1.2.2 SCV G (Phase Two) Monitoring Results**

Client:  
Intertek  
Site:  
South Hook LNG  
Emission Point:  
SCV G (Phase Two)

Field	Units	Oxides of Nitrogen (as NO <sub>2</sub> )	Carbon Monoxide	Oxygen	Moisture	Stack Flow
Emission Limit Value	mg/m <sup>3</sup> , Reference Conditions	107	N/A	N/A	N/A	N/A
Periodic Monitoring Result	Reference Conditions	53.4	523	11.6	1.4	5.6
Uncertainty (95% Confidence Level)	Reference Conditions	5.9	45.5	0.4	N/A	1.1
	Units	mg/m <sup>3</sup>	mg/m <sup>3</sup>	%Vol/Vol	%Vol/Vol	m <sup>3</sup> /s
Reference Conditions			273K, 101.3 kPa, 3% Oxygen on a dry gas basis			
Date	dd/mm/yyyy			09/03/2020		
Sample Period	From hh:mm			13:00		
	To hh:mm			14:00		
Monitoring Method		BS EN 14792:2017	BS EN 15058:2017	BS EN 14789:2017	BS EN 14790:2017	BS EN ISO 16911-1:2013
Accreditation		UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS
Process Status	Load (Tonnes/Hour)			160		
Process Status	Burner Demand (%)			52.4		

**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

### 1.3 Operating Information

South Hook LNG Terminal, situated in Pembrokeshire South West Wales, is a regasification plant for natural gas. The LNG is transported in specially designed vessels to Milford Haven where it is then transferred into storage tanks, where it awaits reheating and distribution into the UK National Grid.

A total of fifteen SCVs have been built across two phases. Each one has at least two five inch ports installed, as set out in BS EN 15259. The ports are located approximately thirteen metres from the base of the stack and can be accessed by ladders that lead to a permanent platform. The sampling platform has lighting, toe boards and handrails. There is sufficient parking on the roadway at the base of the stacks for the mobile laboratory and provision of 240v 16A power supply outlets.

<b>Continuous or Batch Process?</b>	Each SCV operates on a batch process. The number of SCVs operating and the load at which they are set depend upon the required site output.													
<b>What part of the batch process was sampled? (If applicable)</b>	The periodic monitoring is carried out once an SCV has been brought online to the operators required load and has stabilised. This loading remains constant through the one hour test.													
<b>What fuel was used during monitoring? (If applicable)</b>	A small amount of LNG is used as fuel to heat a volume of water. This heat exchange warms up the LNG allowing it to be passed out into the National Grid system.													
<b>What feedstock was used during monitoring? (If applicable)</b>	N/A													
<b>What was the load during monitoring?</b>	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Emission Point</th><th>Load (Ton/hr)</th><th>Burner Demand (%)</th></tr> </thead> <tbody> <tr> <td>SCV 1H</td><td>115</td><td>31.0</td></tr> <tr> <td>SCV 2G</td><td>160</td><td>52.4</td></tr> </tbody> </table>					Emission Point	Load (Ton/hr)	Burner Demand (%)	SCV 1H	115	31.0	SCV 2G	160	52.4
Emission Point	Load (Ton/hr)	Burner Demand (%)												
SCV 1H	115	31.0												
SCV 2G	160	52.4												
<b>What abatement systems are present? Were they in operation?</b>	Each SCV uses water injection to abate NOx emissions. The system was in operation during the periodic monitoring of each SCV.													
<b>Periodic monitoring results and corresponding CEM values</b>	<b>Emission Point</b>	<b>Substance Monitored</b>	<b>CEM Result</b>	<b>Periodic Monitoring Result</b>	<b>Units</b>									
	SCV 1H	Oxides of Nitrogen	48.7	52.3	mg/Nm <sup>3</sup>									
	SCV 1H	Oxygen	11.5	12.2	% Vol									

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**1.4 Monitoring Deviations**

<b>Were all substances in the monitoring objectives monitored? If not why?</b>	All substances set out in the objective were monitored.
<b>Were all substances monitored in accordance to the relevant method? If not why?</b>	All substances set out in the monitoring objectives were measured in accordance to the relevant standards.
<b>Were there any other issues relevant to the monitoring results?</b>	No.

**1.5 Conclusions**

NPL carried out the emissions monitoring at South Hook LNG over a period of week. Two SCVs were monitored for the required determinands.

The client is aware of BS EN 15259 and the requirement to carry out homogeneity testing. These tests were carried out on SCVs 1H and 2A, both units passed.

Reference - XP3538LD/INTERTEK/SHLNG/JULY2011/SCV/HOMOGENEITY.

**1.6 References**

1. STA – Risk Assessment Guide: Industrial-emission monitoring.
2. Environment Agency - Manual Stack emission monitoring performance standard for Organisations.
3. Environment Agency – M1 Technical Guidance Note – Sampling requirements for stack emission monitoring.
4. Environment Agency - MID 15259 - Stationary source emissions - Requirements for the measurement sections and sites and for the measurement objective, plan and report.
5. Guidance on Assessing Measurement Uncertainty in Stack Emissions Monitoring, by Pullen J and Robinson R, Source Testing Association, Quality Guidance Note QGN1.

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## APPENDIX ONE

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**2.1.1 Emissions Testing Personnel Details**

Name	Role	MCERTS Number	Certification Level & Expiry Dates			
			Level 1	Level 2	TE1	TE2
Matthew Ellison	Team Leader	MM-05-682	N/A	Apr-2023	Sep-2023	Sep-2023
Jonny Guy	Technician	MM-16-1388	Apr-2022	N/A	N/A	Sep-2023

**2.1.2 Emissions Testing Procedures**

	Instrumental Methods				Manual Methods		
	Determinand	NOx	CO	O <sub>2</sub>	H <sub>2</sub> O	Stack Flow	Temperature
SRM Standard	BS EN 14792:2017	BS EN 15058:2017	BS EN 14789:2017	BS EN 14790:2017	BS EN ISO 16911-1:2013	BS EN ISO 16911-1:2013	BS EN ISO 16911-1:2013
Instrument	Horiba PG-250	Horiba PG-250	Horiba PG-250	N/A	S-Type Pilot Thermocouple	Type K Thermocouple	
Instrument Serial No.	AS0208	AS0208	AS0208	N/A	AS0640	AS0640	
Principle	Chemiluminescence	NDIR	Zirconia	Saturation chart	Flow	Temperature	
Operational Range	0 - 100 ppm	0 - 200 ppm	0 - 25%	N/A	0 - 240mm H <sub>2</sub> O	N/A	
Certified Range	0 - 125 mg/m <sup>3</sup>	0 - 95 mg/m <sup>3</sup>	0 - 25%	N/A	N/A	N/A	
Uncertainty	10%	6%	6%	20%	10%	1%	
NPL Procedure	QPAS B 538	QPAS B 538	QPAS B 538	QPAS B 540	QPAS B 567	QPAS B 567	
UKAS Accreditation	YES	YES	YES	YES	YES	YES	

The sample gas was extracted from the stack via a chemically inert heated line and drawn through a conditioning unit. All moisture in the gas sample was removed and cooled down to 4°C before it was pumped down another line to the NPL Mobile Source Emissions Laboratory and analysed using a Horiba PG-250. The entire sampling system had been leak tested before testing was carried out to ensure no dilution of the sample gas.

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

The conventional analyser zero and span settings were checked before and after each test run using zero grade nitrogen (ex BOC), a suitable gas mixture (BOC beta gas standard), traceable to national reference standards and a gas dilution system. The certified accuracies of the gas standards are listed below: -

Component	Sample Location	Cylinder ID	Certified Amount	Certified Uncertainty
Carbon Monoxide	Phase 1 & 2 SCVs	183642SG	154.9 ppm	1%
Nitrogen Oxide		183642SG	79.4 ppm	1%
Oxygen		158247SG	15.13%	1%

These measurement uncertainties are expressed at a 95% level of confidence.

The ranges of the Horiba PG-250 analyser used for the testing and the values used to calibrate the instrument before and after the testing are listed below: -

Gaseous Components	Horiba Ranges	Calibrated Values
Carbon Monoxide	0 - 200 ppm	154.9 ppm
Oxides of Nitrogen (as NO <sub>2</sub> )	0 - 100 ppm	79.4 ppm
Oxygen	0 - 25%	15.13%

A leak test was conducted before testing to confirm hydraulic integrity of the gaseous sampling system. This was conducted by sending nitrogen down the entire sample line and ensuring a zero reading was obtained.

The electrical volt/millivolt outputs from the PG-250 gas analyser was collected by data logger software on a PC and downloaded to digital media at the end of each day. Under the program used during the tests, the software records and stores individual readings either every 1 or 10 seconds. From this data, the logger can perform a series of calculations to output 1 minute averaged measurement on a volume/volume or mass/volume basis. After each 1 minute average has been established the data buffer is reset and the process repeats.

#### **2.1.3 Equipment Checklist Reference**

See workfile INTK48MAR20/Equipment Checklist.

#### **2.1.4 Data Capture Location Reference**

All data collected using the NPL computer system on site is backed up at the end of each day onto a memory stick. When the team returns to site this information is then uploaded onto the NPL servers and stored in the relevant location for that job. The link below is where the South Hook emissions data is stored:

P:\Stack Emissions Team\South Hook LNG - Intertek\INTK48MAR20\7. Monitoring Record Sheets



NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## APPENDIX TWO

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.1 - Stack Diagram

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Company	Installation	Stack	Substances Monitored	Permit
Intertek	South Hook LNG	SCVs Phases One & Two	CO, NOx, O <sub>2</sub> and H <sub>2</sub> O	XP3538LD
Stack Cross-Section				
				
Stack Diagram				
				

Position	1	2	3	4	5	6
% of Diameter	4.1	14.9	29.7	70.3	85.1	95.9
Insertion, m	0.06	0.22	0.44	1.04	1.26	1.42

Notes - The circular stack diameter was measured as 1.48 metres, whilst the port offset (distance between the edge of the stack to the end of the port) was measured as 10 centimetres. Access to the top of the stack was by a series of three ladders. Each one had a resting platform in-between with a self closing gate installed. The main platform itself is a permanent structure with toe boards, railings and self closing gate.

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.2 - Flow Criteria Measurements

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Monitoring Objective	Traverse	Site:	South Hook LNG	Stack ID:	SCV1H			
Date	11/03/2020	Site Team:	MRE/JG	Time of Survey:	10:35			
Tape Measure ID	AS0589			Diagram of Sample Location:				
Laser Measurement Device ID	N/A							
Traverse Pitot Type	S-Type							
Traverse Pitot Tube ID	AS0687							
PILOT ASSEMBLY visual inspection (D <sub>base</sub> )	Pass							
Pre test Leak Check >2.5 mm H <sub>2</sub> O	Pass							
Traverse Manometer Type	Digital							
Traverse Manometer ID	AS0640							
Traverse Manometer Range	240							
Traverse Temp. Readout ID	AS0786							
Traverse Thermocouple ID	AS0451a							
Static Pressure	ΔP (mmH2O)							
	12.20							
Swirl Test Conducted	Pass			Comments/Deviations:				
Protractor ID	Yes							
Post-Test Blockage Test (L-T type only)	AS0626							
PILOT ASSEMBLY visual inspection (D <sub>base</sub> )	Pass							
Post Test Leak Check >2.5 mm H <sub>2</sub> O	Pass							
Conditions	Value	Units	Port ID	Reading 1 (m)	Reading 2 (m)	Reading 3 (m)	Average	Port Depth
Stack pressure	752.19	mmHg	A					
Ref O <sub>2</sub> Value	3	%	B					
Moisture Content	1.54	%	C					
CO	165	ppm	D					
CO <sub>2</sub>	5.1	%						
N <sub>2</sub>	82.68	%						
O <sub>2</sub>	12.20	%						
Dry Molecular wt	29.30							
Stack Molecular wt	29.13							
Duct Diameter	1.48	m	Line ID	Reading 1 (m)	Reading 2 (m)	Reading 3 (m)	Average	Duct Diameter
Duct Depth	0.84	m	A					
Duct Width	0.84	m	B					
Area of stack	1.72	m <sup>2</sup>	C					
Pbar	1001.4	mbar	D					
Phar	751	mmHg						
Pilot tube coeff	0.84		Outside Side Division					
Reference Temp	273	K						
Reference Pressure	760	mmHg						
Ambient Temperature	9	°C	Enter manually from previous visit					
			Circular Duct	Rectangular Duct				
			Duct Diameter (m)	Duct Depth (m)				
			Duct Width (m)					
			POST TEST	Traverse Point	ΔP Reading (mm H <sub>2</sub> O)	ΔP Reading (mm H <sub>2</sub> O)	ΔP Average (mm H <sub>2</sub> O)	
			Reading 1					
			Reading 2					

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

SAMPLING LINE: North									
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp T <sub>s</sub> °C	Velocity @ stack gas basis T&P on wet gas basis m/s	Angle of swirl °
1	1.42	3.18	3.18	3.18	3.18	31.19	13.5	5.98	6
2	1.26	3.27	3.27	3.27	3.27	32.09	13.5	6.06	6
3	1.04	3.50	3.50	3.50	3.50	34.29	13.5	6.27	6
4	0.44	4.24	4.24	4.24	4.24	41.58	13.6	6.90	6
5	0.22	4.57	4.57	4.57	4.57	44.78	13.6	7.16	7
6	0.06	4.50	4.50	4.50	4.50	44.08	13.7	7.11	7
SAMPLING LINE: West									
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp T <sub>s</sub> °C	Velocity @ stack gas basis T&P on wet gas basis m/s	Angle of swirl °
1	1.42	2.89	2.89	2.89	2.89	28.29	13.5	5.69	8
2	1.26	3.40	3.40	3.40	3.40	33.29	13.6	6.18	6
3	1.04	3.89	3.89	3.89	3.89	38.09	13.6	6.61	5
4	0.44	4.12	4.12	4.12	4.12	40.38	13.6	6.80	6
5	0.22	4.57	4.57	4.57	4.57	44.78	13.7	7.16	6
6	0.06	4.72	4.72	4.72	4.72	46.28	13.8	7.28	7
Average values		3.9	3.9	3.9	3.9	38.3	13.6	6.6	6.3
Duct / Stack Flow Characteristics:						Average	Units	Flow Criteria Measurements	
Stack Velocity at stack gas T & P and a wet gas basis						6.60	m/s	Is the Flow Ratio 3:1 or less?	1.3
Stack flow @ STP, O <sub>2</sub> (ref) and on a dry gas basis						5.12	m <sup>3</sup> /s		:1
Stack flow @ stack gas T & P and on a wet gas basis						11.35	m <sup>3</sup> /s	Any local negative flow?	NO
Stack flow @ stack gas T & P and on a dry gas basis						11.17	m <sup>3</sup> /s	Flow <15° of duct axis?	YES
Stack flow @ STP and on a wet gas basis						10.70	m <sup>3</sup> /s	Minimum Δp detected > 0.5 mmH <sub>2</sub> O	YES
						5.20	m <sup>3</sup> /s		

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Monitoring Objective	Traverse	Site:	South Hook LNG	Stack ID:	SCV2G			
Date	09/03/2020	Site Team:	MRE/JG	Time of Survey:	12:00			
Tape Measure ID	AS0589			Diagram of Sample Location:				
Laser Measurement Device ID	N/A							
Traverse Pilot Type	S-Type							
Traverse Pilot Tube ID	AS0687							
From Assembly vs Tran Inspection (DPA)	Pass							
Pre test Leak Check <2.5 mm H <sub>2</sub> O	Pass							
Traverse Manometer Type	Digital							
Traverse Manometer ID	AS0640							
Traverse Manometer Range	240							
Traverse Temp. Readout ID	AS0786							
Traverse Thermocouple ID	AS0451a							
Static Pressure	$\Delta p$ (mmH2O)							
	18.15							
Swirl Test Conducted	Pass			Comments/Deviations:				
Protractor ID	Yes							
Post-Test Blockage Test (L-Type only)	AS0626							
Post Test Assembly visual inspection (DPA)	Pass							
Post Test Leak Check <2.5 mm H <sub>2</sub> O	Pass							
Conditions	Value	Units	Port ID	Reading 1 (m)	Reading 2 (m)	Reading 3 (m)	Average	Port Depth
Stack pressure	753.01	mmHg	A					
Ref O <sub>2</sub> Value	3	%	B					
Moisture Content	1.42	%	C					
CO	200	ppm	D					
CO <sub>2</sub>	5.4	%						
N <sub>2</sub>	82.98	%						
O <sub>2</sub>	11.60	%						
Dry Molecular wt	29.33							
Stack Molecular wt	29.17							
Duct Diameter	1.48	m						
Duct Depth	0.84	m						
Duct Width	0.84	m						
Area of stack	1.72	m <sup>2</sup>						
Pbar	1001.9	mbar						
Phar	752	mmHg						
Pilot tube coeff	0.84		Outside Side Division					
Reference Temp	273	K						
Reference Pressure	760	mmHg						
Ambient Temperature	11	°C	Enter manually from previous visit					
			Circular Duct					
			Rectangular Duct					
			Duct Diameter (m)					
			Duct Depth (m)					
			Duct Width (m)					
POST TEST	Traverse Point		ΔP Reading (mm H <sub>2</sub> O)	ΔP Reading (mm H <sub>2</sub> O)	ΔP Average (mm H <sub>2</sub> O)	ΔP Average (mm H <sub>2</sub> O)		
	Reading 1							
	Reading 2							

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

SAMPLING LINE: North											
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp Ts °C	Velocity @ stack gas basis T&P on wet gas basis m/s	Angle of swirl °	√Δp	
1	1.42	3.23	3.23	3.23	3.23	31.7	12.4	6.01	7	1.80	
2	1.26	3.84	3.84	3.84	3.84	37.7	12.3	6.55	7	1.96	
3	1.04	4.00	4.00	4.00	4.00	39.2	12.3	6.68	7	2.00	
4	0.44	4.48	4.48	4.48	4.48	43.9	12.2	7.07	6	2.12	
5	0.22	4.57	4.57	4.57	4.57	44.8	12.4	7.14	6	2.14	
6	0.06	4.86	4.86	4.86	4.86	47.7	12.5	7.37	6	2.21	

SAMPLING LINE: West											
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp Ts °C	Velocity @ stack gas basis T&P on wet gas basis m/s	Angle of swirl °	√Δp	
1	1.42	3.23	3.23	3.23	3.23	31.7	11.9	6.00	5	1.80	
2	1.26	3.32	3.32	3.32	3.32	32.6	12.1	6.09	6	1.82	
3	1.04	3.60	3.60	3.60	3.60	35.3	12.1	6.33	6	1.90	
4	0.44	4.56	4.56	4.56	4.56	44.7	12.2	7.13	5	2.13	
5	0.22	4.67	4.67	4.67	4.67	45.8	11.9	7.21	6	2.16	
6	0.06	4.77	4.77	4.77	4.77	46.8	12.0	7.29	7	2.18	
Average values		4.1	4.1	4.1	4.1	40.1	12.2	6.7	6.2	2.0	

Duct / Stack Flow Characteristics:

Stack Velocity at stack gas T & P and a wet gas basis	6.74 ms <sup>-1</sup>	Is the Flow Ratio 3:1 or less?	1.2
Stack flow @ STP, O <sub>2</sub> (ref) and on a dry gas basis	5.63 m <sup>3</sup> s <sup>-1</sup>		:1
Stack flow @ stack gas T & P and on a wet gas basis	11.59 m <sup>3</sup> s <sup>-1</sup>	Any local negative flow?	NO
Stack flow @ stack gas T & P and on a dry gas basis	11.42 m <sup>3</sup> s <sup>-1</sup>	Flow <15° of duct axis?	YES
Stack flow @ STP and on a wet gas basis	10.99 m <sup>3</sup> s <sup>-1</sup>	Minimum Δp detected > 0.5 mmH <sub>2</sub> O	YES
Stack flow @ STP, O <sub>2</sub> (ref) and on a wet gas basis	5.71 m <sup>3</sup> s <sup>-1</sup>		

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.3 - One Minute Averaged Gaseous Emissions Data

**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

Minute Averaged Gaseous Data from South Hook LNG 11th March 2020			
SCV H - Phase One			
Referenced to 273K, 101.3kPa and 3% Oxygen on a Dry Basis			
Time (hh:mm)	Carbon Monoxide (mg/Nm <sup>3</sup> )	Oxides of Nitrogen (mg/Nm <sup>3</sup> )	Oxygen (%)
12:00	415	52.7	12.3
12:01	426	52.3	12.4
12:02	447	52.2	12.3
12:03	451	51.9	12.2
12:04	464	51.9	12.1
12:05	466	51.6	12.1
12:06	461	51.6	12.0
12:07	466	51.7	12.0
12:08	459	51.5	12.0
12:09	436	51.7	11.9
12:10	437	52.5	12.0
12:11	425	52.6	12.0
12:12	414	53.0	12.1
12:13	412	53.2	12.2
12:14	413	53.2	12.3
12:15	403	53.2	12.3
12:16	398	53.5	12.4
12:17	411	53.3	12.4
12:18	430	52.8	12.4
12:19	445	52.4	12.3
12:20	464	52.1	12.2
12:21	460	52.0	12.1
12:22	477	51.7	12.1
12:23	467	51.6	12.0
12:24	477	51.5	11.9
12:25	453	51.6	11.9
12:26	452	51.9	12.0
12:27	449	51.9	12.0
12:28	439	52.3	12.1
12:29	423	52.9	12.2
12:30	415	52.9	12.3
12:31	419	52.8	12.3
12:32	414	53.1	12.3
12:33	413	52.8	12.3
12:34	433	52.7	12.3
12:35	449	52.1	12.4
12:36	449	51.7	12.3
12:37	458	51.4	12.2
12:38	477	51.2	12.1
12:39	471	51.1	12.1
12:40	471	51.1	12.0
12:41	467	51.1	12.0
12:42	474	50.9	12.0
12:43	459	51.1	12.0
12:44	460	51.3	12.0
12:45	444	51.6	12.0
12:46	417	51.8	12.1
12:47	412	52.4	12.3
12:48	395	52.9	12.4
12:49	366	53.9	12.5
12:50	354	54.2	12.6
12:51	351	54.4	12.7
12:52	359	54.1	12.7
12:53	366	53.7	12.7
12:54	371	53.3	12.7
12:55	395	53.0	12.6
12:56	408	52.5	12.5
12:57	431	51.8	12.4
12:58	433	51.8	12.3
12:59	433	52.0	12.3
13:00	439	51.8	12.3
<b>Min Value</b>	<b>351</b>	<b>50.9</b>	<b>11.9</b>
<b>Max Value</b>	<b>477</b>	<b>54.4</b>	<b>12.7</b>
<b>Average</b>	<b>432</b>	<b>52.3</b>	<b>12.2</b>


**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

Minute Averaged Gaseous Data from South Hook LNG 9th March 2020			
SCV G - Phase Two			
Referenced to 273K, 101.3kPa and 3% Oxygen on a Dry Basis			
Time (hh:mm)	Carbon Monoxide (mg/Nm <sup>3</sup> )	Oxides of Nitrogen (mg/Nm <sup>3</sup> )	Oxygen (%)
13:00	538	53.0	11.6
13:01	529	53.6	11.8
13:02	512	53.6	12.0
13:03	520	53.4	12.1
13:04	546	52.9	12.1
13:05	551	52.9	11.9
13:06	548	53.3	11.8
13:07	549	53.2	11.7
13:08	591	52.9	11.6
13:09	546	53.5	11.4
13:10	507	53.7	11.3
13:11	486	53.8	11.2
13:12	462	53.8	11.1
13:13	497	53.8	11.2
13:14	512	53.8	11.5
13:15	530	53.8	11.7
13:16	493	54.6	11.8
13:17	494	54.1	12.0
13:18	498	53.8	11.9
13:19	537	53.2	11.9
13:20	595	52.4	11.9
13:21	580	53.1	11.8
13:22	570	53.2	11.6
13:23	526	53.5	11.3
13:24	511	53.6	11.2
13:25	488	53.6	11.1
13:26	479	53.5	11.1
13:27	487	53.5	11.1
13:28	536	53.1	11.4
13:29	523	53.6	11.4
13:30	503	53.9	11.6
13:31	521	53.7	11.7
13:32	507	53.6	11.9
13:33	509	53.8	12.0
13:34	503	53.5	12.0
13:35	513	53.3	11.9
13:36	567	52.9	11.9
13:37	568	53.3	11.6
13:38	532	53.6	11.3
13:39	503	53.8	11.1
13:40	474	53.6	11.0
13:41	469	53.5	11.0
13:42	471	53.6	11.1
13:43	512	53.5	11.4
13:44	534	53.1	11.7
13:45	522	53.6	12.0
13:46	486	54.1	12.0
13:47	498	53.9	11.9
13:48	545	52.3	11.9
13:49	561	52.6	11.9
13:50	607	52.4	11.8
13:51	609	52.6	11.5
13:52	542	53.2	11.2
13:53	500	53.1	11.1
13:54	491	53.4	11.1
13:55	496	53.3	11.1
13:56	518	53.4	11.3
13:57	554	53.2	11.7
13:58	528	53.4	11.9
13:59	490	54.0	12.0
14:00	543	52.8	12.1
<b>Min Value</b>	<b>462</b>	<b>52.3</b>	<b>11.0</b>
<b>Max Value</b>	<b>609</b>	<b>54.6</b>	<b>12.1</b>
<b>Average</b>	<b>523</b>	<b>53.4</b>	<b>11.6</b>

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet


## 2.2.4 - Gaseous Emissions Graphical Data

SCV 1H One Minute Averaged Gaseous Emissions Data - 11th March 2020  
(273.15K, 101.325kPa, 3% O<sub>2</sub>, on a Dry Gas basis) using the NPL Conventional Analysis Package



NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

SCV 2G One Minute Averaged Gaseous Emissions Data - 9th March 2020  
(273.15K, 101.325kPa, 3% O<sub>2</sub>, on a Dry Gas basis) using the NPL Conventional Analysis Package



NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.5 - Gas Calibration Log

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

GAS CALIBRATION MEASUREMENTS

Client:	Intertek Ltd	Date:	09/03/2020	SO <sub>2</sub> (ppm)	CO (ppm)	NONNOX (ppm)	N <sub>2</sub> O (ppm)
Site:	South Hook LNG	Job Number:	INTU44MAR20	CBISS auto ranges	0-260	0-200	0-100
Stock ID:	SCVs	Initial N <sub>2</sub> pressure bar:	180	0-700	0-400	0-250 (only NOx)	0-15
Reference oxygen %	3	Leak check method	Flow method	Testo ranges	0-5000	0-10000	0-30000 only NO
<b>GAS CALIBRATION LOG - ANALYSER ADJUSTMENT</b>							
	SO <sub>2</sub>	CO	NO	O <sub>2</sub>	CO <sub>2</sub>	VOCs	N <sub>2</sub> O
Gas Cylinder ID:	183642SG	183642SG	NO	153247SG	153247SG		
Analyser type / Analyser ID	PG250	AS0208	PG250	AS0208	PG250	AS0208	
Cylinder Concentration:	154.9 ppm	79.4 ppm	ppm	15.13 % Vol	15.21 % Vol	ppm C <sub>2</sub> H <sub>6</sub>	ppm
Span Value:	154.9 ppm	79.4 ppm	ppm	15.13 % Vol	15.21 % Vol	ppm C <sub>2</sub> H <sub>6</sub>	ppm
Analyser Range: 0 -	200 ppm	100 ppm	ppm	25 % Vol	20 % Vol	ppm C <sub>2</sub> H <sub>6</sub>	ppm
Check Zero Time			12:22				
Check Zero Reading	ppm	-1.9 ppm	0.0 ppm	ppm	-0.07 % Vol	-0.01 % Vol	ppm C <sub>2</sub> H <sub>6</sub>
Initial Gain	0	0	0	ppm	-8	-2	ppm C <sub>2</sub> H <sub>6</sub>
Time			12:23				
Adjust Zero Reading	ppm	0.0 ppm	0.0 ppm	ppm	0.00 % Vol	0.00 % Vol	ppm C <sub>2</sub> H <sub>6</sub>
Final Gain	-1	0	0	ppm	-9	-3	ppm C <sub>2</sub> H <sub>6</sub>
Time			12:23				
Check Span Reading	ppm	157.1 ppm	81.6 ppm	ppm	12.23	12.28	ppm C <sub>2</sub> H <sub>6</sub>
Initial Gain	1.425	1.129	1.129	ppm	15.29 % Vol	15.87 % Vol	ppm C <sub>2</sub> H <sub>6</sub>
Time			12:35				
Adjust Span Reading	ppm	154.9 ppm	79.4 ppm	ppm	1.003	0.982	ppm C <sub>2</sub> H <sub>6</sub>
Final Gain	1.405	1.095	1.095	ppm	15.13 % Vol	15.21 % Vol	ppm C <sub>2</sub> H <sub>6</sub>
Check Zero Time			12:38				
Check Zero Reading	ppm	0.6 ppm	0.4 ppm	ppm	-0.02 % Vol	0.02 % Vol	ppm C <sub>2</sub> H <sub>6</sub>
Zero Drift	0.0 ppm	0.6 ppm	0.4 ppm	ppm	-0.02 % Vol	0.0 % Vol	ppm C <sub>2</sub> H <sub>6</sub>
Acceptance		Accept <2% of range	Accept <2% of range		Accept <2% of range	Accept <2% of range	0.0 ppm

GAS CALIBRATION LOG - SAMPLING SYSTEM CHECK - FLOW METHOD

Expected Flow:	SO <sub>2</sub>	CO	NO	O <sub>2</sub>	CO <sub>2</sub>	VOCs	N <sub>2</sub> O
Time	l/min	0.40 l/min	0.40 l/min	0.40 l/min	0.40 l/min		0.0 ppm
Leak check	Reading	l/min	0.0 l/min	0.0 l/min	0.0 l/min		
	Pass/Fail	PASS	PASS	PASS	PASS		
<b>GAS CALIBRATION LOG - DUCT CHECK</b>							
Span Value:	SO <sub>2</sub>	CO	NO	O <sub>2</sub>	CO <sub>2</sub>	VOCs	N <sub>2</sub> O
Check Zero Time	ppm	154.9 ppm	79.4 ppm	0.0 ppm	15.13 % Vol	15.21 % Vol	0.0 ppm
Check Span Time	ppm	1.5 ppm	0.1 ppm	ppm	-0.02 % Vol	0.03 % Vol	ppm C <sub>2</sub> H <sub>6</sub>
Zero Drift (%)	ppm	151.9 ppm	80.6 ppm	ppm	14.33	14.33	ppm C <sub>2</sub> H <sub>6</sub>
Span Drift (%)	ppm	1.0	0.1	ppm	15.07 % Vol	14.81 % Vol	ppm C <sub>2</sub> H <sub>6</sub>
Acceptance zero	Accept	1.9	1.5	ppm	0.1	0.2	ppm C <sub>2</sub> H <sub>6</sub>
Acceptance span	Accept	Accept	Accept	Accept	Accept	Accept	Correct for drift

CALIBRATION TO BE CARRIED OUT BY OR UNDER THE SUPERVISION OF MCERTS QUALIFIED PERSONNEL WITH LEVEL TWO AND TE4

Name:	Matthew Ellison
MCERTS ID:	MM-05-082

Personnel Present:	MRE/JG
--------------------	--------

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

GAS CALIBRATION MEASUREMENTS

Client:	Intertek Ltd		Date:	11/03/2020		CBISS auto ranges	SO <sub>2</sub> (ppm)	CO (ppm)	NONNOx (ppm)	N <sub>2</sub> O (ppm)
Site:	South Hook LNG	Job Number:	INTU4MAR20	Initial N <sub>2</sub> pressure bar:	180					
Stack ID:	SCVs		Leak check method	Flow method		Testo ranges	0-200	0-400	0-400	0-400 (only NO)
Reference oxygen %	3	Leak check method	Flow method	Testo ranges		0-5000	0-0000	0-3000	0-3000 (only NO)	
	SO <sub>2</sub>	CO	NO	NO <sub>x</sub>	O <sub>2</sub>	CO <sub>2</sub>	VOCs			N <sub>2</sub> O
Gas Cylinder ID:	183042SG	183642SG			158247SG	158247SG				
Analysers type / Analyser ID	PG250	AS0208	PG250	AS0208	PG250	AS0208	PG250			
Cylinder Concentration:	ppm	154.9 ppm	79.4 ppm	ppm	15.13 % Vol	15.21 % Vol	ppm			ppm
Span Value:	ppm	154.9 ppm	79.4 ppm	ppm	15.13 % Vol	15.21 % Vol	ppm			ppm
Analyser Range: 0 -	ppm	200 ppm	100 ppm	ppm	25 % Vol	20 % Vol	ppm			ppm
Time					11:03					
Check Zero	Reading	ppm	1.1 ppm	0.0 ppm	ppm	-0.13 % Vol	0.00 % Vol	ppm		ppm
	Initial Gain		-1	0		-9	-3			
Time					11:04					
Adjust Zero	Reading	ppm	0.0 ppm	0.0 ppm	0.0 ppm	0.00 % Vol	0.00 % Vol	ppm		ppm
	Final Gain		0	0		-12	-3			
Time					11:10					
Check Span	Reading	ppm	149.3 ppm	78.2 ppm	ppm	15.08 % Vol	14.96 % Vol	ppm		ppm
	Initial Gain		1.405	1.095		0.993	0.962			
Time				1.1:12		11:17	11:17	ppm		ppm
Adjust Span	Reading	ppm	154.9 ppm	79.4 ppm	ppm	15.13 % Vol	15.21 % Vol	ppm		ppm
	Final Gain		1.455	1.110		0.996	0.970			
Check Zero	Time				11:20					
	Reading	ppm	0.1 ppm	0.1 ppm	ppm	0.03 % Vol	0.05 % Vol	ppm		ppm
Zero Drift	0.0 ppm	0.1 ppm	0.1 ppm	0.0 ppm	0.03 % Vol	0.1 % Vol	0.0 ppm	ppm C <sub>2</sub> H <sub>6</sub>	0.0 ppm	ppm
Acceptance			Accept <2% of range	Accept <2% of range		Accept <2% of range	Accept <2% of range			
	SO <sub>2</sub>	CO	NO	O <sub>2</sub>	CO <sub>2</sub>	VOCs				N <sub>2</sub> O
Expected Flow:	l/min	0.40 l/min	0.40 l/min	l/min	l/min	0.40 l/min	0.40 l/min	0.0 ppm C <sub>2</sub> H <sub>6</sub>	0.0 ppm C <sub>2</sub> H <sub>6</sub>	0.0 ppm
Leak check	Time	1:39	11:39	0.0 l/min	0.0 l/min	0.0 l/min	0.0 l/min			
	Reading			PASS	PASS	PASS	PASS			
	SO <sub>2</sub>	CO	NO	O <sub>2</sub>	CO <sub>2</sub>	VOCs				N <sub>2</sub> O
Span Value:	0.0 ppm	154.9 ppm	79.4 ppm	0.0 ppm	15.13 % Vol	15.21 % Vol	0.03 % Vol	0.0 ppm C <sub>2</sub> H <sub>6</sub>	0.0 ppm C <sub>2</sub> H <sub>6</sub>	0.0 ppm
Check Zero	Time				13:20					
	Reading	ppm	-0.1 ppm	0.1 ppm	ppm	0.03 % Vol	0.03 % Vol	ppm C <sub>2</sub> H <sub>6</sub>	ppm C <sub>2</sub> H <sub>6</sub>	ppm
Check Span	Time				13:25					
	Reading	ppm	152.5 ppm	82.3 ppm	ppm	15.16 % Vol	14.90 % Vol	ppm C <sub>2</sub> H <sub>6</sub>	ppm C <sub>2</sub> H <sub>6</sub>	ppm
Zero Drift (%)			0.1	0.1		0.2	0.2			
Span Drift (%)			1.5	3.7		Accept	Accept	Accept	Accept	Accept
Acceptance zero			Accept	Accept		Correct for drift	Correct for drift			
Acceptance span			Accept	Accept		Correct for drift	Correct for drift			

CALIBRATION TO BE CARRIED OUT BY OR UNDER THE SUPERVISION OF MCERTS QUALIFIED PERSONNEL WITH LEVEL TWO AND TE4

Name:	Matthew Ellison
MCERTS ID:	MM-05-682

Personnel Present:	MRE/J/G
--------------------	---------

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.6 - Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## SCV 1H Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of NOx BS EN 14792:2017 - PG-250 AS0208  
Sep-19  
v3.7

Performance characteristics		Value	specification
Response time		70 seconds	<200
Logger sampling interval		60 seconds	0.21 mg/m³
Measurement period		60 minutes	
Number of readings in measurement		60	
Standard deviation of repeatability at zero		0.03	$\leq 5\%$ range
Standard deviation of repeatability at span level		0.15	$\leq 5\%$ range
Standard deviation of reproducibility		0	$\leq 3\%$ range
Deviation from linearity (peak or fl)		0.35 % full scale	$\leq 3\%$ range
Zero drift		0.21 mg/m³	$\leq 5\%$ span value
Span drift		0 mg/m³	$\leq 5\%$ span value
Influence of sample gas flow		2 % full scale/10 Pa	$\leq 5\%$ range
Influence of atmospheric pressure		2 % full scale/3kPa	$\leq 5\%$ range
Influence of ambient temperature		0.18 % full scale/20K	$\leq 5\%$ range
CO (mg/m³)	300	0.4 mg/m³	$\leq 5\%$ range
HCl (mg/m³)	50	1.1 mg/m³	$\leq 4\%$ range (Total)
SO₂ (mg/m³)	1000	0.5 mg/m³	HCl range
N₂O (mg/m³)	20	0.8 mg/m³	SO₂ range
Influence of voltage		0 % full scale/10V	N₂O range
Influence from vibration		0 % full scale	Voltage
Losses in the line (peak)		0 % value	
Converter efficiency		99.5 %	$\leq 5\%$ of value
Uncertainty of calibration gas		1 %	$\leq 2\%$ of value
Performance characteristic		Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero		$U_{0,0}$	0.00 mg/m³
Standard deviation of repeatability at span level		$U_{0,s}$	0.04 mg/m³
Standard deviation of reproducibility		$U_{0,p}$	0.00 mg/m³
Lack of fit		$U_{lin}$	
Drift		$U_{drift}$	0.41 mg/m³
Influence of sample gas flow		$U_{gas}$	0.12 mg/m³
Influence of atmospheric pressure		$U_{atm}$	0.00 mg/m³
Influence of ambient temperature		$U_{temp}$	0.08 mg/m³
CO (mg/m³)		$U_{CO}$	0.02 mg/m³
HCl (mg/m³)		$U_{HCl}$	0.28 mg/m³
SO₂ (mg/m³)		$U_{SO_2}$	0.00 mg/m³
N₂O (mg/m³)		$U_{N_2O}$	0.05 mg/m³
Influence of voltage		$U_{volt}$	0.00 mg/m³
Influence from vibration		$U_{vib}$	0.00 mg/m³
Losses in the line (peak)		$U_{loss}$	0.00 mg/m³
Converter efficiency		$U_{conv}$	0.67 mg/m³
Uncertainty of calibration gas		$U_{cal}$	0.13 mg/m³
Uncertainty in factor		$U_f$	2.26 mg/m³
Measurement uncertainty			
Combined uncertainty			0.88 mg/m³
Expanded uncertainty	$k =$	2	1.76 mg/m³
Uncertainty corrected to std cond			
Expressed uncertainty			5.79 mg/m³
Expanded uncertainty			10.87 % value
Expanded uncertainty			1.65 % ELV

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of CO BS EN 15058:2017 - PG-250 AS0208  
v4.3 Sep-18

Performance characteristics		Value	specification		Effect of drift 3.63 mg/m <sup>3</sup>
			seconds	<200	
Response time	None	mg/m <sup>3</sup> (corrected)	Cal gas conc	193.63 mg/m <sup>3</sup>	
Measured concentration	210.40 mg/m <sup>3</sup> (101.3kPa)	Full Scale	250.00 mg/m <sup>3</sup>	0.00	273.00
Measured concentration	451.90 mg/m <sup>3</sup> (Corrected)			0.00	101.30
			ref	3.00	101.30
			measured	12.23	0.00
			Uncert	0.377	0.00
Factors				1.00	0.00
Uncertainty in factor				1.00	0.00
Correction Factor				0.09	0.09
				2.05	0.09
				uf	0.09
Performance characteristics					
Logger sampling interval	60 seconds				
Number of readings in measurement	60 minutes				
Standard deviation of repeatability at span level	0.04 % full scale				
Standard deviation of repeatability at span level	0.1 % full scale				
Deviation from linearity(lack of fit)	0.2 % full scale				
Zero drift	0.13 mg/m <sup>3</sup>				
Span drift	3.21 mg/m <sup>3</sup>				
Influence of sample gas flow	2 % full scale/10l				
Influence of atmospheric pressure	2 % full scale/30KPa				
Influence of ambient temperature	0.05 % full scale/20K				
NO2 (mg/m <sup>3</sup> )	0.5 mg/m <sup>3</sup>				
N2O (mg/m <sup>3</sup> )	20 mg/m <sup>3</sup>				
CO2 (% vol)	1.0 % vol				
SO2 (mg/m <sup>3</sup> )	0.5 mg/m <sup>3</sup>				
Influence of voltage	1.3 % full scale/10V				
Influence of vibration	0 % full scale				
Losses in the line (leak)	0 % value				
Uncertainty of calibration gas	1 % value				
Performance characteristic					
Standard deviation of repeatability at zero	U <sub>0</sub> 0.00				
Standard deviation of repeatability at span level	U <sub>sp</sub> 0.03				
Standard deviation of reproducibility	U <sub>repro</sub> 0.00				
Lack of fit	U <sub>fit</sub> 0.29				
Drift	U <sub>drift</sub> 2.09				
Influence of sample gas flow	U <sub>gasflow</sub> 0.00				
Influence of atmospheric pressure	U <sub>atm</sub> 0.10				
Influence of ambient temperature	U <sub>temp</sub> 0.01				
NO2 (mg/m <sup>3</sup> )	U <sub>NO2</sub> 0.04				
N2O (mg/m <sup>3</sup> )	U <sub>N2O</sub> 0.06				
CO2 (% vol)	U <sub>CO2</sub> 0.42				
SO2 (mg/m <sup>3</sup> )	U <sub>SO2</sub> -0.02				
Influence of voltage	U <sub>V</sub> 0.52				
Influence of vibration	U <sub>vib</sub> 0.00				
Losses in the line (leak)	U <sub>leak</sub> 0.00				
Uncertainty of calibration gas	U <sub>cal</sub> 1.05				
Uncertainty in factor	U <sub>f</sub> 18.34				
Measurement uncertainty					
Combined uncertainty					
Expanded uncertainty	k = 2				
Uncertainty corrected to std condns					
Expanded uncertainty	expressed with a level of confidence of 95%				
Expanded uncertainty	expressed with a level of confidence of 95%	37.99 mg/m <sup>-3</sup>			
		8.80 % value			

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of O<sub>2</sub> BS EN 14789:2017 - PG-250 AS0208

v3.1 Sep-18

Limit value	N/A	Cal gas conc	15.13 % vol
Measured concentration	12.23 % vol	Full Scale	25.00 % vol

Performance characteristics		Value	specification
Response time		65 seconds	≤200 seconds
Logger sampling interval		60 seconds	
Measurement period		60 minutes	
Number of readings in measurement		60 % vol	<0.2 % vol
Standard deviation of repeatability at zero	0.05	% vol	<0.2 % vol
Standard deviation of repeatability at span level	0.1	% vol	<0.2 % vol
Standard deviation of reproducibility	0.15	% vol	<0.2 % vol
Deviation from linearity(lack of fit)	0.26	% vol	<0.3 % vol
Zero drift	0.03	% vol	<0.2 % vol
Span drift	0.03	% vol	<0.2 % vol
Influence of sample gas flow	0.2	% vol/10/h	<0.2 % vol
Influence of atmospheric pressure	0.2	% vol/3kPa	<0.2 % vol
Influence of ambient temperature	-0.07	% vol/20/K	<0.5 % vol
Cross sensitivity	0.14	% vol	<0.4 % vol
Influence of voltage	0	% vol/10/V	<0.2 % vol
Influence from vibration	0	% vol	<0.2 % vol
Losses in the line (leak)	0	% value	<2% of value
Uncertainty of calibration gas	1	% value	<2% of value
Performance characteristic		Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero		U <sub>0</sub>	0.00
Standard deviation of repeatability at span level		U <sub>rs</sub>	0.00
Standard deviation of reproducibility		U <sub>rp</sub>	0.02
Lack of fit		U <sub>ln</sub>	0.15
Drift		U <sub>drift</sub>	0.03
Influence of sample gas flow		U <sub>gas</sub>	0.000
Influence of atmospheric pressure		U <sub>pres</sub>	0.00
Influence of ambient temperature		U <sub>temp</sub>	0.00
Cross sensitivity		U <sub>interf</sub>	0.08
Influence of voltage		U <sub>volt</sub>	0.00
Influence from vibration		U <sub>vib</sub>	0.00
Losses in the line (leak)		U <sub>loss</sub>	0.00
Uncertainty of calibration gas		U <sub>cal</sub>	0.06
Measurement uncertainty			0.18 % vol
Combined uncertainty	k = 2		0.37 % vol
Expanded uncertainty	expressed with a level of confidence of 95%		0.37 % vol
Expanded uncertainty	expressed with a level of confidence of 95%		3.02 % value

# NATIONAL PHYSICAL LABORATORY

## Continuation Sheet

### Uncertainty calculation for Velocity and Volume Flow Rate Measurement by Pitot tube EN ISO 1691-1

Jan-16

Enter data in orange cells only


v1.3

Gas constant

8.314 J/K/mol

Velocity meas. during calibration	3 m/s
Air density meas. during calibration	1.23 kg/m <sup>3</sup>
DP meas. during calibration	7.9 Pa
Uncertainty of velocity meas. at calibration	2.2 %
Uncertainty of air density meas. at calibration	0.075 %
Uncertainty of DP meas. at calibration	3.95 %
Pit. coefficient, K	0.334
Expanded uncertainty (95%, k=2)	9.0 %
Expanded uncertainty (95%, k=2)	0.08

Duct dimensions



All Pressures should be entered in Pascals, Pa

Measurement Point	Atmospheric Pressure, Pa	Stack Pressure, Pa	Static Pressure, Pa	meas1, Pa	meas2, Pa	meas3, Pa	meas4, Pa	Delta P, Pa	meas5, Pa	Delta P, Pa	meas6, Pa	Delta P, Pa	meas7, Pa	Delta P, Pa	meas8, Pa	Delta P, Pa	Stack Temperature, C	Water Vapour Content, %	Dry gas basis	CO <sub>2</sub> , ppm	N <sub>2</sub> , %	O <sub>2</sub> , %	dry molecular wt, g/mol	stack molecular wt, g/mol	
1	100140	102596	119.6	31.18825849					31.18825449								13.5	13.5	1.54	165	5.1	82.7	12.2	29.30	
2	100140	102596	119.6	32.08791398					32									13.5	13.5	1.54	165	5.1	82.7	12.2	29.30
3	100140	102596	119.6	34.28709187					34									13.5	13.5	1.54	165	5.1	82.7	12.2	29.30
4	100140	102596	119.6	41.5634466					42									13.6	13.6	1.54	165	5.1	82.7	12.2	29.30
5	100140	102596	119.6	44.3931404					45									13.6	13.6	1.54	165	5.1	82.7	12.2	29.30
6	100140	102596	119.6	44.08340383					44									13.7	13.7	1.54	165	5.1	82.7	12.2	29.30
7	100140	102596	119.6	28.28934985					28									13.5	13.5	1.54	165	5.1	82.7	12.2	29.30
8	100140	102596	119.6	33.2874682					33									13.6	13.6	1.54	165	5.1	82.7	12.2	29.30
9	100140	102596	119.6	38.08656182					38									13.6	13.6	1.54	165	5.1	82.7	12.2	29.30
10	100140	102596	119.6	40.36479626					40									13.6	13.6	1.54	165	5.1	82.7	12.2	29.30
11	100140	102596	119.6	44.2891404					45									13.7	13.7	1.54	165	5.1	82.7	12.2	29.30
12	100140	102596	119.6	45.28257591					46									13.8	13.8	1.54	165	5.1	82.7	12.2	29.30
Mean	100140	102596	119.6	38.3					38.3								13.6	13.6	1.5	165.0	5.1	82.7	12.2	29.30	

$$\rho = \frac{molar\ mass}{R \cdot gas\ temperature} e^{-R \cdot gas\ temperature/e}$$

Mean density

1.226 kg/m<sup>3</sup>

Mean velocity

6.57 m/sec

Standard uncertainty of velocity

0.31 m/sec

Expanded uncertainty in velocity

0.61 m/sec

Velocity =  $\sqrt{\frac{2 \Delta_p}{\rho}}$

Flow rate	Circular duct	Rectangular duct
	40684 m <sup>3</sup> /hour	0 m <sup>3</sup> /hour
Volume flow rate expanded uncertainty	3866 m <sup>3</sup> /hour	#DIV/0!
Volume flow rate expanded uncertainty	9.5 % of value	#DIV/0!

Developed for the STN by NPL, David Butterfield & Christos Dimopoulos

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## SCV 2G Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of NOx BS EN 14792:2017 - PG-250 AS0208

v3.7

Sep-19

Performance characteristics		Value	specification
Response time		70 seconds	\$200
Longer sampling interval		60 seconds	
Measurement period		60 minutes	
Number of readings in measurement		60	
Standard deviation of repeatability at zero		0.03	$\leq 2\%$ range
Standard deviation of repeatability at span level		0.15	$\leq 2\%$ range
Standard deviation of reproducibility		0	$\leq 3\%$ range
Deviation from linearity (ack off fit)		0.95	$\leq 2\%$ range
Zero drift		0.21 mg/m³	$\leq 5\%$ span value
Span drift		2.46 mg/m³	$\leq 5\%$ span value
Influence of sample gas flow		2 %/full scale/10Pa	$\leq 2\%$ range
Influence of atmospheric pressure		0.18 %/full scale/20K	$\leq 2\%$ range
Influence of ambient temperature		0.4 mg/m³	$\leq 5\%$ range
CO (mg/m³)		300 mg/m³	
HCl (mg/m³)		50 mg/m³	
SO₂ (mg/m³)		100 mg/m³	
N₂O (mg/m³)		20 mg/m³	
Influence of voltage		0 %/full scale/10V	$\leq 2\%$ range/10V
Influence from vibration		0 %/value	$\leq 2\%$ range
Losses in the line (leak)		0 %/value	$\leq 2\%$ of value
Converter efficiency		99.5 %	29.9% of value
Uncertainty of calibration gas		1 %	st: 2% of value
Performance characteristic		Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero		$u_{\text{U}_0}$	mp/m³
Standard deviation of repeatability at span level		$u_{\text{U}_s}$	
Standard deviation of reproducibility		$u_{\text{U}_{\text{Re}}}$	
Lack of fit		$u_{\text{U}_{\text{Lof}}}$	
Drift		$u_{\text{U}_{\text{Drift}}}$	
Influence of sample gas flow		$u_{\text{U}_{\text{Flow}}}$	
Influence of atmospheric pressure		$u_{\text{U}_{\text{Press}}}$	
Influence of ambient temperature		$u_{\text{U}_{\text{Temp}}}$	
CO (mg/m³)		$u_{\text{U}_{\text{CO}}}$	
HCl (mg/m³)		$u_{\text{U}_{\text{HCl}}}$	
SO₂ (mg/m³)		$u_{\text{U}_{\text{SO2}}}$	
N₂O (mg/m³)		$u_{\text{U}_{\text{N2O}}}$	
Influence of voltage		$u_{\text{U}_{\text{Vt}}}$	
Influence from vibration		$u_{\text{U}_{\text{Vib}}}$	
Losses in the line (leak)		$u_{\text{U}_{\text{Leak}}}$	
Converter efficiency		$u_{\text{U}_{\text{Eff}}}$	
Uncertainty of calibration gas		$u_{\text{U}_{\text{Cal}}}$	
Uncertainty in factor		$u_{\text{U}_{\text{f}}}$	2.13
Measurement uncertainty			
Combined uncertainty			mg/m³
Expanded uncertainty	$k =$	2	2.11 mg/m³
Uncertainty corrected to std cond			
Expressed uncertainty			mg/m³
Expanded uncertainty			5.87 mg/m³
Expanded uncertainty			11.00 % value
Expanded uncertainty			1.98 % ELV

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of CO BS EN 15058:2017 - PG-250 AS0208

Sep-18 v4.3

Performance characteristics			specification		
	Value				
Response time	60	seconds	≤20		
Logger sampling interval	60	seconds			
Measurement period	60	minutes			
Number of readings in measurement	60				
Standard deviation of repeatability at zero	0.04	% full scale	≤±2% range		
Standard deviation of repeatability at span level	0.1	% full scale	≤±2% range		
Standard deviation of reproducibility	0	% full scale	≤±3% range		
Deviation from linearity/lack of fit	0.2	% full scale	≤±3% range		
Zero drift	2.01	mg/m <sup>3</sup>	≤5% span value		
Span drift	4.02	mg/m <sup>3</sup>	≤5% span value		
Influence of sample gas flow	2	% full scale/10	≤±2% range		
Influence of atmospheric pressure	2	% full scale/30Pa	≤±2% range		
Influence of ambient temperature	0.05	% full scale/20K	≤±2% range		
NO <sub>2</sub> (mg/m <sup>3</sup> )	30	0.5	NO <sub>2</sub> range	281.15	282.15
N <sub>2</sub> O (mg/m <sup>3</sup> )	20	1.0	N <sub>2</sub> O range	0	4
CO <sub>2</sub> (% vol)	15	0.5	CO <sub>2</sub> range	0	2
SO <sub>2</sub> (mg/m <sup>3</sup> )	200	-1.3	SO <sub>2</sub> range	10	15
Influence of voltage	0	% full scale/10V	≤±2% range/10V	0	0 mg/m <sup>3</sup>
Influence of vibration	0	% full scale	≤±2% range	1.10	1.10 V
Losses in the line (leak)	0	% value	≤±5% of value		
Uncertainty of calibration gas	1	% value	≤±2% of value		
Performance characteristic			Value of uncertainty quantity	mg/m <sup>3</sup>	
Standard deviation of repeatability at zero	u <sub>g,0</sub>			0.00	
Standard deviation of repeatability at span level	u <sub>g,s</sub>			0.03	
Standard deviation of reproducibility	u <sub>g,r</sub>			0.00	
Lack of fit	u <sub>g,f</sub>			0.29	
Drift	u <sub>g,d</sub>			4.44	
Influence of sample gas flow	u <sub>g,pas</sub>			0.00	
Influence of atmospheric pressure	u <sub>g,pres</sub>			0.48	
Influence of ambient temperature	u <sub>g,temp</sub>			0.00	
NO <sub>2</sub> (mg/m <sup>3</sup> )	u <sub>g,no2</sub>		Use largest of sum of all positive or all negative influences	0.04	
N <sub>2</sub> O (mg/m <sup>3</sup> )	u <sub>g,n2o</sub>		Criteria	0.06	
CO <sub>2</sub> (% vol)	u <sub>g,co2</sub>		sum <4% range	0.42	
SO <sub>2</sub> (mg/m <sup>3</sup> )	u <sub>g,so2</sub>		0.52 largest	-0.02	
Influence of voltage	u <sub>g,vol</sub>		Value to use for inference uncertainty	0.00	
Losses in the line (leak)	u <sub>g,lp</sub>		u <sub>int</sub>	0.00	0.52
Uncertainty of calibration gas	u <sub>g,cal</sub>			1.37	
Uncertainty in factor	u <sub>f</sub>			20.87	
Measurement uncertainty			mg/m <sup>3</sup>		
Combined uncertainty	4.71			4.71	
Expanded uncertainty	9.41			9.41	
Uncertainty corrected to std conds					
Expanded uncertainty expressed with a level of confidence of 95%	45.46	mg/m <sup>3</sup>		45.46	
Expanded uncertainty	8.69	% value		8.69	

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of O<sub>2</sub> BS EN 14789:2017 - PG-250 AS0208

Sep-18

Limit value	 N/A	Cal gas conc	 15.13 % vol
Measured concentration	 11.59 % vol	Full Scale	 25.00 % vol

Performance characteristics		Value	specification
Response time		65 seconds	\$200 seconds
Logger sampling interval		60 seconds	
Measurement period		60 minutes	
Number of readings in measurement		60 % vol	<0.2 % vol
Standard deviation of repeatability at zero	0.05	% vol	<0.2 % vol
Standard deviation of repeatability at span level	0.1	% vol	<0.2 % vol
Standard deviation of reproducibility	0.15	% vol	<0.3 % vol
Deviation from linearity(lack of fit)	0.26	% vol	<0.2 % vol
Zero drift	0.02	% vol	<0.2 % vol
Span drift	0.06	% vol	<0.2 % vol
Influence of sample gas flow	0.2	% vol/10l/h	<0.2 % vol
Influence of atmospheric pressure	0.2	% vol/3kPa	<0.2 % vol
Influence of ambient temperature	-0.07	% vol/20K	<0.5 % vol
Cross sensitivity	0.14	% vol	<0.4 % vol
Influence of voltage	0	% vol/10V	<0.2 % vol
Influence from vibration	0	% vol	<0.2 % vol
Losses in the line (leak)	0	% value	s< 2% of value
Uncertainty of calibration gas	1	% value	s< 2% of value
Performance characteristic		Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero		U <sub>0</sub>	0.00 % vol
Standard deviation of repeatability at span level		U <sub>rs</sub>	0.00 % vol
Standard deviation of reproducibility		U <sub>rp</sub>	0.02 % vol
Lack of fit		U <sub>lf</sub>	0.15 % vol
Drift		U <sub>drift</sub>	0.04 % vol
Influence of sample gas flow		U <sub>gas</sub>	0.000 % vol
Influence of atmospheric pressure		U <sub>pres</sub>	0.02 % vol
Influence of ambient temperature		U <sub>temp</sub>	0.00 % vol
Cross sensitivity		U <sub>intef</sub>	0.08 % vol
Influence of voltage		U <sub>volt</sub>	0.00 % vol
Influence from vibration		U <sub>vib</sub>	0.00 % vol
Losses in the line (leak)		U <sub>leak</sub>	0.06 % vol
Uncertainty of calibration gas		U <sub>cal</sub>	

Measurement uncertainty		0.19 % vol
Combined uncertainty	2	0.37 % vol
Expanded uncertainty	expressed with a level of confidence of 95%	0.37 % vol
Expanded uncertainty	expressed with a level of confidence of 95%	3.21 % value

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for Velocity and Volume Flow Rate Measurement by Pitot tube EN ISO 16911-1


v1.3 Jan-16

Enter data in orange cells only

Gas constant

Velocity meas. during calibration	3 m/s
Air density meas. during calibration	1.23 kg/m <sup>3</sup>
DP meas. during calibration	7.9 Pa
Uncertainty of velocity meas. at calibration	2.2 %
Uncertainty of air density meas. at calibration	0.75 %
Uncertainty of DP meas. at calibration	3.95 %
Pitot coefficient, K	0.334
Expanded uncertainty (95%, k=2) as % of value	9.0 %
Expanded uncertainty (95%, k=2)	0.08

Duct dimensions



All Pressures should be entered in Pascals, Pa

Measurement Point	Atmospheric Pressure, Pa	Stack pressure, Pa	Static Pressure, Pa	meas1, Pa	meas2, Pa	meas3, Pa	meas4, Pa	Delta P, Pa	Stack Temperature, C	Water Vapour Content, %	CO <sub>2</sub> , ppm	CO <sub>2</sub> , %	N <sub>2</sub> , %	O <sub>2</sub> , %	Dry gas basis	
1	100190	1010368	1016807033	178	31.68807033	31.68807033	31.68807033	31.68807033	12.4	1.42	200	5.4	83.0	11.6	29.33	
2	100190	1010368	37.68561235	178	37.68561235	37.68561235	37.68561235	37.68561235	12.3	1.42	200	5.4	83.0	11.6	29.17	
3	100190	1010368	39.18524785	178	39.18524785	39.18524785	39.18524785	39.18524785	12.3	1.42	200	5.4	83.0	11.6	29.33	
4	100190	1010368	43.38344791	178	43.38344791	43.38344791	43.38344791	43.38344791	12.2	1.42	200	5.4	83.0	11.6	29.33	
5	100190	1010368	44.7831404	178	44.7831404	44.7831404	44.7831404	44.7831404	12.4	1.42	200	5.4	83.0	11.6	29.17	
6	100190	1010368	47.68204904	178	47.68204904	47.68204904	47.68204904	47.68204904	11.9	1.42	200	5.4	83.0	11.6	29.33	
7	100190	1010368	51.68807033	178	51.68807033	51.68807033	51.68807033	51.68807033	32	1.42	200	5.4	83.0	11.6	29.17	
8	100190	1010368	52.58773163	178	52.58773163	52.58773163	52.58773163	52.58773163	33	12.1	1.42	200	5.4	83.0	11.6	29.33
9	100190	1010368	35.28671554	178	35.28671554	35.28671554	35.28671554	35.28671554	35	12.1	1.42	200	5.4	83.0	11.6	29.33
10	100190	1010368	44.69337804	178	44.69337804	44.69337804	44.69337804	44.69337804	45	12.2	1.42	200	5.4	83.0	11.6	29.33
11	100190	1010368	45.73221607	178	45.73221607	45.73221607	45.73221607	45.73221607	46	11.9	1.42	200	5.4	83.0	11.6	29.33
12	100190	1010368	46.78238774	178	46.78238774	46.78238774	46.78238774	46.78238774	47	12	1.42	200	5.4	83.0	11.6	29.33
Mean	100190	1010368	40.1	178.0	40.1	#DIV/0!	#DIV/0!	#DIV/0!	40.1	12.2	1.4	200.0	5.4	83.0	11.6	29.33

$$\rho = \text{molar mass} \cdot \text{absolute pressure} \cdot e^{-R \cdot \text{gas temperature} \cdot e}$$

Mean density

$$1.235 \text{ kg/m}^3$$

Mean velocity

$$6.71 \text{ m/sec}$$

Standard uncertainty of velocity

$$0.31 \text{ m/sec}$$

Expanded uncertainty in velocity

$$0.62 \text{ m/sec}$$

Flow rate

$$41335 \text{ m}^3/\text{hour}$$

Volume flow rate expanded uncertainty

$$3329 \text{ m}^3/\text{hour}$$

Volume flow rate expanded uncertainty

$$9.5 \% \text{ of value}$$

Flow rate

$$0 \text{ m}^3/\text{hour}$$

Volume flow rate expanded uncertainty

$$\#DIV/0!$$

Volume flow rate expanded uncertainty

$$\#DIV/0!$$

Flow rate

$$\#DIV/0!$$

Volume flow rate expanded uncertainty

$$\#DIV/0!$$

Volume flow rate expanded uncertainty

$$\#DIV/0!$$

Flow rate

$$0 \text{ m}^3/\text{hour}$$

Volume flow rate expanded uncertainty

$$\#DIV/0!$$

Volume flow rate expanded uncertainty

$$\#DIV/0!$$

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.7 - Calculations Used in Reporting Results

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

### Nozzle Selection

For isokinetic sampling, the pressure difference of the orifice meter must equal the pressure difference of the Pitot tube pressure multiplied by the K-factor. Where:

$$K = \text{Constant} \times C_p^2 \times D_n^4 \times DH_{@} \times \left( \frac{M_d}{M_s} \right) \left( \frac{1 - B_{wm}}{1 - B_{ws}} \right)^2 \left( \frac{T_m + 273}{T_s + 273} \right) \left( \frac{P_s}{P_m} \right)$$

$$DH = K \times DP$$

Where:-

Constant: is a constant dependent on the units used to measure the nozzle ( $8.038 \times 10^{-5}$  for mm)

$D_n$  the nozzle diameter mm

$DH_{@}$  a constant dependent on the sampler control box orifice and gas meter

$B_{ws}$  the percent water vapour in the emission as a fraction i.e. 12% = 0.12

$B_{wm}$  the percentage water vapour in the air around the meter box often assumed to be zero

$C_p$  Pitot tube coefficient dependent on the Pitot tube type

$T_m$  the meter temperature in °C

$T_s$  the stack temperature in °C

$P_s$  the stack pressure

$P_m$  the meter pressure

$M_d$  dry gas molecular weight

$M_s$  apparent stack gas molecular weight

$DH$  pressure drop across the orifice (mm water)

$DP$  differential Pitot pressure (mm water)

From this the correct nozzle size can be determined.

$$D_n = \sqrt{\left( \frac{\text{Constant}.Q_m.P_m}{(T_m + 273)C_p} \right) \left( \frac{1 - B_{wm}}{1 - B_{ws}} \right) \sqrt{\frac{(T_s + 273)M_s}{(P_s . (\Delta P)_{avg})}}}$$

Where the Constant = 0.6071 Metric

$Q_m$  = Orifice flow rate normally 21.2 actual  $\text{lmin}^{-1}$

$$= K_m \sqrt{\frac{(T_m + 273)\Delta H}{P_m M_m}}$$

Where  $K_m$  = Orifice meter coefficient

$$K_m = Q_m \sqrt{\frac{P_m M_m}{\Delta H(T_m + 273)}} = \text{Const} \sqrt{\frac{1}{\Delta H_{@}}}$$

Where Const = 183.7 metric

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

### Moisture Determination Calculations

These calculations are based at 273K and 101.325kPa

To calculate moisture the following equation is used:

$$B_{ws} = \frac{0.001245 \times W_I \times 100}{(0.001245 \times W_I) + 0.359V_m \left( \frac{P_b + \frac{\Delta H_{avg}}{13.6}}{(T_m + 273)} \right)}$$

### Particulate Concentration $C_s$ in stack Gases

At 273K and 101.325kPa and dry gas

$$C_s = \frac{W_t}{V_m} \times \frac{T_m + 273}{273} \times \frac{760}{\left( P_b + \frac{\Delta H_{avg}}{13.6} \right)} \times 1000 \quad \text{mg/Nm}^3$$

### Oxygen Concentration Correction $C_{oxy}$ to Particulate concentration

$$C_{oxy} = C \times \frac{(20.9 - \% O_2 \text{ ref})}{(20.9 - \% O_2 \text{ Meas})} \quad \text{mg/Nm}^3$$

### Dry Molecular Weight of gases

$$M_D = 0.44(\%CO_2) + 0.32(\%O_2) + 0.28(\%CO + \%N)$$

### Stack Molecular Weight of gases

$$M_s = 0.18(B_{ws}) + \frac{M_d}{100}(100 - B_{ws})$$

### Stack Gas Velocity

$$(V_s)_{avg} = 34.96 \times C_p \times \sqrt{(\Delta P)_{avg}} \sqrt{\frac{T_s + 273}{P_s M_s}} \quad \text{m/s}$$

### Mass Emission Rate $M_R$

$$M_R = \frac{C_m \times (V_s)_{avg} \times A \times 3600}{10^6} \quad \text{kg/hr}$$

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**IsoKinicity**

$$I = \frac{2.12 \times 10^8 \times V_m \times Y \times \left( P_b + \left( \frac{\Delta H_{avg}}{13.6} \right) \right) \left( \frac{273 + T_s}{273 + T_m} \right)}{\Theta P_s \pi D_n^2 (V_s) avg (100 - B_{ws})} \%$$

$W_I$	= the weight change of the impingers during sampling in g
$V_m$	= volume of dry gas sample in litres at temperature of the meter box
$B_{ws}$	= the percent water vapour in the emission
$Q$	= length of time sampling in minutes
$Y$	= Gas Meter Calibration correction factor
$V_s$	= Velocity of stack gas m/s
$C_M$	= measured concentration of particulate matter ( $\text{mg/m}^3$ )
$T_m$	= average temperature at dry gas meter ( $^{\circ}\text{C}$ )
$P_b$	= atmospheric pressure (mmHg)
$\%O_{2\text{ref}}$	= % oxygen at standard temperature & pressure
$\%O_{2\text{Meas}}$	= % oxygen measured on site
$C_p$	= Pitot tube coefficient
$DP$	= mean differential Pitot pressure drop (mm $\text{H}_2\text{O}$ )
$DH$	= mean orifice pressure drop (mm $\text{H}_2\text{O}$ )
$D_s$	= diameter of stack (m)
$D_n$	= Nozzle diameter (mm)
$T_s$	= stack temperature ( $^{\circ}\text{C}$ )
$M_d$	= molecular weight of dry stack gas
$B_w$	= moisture fraction
$P_s$	= stack pressure (mmHg)
$A$	= duct c.s.a. ( $\text{m}^2$ )
$M_s$	= molecular weight of wet stack gas
$M_d$	= molecular weight of dry stack gas
$W_t$	= total weight of particulate matter (g)