



# NATIONAL PHYSICAL LABORATORY

Teddington Middlesex UK TW11 0LW Telephone +44 20 8977 3222

## Test Report



This test report is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This test report may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

### OXIDES OF NITROGEN (AS NO<sub>2</sub>) AND CARBON MONOXIDE COMPLIANCE TESTING AT SOUTH HOOK LNG TERMINAL ON BEHALF OF INTERTEK LIMITED

Permit Number: **XP3538LD**

Operator Name: **South Hook LNG (on behalf of Intertek Limited)**

Installation Name: **South Hook LNG Terminal**

Dates of Monitoring Visit: **18th - 22nd November 2019**

Contract Reference: **E08040221**

Client Contact: **Adrian Walsh**

Client Organisation: **Intertek Limited**

Address: **Unit 14 - Waterston Trading Estate  
Main Road, Waterston  
Milford Haven  
SA73 3SL**

Monitoring Organisation: **National Physical Laboratory**

Address: **Hampton Road  
Teddington  
Middlesex  
TW11 0LW**

Date of Report: **20th December 2019**

Report Author: **Matthew Ellison**

Reference: XP3538LD/INTERTEK/SHLNG/NOV2019/SCV/PPC/Q4/V1

Report Approver: **Kevin Blakley** NPL Authorised Signatory  
MCERTS Registration: **MM-03-317** Name: Mr R Robinson (for NPLML)  
Level & TEs Held: **Level 2, TE1, TE2, TE3 & TE4** Signature:

Signature:

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**CONTENTS PAGE**

	<b>PAGE</b>
<b>Part One: Executive Summary</b>	
1.1    Monitoring Objectives	3
1.2.1   SCV 1G (Phase One) Monitoring Results	4
1.2.2   SCV 2E (Phase Two) Monitoring Results	5
1.2.3   SCV 2F (Phase Two) Monitoring Results	6
1.3    Operating Information	7
1.4    Monitoring Deviations	8
1.5    Conclusions	8
1.6    References	8
<b>Part Two: Supporting Information</b>	
<b>Appendix One</b>	
2.1.1   Emissions Testing Personnel Details	10
2.1.2   Emissions Testing Procedures	10
2.1.3   Equipment Checklist Reference	11
2.1.4   Data Capture Location Reference	11
<b>Appendix Two</b>	
2.2.1   Stack Diagram	13
2.2.2   Flow Criteria Measurements	15
2.2.3   One Minute Averaged Gaseous Emissions Data	22
2.2.4   Gaseous Emissions Graphical Data	26
2.2.5   Gas Calibration Log	30
2.2.6   Uncertainty Calculations	33
2.2.7   Calculations Used in Reporting Results	49

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**1.1 Monitoring Objectives**

NPL were awarded a contract by Intertek Limited to carry out emissions compliance testing at South Hook LNG plant near Milford Haven, Pembrokeshire. The scope of work includes carrying out emissions monitoring on the SCV flue stacks phases one and two.

There are a total of fifteen SCV units on the South Hook site that require monitoring, eight on phase one and seven on phase two. Three units were measured for oxides of nitrogen and carbon monoxide.

In addition, oxygen measurements were taken to allow a correction to reference conditions. Water vapour and flow measurements were also taken to determine the moisture content and velocity of the flue gas.

NPL carried out the monitoring visit between the 18th and 22nd November 2019. The report documents the results obtained.

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**1.2.1 SCV G (Phase One) Monitoring Results**

Client: Intertek  
Site: South Hook LNG  
Emission Point: SCV G (Phase One)

Field	Units	Oxides of Nitrogen (as NO <sub>2</sub> )	Carbon Monoxide	Oxygen	Moisture
Emission Limit Value	mg/m <sup>3</sup> , Reference Conditions	107	N/A	N/A	N/A
Periodic Monitoring Result	Reference Conditions	52.3	571	12.0	1.3
Uncertainty (95% Confidence Level)	Reference Conditions	5.9	50.7	0.4	N/A
	Units	mg/m <sup>3</sup>	mg/m <sup>3</sup>	% Vol/Vol	% Vol/Vol
Average Stack Flow	m <sup>3</sup> /s at Reference Conditions		4.7		
Reference Conditions		273K, 101.3 kPa, 3% Oxygen on a dry gas basis			
Date	dd/mm/yyyy		22/11/2019		
Sample Period	From hh:mm		10:00		
	To hh:mm		11:00		
Monitoring Method		BS EN 14792:2017	BS EN 15058:2017	BS EN 14789:2017	BS EN 14790:2017
Accreditation		UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS
Process Status	Load (Tonnes/Hour)		100		
Process Status	Burner Demand (%)		22.3		

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**1.2.2 SCV E (Phase Two) Monitoring Results**

Client: Intertek  
Site: South Hook LNG  
Emission Point: SCV E (Phase Two)

Field	Units	Oxides of Nitrogen (as NO <sub>2</sub> )	Carbon Monoxide	Oxygen	Moisture
Emission Limit Value	mg/m <sup>3</sup> , Reference Conditions	107	N/A	N/A	N/A
Periodic Monitoring Result	Reference Conditions	53.9	308	10.7	1.8
Uncertainty (95% Confidence Level)	Reference Conditions	5.7	23.0	0.4	N/A
	Units	mg/m <sup>3</sup>	mg/m <sup>3</sup>	%Vol/Vol	%Vol/Vol
Average Stack Flow	m <sup>3</sup> /s at Reference Conditions	6.5			
Reference Conditions		273K, 101.3 kPa, 3% Oxygen on a dry gas basis			
Date	dd/mm/yyyy		18/11/2019		
Sample Period	From hh:mm		15:30		
	To hh:mm		16:30		
Monitoring Method		BS EN 14792:2017	BS EN 15058:2017	BS EN 14789:2017	BS EN 14790:2017
Accreditation		UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS
Process Status	Load (Tonnes/Hour)			150	
Process Status	Burner Demand (%)			45.1	

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**1.2.3 SCV F (Phase Two) Monitoring Results**

Client: Intertek  
Site: South Hook LNG  
Emission Point: SCV F (Phase Two)

Field	Units	Oxides of Nitrogen (as NO <sub>2</sub> )	Carbon Monoxide	Oxygen	Moisture
Emission Limit Value	mg/m <sup>3</sup> , Reference Conditions	107	N/A	N/A	N/A
Periodic Monitoring Result	Reference Conditions	55.5	68.9	9.2	7.5
Uncertainty (95% Confidence Level)	Reference Conditions	5.3	5.6	0.4	N/A
	Units	mg/m <sup>3</sup>	mg/m <sup>3</sup>	% Vol/Vol	% Vol/Vol
Average Stack Flow	m <sup>3</sup> /s at Reference Conditions		7.5		
Reference Conditions		273K, 101.3 kPa, 3% Oxygen on a dry gas basis			
Date	dd/mm/yyyy		18/11/2019		
Sample Period	From hh:mm		13:00		
	To hh:mm		14:00		
Monitoring Method		BS EN 14792:2017	BS EN 15058:2017	BS EN 14789:2017	BS EN 14790:2017
Accreditation		UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS
Process Status	Load (Tonnes/Hour)			150	
Process Status	Burner Demand (%)			47.0	

**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

**1.3 Operating Information**

South Hook LNG Terminal, situated in Pembrokeshire South West Wales, is a regasification plant for natural gas. The LNG is transported in specially designed vessels to Milford Haven where it is then transferred into storage tanks, where it awaits reheating and distribution into the UK National Grid.

A total of fifteen SCVs have been built across two phases. Each one has at least two five inch ports installed, as set out in BS EN 15259. The ports are located approximately thirteen metres from the base of the stack and can be accessed by ladders that lead to a permanent platform. The sampling platform has lighting, toe boards and handrails. There is sufficient parking on the roadway at the base of the stacks for the mobile laboratory and provision of 240v 16A power supply outlets.

<b>Continuous or Batch Process?</b>	Each SCV operates on a batch process. The number of SCVs operating and the load at which they are set depend upon the required site output.		
<b>What part of the batch process was sampled? (If applicable)</b>	The periodic monitoring is carried out once an SCV has been brought online to the operators required load and has stabilised. This loading remains constant through the one hour test.		
<b>What fuel was used during monitoring? (If applicable)</b>	A small amount of LNG is used as fuel to heat a volume of water. This heat exchange warms up the LNG allowing it to be passed out into the National Grid system.		
<b>What feedstock was used during monitoring? (If applicable)</b>	N/A		
<b>What was the load during monitoring?</b>	Emission Point	Load (Ton/hr)	Burner Demand (%)
	SCV 1G	100	22.3
	SCV 2E	150	45.1
	SCV 2F	150	47.0
<b>What abatement systems are present? Were they in operation?</b>	Each SCV uses water injection to abate NOx emissions. The system was in operation during the periodic monitoring of each SCV.		
<b>Periodic monitoring results and corresponding CEM values</b>	There are no CEMs installed on the SCVs monitored.		

**NATIONAL PHYSICAL LABORATORY**  
**Continuation Sheet**

**1.4 Monitoring Deviations**

<b>Were all substances in the monitoring objectives monitored? If not why?</b>	All substances set out in the objective were monitored.
<b>Were all substances monitored in accordance to the relevant method? If not why?</b>	All substances set out in the monitoring objectives were measured in accordance to the relevant standards.
<b>Were there any other issues relevant to the monitoring results?</b>	No.

**1.5 Conclusions**

NPL carried out the emissions monitoring at South Hook LNG over a period of one week. Three SCVs were monitored for the required determinands.

The client is aware of BS EN 15259 and the requirement to carry out homogeneity testing. These tests were carried out on SCVs 1H and 2A, both units passed.

Reference - XP3538LD/INTERTEK/SHLNG/JULY2011/SCV/HOMOGENEITY.

**1.6 References**

1. STA – Risk Assessment Guide: Industrial-emission monitoring.
2. Environment Agency - Manual Stack emission monitoring performance standard for Organisations.
3. Environment Agency – M1 Technical Guidance Note – Sampling requirements for stack emission monitoring.
4. Environment Agency – M2 Technical Guidance Note – Monitoring of stack emissions to air.
5. Environment Agency - MID 15259 - Stationary source emissions - Requirements for the measurement sections and sites and for the measurement objective, plan and report.
6. Guidance on Assessing Measurement Uncertainty in Stack Emissions Monitoring, by Pullen J and Robinson R, Source Testing Association, Quality Guidance Note QGN1.

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## APPENDIX ONE

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**2.1.1 Emissions Testing Personnel Details**

Name	Role	MCERTS Number	Level 1	Level 2	Certification Level & Expiry Dates			
			N/A	Apr-2023	TE1 Sep-2023	TE2 Sep-2023	TE3 Dec-2023	TE4 Sep-2023
Matthew Ellison	Team Leader	MM-05-682	N/A	N/A	N/A	N/A	N/A	N/A
Jonny Guy	Site Assistant	MM-16-1388	Sep-2021	N/A	N/A	N/A	N/A	N/A

**2.1.2 Emissions Testing Procedures**

Instrumental Methods			Manual Methods		
Determinand	NOx	CO	O <sub>2</sub>	H <sub>2</sub> O	Stack Flow
SRM Standard	BS EN 14792:2017	BS EN 15058:2017	BS EN 14789:2017	BS EN 14790:2017	BS EN ISO 16911:2013
Instrument	Horiba PG-250	Horiba PG-250	Horiba PG-250	N/A	Pitot
Instrument Serial No.	AS0208	AS0208	AS0208	N/A	Type K Thermocouple
Principle	Chemiluminescence	NDIR	Zirconia	AS0638	AS0638
Operational Range	0 - 250 ppm	0 - 500 ppm	0 - 25%	N/A	N/A
Certified Range	0 - 125 mg/m <sup>3</sup>	0 - 95 mg/m <sup>3</sup>	0 - 25%	N/A	N/A
Uncertainty	10%	6%	6%	20%	10%
NPL Procedure	QPAS B 538	QPAS B 538	QPAS B 538	QPAS B 540	QPAS B 567
UKAS Accreditation	YES	YES	YES	YES	YES

The sample gas was extracted from the stack via a chemically inert heated line and drawn through a conditioning unit. All moisture in the gas sample was removed and cooled down to 4°C before it was pumped down another line to the NPL Mobile Source Emissions Laboratory and analysed using a Horiba PG-250. The entire sampling system had been leak tested before testing was carried out to ensure no dilution of the sample gas.

**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

The conventional analyser zero and span settings were checked before and after each test run using zero grade nitrogen (ex BOC), a suitable gas mixture (BOC beta gas standard), traceable to national reference standards and a gas dilution system. The certified accuracies of the gas standards are listed below: -

Component	Sample Location	Cylinder ID	Certified Amount	Certified Uncertainty
Carbon Monoxide	Phase 1 & 2 SCVs	146655SG	152.1 ppm	1%
Nitrogen Oxide		146655SG	78.3 ppm	1%
Oxygen		253223SG	15.30%	1%

These measurement uncertainties are expressed at a 95% level of confidence.

The ranges of the Horiba PG-250 analyser used for the testing and the values used to calibrate the instrument before and after the testing are listed below: -

Gaseous Components	Horiba Ranges	Calibrated Values
Carbon Monoxide	0 - 200 ppm	152.1 ppm
Oxides of Nitrogen (as NO <sub>2</sub> )	0 - 100 ppm	78.3 ppm
Oxygen	0 - 25%	15.30%

A leak test was conducted before testing to confirm hydraulic integrity of the gaseous sampling system. This was conducted by blocking the inlet of the sampling probe and checking the flow rate at the analyser.

The electrical volt/millivolt outputs from the PG-250 gas analyser was collected by data logger software on a PC and downloaded to digital media at the end of each day. Under the program used during the tests, the software records and stores individual readings either every 1 or 10 seconds. From this data, the logger can perform a series of calculations to output 1 minute averaged measurement on a volume/volume or mass/volume basis. After each 1 minute average has been established the data buffer is reset and the process repeats.

#### **2.1.3 Equipment Checklist Reference**

See workfile INTK47NOV19/Equipment Checklist.

#### **2.1.4 Data Capture Location Reference**

All data collected using the NPL computer system on site is backed up at the end of each day onto a memory stick. When the team returns to site this information is then uploaded onto the NPL servers and stored in the relevant location for that job. The link below is where the South Hook emissions data is stored:

P:\Stack Emissions Team\South Hook LNG - Intertek\INTK47NOV19\7. Monitoring Record Sheets


NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## APPENDIX TWO

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.1 - Stack Diagram

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet



Position	1	2	3	4	5	6	7	8	9	10
% of Diameter	3.2	8.2	14.6	22.6	34.2	65.8	77.4	85.4	91.8	96.8
Insertion, m	0.05	0.12	0.22	0.33	0.51	0.97	1.15	1.26	1.36	1.43
Insertion plus offset, m	0.15	0.22	0.32	0.43	0.61	1.07	1.25	1.36	1.46	1.53

Notes - The circular stack diameter was measured as 1.48 metres, whilst the port offset (distance between the edge of the stack to the end of the port) was measured as 23 centimetres. Access to the top of the stack was by a series of three ladders. Each one had a resting platform in-between with a self closing gate installed. The main platform itself is a permanent structure with toe boards, railings and self closing gate.

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.2 - Flow Criteria Measurements

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet


Monitoring Objective	Traverse	Site:	South Hook LNG	Stack ID:	SCV 1G
Date	22/11/2019	Site Team:	MRE/JUG	Time of Survey:	09:00
Tape Measure ID	AS0589				
Laser Measurement Device ID	N/A				
Traverse Pilot Type	S-Type				
Traverse Pilot Tube ID	AS0687				
PITOT ASSEMBLY VISUAL INSPECTION (Duct)					
Pre Test Leak Check <2.5 mm H2O	Pass				
Traverse Manometer Type	Digital				
Traverse Manometer ID	AS0638				
Traverse Manometer Range	240				
Traverse Temp. Readout ID	AS0638				
Traverse Thermocouple ID	AS0451a				
Static Pressure	$\Delta p$ (mmH2O)				
Swirl Test Conducted	Pass				
Protraction ID	AS0626				
Post Test Blockage Test (L-type only)					
PITOT ASSEMBLY VISUAL INSPECTION (Duct)					
Post Test Leak Check <2.5 mm H2O	Pass				
Conditions	Value	Units	Port ID	Reading 1 (m)	Reading 2 (m)
Stack pressure	765.86	mmHg	A		Average
Ref O <sub>2</sub> Value	3	%	B		Port Depth
Moisture Content	1.33	%	C		
CO	220	ppm	D		
CO <sub>2</sub>	5.1	%			
N <sub>2</sub>	82.88	%			
O <sub>2</sub>	12.00	%			
Dry Molecular wt	29.30				
Stack Molecular wt	29.15				
Duct Diameter	1.48	m	Line ID	Reading 1 (m)	Reading 3 (m)
Duct Depth		m			Average
Duct Width		m			Duct Diameter
Area of stack	1.72	m <sup>2</sup>			
Pbar	1011.6	mbar			
Pbar	759	mmHg			
Pilot tube coeff	0.834		Outside Side Division		
Reference Temp	273	K			
Reference Pressure	760	mmHg			
Ambient Temperature	5	°C	Enter manually from previous visit		
			Circular Duct		
			Rectangular Duct		
			Duct Diameter (m)	1.48	12.40
			Duct Depth (m)		
			Duct Width (m)		
POST TEST	Traverse Point		Δp Reading (mm H <sub>2</sub> O)	Δp Reading (mm H <sub>2</sub> O)	Δp Average (mm H <sub>2</sub> O)
Procedure	Reading 1				
	Reading 2				

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

SAMPLING LINE: North								
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s
1	1.42	3.10	3.10	3.10	3.10	30.39	11.5	5.82
2	1.26	3.40	3.40	3.40	3.40	33.33	11.4	6.09
3	1.04	3.40	3.40	3.40	3.40	33.33	11.4	6.09
4	0.44	3.60	3.60	3.60	3.60	35.29	11.3	6.27
5	0.22	3.20	3.20	3.20	3.20	31.37	11.2	5.91
6	0.06	3.60	3.60	3.60	3.60	35.29	11.2	6.27
SAMPLING LINE: West								
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s
1	1.42	2.20	2.20	2.20	2.20	21.57	11.4	4.90
2	1.26	2.30	2.30	2.30	2.30	22.55	11.3	5.01
3	1.04	2.70	2.70	2.70	2.70	26.47	11.3	5.43
4	0.44	3.10	3.10	3.10	3.10	30.39	11.3	5.82
5	0.22	3.50	3.50	3.50	3.50	34.31	11.4	6.18
6	0.06	3.50	3.50	3.50	3.50	34.31	11.4	6.18
Average values		3.1	3.1	3.1	3.1	30.7	11.3	5.8
Duct / Stack Flow Characteristics:								
Stack Velocity at stack gas T & P and a wet gas basis							Average Units	Flow Criteria Measurements
Stack flow @ STP, O <sub>2</sub> (ref) and on a dry gas basis							5.83 m <sup>3</sup> s <sup>-1</sup>	Is the Flow Ratio 3:1 or less? 1.3
Stack flow @ stack gas T & P and on a wet gas basis							4.72 m <sup>3</sup> s <sup>-1</sup>	Any local negative flow? NO
Stack flow @ stack gas T & P and on a dry gas basis							10.03 m <sup>3</sup> s <sup>-1</sup>	Flow <-15° of duct axis? YES
Stack flow @ STP and on a wet gas basis							9.89 m <sup>3</sup> s <sup>-1</sup>	Minimum Δp detected > 0.5 mmH <sub>2</sub> O YES
Stack flow @ STP, O <sub>2</sub> (ref) and on a wet gas basis							9.63 m <sup>3</sup> s <sup>-1</sup>	
							4.79 m <sup>3</sup> s <sup>-1</sup>	

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Monitoring Objective	Traverse	Site:	South Hook LNG	Stack ID:	SCV 2E
Date	18/11/2019	Site Team:	MRE/JG	Time of Survey:	15:00
Tape Measure ID	AS0569				
Laser Measurement Device ID	N/A				
Traverse Pilot Type	S-Type				
Traverse Pilot Tube ID	AS0687				
PITOT ASSEMBLY VISUAL INSPECTION (P.D.L.)					
Pre Test Leak Check <2.5 mm H2O	Pass				
Traverse Manometer Type	Digital				
Traverse Manometer ID	AS0638				
Traverse Manometer Range	240				
Traverse Temp. Readout ID	AS0638				
Traverse Thermocouple ID	AS0451a				
Δp (mmH2O)					
Static Pressure	21.50				
Swirl Test Conducted	Pass				
Protractor ID	AS0626				
Post-Test Blockage Test (L-type only)	Pass				
PITOT ASSEMBLY VISUAL INSPECTION (P.D.L.)					
Post Test Leak Check <2.5 mm H2O	Pass				
Conditions	Value	Units	Port ID	Reading 1 (m)	Reading 2 (m)
Stack pressure	759.48	mmHg	A		Reading 3 (m)
Ref O <sub>2</sub> Value	3	%	B		Average
Moisture Content	1.79	%	C		Port Depth
CO	100	ppm	D		
CO <sub>2</sub>	6	%			
N <sub>2</sub>	83.59	%			
O <sub>2</sub>	10.40	%			
Dry Molecular wt	29.38				
Stack Molecular wt	29.17				
Duct Diameter	1.48	m			
Duct Depth		m			
Duct Width		m			
Area of stack	1.72	m <sup>2</sup>			
Pbar	1010.2	mbar			
Pbar	758	mmHg			
Pilot tube cost	0.834				
Reference Temp	273	K			
Reference Pressure	760	mmHg			
Ambient Temperature	7	°C			
			Enter manually from previous visit		
			Circular Duct		
			Rectangular Duct		
			Duct Diameter (m)		
			Duct Depth (m)		
			Duct Width (m)		
POST-TEST	Traverse Point	Δp Reading (mm H <sub>2</sub> O)	Δp Reading (mm H <sub>2</sub> O)	Δp Average (mm H <sub>2</sub> O)	
PROCEDURE	Reading 1				
	Reading 2				



Comments/Deviations:

Flow and Temperature Measurements

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

SAMPLING LINE: North										
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s	Angle of Swirl °	Δp
1	1.42	3.90	3.90	3.90	3.90	38.23	16.5	6.58	7	1.97
2	1.26	2.50	2.50	2.50	2.50	24.51	16.5	5.27	7	1.58
3	1.04	5.00	5.00	5.00	5.00	49.02	16.7	7.46	7	2.24
4	0.44	5.20	5.20	5.20	5.20	50.98	17.5	7.62	6	2.28
5	0.22	4.30	4.30	4.30	4.30	42.15	17.8	6.93	5	2.07
6	0.06	4.00	4.00	4.00	4.00	39.21	18.0	6.68	6	2.00

SAMPLING LINE: West										
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s	Angle of Swirl °	Δp
1	1.42	3.40	3.40	3.40	3.40	33.33	13.8	6.12	7	1.84
2	1.26	4.00	4.00	4.00	4.00	39.21	13.9	6.64	6	2.00
3	1.04	4.30	4.30	4.30	4.30	42.15	14.1	6.88	7	2.07
4	0.44	5.20	5.20	5.20	5.20	50.98	14.8	7.58	6	2.28
5	0.22	5.20	5.20	5.20	5.20	50.98	15.4	7.59	7	2.28
6	0.06	5.70	5.70	5.70	5.70	55.88	16.0	7.95	7	2.39
Average values	4.4	4.4	4.4	4.4	4.4	43.1	15.9	6.9	6.5	2.1

Duct & Stack Flow Characteristics:


Stack Velocity at stack gas T & P and a wet gas basis	6.94	Average Units	Flow Criteria Measurements
Stack flow @ STP, O <sub>2</sub> (ref) and on a dry gas basis	6.49	m <sup>-3</sup> s <sup>-1</sup>	Is the Flow Ratio 3:1 or less? : 1
Stack flow @ stack gas T & P and on a wet gas basis	11.94	m <sup>-3</sup> s <sup>-1</sup>	Any local negative flow? : NO
Stack flow @ stack gas T & P and on a dry gas basis	11.72	m <sup>-3</sup> s <sup>-1</sup>	Flow $\geq 15^\circ$ of duct axis? : YES
Stack flow @ STP and on a wet gas basis	11.27	m <sup>-3</sup> s <sup>-1</sup>	Minimum Δp detected > 0.5 mmH <sub>2</sub> O : YES
Stack flow @ STP, O <sub>2</sub> (ref) and on a wet gas basis	6.61	m <sup>-3</sup> s <sup>-1</sup>	

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Monitoring Objective	Traverse	Site:	South Hook LNG	Stack ID:	SCV/2F
Date	18/11/2019	Site Team:	MRE/JUG	Time of Survey:	14:00
Tape Measure ID	AS0569				
Laser Measurement Device ID	N/A				
Traverse Pilot Type	S-Type				
Traverse Pilot Tube ID	AS0687				
PITOT ASSEMBLY visual inspection (Btu)					
Pre Test Leak Check <2.5 mm H2O	Pass				
Traverse Manometer Type	Digital				
Traverse Manometer ID	AS0638				
Traverse Manometer Range	240				
Traverse Temp. Readout ID	AS0638				
Traverse Thermocouple ID	AS0451a				
Δp (mmH2O)					
Static Pressure	22.97				
Swirl Test Conducted	Pass				
Protractor ID	Yes				
Post-test Blockage Test (L-type only)	AS0626				
PITOT ASSEMBLY visual inspection (Duct)					
Post Test Leak Check <2.5 mm H2O	Pass				
Conditions	Value	Units	Port ID	Reading 1 (m)	Reading 3 (m)
Stack pressure	759.59	mmHg	A		Average
Ref O <sub>2</sub> Value	3	%	B		Port Depth
Moisture Content	1.5	%	C		
CO	30	ppm	D		
CO <sub>2</sub>	6.5	%			
N <sub>2</sub>	84.00	%			
O <sub>2</sub>	9.50	%			
Dry Molecular wt	29.42				
Stack Molecular wt	29.25				
Duct Diameter	1.48	m	Line ID	Reading 1 (m)	Reading 3 (m)
Duct Depth		m			Duct Diameter
Duct Width		m			
Area of stack	1.72	m <sup>2</sup>			
Pbar	1010.2	mbar			
Pbar	758	mmHg			
Pilot tube coeff	0.834		Outside Side Division		
Reference Temp	273	K			
Reference Pressure	760	mmHg			
Ambient Temperature	6	°C	Enter manually from previous visit		
			Circular Duct		
			Rectangular Duct		
			Duct Diameter (m)		
			Duct Depth (m)		
			Duct Width (m)		
Post-test	Traverse Point	Δp Reading (mm H <sub>2</sub> O)	Δp Reading (mm H <sub>2</sub> O)	Δp Average (mm H <sub>2</sub> O)	
	Reading 1				
	Reading 2				

Comments/Deviations:

Flow and Temperature Measurement



NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

SAMPLING LINE: North								
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s
1	1.42	4.40	4.40	4.40	4.40	43.13	13.2	6.94
2	1.26	5.50	5.50	5.50	5.50	53.92	13.4	7.77
3	1.04	5.20	5.20	5.20	5.20	50.98	13.3	7.55
4	0.44	4.40	4.40	4.40	4.40	43.13	13.1	6.94
5	0.22	4.80	4.80	4.80	4.80	47.05	12.9	7.25
6	0.06	4.70	4.70	4.70	4.70	46.07	12.9	7.17
SAMPLING LINE: West								
Traverse Point	Distance into duct (m)	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Spot Reading mm H <sub>2</sub> O	Δp Average mm H <sub>2</sub> O	Δp Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s
1	1.42	3.80	3.80	3.80	3.80	37.25	13.3	6.45
2	1.26	3.80	3.80	3.80	3.80	37.25	13.2	6.45
3	1.04	4.10	4.10	4.10	4.10	40.19	13.1	6.70
4	0.44	5.10	5.10	5.10	5.10	50.00	13.0	7.47
5	0.22	6.40	6.40	6.40	6.40	62.74	13.0	8.37
6	0.06	5.90	5.90	5.90	5.90	57.84	13.5	8.04
Average values		4.8	4.8	4.8	4.8	47.5	13.2	7.3
Duct / Stack Flow Characteristics:					Average		Units	Flow Criteria Measurements
Stack Velocity at stack gas T & P and a wet gas basis					7.26	m <sup>-1</sup>	Is the Flow Ratio 3:1 or less?	1.3
Stack flow @ STP, O <sub>2</sub> (ref) and on a dry gas basis					7.47	m <sup>3</sup> s <sup>-1</sup>	Any local negative flow?	1
Stack flow @ stack gas T & P and on a wet gas basis					12.48	m <sup>3</sup> s <sup>-1</sup>	Flow <-15° of duct axis?	NO
Stack flow @ stack gas T & P and on a dry gas basis					12.30	m <sup>3</sup> s <sup>-1</sup>	Minimum Δp detected > 0.5 mmH <sub>2</sub> O	YES
Stack flow @ STP, O <sub>2</sub> (ref) and on a wet gas basis					11.90	m <sup>3</sup> s <sup>-1</sup>		
					7.58	m <sup>3</sup> s <sup>-1</sup>		

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.3 - One Minute Averaged Gaseous Emissions Data

**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

Minute Averaged Gaseous Data from South Hook LNG 22nd November 2019 SCV G - Phase One			
Referenced to 273K, 101.3kPa and 3% Oxygen on a Dry Basis			
Time (hh:mm)	Carbon Monoxide (mg/Nm <sup>3</sup> )	Oxides of Nitrogen (mg/Nm <sup>3</sup> )	Oxygen (%)
10:00	579	51.7	12.1
10:01	572	52.4	12.0
10:02	570	52.1	11.9
10:03	570	51.5	12.0
10:04	573	52.2	11.9
10:05	562	52.1	11.9
10:06	559	51.7	11.9
10:07	567	51.8	12.0
10:08	568	51.9	12.1
10:09	570	52.3	12.0
10:10	582	52.0	12.1
10:11	579	51.6	12.1
10:12	578	51.8	12.1
10:13	573	52.2	12.1
10:14	559	52.5	12.0
10:15	564	52.6	12.1
10:16	577	52.9	12.1
10:17	579	52.8	12.0
10:18	570	52.9	12.1
10:19	578	53.0	12.1
10:20	589	52.4	12.1
10:21	581	52.9	12.1
10:22	577	52.9	12.1
10:23	572	53.3	12.0
10:24	569	52.9	11.9
10:25	557	52.7	11.9
10:26	567	52.4	11.9
10:27	568	52.0	12.0
10:28	554	52.1	12.0
10:29	560	51.9	12.1
10:30	558	52.2	12.1
10:31	554	52.5	12.1
10:32	581	52.2	12.1
10:33	575	52.2	12.1
10:34	560	52.6	12.0
10:35	559	52.4	12.0
10:36	566	52.2	12.0
10:37	562	52.4	12.1
10:38	571	52.5	12.1
10:39	576	52.7	12.1
10:40	570	53.3	12.0
10:41	582	52.7	12.1
10:42	576	53.0	12.1
10:43	582	52.9	12.1
10:44	586	53.0	12.1
10:45	582	52.9	12.0
10:46	568	52.9	12.0
10:47	563	52.3	12.0
10:48	551	51.8	12.0
10:49	547	51.7	12.0
10:50	551	51.8	12.0
10:51	570	51.3	12.0
10:52	567	51.5	12.1
10:53	574	51.5	12.1
10:54	580	51.4	12.1
10:55	579	51.9	12.1
10:56	577	52.2	12.1
10:57	587	52.1	12.1
10:58	589	52.1	12.1
10:59	589	52.1	12.1
11:00	591	51.9	12.1
<b>Min Value</b>	<b>547</b>	<b>51.3</b>	<b>11.9</b>
<b>Max Value</b>	<b>591</b>	<b>53.3</b>	<b>12.1</b>
<b>Average</b>	<b>571</b>	<b>52.3</b>	<b>12.0</b>

**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

Minute Averaged Gaseous Data from South Hook LNG 18th November 2019 SCV E - Phase Two Referenced to 273K, 101.3kPa and 3% Oxygen on a Dry Basis			
Time (hh:mm)	Carbon Monoxide (mg/Nm <sup>3</sup> )	Oxides of Nitrogen (mg/Nm <sup>3</sup> )	Oxygen (%)
15:30	240	53.8	10.4
15:31	390	51.5	11.2
15:32	464	50.3	11.7
15:33	480	50.2	11.8
15:34	484	51.4	12.0
15:35	497	50.9	12.0
15:36	464	52.0	11.8
15:37	345	53.3	11.1
15:38	238	55.1	10.4
15:39	173	56.0	9.8
15:40	150	56.0	9.5
15:41	136	56.3	9.4
15:42	140	56.3	9.4
15:43	149	56.1	9.4
15:44	192	56.0	10.0
15:45	297	54.1	10.8
15:46	425	51.8	11.4
15:47	463	51.2	11.9
15:48	481	50.8	12.0
15:49	459	51.9	12.0
15:50	480	52.0	12.0
15:51	423	52.7	11.6
15:52	315	54.4	10.9
15:53	216	55.9	10.2
15:54	166	56.5	9.7
15:55	152	56.5	9.5
15:56	147	56.3	9.4
15:57	147	56.5	9.4
15:58	152	56.6	9.5
15:59	213	56.0	10.2
16:00	328	54.1	10.9
16:01	436	51.9	11.5
16:02	494	50.5	11.9
16:03	484	50.6	12.0
16:04	497	50.7	12.0
16:05	481	51.2	11.9
16:06	378	52.8	11.3
16:07	260	55.1	10.6
16:08	193	56.1	10.0
16:09	158	56.3	9.6
16:10	149	56.3	9.4
16:11	147	56.3	9.4
16:12	145	56.2	9.4
16:13	181	56.1	9.8
16:14	266	54.7	10.5
16:15	389	52.4	11.3
16:16	467	51.0	11.8
16:17	486	50.8	12.0
16:18	483	51.0	12.1
16:19	501	50.7	12.1
16:20	477	51.3	11.8
16:21	338	53.8	11.1
16:22	232	55.7	10.4
16:23	177	56.3	9.8
16:24	148	56.5	9.4
16:25	143	56.4	9.3
16:26	139	56.3	9.4
16:27	153	56.0	9.4
16:28	208	55.5	10.0
16:29	325	53.3	10.8
16:30	460	50.6	11.5
<b>Min Value</b>	<b>136</b>	<b>50.2</b>	<b>9.3</b>
<b>Max Value</b>	<b>501</b>	<b>56.6</b>	<b>12.1</b>
<b>Average</b>	<b>308</b>	<b>53.9</b>	<b>10.7</b>


**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

Minute Averaged Gaseous Data from South Hook LNG 18th November 2019 SCV F - Phase Two			
Referenced to 273K, 101.3kPa and 3% Oxygen on a Dry Basis			
Time (hh:mm)	Carbon Monoxide (mg/Nm <sup>3</sup> )	Oxides of Nitrogen (mg/Nm <sup>3</sup> )	Oxygen (%)
13:00	84.4	56.2	9.5
13:01	80.8	56.0	9.5
13:02	76.2	55.8	9.4
13:03	65.0	55.5	9.1
13:04	59.4	55.3	9.0
13:05	57.5	55.0	9.0
13:06	55.8	55.1	8.9
13:07	52.0	55.0	8.8
13:08	50.3	54.8	8.8
13:09	53.0	54.7	8.8
13:10	56.9	55.1	9.0
13:11	58.9	55.8	9.1
13:12	66.7	55.9	9.2
13:13	76.7	55.9	9.4
13:14	82.6	56.1	9.5
13:15	82.6	56.1	9.5
13:16	86.4	56.1	9.6
13:17	82.8	56.1	9.5
13:18	82.6	56.2	9.5
13:19	78.8	56.0	9.4
13:20	75.1	56.3	9.4
13:21	73.4	56.0	9.3
13:22	65.9	56.0	9.2
13:23	61.9	56.0	9.1
13:24	59.1	56.0	9.1
13:25	56.4	55.6	9.0
13:26	51.3	55.6	8.9
13:27	48.0	55.3	8.8
13:28	49.7	55.1	8.8
13:29	55.6	55.0	8.9
13:30	59.8	55.3	9.0
13:31	67.6	55.5	9.2
13:32	74.9	55.5	9.3
13:33	76.1	55.5	9.3
13:34	79.2	55.5	9.4
13:35	83.0	55.5	9.5
13:36	83.4	55.9	9.5
13:37	81.0	55.6	9.5
13:38	82.9	55.6	9.4
13:39	82.4	55.3	9.4
13:40	79.3	55.3	9.4
13:41	70.9	55.6	9.3
13:42	66.6	55.4	9.2
13:43	61.4	55.4	9.1
13:44	56.6	55.1	9.0
13:45	56.0	55.0	9.0
13:46	55.0	54.7	8.9
13:47	52.2	54.6	8.9
13:48	49.5	54.7	8.8
13:49	48.7	55.0	8.8
13:50	58.0	55.0	8.9
13:51	66.6	55.5	9.1
13:52	70.6	55.5	9.2
13:53	80.9	55.6	9.4
13:54	84.0	55.9	9.5
13:55	82.2	56.3	9.5
13:56	84.9	55.9	9.5
13:57	84.7	56.0	9.5
13:58	82.9	55.9	9.5
13:59	82.7	55.8	9.5
14:00	71.1	55.5	9.3
<b>Min Value</b>	<b>48.0</b>	<b>54.6</b>	<b>8.8</b>
<b>Max Value</b>	<b>86.4</b>	<b>56.3</b>	<b>9.6</b>
<b>Average</b>	<b>68.9</b>	<b>55.5</b>	<b>9.2</b>


NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.4 - Gaseous Emissions Graphical Data


SCV 1G One Minute Averaged Gaseous Emissions Data - 22nd November 2019  
(273.15K, 101.325kPa, 3% O<sub>2</sub>, on a Dry Gas basis) using the NPL Conventional Analysis Package



**SCV 2E One Minute Averaged Gaseous Emissions Data - 18th November 2019  
(273.15K, 101.325kPa, 3% O<sub>2</sub>, on a Dry Gas basis) using the NPL Conventional Analysis Package**



**SCV 2F One Minute Averaged Gaseous Emissions Data - 18th November 2019  
(273.15K, 101.325kPa, 3% O<sub>2</sub>, on a Dry Gas basis) using the NPL Conventional Analysis Package**



NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.5 - Gas Calibration Log

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**GAS CALIBRATION MEASUREMENTS**

Client:	Intertek Ltd	Date:	18/11/2019	SO <sub>2</sub> (ppm)		CO (ppm)	NONOX (ppm)		N <sub>2</sub> O (ppm)
	South Hook LNG	Job Number:	INTK47NOV19	CBSS auto ranges		0-200	0-60	0-100	0-15
Site:	SCVs	Initial N <sub>2</sub> pressure bar:	180	Testo ranges		0-700	0-400	0-100	0-100 (only NO)
Stack ID:	3	Leak check method	Flow method	Testo ranges		0-5000	0-10000	0-10000	0-1000 (only NO)
<b>GAS CALIBRATION LOG - ANALYSER ADJUSTMENT</b>									
Gas Cylinder ID:	SO <sub>2</sub>	CO	NO	O <sub>2</sub>	CO <sub>2</sub>	VOCs	N <sub>2</sub> O		
Initial Reg. Pressure bar	146655SG	146655SG	146655SG	253223SG	253223SG				
Analyser Type / Analyser ID	PG250	AS0208	PG250	AS0208	PG250	AS0208			
Cylinder Concentration:	152.1 ppm	78.3 ppm	ppm	15.30 % Vol	15.24 % Vol	ppm C <sub>3</sub> H <sub>8</sub>			
Span Value:	152.1 ppm	78.3 ppm	ppm	15.30 % Vol	15.24 % Vol	ppm C <sub>3</sub> H <sub>8</sub>			
Analyzer Range: 0 -	200 ppm	100 ppm	ppm	25 % Vol	20 % Vol	ppm C <sub>3</sub> H <sub>8</sub>			
Time			12:30						
Check Zero	ppm	-2.4 ppm	ppm	-0.1 ppm	ppm	-0.13 % Vol	-0.02 % Vol		
Initial Gain	0	0	0	-15	-15	-2			
Time			12:31						
Adjust Zero	ppm	0.0 ppm	0.0 ppm	0.00 % Vol	0.00 % Vol	-3	0.00 % Vol		
Reading	0	0	0	-17	-17	-3	ppm C <sub>3</sub> H <sub>8</sub>		
Final Gain									
Time			12:46				ppm C <sub>3</sub> H <sub>8</sub>		
Check Span	ppm	147.4 ppm	72.4 ppm	ppm	15.36 % Vol	12:37	16.75 % Vol		
Reading	1.419	1.011	1.011	0.990	1.009	1.009	ppm C <sub>3</sub> H <sub>8</sub>		
Initial Gain									
Time			12:47				ppm C <sub>3</sub> H <sub>8</sub>		
Adjust Span	ppm	152.1 ppm	78.3 ppm	ppm	15.30 % Vol	12:38	15.24 % Vol		
Reading	1.463	1.090	1.090	0.980	0.965	0.965	ppm C <sub>3</sub> H <sub>8</sub>		
Final Gain									
Time			12:52				ppm C <sub>3</sub> H <sub>8</sub>		
Check Zero	ppm	1.0 ppm	0.3 ppm	ppm	0.01 % Vol	0.01 % Vol	0.01 % Vol		
Zero Drift	0.0 ppm	1.0 ppm	0.3 ppm	0.0 ppm	0.01 % Vol	0.0 % Vol	0.0 ppm		
Acceptance									
<b>GAS CALIBRATION LOG - SAMPLING SYSTEM CHECK - FLOW METHOD</b>									
Expected Flow:	SO <sub>2</sub>	CO	NO	O <sub>2</sub>	CO <sub>2</sub>	VOCs	N <sub>2</sub> O		
Time	Y/min	0.40 l/min	0.40 l/min	Y/min	0.40 l/min				
Leak check	Reading	Y/min	0.0 l/min	0.0 l/min	Y/min	12:54			
	Pass/Fail	PASS	PASS	PASS	PASS	PASS			
<b>GAS CALIBRATION LOG - DRAFT CHECK</b>									
Span Value:	SO <sub>2</sub>	CO	NO	O <sub>2</sub>	CO <sub>2</sub>	VOCs	N <sub>2</sub> O		
Time	0.0 ppm	152.1 ppm	78.3 ppm	0.0 ppm	15.30 % Vol	0.04 % Vol	0.0 ppm		
Check Zero	Reading	ppm	0.6 ppm	0.0 ppm	0.02 % Vol	0.04 % Vol	0.0 ppm		
Check Span	Time	16:47	16:47	ppm	16:54	16:54	ppm C <sub>3</sub> H <sub>8</sub>		
Zero Drift (%)	Reading	145.7 ppm	76.3 ppm	0.0 ppm	15.23 % Vol	15.09 % Vol	ppm C <sub>3</sub> H <sub>8</sub>		
Span Drift (%)	Time	0.4	0.0	0.1	0.3	0.3	ppm C <sub>3</sub> H <sub>8</sub>		
Acceptance zero	Accept	4.2	2.6	0.5	1.0	1.0	ppm C <sub>3</sub> H <sub>8</sub>		
Acceptance span	Accept	Accept	Accept	Accept	Accept	Accept	ppm C <sub>3</sub> H <sub>8</sub>		
	Correct for drift	Correct for drift	Correct for drift	Correct for drift	Correct for drift	Correct for drift	ppm C <sub>3</sub> H <sub>8</sub>		

CALIBRATION TO BE CARRIED OUT BY OR UNDER THE SUPERVISION OF MCERTS QUALIFIED PERSONNEL WITH LEVEL TWO AND TEF4

Name:	Matthew Ellison
MCERTS ID:	MM-05-082

Personnel Present:	MREJUG
--------------------	--------

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**GAS CALIBRATION MEASUREMENTS**

Client:	Intertek Ltd		Date:	22/11/2019		CBISS auto ranges	Testo ranges	CO (ppm)		NO/NOx (ppm)				
	Site:	South Hook LNG	Job Number:	INTKATNOV19 180				0-200	0-400	0-100	0-250 only NOx			
Stack ID:	SCV's	Initial N <sub>2</sub> pressure bar:	0.700		Flow method:	0-5000		0-10000		0-100 only NOx				
Reference oxygen %	3	Leak check method	<b>GAS CALIBRATION LOG - ANALYSER/ADJUSTMENT</b>											
	SO <sub>2</sub>	CO	NO	NOx	O <sub>2</sub>	CO <sub>x</sub>	VOCs	N <sub>2</sub> O (ppm)						
Gas Cylinder ID:	146655SG	146655SG	146655SG	146655SG	253223SG	253223SG	VOCs	N <sub>2</sub> O (ppm)						
Initial Reg. Pressure bar								N <sub>2</sub> O (ppm)						
Analyser Type/ Analyser ID	PG250	ASD208	PG250	ASD208	PG250	ASD208	PG250	N <sub>2</sub> O (ppm)						
Cylinder Concentration:	ppm	152.1 ppm	ppm	78.3 ppm	ppm	15.30 % Vol	ppm	N <sub>2</sub> O (ppm)						
Span Value:	ppm	152.1 ppm	ppm	78.3 ppm	ppm	15.30 % Vol	ppm	N <sub>2</sub> O (ppm)						
Analyser Range: 0 -	ppm	200 ppm	ppm	100 ppm	ppm	25 % Vol	ppm	N <sub>2</sub> O (ppm)						
Time					09:29			N <sub>2</sub> O (ppm)						
Check Zero	Reading	-0.1 ppm	0.1 ppm	0.1 ppm	ppm	0.04 % Vol	ppm	N <sub>2</sub> O (ppm)						
Initial Gain	Reading	0	0	0	ppm	-14	ppm	N <sub>2</sub> O (ppm)						
Time					09:30			N <sub>2</sub> O (ppm)						
Adjust Zero	Reading	0.0 ppm	0.0 ppm	0.0 ppm	ppm	0.00 % Vol	ppm	N <sub>2</sub> O (ppm)						
Final Gain	Reading	0	0	0	ppm	-14	ppm	N <sub>2</sub> O (ppm)						
Time					09:41			N <sub>2</sub> O (ppm)						
Check Span	Reading	ppm	156.4 ppm	76.8 ppm	ppm	15.07 % Vol	ppm	N <sub>2</sub> O (ppm)						
Initial Gain	Reading	1.505	1.505	1.107	ppm	0.988	ppm	N <sub>2</sub> O (ppm)						
Time					09:42			N <sub>2</sub> O (ppm)						
Adjust Span	Reading	ppm	152.1 ppm	78.3 ppm	ppm	15.30 % Vol	ppm	N <sub>2</sub> O (ppm)						
Final Gain	Reading	1.465	1.465	1.129	ppm	1.003	ppm	N <sub>2</sub> O (ppm)						
Time					09:45			N <sub>2</sub> O (ppm)						
Check Zero	Reading	ppm	0.7 ppm	0.4 ppm	ppm	0.00 % Vol	ppm	N <sub>2</sub> O (ppm)						
Zero Drift	Reading	0.0 ppm	0.7 ppm	0.4 ppm	ppm	0.00 % Vol	ppm	N <sub>2</sub> O (ppm)						
Acceptance	Accept	Accept <2% of range		Accept <2% of range		Accept <2% of range		N <sub>2</sub> O (ppm)						
	SO <sub>2</sub>	CO	NO	NOx	O <sub>2</sub>	CO <sub>x</sub>	VOCs	N <sub>2</sub> O (ppm)						
Expected Flow:	l/min	0.40 l/min	0.40 l/min	Y/min	Y/min	0.40 l/min		N <sub>2</sub> O (ppm)						
Time					09:47	09:47	09:47	N <sub>2</sub> O (ppm)						
Leak check	Reading	0.0 l/min	0.0 l/min	0.0 l/min	ppm	0.01 % Vol	ppm	N <sub>2</sub> O (ppm)						
	Pass/fail	PASS	PASS	PASS	PASS	PASS	PASS	N <sub>2</sub> O (ppm)						
	<b>GAS CALIBRATION LOG - DRIFT CHECK</b>													
Span Value:	0.0 ppm	152.1 ppm	ppm	78.3 ppm	0.0 ppm	15.30 % Vol	ppm	N <sub>2</sub> O (ppm)						
Check Zero	Reading	ppm	1.5 ppm	0.1 ppm	ppm	0.01 % Vol	ppm	N <sub>2</sub> O (ppm)						
Reg Pressure (N <sub>2</sub> )								N <sub>2</sub> O (ppm)						
Time					11:12	11:12	11:20	N <sub>2</sub> O (ppm)						
Check Span	Reading	ppm	149.6 ppm	80.0 ppm	ppm	14.93 % Vol	ppm	N <sub>2</sub> O (ppm)						
Reg Pressure								N <sub>2</sub> O (ppm)						
Zero Drift (%)					1.0	0.1	0.1	N <sub>2</sub> O (ppm)						
Span Drift (%)					1.6	2.2	0.3	N <sub>2</sub> O (ppm)						
Acceptance zero					Accept	Accept	Accept	N <sub>2</sub> O (ppm)						
Acceptance span					Accept	Correct for drift	Accept	N <sub>2</sub> O (ppm)						

CALIBRATION TO BE CARRIED OUT BY OR UNDER THE SUPERVISION OF MCERTS QUALIFIED PERSONNEL WITH LEVEL TWO AND TE4

Name:	Matthew Ellison
MCERTS ID:	MM-Q5-482

Personnel Present:	MRE/JGC
--------------------	---------

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.6 - Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## SCV 1G Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of NOx BS EN 14792:2017 - PG-250 AS0208  
v3.7 Sep-19

Performance characteristics	Value	specification		
		seconds	\$200	
Response time	70			
Logger sampling interval	60	seconds		
Measurement period	61	minutes		
Number of readings in measurement	61			
Standard deviation of repeatability at zero	0.03	% full scale	\$\leq\$2% range	
Standard deviation of reproducibility at span level	0.15	% full scale	\$\leq\$3.2% range	
Standard deviation of reproducibility	0	% full scale	\$\leq\$2% range	
Deviation from linearity/lack of fit	0.35	% full scale	\$\leq\$2% range	
Zero drift	0.21	mg/m³	\$\leq\$5% span value	
Span drift	0	mg/m³	\$\leq\$5% span value	
Influence of sample gas flow	2	% full scale/10	\$\leq\$2% range	
Influence of atmospheric pressure	0.18	% full scale/20K	\$\leq\$2% range	
Influence of ambient temperature	0.4	mg/m³	\$\leq\$15% range	
CO (mg/m³)	300	mg/m³	CO range	100
HCl (mg/m³)	50	mg/m³	HCl range	0
SO₂ (mg/m³)	1000	% vol	SO₂ range (Total)	0
NO (mg/m³)	20	mg/m³	NO range	0
Influence of voltage	0	% full scale/10V	\$\leq\$2% range/10V	0
Influence of vibration	0	% full scale	\$\leq\$2% range	
Losses in the line (leak)	0	% value	\$\leq\$5% of value	
Convector efficiency	95.5	%	295.6	
Uncertainty of calibration gas	1	% value	\$\leq\$2% of value	
Performance characteristic		Uncertainty	Value of uncertainty quantity	mg/m³
Standard deviation of repeatability at zero		U₀	0.00	
Standard deviation of reproducibility at span level		Uₛ	0.04	
Standard deviation of reproducibility		U₀	0.00	
Lack of fit		Uₜ	0.41	
Drift		U₉	0.12	
Influence of sample gas flow		U₈	0.00	
Influence of atmospheric pressure		U₇	0.42	
Influence of ambient temperature		U₆	0.03	
CO (mg/m³)		U₅	0.28	Use largest of sum of all positive or all negative influences
HCl (mg/m³)		U₄	0.00	Criteria
SO₂ (mg/m³)		U₃	0.00	sum <4% range
NO (mg/m³)		U₂	0.05	
Influence of voltage		U₁	0.33 largest	0.16 %
Influence of vibration		U₉	0.00	Value to use for interference uncertainty
Losses in the line (leak)		U₈	0.33	U₉
Convector efficiency		U₇	0.00	
Uncertainty of calibration gas		U₆	0.68	
Uncertainty in factor		U₅	0.13	
Measurement uncertainty				2.18
Combined uncertainty				
Expanded uncertainty	k =	2	0.37	mg/m³
Uncertainty corrected to std cond			1.95	mg/m³
Expanded uncertainty			5.95	mg/m³
Expanded uncertainty			5.85 mg/m³	
Expressed uncertainty			11.19 % value	
Expanded uncertainty			1.82 % ELV	

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of CO BS EN 15058:2017 - PG-250 AS0208  
v4.3 Sep-18

Performance characteristics		Value	Specification		Correction for reference conditions	
			seconds	seconds	ref	O <sub>2</sub> , %
Response time		60	60	5200	measured	12.04
Longer sampling interval		60	60		Uncert	0.37
Measurement period		61	61		Factors	2.01
Number of readings in measurement		0.04	0.04		Uncertainty in factor	1.00
Standard deviation of repeatability at zero		0.1	% full scale		Correction Factor	0.00
Standard deviation of repeatability at span level		0.04	% full scale			0.08
Standard deviation of reproducibility		0.1	% full scale			
Deviation from linearity (lack of fit)		0	% full scale			
Zero drift		0.2	% full scale			
Span drift		2.01	mg/m <sup>3</sup>			
Influence of sample gas flow		3.35	mg/m <sup>3</sup>			
Influence of atmospheric pressure		2	% full scale/10			
Influence of ambient temperature		2	% full scale/3kPa			
NO <sub>2</sub> (mg/m <sup>3</sup> )	30	0.05	% full scale/20K			
NO (mg/m <sup>3</sup> )	20	0.5	mg/m <sup>3</sup>			
CO <sub>2</sub> (% vol)	15	1.0	mg/m <sup>3</sup>			
SO <sub>2</sub> (mg/m <sup>3</sup> )	200	0.5	% vol			
Influence of voltage		-1.3	mg/m <sup>3</sup>			
Influence of vibration		0	% full scale/10V			
Losses in the line (leak)		0	% full scale			
Uncertainty of calibration gas		0	% value			
Uncertainty of callibration gas		1	% value			
Performance characteristic			Uncertainty	Value of uncertainty quantity		
Standard deviation of repeatability at zero		U <sub>zero</sub>		mg/m <sup>3</sup>		
Standard deviation of repeatability at span level		U <sub>span</sub>		0.00		
Standard deviation of reproducibility		U <sub>repro</sub>		0.03		
Lack of fit		U <sub>lack</sub>		0.00		
Drift		U <sub>drift</sub>		0.29		
Influence of sample gas flow		U <sub>flow</sub>		4.05		
Influence of atmospheric pressure		U <sub>atmos</sub>		0.00		
Influence of ambient temperature		U <sub>temp</sub>		0.51		
NO <sub>2</sub> (mg/m <sup>3</sup> )		U <sub>NO2</sub>		0.01		
NO (mg/m <sup>3</sup> )		U <sub>NO</sub>		0.04		
CO <sub>2</sub> (% vol)		U <sub>CO2</sub>		Use largest of sum of all positive or all negative influences		
SO <sub>2</sub> (mg/m <sup>3</sup> )		U <sub>SO2</sub>		Criteria sum >4% range		
Influence of voltage		U <sub>volt</sub>		0.52 all +ve		
Influence of vibration		U <sub>vib</sub>		0.02 all -ve		
Losses in the line (leak)		U <sub>leak</sub>		0.52 largest		
Uncertainty of calibration gas		U <sub>cal</sub>		Value to use for interference uncertainty		
Uncertainty in factor		U <sub>f</sub>		U <sub>in</sub>		
Measurement uncertainty				0.52		
Combined uncertainty				0.21 %		
Expanded uncertainty	k = 2					
Uncertainty corrected to std condns						
Expanded uncertainty						
Expanded uncertainty						

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of O<sub>2</sub> BS EN 14789:2017 - PG-250 AS0208

Sep-18

v3.1

Limit value	N/A	Cal gas conc	15.01 % vol
Measured concentration	12.04 % vol	Full Scale	25.01 % vol

Performance characteristics		Value	specification	Effect of drift
Response time		65 seconds	≤200 seconds	0.04 % vol
Logger sampling interval		60 seconds		
Measurement period		61 minutes		
Number of readings in measurement	61	% vol	<0.2 % vol	
Standard deviation of repeatability at zero	0.05	% vol	<0.2 % vol	
Standard deviation of repeatability at span level	0.1	% vol	<0.2 % vol	
Standard deviation of reproducibility	0.15	% vol	<0.2 % vol	
Deviation from linearity(lack of fit)	0.26	% vol	<0.3 % vol	
Zero drift	0.01	% vol	<0.2 % vol	
Span drift	0.04	% vol	<0.2 % vol	
Influence of sample gas flow	0.2	% vol/10l/h	0.40	0.4 l/min
Influence of atmospheric pressure	-0.07	% vol/3kPa	101.6	101.6 kPa
Influence of ambient temperature	-0.07	% vol/20K	280	280 K
Cross sensitivity	0.14	% vol	<0.4 % vol	
Influence of voltage	0	% vol/10V	<0.2 % vol	
Influence from vibration	0	% vol	<0.2 % vol	
Losses in the line (leak)	0	% value	≤± 2% of value	
Uncertainty of calibration gas	1	% value	≤± 2% of value	
Performance characteristic		Uncertainty	Value of uncertainty quantity	Value of uncertainty quantity
Standard deviation of repeatability at zero		U <sub>10</sub>	0.00	0.00 % vol
Standard deviation of repeatability at span level		U <sub>s</sub>	0.00	0.00 % vol
Standard deviation of reproducibility		U <sub>p</sub>	0.02	0.02 % vol
Lack of fit		U <sub>lt</sub>	0.15	0.15 % vol
Drift		U <sub>dr</sub>	0.02	0.02 % vol
Influence of sample gas flow		U <sub>spes</sub>	0.000	0.000 % vol
Influence of atmospheric pressure		U <sub>pres</sub>	0.02	0.02 % vol
Influence of ambient temperature		U <sub>temp</sub>	-0.01	-0.01 % vol
Cross sensitivity		U <sub>intef</sub>	0.08	0.08 % vol
Influence of voltage		U <sub>volt</sub>	0.00	0.00 % vol
Influence from vibration		U <sub>vib</sub>	0.00	0.00 % vol
Losses in the line (leak)		U <sub>leak</sub>	0.00	0.00 % vol
Uncertainty of calibration gas		U <sub>cal</sub>	0.06	0.06 % vol

Measurement uncertainty		0.18	% vol
Combined uncertainty		0.37	% vol
Expanded uncertainty	2		
Expressed uncertainty expressed with a level of confidence of 95%		0.37 % vol	
Expanded uncertainty expressed with a level of confidence of 95%		3.07 % value	

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for Velocity and Volume Flow Rate Measurement by Pitot tube EN ISO 16911-1

v1.3 Jan-16

Enter data in orange cells only

Constants

Gas constant	8.314 J/(K·mol)
Velocity meas. during calibration	3 m/s
Air density meas. during calibration	1.23 kg/m <sup>3</sup>
DP meas. during calibration	7.9 Pa
Uncertainty of velocity meas. at calibration	0.075 %
Uncertainty of air density meas. at calibration	0.90 %
Pilot coefficient, K	0.834
Expanded uncertainty (95%, k=2) as % of value	0.08
Expanded uncertainty (95%, k=2)	0.08

Duct dimensions	Circular	Rectangular
Diameter	1.48 m	a
Area	1.17 m <sup>2</sup>	b
		Area

All Pressures should be entered in Pascals, Pa

Measurement Point	Atmospheric Pressure, Pa	Stack Pressure, Pa	meas1, Pa	meas2, Pa	meas3, Pa	meas4, Pa	meas5, Pa	Delta P, Pa	Stack Temperature, C	Water Vapour Content, %
1	101281.1535	121.1535	1018895751	121.1535	121.1535	121.1535	121.1535	30.38955751	11.5	11.5
2	101281.1535	101281.1535						33	11.4	11.4
3	101281.1535	121.1535	33.33059211	121.1535	121.1535	121.1535	121.1535	33	11.3	11.3
4	101281.1535	101281.1535						35	11.2	11.2
5	101281.1535	121.1535	36.29121517	121.1535	121.1535	121.1535	121.1535	31	11.2	11.2
6	101281.1535	101281.1535						35	11.2	11.2
7	101281.1535	121.1535	31.36986904	121.1535	121.1535	121.1535	121.1535	22	11.4	11.4
8	101281.1535	101281.1535						23	11.3	11.3
9	101281.1535	121.1535	22.54716252	121.1535	121.1535	121.1535	121.1535	26	11.3	11.3
10	101281.1535	101281.1535						30	11.3	11.3
11	101281.1535	121.1535	30.38955751	121.1535	121.1535	121.1535	121.1535	34	11.4	11.4
12	101281.1535	101281.1535						34	11.4	11.4
Mean	101281.1535	121.1535	34.310909564	121.1535	121.1535	121.1535	121.1535	30.7	11.3	11.3
	101281.1535	101281.1535						30.7	11.3	11.3

$$\rho = \frac{molar\ mass \cdot absolute\ pressure}{R \cdot gas\ temperature}$$

$$Mean\ density = \frac{\rho}{V} = \frac{2.14_p}{\rho}$$

Mean velocity	5.83 m/sec
Standard uncertainty of velocity	0.28 m/sec
Expanded uncertainty in velocity	0.56 m/sec

Flow rate	36120 m <sup>3</sup> /hour	Rectangular duct
Volume flow rate expanded uncertainty	3522 m <sup>3</sup> /hour	m3/hour
Volume flow rate expanded uncertainty	9.8 % of value	#DIV/0!
		% of value

Developed for the STA by NPL, David Butterfield & Chris Dimopoulos  
Checked by: 

Characteristics of pressure sensor used for Delta P	
Repeatability of Delta P transducer	Enter uncertainties as (95% k=2) where relevant
Range of Delta P transducer	-1 % of value
Resolution of Delta P transducer	245.1 Pa
Drift of Delta P transducer	1.96 Pa
Lack of fit of measurement system	0.1 % of range between calibrations
Uncertainty in Delta P transducer	0.1 % of range
Uncertainty in temperature readout system	10.0 Pa
Uncertainty in atmospheric pressure transducer	Enter uncertainties as (95% k=2) where relevant
Uncertainty in duct area measurement	20 % relative
	Enter uncertainties as (95% k=2) where relevant
Water vapour measurement	Water vapour measurement
CO content measurement	CO content measurement
CO <sub>2</sub> content measurement	CO <sub>2</sub> content measurement
O <sub>2</sub> content measurement	O <sub>2</sub> content measurement

Dry gas basis	
CO ppm	CO <sub>2</sub> , %
CO <sub>2</sub> , %	N <sub>2</sub> , %
O <sub>2</sub> , %	O <sub>2</sub> , %
dry molecular wt, g/mol	dry molecular wt, g/mol
stack molecular wt, g/mol	stack molecular wt, g/mol

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## SCV 2E Uncertainty Calculations

**NATIONAL PHYSICAL LABORATORY**  
Continuation Sheet

Uncertainty calculation for gaseous measurement of NOx BS EN 14792:2017 - PG-250 AS0208  
Sep-19  
V3.7

Limit value	107 mg/m <sup>3</sup> (corrected)	Cal gas conc	160.75 mg/m <sup>3</sup>
Measured concentration	30.79 mg/m <sup>3</sup> (101.3kPa, 273K)	Full Scale	205.30 mg/m <sup>3</sup>
Measured concentration	53.65 mg/m <sup>3</sup> (Corrected)		
(NO)/NO <sub>2</sub> ratio	100.00	Gas	NO
		Full Scale	100 ppm
		Cal gas conc	78.3 ppm
		Conversion	2.05

Performance characteristics		Value	70	seconds		specification
Response time			60	seconds		$\leq 200$
Logging sampling interval			61	minutes		
Measurement period			61			
Number of readings in measurement			0.03	% full scale		$\leq 4\%$ range
Standard deviation of repeatability at zero level			0.15	% full scale		$\leq 2\%$ range
Standard deviation of repeatability at open level			0	% full scale		$\leq 3.3\%$ range
Standard deviation of reproducibility			0.35	% full scale		$\leq 2\%$ range
Deviation from linearity/lack of fit			0	mg/m <sup>3</sup>		$\leq 5\%$ span value
Zero drift			0	mg/m <sup>3</sup>		$\leq 5\%$ span value
Span drift			2	% full scale/10		$\leq 2\%$ range
Influence of sample gas flow			2	% full scale/3kPa		$\leq 5\%$ range
Influence of ambient temperature			0.18	% full scale/20K		$\leq 5\%$ range
Influence of ambient pressure			0.4	mg/m <sup>3</sup>		
CO (mg/m <sup>3</sup> )		300	0.4	mg/m <sup>3</sup>		
HCl (mg/m <sup>3</sup> )		50	1.1	mg/m <sup>3</sup>		
SO <sub>2</sub> (mg/m <sup>3</sup> )		1000	0.5	% vol		$\leq 4\%$ range (Total)
NO <sub>x</sub> (mg/m <sup>3</sup> )		20	0.8	mg/m <sup>3</sup>		
Influence of voltage			0	% full scale/10V		$\leq 2\%$ range/10V
Influence from vibration			0	% full scale		$\leq 5\%$ range
Losses in line (leak)			0	% value		$\leq 5\%$ of value
Converger efficiency			95.5	%		$\geq 95\%$ of value
Linearity of calibration gas			1	% value		$\leq 2\%$ of value

Performance characteristic	Uncertainty	Value of uncertainty quantity	mg/m <sup>3</sup>
Standard deviation of repeatability at zero	$U_0$		0.00
Standard deviation of repeatability at span level	$U_{\text{fz}}$		0.04
Standard deviation of reproducibility	$U_0$		0.00
Lack of fit	$U_{\text{lin}}$		0.41
Drift	$U_{\text{Drif}}$		0.00
Influence of sample gas flow	$U_{\text{spres}}$		0.00
Influence of atmospheric pressure	$U_{\text{pres}}$		0.68
Influence of ambient temperature	$U_{\text{temp}}$		0.03
CO (mg/m <sup>3</sup> )	$U_{\text{CO}}$	0.28	
HCl (mg/m <sup>3</sup> )	$U_{\text{HCl}}$	0.00	
SO <sub>2</sub> (mg/m <sup>3</sup> )	$U_{\text{SO}_2}$	0.00	
N <sub>2</sub> O (mg/m <sup>3</sup> )	$U_{\text{N}_2\text{O}}$	0.05	
Influence of voltage	$U_{\text{voltage}}$	-	
Influence from vibration	$U_{\text{vib}}$	0.00	
Losses in the line (leak)	$U_{\text{leak}}$	0.00	
Convector efficiency	$U_{\text{conv}}$	0.80	
Uncertainty of calibration gas	$U_{\text{cal}}$	0.15	

Measurement uncertainty				
Combined uncertainty				
Expanded uncertainty				
Uncertainty corrected to std condns				
Uncertainty expressed with a level of confidence of 95%				
Expanded uncertainty				
Expanded uncertainty				
Expanded uncertainty				

Correction for reference conditions		O <sub>2</sub> , %	Moisture, %	Pressure, kPa	Temperature, K
Factors	ref	3.00	0.00	101.30	273.00
	measured	10.71	0.00	101.30	273.00
	Uncert.	0.37	1.00	0.00	1.00
Uncertainty in factor	1.75	1.00	1.00	0.00	0.00
	Correction Factor	1.75 ul	0.01	0.06	0.00

ranges	min	max	value at calib.
	0.40	0.4	0.4 /min
	10.1	102	10.1 kPa
	2.79	283	280 K
	100	300	0 mcm/m3
	0	0	0 mcm/m3
	0	5	0 nm/s
	0	2	0 mcm/m3
	110	110	110 V

Use largest of sum of all positive or all negative Influences		Criteria	sum <4% range	0.16 %
0.33	all +yes			
0.00	all -yes			
0.33	largest			
Value to use for influence uncertainty				
$U_{\text{infl}}$				0.33

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of CO BS EN 15058:2017 - PG-250 AS0208

Sep-18

v4.3

Limit value		None	mg/m <sup>3</sup> (corrected)	Cal gas conc	190.13 mg/m <sup>3</sup>	Correction for reference conditions	O <sub>2</sub> %	Moisture, %	Pressure, kPa	Temperature, K
<b>Measured concentration</b>		176.23 mg/m <sup>3</sup> (101.3kPa, 273K)	308.22 mg/m <sup>3</sup> (Corrected)	Full Scale	250.00 mg/m <sup>3</sup>	ref	3.00	0.00	101.30	273.00
<b>Measured concentration</b>		176.23 mg/m <sup>3</sup> (101.3kPa, 273K)	308.22 mg/m <sup>3</sup> (Corrected)	Gas	CO	measured	10.71	0.00	101.30	273.00
				Full Scale	200 ppm	Uncert	0.37	1.00	0.00	0.00
				Cal gas conc	152.1 ppm	Factors	1.75	1.00	1.00	1.00
				Conversion	1.25	Uncertainty in factor	0.06	0.01	0.00	0.00
				Correction Factor	1.75 uf	Correction Factor	1.75 uf	0.06		
<b>Performance characteristics</b>		Value	specification							
Response time		60 seconds	\$200 seconds							
Longer sampling interval		60 seconds								
Measurement period		61 minutes								
Number of readings in measurement		61								
Standard deviation of repeatability at zero		0.04	% full scale							
Standard deviation of repeatability at span level		0.1	% full scale							
Standard deviation of reproducibility		0	% full scale							
Deviation from linearity(lack of fit)		0.2	% full scale							
Zero drift		0.8 mg/m <sup>3</sup>								
Span drift		0 mg/m <sup>3</sup>								
Influence of sample gas flow		2 % full scale/10V								
Influence of atmospheric pressure		2 % full scale/20kPa								
Influence of ambient temperature		0.05 % full scale/20K								
NO <sub>2</sub> (mg/m <sup>3</sup> )	30	0.5 mg/m <sup>3</sup>								
NEO (mg/m <sup>3</sup> )	20	1.0 mg/m <sup>3</sup>								
CO <sub>2</sub> (% vol)	15	0.5 % vol								
SO <sub>2</sub> (mg/m <sup>3</sup> )	200	-1.3 mg/m <sup>3</sup>								
Influence of voltage		0 % full scale/10V								
Influence from vibration		0 % full scale								
Losses in the line (leak)		0 % value								
Uncertainty of calibration gas		1 % value								
<b>Performance characteristic</b>		Uncertainty	Value of uncertainty quantity							
Standard deviation of repeatability at zero		U <sub>0</sub>	mg/m <sup>3</sup>							
Standard deviation of repeatability at span level		U <sub>rs</sub>	0.00							
Standard deviation of reproducibility		U <sub>rp</sub>	0.03							
Lack of fit		U <sub>lf</sub>	0.00							
Drift		U <sub>drift</sub>	0.29							
Influence of sample gas flow		U <sub>gas</sub>	0.46							
Influence of atmospheric pressure		U <sub>atmos</sub>	0.00							
Influence of ambient temperature		U <sub>temp</sub>	0.82							
NO <sub>2</sub> (mg/m <sup>3</sup> )		U <sub>NO2</sub>	0.01							
NEO (mg/m <sup>3</sup> )		U <sub>NEO</sub>	0.04	Use largest of sum of all positive or all negative influences						
CO <sub>2</sub> (% vol)		U <sub>CO2</sub>	0.06	Criteria						
SO <sub>2</sub> (mg/m <sup>3</sup> )		U <sub>SO2</sub>	0.42	sum <4% range						
Influence of voltage		U <sub>Volt</sub>	0.46	0.52 all +ve						
Influence from vibration		U <sub>vib</sub>	0.00	0.02 all -ve						
Losses in the line (leak)		U <sub>leak</sub>	0.00	0.52 largest						
Uncertainty of calibration gas		U <sub>cal</sub>	0.88	Value to use for interference uncertainty						
Uncertainty in factor		uf	11.22	U <sub>lin</sub>						

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of O<sub>2</sub> BS EN 14789:2017 - PG-250 AS0208

Sep-18

v3.1

Limit value	N/A	Cal gas conc	15.01 % vol
Measured concentration	10.71 % vol	Full Scale	25.00 % vol

Performance characteristics	Value	specification
Response time	0.5 seconds	≤200 seconds
Logger sampling interval	60 seconds	
Measurement period	61 minutes	
Number of readings in measurement	61	
Standard deviation of repeatability at zero	0.05 % vol	<0.2 % vol
Standard deviation of repeatability at span level	0.1 % vol	<0.2 % vol
Standard deviation of reproducibility	0.15 % vol	<0.2 % vol
Deviation from linearity(fack of fit)	0.26 % vol	<0.3 % vol
Zero drift	0.02 % vol	<0.2 % vol
Span drift	0.07 % vol	<0.2 % vol
Influence of sample gas flow	0.2 % vol/10/lit	<0.2 % vol
Influence of atmospheric pressure	0.2 % vol/3kPa	<0.2 % vol
Influence of ambient temperature	-0.07 % vol/20K	<0.5 % vol
Cross sensitivity	0.14 % vol	<0.4 % vol
Influence of voltage	0 % vol/10V	<0.2 % vol
Influence from vibration	0 % vol	<0.2 % vol
Losses in the line (leak)	0 % value	≤± 2% of value
Uncertainty of calibration gas	1 % value	≤± 2% of value
Performance characteristic	Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero	U <sub>0.0</sub>	0.00 % vol
Standard deviation of repeatability at span level	U <sub>s</sub>	0.00 % vol
Standard deviation of reproducibility	U <sub>p</sub>	0.02 % vol
Lack of fit	U <sub>fit</sub>	0.15 % vol
Drift	U <sub>drift</sub>	0.04 % vol
Influence of sample gas flow	U <sub>spec</sub>	0.000 % vol
Influence of atmospheric pressure	U <sub>atmos</sub>	0.03 % vol
Influence of ambient temperature	U <sub>temp</sub>	-0.01 % vol
Cross sensitivity	U <sub>cross</sub>	0.08 % vol
Influence of voltage	U <sub>volt</sub>	0.00 % vol
Influence from vibration	U <sub>vib</sub>	0.00 % vol
Losses in the line (leak)	U <sub>leak</sub>	0.00 % vol
Uncertainty of calibration gas	U <sub>cal</sub>	0.05 % vol

Measurement uncertainty	Combined uncertainty	Expanded uncertainty
Combined uncertainty	0.19 % vol	0.37 % vol
Expanded uncertainty	2	
Expressed with a level of confidence of 95%		0.37 % vol
Expressed with a level of confidence of 95%		3.50 % value

# NATIONAL PHYSICAL LABORATORY

## Continuation Sheet

### Uncertainty calculation for Velocity and Volume Flow Rate Measurement by Pitot tube EN ISO 16911-1

v1.3 Jan-16

Enter data in orange cells only

Constants

Gas constant	8.314 J/(K·mol)
Velocity meas. during calibration	3 m/s
Air density meas. during calibration	1.23 kg/m <sup>3</sup>
DP meas. during calibration	7.9 Pa
Uncertainty of velocity meas. at calibration	0.075 %
Uncertainty of air density meas. at calibration	0.22 %
Uncertainty of DP meas. at calibration	0.395 %
Pitot coefficient, K	0.834
Expanded uncertainty (95%, k=2) as % of value	9.0 %
Expanded uncertainty (95%, k=2)	0.08
Duct dimensions	
Circular Diamter	1.48 m
Area	2.77 m <sup>2</sup>
Area	0.0 m <sup>2</sup>

Characteristics of pressure sensor used for Delta P

Repeatability of Delta P transducer	Enter uncertainties as (95% k=2) where relevant 1 % of value
Range of Delta P transducer	245.1 Pa
Resolution of Delta P transducer	1.96 Pa
Drift of Delta P transducer	0.1 % of range between calibrations
Drift of fit of measurement system	0.1 % of range
Uncertainty in Delta P transducer	10.0 Pa
Uncertainty in temperature readout system	Enter uncertainties as (95% k=2) where relevant 1 °C
Uncertainty in atmospheric pressure transducer	Water vapour measurement CO content measurement CO <sub>2</sub> content measurement O <sub>2</sub> content measurement
Uncertainty in duct area measurement	0.8 % 20 % relative 6 % relative 10 % relative 6 % relative

All Pressures should be entered in Pascals, Pa

Measurement Point	Atmospheric Pressure, Pa	Stack Pressure, Pa	Static Pressure, Pa	meas1, Pa	meas2, Pa	meas3, Pa	meas4, Pa	meas5, Pa	Delta P, Pa	Stack Temperature, C	Water Vapour Content, %	Water Vapour Content, %	Dry gas basis
1	101230.15	210.915	36.23214977						38.23214977	16.5	1.79	100	6
2	101230.15	210.915	24.50778331						25	16.5	1.79	100	6
3	101230.15	210.915	49.0557663						49	17.5	1.79	100	6
4	101230.15	210.915	50.97619969						51	17.5	1.79	100	6
5	101230.15	210.915	42.1533959						42	17.8	1.79	100	6
6	101230.15	210.915	29.21246513						39	18	1.79	100	6
7	101230.15	210.915	33.33059211						33	13.8	1.79	100	6
8	101230.15	210.915	39.2124613						39	13.9	1.79	100	6
9	101230.15	210.915	42.1533959						42	14.1	1.79	100	6
10	101230.15	210.915	50.97619969						51	14.8	1.79	100	6
11	101230.15	210.915	50.97619969						51	15.4	1.79	100	6
12	101230.15	210.915	55.8775735						56	16	1.79	100	6
Mean	101230.15	210.91	43.1	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	43.1	15.9	1.8	100.0	60
	101230.15	210.91	43.1	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	43.1	15.9	1.8	100.0	60

$$\rho = \frac{molar\ mass \cdot absolute\ pressure}{R \cdot gas\ temperature}$$

$$Mean\ density = \frac{\rho}{V} = \frac{2.14_p}{\rho}$$

$$Standard\ uncertainty\ of\ velocity = \sqrt{\frac{2.14_p}{\rho}}$$

$$Expanded\ uncertainty\ in\ velocity = \sqrt{2.14_p^2 + \frac{2.14_p^2}{\rho}}$$

Mean velocity	6.94 m/sec
Standard uncertainty of velocity	0.32 m/sec
Expanded uncertainty in velocity	0.64 m/sec

Flow rate	42992 m <sup>3</sup> /hour
Volume flow rate expanded uncertainty	4054 m <sup>3</sup> /hour
Volume flow rate expanded uncertainty	9.4 % of value

Developed for the STA by NPL, David Butterfield & Chris Dimopoulos

Checked by: *[Signature]*

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## SCV 2F Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of NOx BS EN 14792:2017 - PG-250 AS2028  
v3.7 Sep-19

Limit value	1.07 mg/m <sup>3</sup> (corrected)	Cal gas conc	160.75 mg/m <sup>3</sup>
Measured concentration	36.37 mg/m <sup>3</sup> (101.3kPa, 273K)	Full Scale	205.30 mg/m <sup>3</sup>
Measured concentration	55.55 mg/m <sup>3</sup> (Corrected)		
NO/NO <sub>2</sub> ratio	100.00	Gas	NO
		Full Scale	100 ppm
		Cal gas conc	78.31 ppm
		Conversion	2.05

Performance characteristics		Value	specification	
			70 seconds	≤200
Response time				
Longer sampling interval	seconds	60		
Measurement period	minutes	61		
Number of readings in measurement	61			
Standard deviation of repeatability at zero	% full scale	0.03	≤±2% range	
Standard deviation of repeatability at span level	% full scale	0.15	≤±2% range	
Standard deviation of reproducibility	% full scale	0	≤±3.2% range	
Deviation from linearity(lack of fit)	% full scale	0.35	≤±2% range	
Zero drift	mg/m <sup>3</sup>	0	≤5% span value	
Span drift	mg/m <sup>3</sup>	0	≤5% span value	
Influence of sample gas flow	% full scaled@10	2	≤±2% range	
Influence of atmospheric pressure	% full scale/3kPa	2	≤±2% range	
Influence of ambient temperature	% full scale/20°C	0.18	≤±25% range	
CO (mg/m <sup>3</sup> )	mg/m <sup>3</sup>	0.4	CO range	
HCl (mg/m <sup>3</sup> )	mg/m <sup>3</sup>	1.1	HCl range	
SO <sub>2</sub> (mg/m <sup>3</sup> )	mg/m <sup>3</sup>	0.5	SO <sub>2</sub> range (Total)	
N <sub>2</sub> O (mg/m <sup>3</sup> )	mg/m <sup>3</sup>	0.8	N <sub>2</sub> O range	
Influence of voltage	0	0	Voltage	
Influence from vibration	0	0	Voltage	
Losses in the line (leak)	% value	0	≤±2% range	
Converter efficiency	% value	95.5	≤±5% of value	
Uncertainty of calibration gas	% value	1	≤±5% of value	

Performance characteristic		Uncertainty	Value of uncertainty quantity	mg/m <sup>3</sup>
Standard deviation of repeatability at zero		U <sub>10</sub>		0.00
Standard deviation of repeatability at span level		U <sub>1a</sub>		0.04
Standard deviation of reproducibility		U <sub>10</sub>		0.00
Lack of fit		U <sub>1H</sub>		0.41
Drift		U <sub>1Dr</sub>		0.00
Influence of sample gas flow		U <sub>1gas</sub>		0.00
Influence of atmospheric pressure		U <sub>1pres</sub>		0.68
Influence of ambient temperature		U <sub>1temp</sub>		0.03
CO (mg/m <sup>3</sup> )		U <sub>1CO</sub>	0.28	Use largest of sum of all positive or all negative influences
HCl (mg/m <sup>3</sup> )		U <sub>1HCl</sub>	0.00	Criteria
SO <sub>2</sub> (mg/m <sup>3</sup> )		U <sub>1SO2</sub>	0.00	sum <4% range
N <sub>2</sub> O (mg/m <sup>3</sup> )		U <sub>1N2O</sub>	0.05	0.33 largest
Influence of voltage		U <sub>1V</sub>	0.00	Value to use for interference uncertainty
Influence from vibration		U <sub>1vb</sub>	0.00	U <sub>ini</sub>
Losses in the line (leak)		U <sub>1lk</sub>	0.00	0.33
Converter efficiency		U <sub>1conv</sub>	0.95	
Uncertainty of calibration gas		U <sub>1cal</sub>	0.18	
Uncertainty in factor		U <sub>f</sub>	1.78	
Measurement uncertainty				
Combined uncertainty				
Expanded uncertainty				
K =	2			
Uncertainty corrected to std cond				
Expanded uncertainty	expressed with a level of confidence of 95%		5.31 mg/m <sup>3</sup>	
Expanded uncertainty	expressed with a level of confidence of 95%		9.57 % value	
Expanded uncertainty	expressed with a level of confidence of 95%		2.41 % ELV	

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of CO BS EN 15053:2017 - PG-250 AS0208  
v4.3 Sep-18

Performance characteristics		Value	mg/m <sup>3</sup> (corrected)	Cal gas conc	190.13 mg/m <sup>3</sup>	Correction for reference conditions	O <sub>2</sub> %	Moisture %	Pressure, kPa	Temperature, K
Measured concentration		None	45.10 mg/m <sup>3</sup> (101.3kPa, 273K)	Full Scale	250.00 mg/m <sup>3</sup>	ref	3.00	0.00	101.30	273.00
Measured concentration		68.87 mg/m <sup>3</sup> (Corrected)				measured	9.21	0.00	101.30	273.00
						Uncert	0.37	1.00	0.00	0.00
						Uncertainty in factor	1.53	1.00	1.00	1.00
						Correction Factor	1.53	0.01	0.05	0.05
						uf				
Performance characteristics		Value	60	seconds	\$200	Specification				
Response time		60	seconds							
Longer sampling interval		60	seconds							
Measurement period		61	minutes							
Number of readings in measurement		61								
Standard deviation of repeatability at zero		0.04		% full scale		\$±2% range				
Standard deviation of repeatability at span level		0.1		% full scale		\$±1% range				
Standard deviation of reproducibility		0		% full scale		\$±3% range				
Deviation from linearity(lack of fit)		0.2		% full scale		\$±2% range				
Zero drift		0.8		mg/m <sup>3</sup>		<5% span value				
Span drift		0		mg/m <sup>3</sup>		<5% span value				
Influence of sample gas flow		2		% full scale/10V		\$±1% range				
Influence of atmospheric pressure		2		% full scale/3kPa		\$±2% range				
Influence of ambient temperature		0.05		% full scale/20K		\$±2% range				
NO <sub>2</sub> (mg/m <sup>3</sup> )		30	0.5	mg/m <sup>3</sup>		NO <sub>2</sub> range	0	4	0.4 l/min	280 K
NO <sub>x</sub> (mg/m <sup>3</sup> )		20	1.0	mg/m <sup>3</sup>		N <sub>2</sub> O range	0	2	0 mg/m <sup>3</sup>	
CO <sub>2</sub> (% vol)		15	0.5	% vol		CO <sub>2</sub> range (Total)	10	15	0 % vol	
SC <sub>2</sub> (mg/m <sup>3</sup> )		200	-1.3	mg/m <sup>3</sup>		SC <sub>2</sub> range	0	5	0 mg/m <sup>3</sup>	
Influence of voltage		0		% full scale/10V		Voltage	110	110	110 l/V	
Influence from vibration		0		% full scale		\$±2% range				
Losses in the line (leak)		0		% value		\$± 5% of value				
Uncertainty of calibration gas		1		% value		\$± 2% of value				
Performance characteristic				Uncertainty		Value of uncertainty quantity	mg/m <sup>3</sup>			
Standard deviation of repeatability at zero				U <sub>00</sub>			0.00			
Standard deviation of repeatability at span level				U <sub>0s</sub>			0.03			
Standard deviation of reproducibility				U <sub>0p</sub>			0.00			
Lack of fit				U <sub>0ll</sub>			0.29			
Drift				U <sub>0dr</sub>			0.46			
Influence of sample gas flow				U <sub>0ps</sub>			0.00			
Influence of atmospheric pressure				U <sub>0aps</sub>			0.82			
Influence of ambient temperature				U <sub>0temp</sub>			0.01			
NO <sub>2</sub> (mg/m <sup>3</sup> )				U <sub>0n2</sub>			0.04		Use largest of sum of all positive or all negative influences	
NO <sub>x</sub> (mg/m <sup>3</sup> )				U <sub>0nx</sub>			0.06		Criteria sum >4% range	
CO <sub>2</sub> (% vol)				U <sub>0co2</sub>			0.42		0.02 all -ve	
SC <sub>2</sub> (mg/m <sup>3</sup> )				U <sub>0sc2</sub>			-0.02		0.52 largest	
Influence of voltage				U <sub>0v</sub>			0.00		Value to use for interference uncertainty	
Influence from vibration				U <sub>0vib</sub>			0.00		U <sub>in</sub>	
Losses in the line (leak)				U <sub>0leak</sub>			0.00		0.52	
Uncertainty of calibration gas				U <sub>0cal</sub>			0.23			
Uncertainty in factor				U <sub>f</sub>			2.21			
Measurement uncertainty										
Combined uncertainty										
k =		2								
Uncertainty corrected to std conds										
Expanded uncertainty										
Expressed uncertainty										
Expressed uncertainty										

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for gaseous measurement of O<sub>2</sub> BS EN 14789:2017 - PG-250 AS0208  
Sep-18  
v3.1

Limit value	N/A	Cal gas conc	15.01 % vol
Measured concentration	9.21 % vol	Full Scale	25.00 % vol

Performance characteristics		Value	specification	Effect of drift
Response time		65 seconds	≤200 seconds	0.06 % vol
Logger sampling interval		60 seconds		
Measurement period		61 minutes		
Number of readings in measurement	61	% vol	<0.2 % vol	
Standard deviation of repeatability at zero	0.05	% vol	<0.2 % vol	
Standard deviation of repeatability at span level	0.1	% vol	<0.2 % vol	
Standard deviation of reproducibility	0.15	% vol	<0.2 % vol	
Deviation from linearity(lack of fit)	0.26	% vol	<0.3 % vol	
Zero drift	0.02	% vol	<0.2 % vol	
Span drift	0.07	% vol	<0.2 % vol	
Influence of sample gas flow	0.2	% vol/10/lit	0.40	0.4 l/min
Influence of atmospheric pressure	0.2	% vol/3kPa	101.1	101.1 kPa
Influence of ambient temperature	-0.07	% vol/20K	27.9	28.0 K
Cross sensitivity	0.14	% vol	<0.4 % vol	
Influence of voltage	0	% vol/10V	110	110 V
Influence from vibration	0	% vol	<0.2 % vol	
Losses in the line (leak)	0	% value	≤± 2% of value	
Uncertainty of calibration gas	1	% value	≤± 2% of value	
Performance characteristic		Uncertainty	Value of uncertainty quantity	% vol
Standard deviation of repeatability at zero		U <sub>10</sub>		0.00
Standard deviation of repeatability at span level		U <sub>s</sub>		0.00
Standard deviation of reproducibility		U <sub>p</sub>		0.02
Lack of fit		U <sub>lt</sub>		0.15
Drift		U <sub>dr</sub>		0.04
Influence of sample gas flow		U <sub>spec</sub>		0.000
Influence of atmospheric pressure		U <sub>atmos</sub>		0.03
Influence of ambient temperature		U <sub>temp</sub>		-0.01
Cross sensitivity		U <sub>interf</sub>		0.08
Influence of voltage		U <sub>volt</sub>		0.00
Influence from vibration		U <sub>vib</sub>		0.00
Losses in the line (leak)		U <sub>leak</sub>		0.00
Uncertainty of calibration gas		U <sub>cal</sub>		0.05
Measurement uncertainty				
Combined uncertainty			0.18	% vol
Expanded uncertainty	2		0.37	% vol
Expanded uncertainty expressed with a level of confidence of 95%			0.37 % vol	
Expanded uncertainty expressed with a level of confidence of 95%			4.00 % value	

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

Uncertainty calculation for Velocity and Volume Flow Rate Measurement by Pitot tube EN ISO 16911-1

Jan-16

v1.3

Enter data in orange cells only

Constants

$\rho = \frac{molar\ mass \cdot absolute\ pressure}{R \cdot gas\ temperature}$

Characteristics of pressure sensor used for Delta P

Gas constant	8314 J/(K·mol)	Enter uncertainties as (95% k=2) where relevant
Repeatability of Delta P transducer	3 m/s	1 % of value
Range of Delta P transducer	1.23 kg/m <sup>3</sup>	245 Pa
Air density meas. during calibration	1.23 kg/m <sup>3</sup>	1.96 Pa
DP meas. during calibration	7.9 Pa	0.1 % of range between calibrations
Uncertainty of velocity meas. at calibration	2.2 %	0.1 % of range
Uncertainty of DP meas. at calibration	0.075 %	10.0 Pa
Pilot coefficient, K	0.834	Uncertainty in stack pressure readout system
Expanded uncertainty (95%, k=2) as % of value	9.0 %	Enter uncertainties as (95% k=2) where relevant
Expanded uncertainty (95%, k=2)	0.08	Water vapour measurement
Duct dimensions		CO content measurement
Circular		CO <sub>2</sub> content measurement
Diameter	1.48 m	Water vapour measurement
Area	2.7 m <sup>2</sup>	CO content measurement
	0.0 m <sup>2</sup>	CO <sub>2</sub> content measurement

All Pressures should be entered in Pascals, Pa

Measurement Point	Atmospheric Pressure, Pa	Stack Pressure, Pa	Static Pressure, Pa	meas2, Pa	meas3, Pa	meas4, Pa	meas5, Pa	Delta P, Pa	Stack Temperature, C	Water Vapour Content, %	Water Vapour Content, %	CO, ppm	CO <sub>2</sub> , ppm	N <sub>2</sub> , %	O <sub>2</sub> , %	Dry gas basis
1	101245.3	1013370743	225.3	43.13370743	54				43.13370743	13.2		1.5	30	6.5	84.0	9.5
2	101020	101245.3	225.3	53.9173429						13.4		1.5	30	6.5	84.0	9.5
3	101020	101245.3	225.3	50.9761969						13.3		1.5	30	6.5	84.0	9.5
4	101020	101245.3	225.3	43.13370743						13.1		1.5	30	6.5	84.0	9.5
5	101020	101245.3	225.3	47.05485556						12.9		1.5	30	6.5	84.0	9.5
6	101020	101245.3	225.3	46.07464203						12.9		1.5	30	6.5	84.0	9.5
7	101020	101245.3	225.3	37.25183824						13.3		1.5	30	6.5	84.0	9.5
8	101020	101245.3	225.3	37.75183824						13.2		1.5	30	6.5	84.0	9.5
9	101020	101245.3	225.3	40.19277283						13.1		1.5	30	6.5	84.0	9.5
10	101020	101245.3	225.3	49.99588816						13		1.5	30	6.5	84.0	9.5
11	101020	101245.3	225.3	67.739383808						13		1.5	30	6.5	84.0	9.5
12	101020	101245.3	225.3	57.285889392						13.5		1.5	30	6.5	84.0	9.5
Mean	101020	101245	225.3	47.5	#DIV/0!	#DIV/0!	#DIV/0!	47.5		13.2		1.5	30.0	6.5	84.0	9.5

$\rho = \frac{molar\ mass \cdot absolute\ pressure}{R \cdot gas\ temperature}$

Mean density

1.245 kg/m<sup>3</sup>

$$Velocity = K \cdot \sqrt{\frac{2 \cdot \Delta P}{\rho}}$$

Mean velocity

7.26 m/sec

Standard uncertainty of velocity

0.33 m/sec

4.6 % of value

Expanded uncertainty in velocity

0.67 m/sec

9.2 % of value

Flow rate	44963 m <sup>3</sup> /hour	Rectangular duct
Volume flow rate expanded uncertainty	4208 m <sup>3</sup> /hour	0 m3/hour
Volume flow rate expanded uncertainty	0.67 m/sec	#DIV/0!
Volume flow rate expanded uncertainty	9.4 % of value	#DIV/0!

Developed for the STA by NPL David Butterfield & Chris Dimopoulos

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

## 2.2.7 - Calculations Used in Reporting Results

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

### Nozzle Selection

For isokinetic sampling, the pressure difference of the orifice meter must equal the pressure difference of the Pitot tube pressure multiplied by the K-factor. Where:

$$K = \text{Constant} \times C_p^2 \times D_n^4 \times DH_{@} \times \left( \frac{M_d}{M_s} \right) \left( \frac{1 - B_{wm}}{1 - B_{ws}} \right)^2 \left( \frac{T_m + 273}{T_s + 273} \right) \left( \frac{P_s}{P_m} \right)$$

$$DH = K \times Dp$$

Where:-

Constant: is a constant dependent on the units used to measure the nozzle ( $8.038 \times 10^{-5}$  for mm)

$D_n$  the nozzle diameter mm

$DH_{@}$  a constant dependent on the sampler control box orifice and gas meter

$B_{ws}$  the percent water vapour in the emission as a fraction i.e.  $12\% = 0.12$

$B_{wm}$  the percentage water vapour in the air around the meter box often assumed to be zero

$C_p$  Pitot tube coefficient dependent on the Pitot tube type

$T_m$  the meter temperature in  $^{\circ}\text{C}$

$T_s$  the stack temperature in  $^{\circ}\text{C}$

$P_s$  the stack pressure

$P_m$  the meter pressure

$M_d$  dry gas molecular weight

$M_s$  apparent stack gas molecular weight

$DH$  pressure drop across the orifice (mm water)

$DP$  differential Pitot pressure (mm water)

From this the correct nozzle size can be determined.

$$D_n = \sqrt{\left( \frac{\text{Constant.Q}_m \cdot P_m}{(T_m + 273)C_p} \right) \left( \frac{1 - B_{wm}}{1 - B_{ws}} \right) \sqrt{\frac{(T_s + 273)M_s}{(P_s \cdot (\Delta P)_{avg})}}}$$

Where the Constant = 0.6071 Metric

$Q_m$  = Orifice flow rate normally 21.2 actual  $\text{lmin}^{-1}$

$$= K_m \sqrt{\frac{(T_m + 273)\Delta H}{P_m M_m}}$$

Where  $K_m$  = Orifice meter coefficient

$$K_m = Q_m \sqrt{\frac{P_m M_m}{\Delta H(T_m + 273)}} = \text{Const} \sqrt{\frac{1}{\Delta H_{@}}}$$

Where Const = 183.7 metric

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**Moisture Determination Calculations**

These calculations are based at 273K and 101.325kPa

To calculate moisture the following equation is used:

$$B_{ws} = \frac{0.001245 \times W_I \times 100}{(0.001245 \times W_I) + 0.359V_m \left( \frac{P_b + \frac{\Delta H_{avg}}{13.6}}{(T_m + 273)} \right)}$$

**Particulate Concentration  $C_s$  in stack Gases**

At 273K and 101.325kPa and dry gas

$$C_s = \frac{W_t}{V_m} \times \frac{T_m + 273}{273} \times \frac{760}{\left( P_b + \frac{\Delta H_{avg}}{13.6} \right)} \times 1000 \quad \text{mg/Nm}^3$$

**Oxygen Concentration Correction  $C_{oxy}$  to Particulate concentration**

$$C_{oxy} = C \times \frac{(20.9 - \%O_2 \text{ref})}{(20.9 - \%O_2 \text{Meas})} \quad \text{mg/Nm}^3$$

**Dry Molecular Weight of gases**

$$M_D = 0.44(\%CO_2) + 0.32(\%O_2) + 0.28(\%CO + \%N)$$

**Stack Molecular Weight of gases**

$$M_s = 0.18(B_{ws}) + \frac{M_d}{100}(100 - B_{ws})$$

**Stack Gas Velocity**

$$(V_s)_{avg} = 34.96 \times C_p \times \sqrt{(\Delta P)_{avg}} \sqrt{\frac{T_s + 273}{P_s M_s}} \quad \text{m/s}$$

**Mass Emission Rate  $M_R$**

$$M_R = \frac{C_m \times (V_s)_{avg} \times A \times 3600}{10^6} \quad \text{kg/hr}$$

NATIONAL PHYSICAL LABORATORY  
Continuation Sheet

**IsoKinicity**

$$I = \frac{2.12 \times 10^8 \times V_m \times Y \times \left( P_b + \left( \frac{\Delta H_{avg}}{13.6} \right) \right) \left( \frac{273 + T_s}{273 + T_m} \right)}{\Theta P_s \pi D_n^2 (V_s) avg (100 - B_{ws})} \%$$

$W_I$	= the weight change of the impingers during sampling in g
$V_m$	= volume of dry gas sample in litres at temperature of the meter box
$B_{ws}$	= the percent water vapour in the emission
$Q$	= length of time sampling in minutes
$Y$	= Gas Meter Calibration correction factor
$V_s$	= Velocity of stack gas m/s
$C_M$	= measured concentration of particulate matter ( $\text{mg}/\text{m}^3$ )
$T_m$	= average temperature at dry gas meter ( $^\circ\text{C}$ )
$P_b$	= atmospheric pressure (mmHg)
$\%O_{2ref}$	= % oxygen at standard temperature & pressure
$\%O_{2Meas}$	= % oxygen measured on site
$C_p$	= Pitot tube coefficient
$DP$	= mean differential Pitot pressure drop (mm $H_2O$ )
$DH$	= mean orifice pressure drop (mm $H_2O$ )
$D_s$	= diameter of stack (m)
$D_n$	= Nozzle diameter (mm)
$T_s$	= stack temperature ( $^\circ\text{C}$ )
$M_d$	= molecular weight of dry stack gas
$B_W$	= moisture fraction
$P_s$	= stack pressure (mmHg)
$A$	= duct c.s.a. ( $\text{m}^2$ )
$M_s$	= molecular weight of wet stack gas
$M_d$	= molecular weight of dry stack gas
$W_t$	= total weight of particulate matter (g)