



Test Report



This test report is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This test report may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

**OXIDES OF NITROGEN (AS NO₂) AND CARBON MONOXIDE COMPLIANCE TESTING
AT SOUTH HOOK LNG TERMINAL ON BEHALF OF INTERTEK LIMITED**

Permit Number: **XP3538LD**

Operator Name: **South Hook LNG (on behalf of Intertek Limited)**

Installation Name: **South Hook LNG Terminal**

Dates of Monitoring Visit: **11th October & 5th December 2016**

Contract Reference: **E08040221**

Client Contact: **Adrian Walsh**

Client Organisation: **Intertek Limited**

Address: **Unit 14 - Waterston Trading Estate
Main Road, Waterston
Milford Haven
SA73 3SL**

Monitoring Organisation: **National Physical Laboratory**

Address: **Hampton Road
Teddington
Middlesex
TW11 0LW**

Date of Report: **21st December 2016**

Report Author: **Matthew Ellison**

Reference: XP3538LD/INTERTEK/SHLNG/OCT2016/SCV/PPC/Q4/V1

Report Approver: Kevin Blakley
MCERTS Registration: MM-03-317
Level & TEs Held: Level 2, TE1, TE2, TE3 & TE4
Signature:

NPL Authorised Signatory
Name: Mr R Robinson (for NPLML)
Signature:

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

CONTENTS PAGE

Part One: Executive Summary		PAGE
1.1	Monitoring Objectives	3
1.2.1	SCV 1H (Phase One) Monitoring Results	4
1.2.2	SCV 2A (Phase Two) Monitoring Results	5
1.3	Operating Information	6
1.4	Monitoring Deviations	7
1.5	Conclusions	7
1.6	References	7
Part Two: Supporting Information		
Appendix One		
2.1.1	Emissions Testing Personnel Details	9
2.1.2	Emissions Testing Procedures	9
2.1.3	Equipment Checklist Reference	10
2.1.4	Data Capture Location Reference	10
Appendix Two		
2.2.1	Stack Diagram	12
2.2.2	Flow Criteria Measurements	14
2.2.3	One Minute Averaged Gaseous Emissions Data	19
2.2.4	Gaseous Emissions Graphical Data	22
2.2.5	Gas Calibration Log	25
2.2.6	Uncertainty Calculations	28
2.2.7	Calculations Used in Reporting Results	39

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

1.1 Monitoring Objectives

NPL were awarded a contract by Intertek Limited to carry out emissions compliance testing at South Hook LNG plant near Milford Haven, Pembrokeshire. The scope of work includes carrying out emissions monitoring on the SCV flue stacks phases one and two.

There are a total of fifteen SCV units on the South Hook site that require monitoring, eight on phase one and seven on phase two. Two units were measured for oxides of nitrogen and carbon monoxide.

In addition, oxygen measurements were taken to allow a correction to reference conditions. Water vapour and flow measurements were also taken to determine the moisture content and velocity of the flue gas.

NPL carried out the monitoring visit between the 11th October and 5th December 2016. The report documents the results obtained.

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

1.2.1 SCV H (Phase One) Monitoring Results

Client: Intertek
Site: South Hook LNG
Emission Point: SCV H (Phase One)

Field	Units	Oxides of Nitrogen (as NO _x)	Carbon Monoxide	Oxygen	Moisture
Emission Limit Value	mg/m ³ , Reference Conditions	107	N/A	N/A	N/A
Periodic Monitoring Result	Reference Conditions	44.4	135	9.3	2.5
Uncertainty (95% Confidence Level)	Reference Conditions	13.8	12.4	0.5	N/A
Average Stack Flow	Units	mg/m ³	mg/m ³	%Vol/Vol	%Vol/Vol
Reference Conditions	m ³ /s at Reference Conditions	8.6			
Date		273K, 101.3 kPa, 3% Oxygen on a dry gas basis			
Sample Period	dd/mm/yyyy	05/12/2016			
	From hh:mm	12:00			
	To hh:mm	13:00			
Monitoring Method		BS EN 14792	BS EN 15058	BS EN 14789	BS EN 14790
Accreditation		UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS
Process Status	Load (Tonnes/Hour)	150			

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

1.2.2 SCV A (Phase Two) Monitoring Results

Client: Intertek
Site: South Hook LNG
Emission Point: SCV A (Phase Two)

Field	Units	Oxides of Nitrogen (as NO ₂)	Carbon Monoxide	Oxygen	Moisture
Emission Limit Value	mg/m ³ , Reference Conditions	107	N/A	N/A	N/A
Periodic Monitoring Result	Reference Conditions	49.4	127	8.8	2.4
Uncertainty (95% Confidence Level)	Reference Conditions	13.3	10.8	0.5	N/A
Average Stack Flow	Units	mg/m ³	mg/m ³	%Vol/Vol	%Vol/Vol
Reference Conditions	m ³ /s at Reference Conditions	7.7			
Date	dd/mm/yyyy	273K, 101.3 kPa, 3% Oxygen on a dry gas basis			
Sample Period	From hh:mm	11/10/2016			
	To hh:mm	12:00			
Monitoring Method		13:00			
Accreditation		BS EN 14792	BS EN 15058	BS EN 14789	BS EN 14790
Process Status	Load (Tonnes/Hour)	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS	UKAS & MCERTS
		160			

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

1.3 **Operating Information**

South Hook LNG Terminal, situated in Pembrokeshire South West Wales, is a regasification plant for natural gas. The LNG is transported in specially designed vessels to Milford Haven where it is then transferred into storage tanks, where it awaits reheating and distribution into the UK National Grid.

A total of fifteen SCVs have been built across two phases. Each one has at least two five inch ports installed, as set out in BS EN 13284-1. The ports are located approximately thirteen metres from the base of the stack and can be accessed by ladders that lead to a permanent platform. The sampling platform has lighting, toe boards and handrails. There is sufficient parking on the roadway at the base of the stacks for the mobile laboratory and provision of 240v 16A power supply outlets.

Continuous or Batch Process?	Each SCV operates on a batch process. The number of SCVs operating and the load at which they are set depend upon the required site output.				
What part of the batch process was sampled? (If applicable)	The periodic monitoring is carried out once an SCV has been brought online to the operators required load and has stabilised. This loading remains constant through the one hour test.				
What fuel was used during monitoring? (If applicable)	A small amount of LNG is used as fuel to heat a volume of water. This heat exchange warms up the LNG allowing it to be passed out into the National Grid system.				
What feedstock was used during monitoring? (If applicable)	N/A				
What was the load during monitoring?	Emission Point		Load (Tonnes per Hour)		
	SCV 1H		150		
	SCV 2A		160		
What abatement systems are present? Were they in operation?	Each SCV uses water injection to abate NOx emissions. The system was in operation during the periodic monitoring of each SCV.				
Periodic monitoring results and corresponding CEM values	Emission Point	Substance Monitored	CEM Result	Periodic Monitoring Result	Units
	SCV 1H	Oxides of Nitrogen	48.8	44.4	mg/Nm ³
	SCV 2A	Oxides of Nitrogen	54.8	49.4	mg/Nm ³

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

1.4 **Monitoring Deviations**

Were all substances in the monitoring objectives monitored? If not why?	All substances set out in the objective were monitored.
Were all substances monitored in accordance to the relevant method? If not why?	All substances set out in the monitoring objectives were measured in accordance to the relevant standards.
Were there any other issues relevant to the monitoring results?	No.

1.5 **Conclusions**

NPL carried out the emissions monitoring at South Hook LNG over a period of two weeks. Two SCVs were monitored for the required determinands.

The client is aware of BS EN 15259 and the requirement to carry out homogeneity testing. These tests were carried out on SCVs 1H and 2A, both units passed.

Reference - XP3538LD/INTERTEK/SHLNG/JULY2011/SCV/HOMOGENEITY.

1.6 **References**

1. STA – Risk Assessment Guide: Industrial-emission monitoring – Version 12 - June 2012.
2. Environmental Agency - Manual Stack emission monitoring performance standard for Organisations – Version 8.1 - October 2012.
3. Environmental Agency – M1 Technical Guidance Note – Sampling requirements for stack emission monitoring – Version 6 –January 2010.
4. Environmental Agency – M2 Technical Guidance Note – Monitoring of stack emissions to air – Version 10 – October 2013.
5. Environment Agency - MID 14792 - Stationary source emissions - Determination of mass concentration of nitrogen oxides (NOx) - Reference Method: Chemiluminescence - Version 1.1 - May 2011.
6. Environment Agency - MID 15259 - Stationary source emissions - Requirements for the measurement sections and sites and for the measurement objective, plan and report - Version 1.2 - January 2012.
7. Guidance on Assessing Measurement Uncertainty in Stack Emissions Monitoring, by Pullen J and Robinson R, Source Testing Association, Quality Guidance Note QGN1.

APPENDIX ONE

2.1.1 Emissions Testing Personnel Details

Name	Role	MCERTS Number	Certification Level & Expiry Dates					
			Level 1	Level 2	TE1	TE2	TE3	TE4
Matthew Ellison	Team Leader	MM-05-682	N/A	Sep-2018	Sep-2018	Sep-2018	Dec-2018	Sep-2018
Bob Lipscombe	Site Assistant	MM-07-879	Sep-2018	N/A	N/A	N/A	N/A	N/A
Kevin Blakley	Team Leader	MM-03-317	N/A	Sep-2018	Mar-2021	Mar-2021	Sep-2018	May-2019

2.1.2 Emissions Testing Procedures

Determinand	Instrumental Methods			Manual Methods		
	NOx	CO	O ₂	H ₂ O	Stack Flow	Temperature
SRM Standard	BS EN 14792	BS EN 15058	BS EN 14789	BS EN 14790	BS ISO 16911	BS ISO 16911
Instrument	PG-250 SRM	PG-250 SRM	PG-250 SRM	Impingers with DGM	Pitot	Type K Thermocouple
Instrument Serial No.	AS0450	AS0450	AS0450	N/A	AS0638	AS0638
Principle	Chemiluminescence	NDIR	Zirconia & Paramagnetic	Gravimetric	Flow	Temperature
Operational Range	0 - 250 ppm	0 - 200 ppm	0 - 25%	N/A	N/A	N/A
Certified Range	0 - 125 mg/m ³	0 - 75 mg/m ³	0 - 25%	N/A	N/A	N/A
Uncertainty	10%	6%	6%	20%	N/A	N/A
NPL Procedure	QPAS B 538	QPAS B 538	QPAS B 538	QPAS B 540	QPAS B 567	QPAS B 567
UKAS Accreditation	YES	YES	YES	YES	YES	YES

The sample gas was extracted from the stack via a chemically inert heated line and drawn through a conditioning unit. All moisture in the gas sample was removed and cooled down to 4°C before it was pumped down another line to the NPL Mobile Source Emissions Laboratory and analysed using a Horiba PG-250 SRM. The entire sampling system had been leak tested before testing was carried out to ensure no dilution of the sample gas.

Moisture tests were conducted using a mini impinger setup, a dry gas meter and a small 110v pump. The entire setup was transported to the sampling platform and operated at the same time as the instrumental methods. A flow and temperature profile was carried as well using an S-type pitot and K-type thermocouple, both of which were built into a 2m traversing probe.

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

The conventional analyser zero and span settings were checked before and after each test run using zero grade nitrogen (ex BOC), a suitable gas mixture (BOC beta gas standard), traceable to national reference standards and a gas dilution system. The certified accuracies of the gas standards are listed below: -

Component	Sample Location	Cylinder ID	Certified Amount	Certified Uncertainty
Carbon Monoxide	Phase 1 & 2 SCVs	102400	153.1 ppm	1%
		164171	155.4 ppm	1%
Nitrogen Oxide		242465	149.3 ppm	1%
		164171	151.7 ppm	1%
Oxygen		102488	14.93%	1%
		132170	14.99%	1%

These measurement uncertainties are expressed at a 95% level of confidence.

The ranges of the Horiba PG-250 SRM analyser used for the testing and the values used to calibrate the instrument before and after the testing are listed below: -

Gaseous Components	Horiba Ranges	Calibrated Values
Carbon Monoxide	0 - 200 ppm	153.1 ppm
Oxides of Nitrogen (as NO ₂)	0 - 250 ppm	149.3 ppm
Oxygen	0 - 25%	14.93%

A leak test was conducted before testing to confirm hydraulic integrity of the gaseous sampling system. This was conducted by sending nitrogen down the entire sample line and ensuring a zero reading was obtained.

The electrical volt/millivolt outputs from the PG-250 SRM gas analyser was collected by data logger software on a PC and downloaded to digital media at the end of each day. Under the program used during the tests, the software records and stores individual readings either every 1 or 10 seconds. From this data, the logger can perform a series of calculations to output 1 minute averaged measurement on a volume/volume or mass/volume basis. After each 1 minute average has been established the data buffer is reset and the process repeats.

2.1.3 Equipment Checklist Reference

See workfile INTK34OCT16/Equipment Checklist.

2.1.4 Data Capture Location Reference


All data collected using the NPL computer system on site is backed up at the end of each day onto a memory stick. When the team returns to site this information is then uploaded onto the NPL servers and stored in the relevant location for that job. The link below is where the South Hook emissions data is stored:

P:\Stack Emissions Team\South Hook LNG - Intertek\INTK34OCT16\7. Monitoring Record Sheets

APPENDIX TWO

2.2.1 - Stack Diagram

NATIONAL PHYSICAL LABORATORY
Continuation Sheet



Position	1	2	3	4	5	6	7	8	9	10
% of Diameter	3.2	8.2	14.6	22.6	34.2	65.8	77.4	85.4	91.8	96.8
Insertion, m	0.05	0.12	0.22	0.33	0.51	0.97	1.15	1.26	1.36	1.43
Insertion plus offset, m	0.15	0.22	0.32	0.43	0.61	1.07	1.25	1.36	1.46	1.53

Notes - The circular stack diameter was measured as 1.48 metres, whilst the port offset (distance between the edge of the stack to the end of the port) was measured as 23 centimetres. Access to the top of the stack was by a series of three ladders. Each one had a resting platform in-between with a self closing gate installed. The main platform itself is a permanent structure with toe boards, railings and self closing gate.

2.2.2 - Flow Criteria Measurements

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Monitoring Objective	Traverse	Stack ID:	South Hook LNG	SCV 1H
Date	05/12/2016	Site Team:	RPL/KCB	Time of Survey:
Time of Survey:	11:30	Diagram of Sample Location:		
Traverse Measure ID	AS0659			
Laser Measurement Device ID	N/A			
Traverse Pilot Type	S-Type			
Traverse Pilot Tube ID	AS0211			
Pilot Assembly Visual Inspection (Pre)	Pass			
Pilot Leak Check @ ..	Pass			
Traverse Manometer Type	Digital			
Traverse Manometer ID	AS0638			
Damping Device used	No			
Traverse Temp. Readout ID	AS0638			
Traverse Thermocouple ID	AS0451a			
Static Pressure	23.73	Comments/Deviations:		
Static Pressure	23.73	Velocity & Temperature Measurements		
Swirl Test Conducted	Pass			
Protractor ID	Yes			
Protractor ID	AS0626			
Post-Test Blockage Test (L-Type only)	Pass			
Post-Test Pilot Leak Check @ ..	Pass			
Pilot Assembly Visual Inspection (Post)	Pass			
Conditions	Value	Units	Duct Dimensions	
Stack pressure	754.99	mmHg	Port ID	Reading 1 (m)
Ref O ₂ Value	3	%	A	Reading 2 (m)
Moisture Content	2.5	%	B	Reading 3 (m)
CO	135	ppm	C	Average
CO ₂	6.8	%	D	
N ₂	83.89	%	Line ID	Reading 1 (m)
O ₂	9.30	%	A	Reading 2 (m)
Dry Molecular wt	29.46		B	Reading 3 (m)
Stack Molecular wt	29.18		Average	Duct Diameter
Duct Diameter	1.46	m	Line ID	Reading 1 (m)
Duct Depth	1.72	m	A	Reading 2 (m)
Duct Width	1.72	m ²	B	Reading 3 (m)
Area of stack	1004	mbar	C	Average
Pbar	753	mmHg	D	
Pbar	0.819	mmHg	Outside Side Division	Reading 1 (m)
Pilot tube coeft	273	K	Average	Reading 2 (m)
Reference Temp	760	mmHg	Static Measurement	Reading 3 (m)
Reference Pressure	12	°C	Measurement Line	Δp (mmH2O)
Ambient Temperature			Measurement Line	Reading 1 (Reading 2)
			Rectangular Duct	Reading 1
			Circular Duct	Reading 2
			1.48	24.10
			Duct Diameter (m)	23.30
			Duct Depth (m)	
			Duct Width (m)	

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

SAMPLING LINE: North											
Traverse Point	Distance into duct (m)	Ap Spot Reading mm H ₂ O	Ap Spot Reading mm H ₂ O	Ap Spot Reading mm H ₂ O	Ap Average mm H ₂ O	Ap Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s	Angle of Swirl °	∇Ap	
1	1.42	5.2	5.2	5.2	5.2	50.98	21.3	7.55	4	2.28	
2	1.26	5.8	5.8	5.8	5.8	56.86	21.3	7.97	4	2.41	
3	1.04	7.0	7.0	7.0	7.0	68.82	21.2	8.76	3	2.65	
4	0.44	7.9	7.9	7.9	7.9	77.44	21.2	9.30	3	2.81	
5	0.22	7.5	7.5	7.5	7.5	73.52	21.1	9.06	4	2.74	
6	0.06	6.5	6.5	6.5	6.5	63.72	21.0	8.44	3	2.55	
SAMPLING LINE: West											
Traverse Point	Distance into duct (m)	Ap Spot Reading mm H ₂ O	Ap Spot Reading mm H ₂ O	Ap Spot Reading mm H ₂ O	Ap Average mm H ₂ O	Ap Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s	Angle of Swirl °	∇Ap	
1	1.42	4.8	4.8	4.8	4.8	47.05	20.9	7.25	3	2.19	
2	1.26	5.2	5.2	5.2	5.2	50.98	20.9	7.54	4	2.28	
3	1.04	6.6	6.6	6.6	6.6	64.70	20.9	8.50	3	2.57	
4	0.44	8.2	8.2	8.2	8.2	80.39	20.8	9.47	4	2.86	
5	0.22	7.4	7.4	7.4	7.4	72.54	20.6	8.99	4	2.72	
6	0.06	9.1	9.1	9.1	9.1	89.21	20.4	9.97	3	3.02	
Average values		6.8	6.8	6.8	6.8	66.3	21.0	8.6	3.5	2.6	
Duct / Stack Flow Characteristics:											
Stack Velocity at stack gas T & P and a wet gas basis		8.57								Units	
Stack flow @ STP, O ₂ (ref) and on a dry gas basis		8.59								ms ⁻¹	1.4
Stack flow @ stack gas T & P and on a wet gas basis		14.73								m ³ s ⁻¹	:1
Stack flow @ STP and on a dry gas basis		14.37								m ³ s ⁻¹	NO
Stack flow @ STP and on a wet gas basis		13.59								m ³ s ⁻¹	YES
Stack flow @ STP, O ₂ (ref) and on a wet gas basis		8.81								m ³ s ⁻¹	YES
Flow Criteria Measurements											
Is the Flow Ratio 3:1 or less?										YES	
Any local negative flow?										NO	
Flow <15° of duct axis?										YES	
Minimum Δp detected > 0.5 mmH ₂ O										YES	

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

SAMPLING LINE: North										
Traverse Point	Distance into duct (m)	Ap Spot Reading mm H ₂ O	Ap Spot Reading mm H ₂ O	Ap Spot Reading mm H ₂ O	Ap Average mm H ₂ O	Ap Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s	Angle of Swirl °	∇Ap
1	1.42	3.2	3.2	4.8	3.2	31.37	21.1	5.88	5	1.79
2	1.26	4.8	4.8	4.7	4.8	47.05	21.1	7.20	6	2.19
3	1.04	4.7	4.7	6.2	4.7	46.07	21.0	7.12	6	2.17
4	0.44	6.2	6.2	5.3	6.2	60.78	21.0	8.18	5	2.49
5	0.22	5.3	5.3	5.3	5.3	51.96	21.0	7.56	5	2.30
6	0.06	5.3	5.3	4.9	5.3	51.96	20.8	7.56	6	2.30
Average values										
		4.9	4.9	4.9	4.9	48.0	20.8	7.2	5.3	2.2
SAMPLING LINE: West										
Traverse Point	Distance into duct (m)	Ap Spot Reading mm H ₂ O	Ap Spot Reading mm H ₂ O	Ap Spot Reading mm H ₂ O	Ap Average mm H ₂ O	Ap Pa	Stack Temp Ts °C	Velocity @ stack gas T&P on wet gas basis m/s	Angle of Swirl °	∇Ap
1	1.42	3.9	3.9	3.5	3.9	38.23	20.9	6.49	5	1.97
2	1.26	3.5	3.5	3.7	3.5	34.31	20.0	6.13	4	1.87
3	1.04	3.7	3.7	5.6	3.7	36.27	20.9	6.32	5	1.92
4	0.44	5.6	5.6	6.4	5.6	54.90	20.9	7.77	5	2.37
5	0.22	6.4	6.4	6.2	6.4	62.74	20.8	8.31	6	2.53
6	0.06	6.2	6.2	4.9	6.2	60.78	20.6	8.17	5	2.49
Average values										
		4.9	4.9	4.9	4.9	48.0	20.8	7.2	5.3	2.2
Duct / Stack Flow Characteristics:										
Stack Velocity at stack gas T & P and a wet gas basis										
Stack flow @ STP, O ₂ (ref) and on a dry gas basis										
Stack flow @ stack gas T & P and on a wet gas basis										
Stack flow @ stack gas T & P and on a dry gas basis										
Stack flow @ STP and on a wet gas basis										
Stack flow @ STP, O ₂ (ref) and on a wet gas basis										
Flow Criteria Measurements										
Is the Flow Ratio 3:1 or less?										
Any local negative flow?										
Flow <16° of duct axis?										
Minimum Δp detected > 0.5 mmH ₂ O										

2.2.3 - One Minute Averaged Gaseous Emissions Data

NATIONAL PHYSICAL LABORATORY
Continuation Sheet


Minute Averaged Gaseous Data from South Hook LNG 5th December 2016			
SCV H - Phase One			
Referenced to 273K, 101.3kPa and 3% Oxygen on a Dry Basis			
Time (hh:mm)	Carbon Monoxide (mg/Nm ³)	Oxides of Nitrogen (mg/Nm ³)	Oxygen (%)
12:00	131	44.9	9.3
12:01	141	44.8	9.4
12:02	140	45.1	9.5
12:03	143	45.2	9.5
12:04	157	45.0	9.6
12:05	168	45.1	9.7
12:06	176	45.1	9.8
12:07	188	44.8	9.8
12:08	190	45.0	9.9
12:09	185	45.1	9.9
12:10	172	45.0	9.8
12:11	167	44.8	9.6
12:12	157	44.8	9.6
12:13	150	44.4	9.5
12:14	135	44.5	9.4
12:15	128	44.3	9.3
12:16	126	44.2	9.2
12:17	120	44.4	9.2
12:18	115	44.4	9.1
12:19	108	44.3	9.1
12:20	107	44.3	9.0
12:21	109	44.2	9.0
12:22	104	44.4	9.0
12:23	107	44.2	9.0
12:24	101	44.3	9.0
12:25	104	44.3	8.9
12:26	109	44.3	9.0
12:27	116	44.3	9.1
12:28	120	44.3	9.2
12:29	128	44.4	9.3
12:30	131	44.6	9.3
12:31	135	44.1	9.3
12:32	144	44.7	9.4
12:33	150	44.6	9.5
12:34	152	44.7	9.5
12:35	153	44.7	9.5
12:36	168	44.9	9.7
12:37	171	44.9	9.7
12:38	170	44.7	9.7
12:39	164	44.8	9.7
12:40	156	44.6	9.6
12:41	146	44.3	9.5
12:42	141	44.2	9.4
12:43	138	44.1	9.4
12:44	135	44.0	9.3
12:45	123	43.8	9.2
12:46	121	43.6	9.2
12:47	117	43.9	9.1
12:48	114	44.2	9.1
12:49	107	44.0	9.0
12:50	111	43.8	9.0
12:51	111	43.7	9.0
12:52	111	43.7	9.0
12:53	108	43.5	9.0
12:54	108	43.7	9.0
12:55	117	43.8	9.0
12:56	117	43.8	9.0
12:57	119	43.8	9.1
12:58	127	44.0	9.2
12:59	141	44.3	9.3
13:00	151	44.0	9.4
Min Value	101	43.5	8.9
Max Value	190	45.2	9.9
Average	135	44.4	9.3

NATIONAL PHYSICAL LABORATORY
Continuation Sheet


Minute Averaged Gaseous Data from South Hook LNG 11th October 2016 SCV A - Phase Two Referenced to 273K, 101.3kPa and 3% Oxygen on a Dry Basis			
Time (hh:mm)	Carbon Monoxide (mg/Nm ³)	Oxides of Nitrogen (mg/Nm ³)	Oxygen (%)
12:00	134	49.6	8.9
12:01	134	49.6	8.9
12:02	134	49.4	8.9
12:03	130	49.3	8.9
12:04	128	49.2	8.9
12:05	122	49.5	8.8
12:06	117	49.3	8.8
12:07	124	49.2	8.7
12:08	122	49.4	8.8
12:09	122	49.4	8.7
12:10	119	49.2	8.7
12:11	122	49.3	8.7
12:12	122	49.1	8.7
12:13	119	49.3	8.7
12:14	124	49.5	8.8
12:15	116	49.3	8.8
12:16	120	49.4	8.8
12:17	125	49.3	8.8
12:18	122	49.4	8.8
12:19	130	49.3	8.9
12:20	128	49.5	8.9
12:21	126	49.5	8.9
12:22	132	49.2	8.9
12:23	131	49.1	8.9
12:24	122	49.4	8.8
12:25	126	49.4	8.8
12:26	122	49.3	8.7
12:27	122	49.3	8.7
12:28	122	49.4	8.7
12:29	119	49.3	8.7
12:30	126	49.2	8.7
12:31	124	49.2	8.7
12:32	124	49.4	8.7
12:33	125	49.4	8.8
12:34	126	49.2	8.8
12:35	134	49.3	8.9
12:36	134	49.2	8.9
12:37	145	49.2	9.0
12:38	141	49.3	9.0
12:39	142	49.2	9.0
12:40	135	49.3	8.9
12:41	136	49.2	8.9
12:42	138	49.1	8.9
12:43	125	49.3	8.9
12:44	126	49.5	8.8
12:45	127	49.3	8.8
12:46	125	49.6	8.8
12:47	124	49.5	8.7
12:48	129	49.6	8.8
12:49	130	49.5	8.8
12:50	130	49.2	8.8
12:51	127	49.4	8.8
12:52	130	49.3	8.8
12:53	129	49.5	8.8
12:54	133	49.4	8.8
12:55	131	49.6	8.9
12:56	130	49.6	8.9
12:57	135	49.6	8.9
12:58	125	49.7	8.8
12:59	122	49.6	8.8
13:00	123	49.6	8.8
Min Value	116	49.1	8.7
Max Value	145	49.7	9.0
Average	127	49.4	8.8

2.2.4 - Gaseous Emissions Graphical Data

SCV 1H One Minute Averaged Gaseous Emissions Data - 5th December 2016
(273.15K, 101.325kPa, 3% O₂, on a Dry Gas basis) using the NPL Conventional Analysis Package



SCV 2A One Minute Averaged Gaseous Emissions Data - 11th October 2016
(273.15K, 101.325kPa, 3% O₂, on a Dry Gas basis) using the NPL Conventional Analysis Package



2.2.5 - Gas Calibration Log

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

GAS CALIBRATION MEASUREMENTS

Client:	Intertek Ltd	Date:	05/12/2016	Horiba ID:	AS0450
Site:	South Hook LNG	Job Number:	INTK34OCT16	FID ID:	N/A
Stack ID:	SCV 1H	Mobile Lab ID:	VU64 VTY	Sonimix ID:	N/A
Reference oxygen %	3	Nitrogen cylinder ID:	Zero Grade	Initial N ₂ pressure bar	150

GAS CALIBRATION LOG - DIRECT CALIBRATION

		SO ₂	CO	NOx	O ₂	CO ₂	VOCs
Gas Cylinder ID:			164171	164171	132170	132170	
Initial Reg. Pressure bar			160	160	30	30	
Cylinder Concentration:			155.4 ppm	151.7 ppm	14.99 % Vol	12.02 % Vol	
Span Value:			155.4 ppm	151.7 ppm	14.99 % Vol	12.02 % Vol	
Analyser Range:0 -			200 ppm	250 ppm	25 % Vol	25 % Vol	
Check Zero	Time	10:41					
	Reading		-1.0 ppm	0.1 ppm	0.23 % Vol	-0.08 % Vol	
	Initial Gain		-1.0	-2.0	-4.00	2.00	
Adjust Zero	Time	10:43					
	Reading		0.0 ppm	0.0 ppm	-0.01 % Vol	0.00 % Vol	
	Final Gain		3	-1	1	0	
Check Span	Time		10:58	10:58	10:47	10:47	
	Reading		134.2 ppm	156.2 ppm	14.75 % Vol	12.48 % Vol	
	Initial Gain		1.251	1.005	0.854	0.954	
Adjust Span	Time		11:01	11:01	10:48	10:48	
	Reading		155.9 ppm	149.3 ppm	14.99 % Vol	12.02 % Vol	
	Final Gain		1.267	1.042	0.872	0.937	
Check Zero	Time	11:05					
	Reading		-0.1 ppm	0.1 ppm	-0.15 % Vol	0.00 % Vol	

GAS CALIBRATION LOG - SYSTEM CAL (Including heated line and chiller unit)

		SO ₂	CO	NOx	O ₂	CO ₂	VOCs
Span Value:			155.4 ppm	151.7 ppm	15.0 % Vol	12.0 % Vol	
Check Zero	Time	11:10					
	Reading		0.4 ppm	0.2 ppm	-0.04 % Vol	0.00 % Vol	
	Pass/fail		PASS	PASS	PASS	PASS	
Check Span	Time		11:20	11:20	11:14	11:14	
	Reading		155.9 ppm	149.6 ppm	14.85 % Vol	11.82 % Vol	
	Response Time/s		81	92	86	102	
Pass/fail			PASS	PASS	PASS	PASS	

GAS CALIBRATION LOG - POST CAL (Including heated line and chiller unit)

		SO ₂	CO	NOx	O ₂	CO ₂	VOCs
Span Value:			155.4 ppm	151.7 ppm	14.99 % Vol	12.02 % Vol	
Check Zero	Time	17:08					
	Reading		-4.3 ppm	0.1 ppm	-0.11 % Vol	0.06 % Vol	
Check Span	Time		17:16	17:16	17:12	17:12	
	Reading		152.6 ppm	150.8 ppm	14.74 % Vol	11.80 % Vol	
	Reg Pressure		140	140	15	15	
Zero Drift check @2x repeatability zero			-0.05 ppm	0.1 ppm	-0.14 % Vol	0.00 % Vol	
Acceptance			Accept	Accept	Accept	Accept	
Zero Drift (%)			2.8	0.1	0.7	0.5	
Span Drift (%)			1.8	0.6	1.7	1.8	
Acceptance zero			Correct for drift	Accept	Accept	Accept	
Acceptance span			Accept	Accept	Accept	Accept	

CALIBRATION TO BE CARRIED OUT BY OR UNDER THE SUPERVISION OF MCERTS QUALIFIED PERSONNEL WITH LEVEL TWO AND TE4

Name:	Kevin Blakley
MCERTS ID:	MM-03-317

Personnel Present:	KCB/RPL
--------------------	---------

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

GAS CALIBRATION MEASUREMENTS

Client:	Intertek	Date:	11/10/2016	Horiba ID:	AS0450
Site:	South Hook LNG	Job Number:	INTK34OCT16	FID ID:	N/A
Stack ID:	SCV 2A	Mobile Lab ID:	VU64 VTY	Sonimix ID:	N/A
Reference oxygen %	3	Nitrogen cylinder ID:	Zero Grade	Initial N ₂ pressure bar	150

GAS CALIBRATION LOG - DIRECT CALIBRATION

		SO ₂	CO	NO _x	O ₂	CO ₂	VOCs
Gas Cylinder ID:			102400	242465	102488	102488	
Initial Reg. Pressure bar			170	170	60	60	
Cylinder Concentration:			153.1 ppm	149.3 ppm	14.93 % Vol	12.06 % Vol	
Span Value:			153.1 ppm	149.3 ppm	14.93 % Vol	12.06 % Vol	
Analyser Range:0 -			200 ppm	250 ppm	25 % Vol	20 % Vol	
Check Zero	Time	10:00					
	Reading		-1.7 ppm	-0.1 ppm	0.25 % Vol	-0.07 % Vol	
	Initial Gain		-6.0	-2.0	-3.00	1.00	
Adjust Zero	Time	10:02					
	Reading		0.3 ppm	0.0 ppm	0.00 % Vol	0.00 % Vol	
	Final Gain		1	-2	3	0	
Check Span	Time		10:13	10:13	10:22	10:22	
	Reading		156.3 ppm	146.6 ppm	14.84 % Vol	12.53 % Vol	
	Initial Gain		1.142	0.986	0.868	0.950	
Adjust Span	Time		10:15	10:15	10:24	10:24	
	Reading		153.2 ppm	149.3 ppm	14.93 % Vol	12.06 % Vol	
	Final Gain		1.144	1.001	0.873	0.933	
Check Zero	Time	10:27					
	Reading		0.2 ppm	0.1 ppm	0.02 % Vol	0.06 % Vol	

GAS CALIBRATION LOG - SYSTEM CAL (Including heated line and chiller unit)

		SO ₂	CO	NO _x	O ₂	CO ₂	VOCs
Span Value:			153.1 ppm	149.3 ppm	14.9 % Vol	12.1 % Vol	
Check Zero	Time	10:34					
	Reading		-0.5 ppm	0.0 ppm	0.01 % Vol	0.04 % Vol	
	Pass/fail		PASS	PASS	PASS	PASS	
Check Span	Time		10:40	10:40	10:46	10:46	
	Reading		155.6 ppm	150.6 ppm	14.84 % Vol	11.97 % Vol	
	Response Time/s		76	82	85	85	
Pass/fail			PASS	PASS	PASS	PASS	

GAS CALIBRATION LOG - POST CAL (Including heated line and chiller unit)

		SO ₂	CO	NO _x	O ₂	CO ₂	VOCs
Span Value:			153.1 ppm	149.3 ppm	14.93 % Vol	12.06 % Vol	
Check Zero	Time	16:04					
	Reading		1.3 ppm	0.1 ppm	-0.12 % Vol	0.09 % Vol	
Check Span	Time		16:10	16:10	16:16	16:16	
	Reading		148.1 ppm	147.9 ppm	14.67 % Vol	11.92 % Vol	
	Reg Pressure		165	165	55	55	
Zero Drift check @2x repeatability zero			-0.10 ppm	0.1 ppm	0.02 % Vol	0.06 % Vol	
Acceptance			Accept	Accept	Accept	Accept	
Zero Drift (%)			0.8	0.1	0.8	0.7	
Span Drift (%)			3.3	0.9	1.7	1.2	
Acceptance zero			Accept	Accept	Accept	Accept	
Acceptance span			Correct for drift	Accept	Accept	Accept	

CALIBRATION TO BE CARRIED OUT BY OR UNDER THE SUPERVISION OF MCERTS QUALIFIED PERSONNEL WITH LEVEL TWO AND TE4

Name:	Matthew Ellison
MCERTS ID:	MM-05-682

Personnel Present:	MRE/KCB
--------------------	---------

2.2.6 - Uncertainty Calculations

SCV 1H Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Uncertainty calculation for Gaseous Measurement NOx EN14792 - PG-250 (AS0450)- SRM Version
v3.6
SCV-1H
Feb-15

Limit value	107 mg/m ³ (corrected) NOx	Cal gas conc	311.44 mg.m ⁻³ (NO _x)
Measured concentration	28.79 mg/m ³ (101.3kPa, 273K)	Full Scale	513.25 mg/m ³ (NO _x)
Measured concentration	44.38 mg/m ³ (Corrected)	Gas	NO
NO/NO ₂ ratio	100.00	Cal gas conc	250 ppm
		Conversion	151.7 ppm
			2.053

Correction for reference conditions			
	O ₂ , %	Moisture, %	Temperature, K
ref	3.00	0.00	101.30
measured	9.32	0.00	101.30
Uncert	0.48	1.00	0.00
Factors	1.54	0.06	1.00
Uncertainty in factor	0.01	0.01	0.00
Correction Factor	1.54	1.54	0.07

Performance characteristics	Value	specification
Response time	92	≤200
Longer sampling interval	60	
Measurement period	61	
Number of readings in measurement	61	
Repeatability at zero	0.03	≤41 % range
Repeatability at span level	0.09	≤42 % range
Deviation from linearity(lack of fit)	0.56	≤42 % range
Zero drift	0.21	≤5% span value
Span drift	1.55	≤5% span value
volume or pressure flow dependence	0.36	% full scale
atmospheric pressure dependence	0.16	% full scale /kPa
ambient temperature dependence	-2.16	% full scale /°K
Cross sensitivity	0.43	% full scale /°K
dependence on voltage	0.3	% full scale
losses in the line (leak)	0	% full scale /°V
Converter efficiency	96.1	% value
Uncertainty of calibration gas	1	% value

Effect of drift		0.38 mg/m ³	
	min	max	value at calib
flow	0.4	0.4	0.4 l/min
pressure	101.0	102	101 kPa
temp	280	285	280 K
voltage	110	110	110 V

Performance characteristic	Uncertainty	Value of uncertainty quantity	
		for mean	use rep at span
Standard deviation of repeatability at zero	U ₀		0.06
Standard deviation of repeatability at span level	U _s		1.66
Lack of fit	U _{li}		0.22
Drift	U _{dr}		0.00
volume or pressure flow dependence	U _{vol}		-0.24
atmospheric pressure dependence	U _{atm}		-3.60
ambient temperature dependence	U _{atm}		0.00
Cross sensitivity	U _{int}		0.00
Dependence on voltage	U _{volt}		0.17
losses in the line (leak)	U _{leak}		0.65
Uncertainty of calibration gas	U _{conv}		1.88
converter efficiency	U _{eff}		
Uncertainty in factor	U _f		

Measurement uncertainty			
Combined uncertainty	k =	4.30	mg/m ³
Expanded uncertainty		8.59	mg/m ³
Uncertainty corrected to std conds		13.77	mg/m ³
Expanded uncertainty expressed with a level of confidence of 95%		13.77	mg/m ³
Expanded uncertainty expressed with a level of confidence of 95%		31.03	% value
Expanded uncertainty expressed with a level of confidence of 95%		8.03	% ELV

Requirement in standard is for uncertainty to be < 10% of ELV (dry before O₂ correction)

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Uncertainty calculation for Gaseous Measurement CO EN15058 - PG-250 AS0450 - SRM Version
v4.1
SCV 1H
Feb-15

Limit value	N/A [mg/m ³] (corrected)	Cal gas conc	194.25 mg/m ³
Measured concentration	87.83 mg/m ³ (101.3kPa, 273K)	Full Scale	250.00 mg/m ³
Measured concentration	135.41 mg/m ³ (Corrected)	Gas	CO
		Full Scale	200 ppm
		Cal gas conc	155.4 ppm
		Conversion	1.25

Correction for reference conditions				
	ref	measured	Uncert	
102, %	3	0.00	0.00	101.30
Moisture, %	9.32	0.00	0.00	101.30
Pressure, kPa	0.49	1.00	0.00	273.00
Temperature, K	1.54	1.00	0.00	101.30
Factors	0.06	0.01	0.00	1.00
Uncertainty in factor	1.54 uf			0.07
Correction Factor				

Performance characteristics	Value	specification
Response time	81	seconds
Logger sampling interval	60	seconds
Measurement period	61	minutes
Number of readings in measurement	61	
Repeatability at zero	0.09	% full scale
Repeatability at span level	0.08	% full scale
Deviation from linearity (lack of fit)	0.71	% full scale
Zero drift	0	mg/m ³
Span drift	0	<5% span value
volume or pressure flow dependence	-0.27	% full scale
atmospheric pressure dependence	0.09	% full scale/2kPa
ambient temperature dependence	1.48	% full scale/10K
Cross sensitivity	0.1	% full scale
dependence on voltage	0.33	% full scale/10V
losses in the line (leak)	0	% value
Uncertainty of calibration gas	1	% value

Effect of drift	0.00 mg/m ³
-----------------	------------------------

ranges	min	max	value at calib
flow	0.4	0.4	0.4 l/min
pressure	101.0	102	101 kPa
temp	280	285	280 K
Voltage	110	110	110 V

Performance characteristic	Uncertainty	Value of uncertainty quantity	mg/m ³
Standard deviation of repeatability at zero	u ₀	for mean	use rep at span
Standard deviation of repeatability at span level	u ₁	for mean	0.03
Lack of fit	u ₂		1.02
Drift	u ₃		0.00
volume or pressure flow dependence	u ₄		0.00
atmospheric pressure dependence	u ₅		0.06
ambient temperature dependence	u ₆		1.07
Cross sensitivity	u ₇		0.12
Dependence on voltage	u ₈		0.00
losses in the line (leak)	u ₉		0.00
Uncertainty of calibration gas	u ₁₀		0.51
Uncertainty in factor	uf		5.73

Measurement uncertainty			
Combined uncertainty	k =	1.57	mg/m ³
Expanded uncertainty		3.14	mg/m ³
Uncertainty corrected to std conds			
Expanded uncertainty expressed with a level of confidence of 95%		12.43	mg/m ³
Expanded uncertainty expressed with a level of confidence of 95%		12.43	9.18 % value

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Uncertainty calculation for Gaseous Measurement O2 EN14789 - PG-250 AS0450 SRMI
v3.0
SCV 1H
Feb-15

Limit value	N/A	Cal gas conc	14.99 % vol
Measured concentration	9.32 % vol	Full Scale	25 % vol

Performance characteristics	Value	specification
Response time	86	seconds
Logger sampling interval	60	seconds
Measurement period	61	minutes
Number of readings in measurement	61	
Repeatability at zero	0.03	% range
Repeatability at span level	0.03	% range
Deviation from linearity(lack of fit)	-0.05	% vol
Zero drift	0.11	% vol
Span drift	0.25	% vol
volume or pressure flow dependence	0.08	% range
atmospheric pressure dependence	0.91	% full scale /kPa
ambient temperature dependence	0.19	% vol/10K
Cross Sensitivity	0.2	% vol
dependence on voltage	0.05	% vol/10V
losses in the line (leak)	0	% value
Uncertainty of calibration gas	1	% value

Effect of drift
0.27 % vol

ranges	min	max	value at calib
flow	0.4	0.4	0.4 l/min
pressure	101.0	102	101 kPa
temp	280	285	280 K
Voltage	110	110	110 V

Performance characteristic	Uncertainty	Value of uncertainty quantity	% vol
Standard deviation of repeatability at zero	U_{r0}	for mean	
Standard deviation of repeatability at span level	U_{rs}	for mean	
Lack of fit	U_{fit}		0.00
Drift	U_{dr}		-0.03
volume or pressure flow dependence	U_{vpr}		0.15
atmospheric pressure dependence	U_{aprs}		0.00
ambient temperature dependence	U_{atmp}		0.13
Cross Sensitivity	U_{cst}		0.05
Dependence on voltage	U_{vdep}		0.11
losses in the line (leak)	U_{leak}		0.00
Uncertainty of calibration gas	U_{cal}		0.00

Measurement uncertainty			
Combined uncertainty	k = 2	0.24	% vol
Expanded uncertainty		0.49	% vol
Expanded uncertainty expressed with a level of confidence of 95%		0.49 % vol	
Expanded uncertainty expressed with a level of confidence of 95%		5.23 % value	

Requirement in standard is for uncertainty to be < 6% of value (dry)

Uncertainty calculation for Velocity and Volume Flow Rate Measurement by Pitot tube EN ISO 16911-1

v1.3 Jan-16

Enter data in orange cells only SCV JH

Constants	
Gas constant	8.314 J/(K·mol)
Velocity meas. during calibration	2.993 m/s
Air density meas. during calibration	1.215 kg/m ³
DP meas. during calibration	57.32 Pa
Uncertainty of velocity meas. at calibration	1.1 %
Uncertainty of air density meas. at calibration	0.075 %
Uncertainty of DP meas. at calibration	0.67 %
Pitot coefficient, K	0.819
Expanded uncertainty (95%, k=2) axis of value	2.6 %
Expanded uncertainty (95%, k=2)	0.02

Duct dimensions	
Rectangular	
Diameter	1.48 m
Area	1.72 m ²
	0.0 m ²

Characteristics of pressure sensor used for Delta P	
Repeatability of Delta P transducer	0.05 % of value
Range of Delta P transducer	2500 Pa
Resolution of Delta P transducer	0.020 Pa
Drift of Delta P transducer	0.1 % of range
Lack of fit of measurement system	0.1 % of range
Uncertainty in Delta P transducer	0.035 Pa
Uncertainty in temperature readout system	0.2 °C
Uncertainty in atmospheric pressure transducer	173.4 Pa
Uncertainty in duct area measurement	1.6 %

Uncertainty in stack gas composition	
Water vapour measurement	20 % relative
CO content measurement	6 % relative
CO ₂ content measurement	10 % relative
O ₂ content measurement	6 % relative

All Pressures should be entered in Pascals, Pa

Measurement Point	Atmospheric Pressure, Pa	Stack Pressure, Pa	messt1, Pa	messt2, Pa	messt3, Pa	messt4, Pa	messt5, Pa	messt6, Pa	Delta P, Pa	Stack Temperature, C	Water Vapour Content, %	CO ₂ ppm	CO ₂ , %	O ₂ , %	N ₂ , %	Dry gas basis	dry molecular wt, g/mol	stack molecular wt, g/mol
1	100000	184.37	44.1401856						44	21.3		2.47	135	6.8	85.9	9.3	29.46	29.18
2	100000	184.37	44.1401856						40	21.2		2.47	135	6.8	85.9	9.3	29.46	29.18
3	100000	184.37	44.1401856						68	21.2		2.47	135	6.8	85.9	9.3	29.46	29.18
4	100000	184.37	53.91713429						54	21.1		2.47	135	6.8	85.9	9.3	29.46	29.18
5	100000	184.37	67.6449574						64	21.1		2.47	135	6.8	85.9	9.3	29.46	29.18
6	100000	184.37	65.72924951						65	20.9		2.47	135	6.8	85.9	9.3	29.46	29.18
7	100000	184.37	48.0326569						41	20.9		2.47	135	6.8	85.9	9.3	29.46	29.18
8	100000	184.37	41.17308437						20.9	20.8		2.47	135	6.8	85.9	9.3	29.46	29.18
9	100000	184.37	51.95651122						52	20.6		2.47	135	6.8	85.9	9.3	29.46	29.18
10	100000	184.37	59.79960348						62	20.4		2.47	135	6.8	85.9	9.3	29.46	29.18
11	100000	184.37	61.79962955						62	20.4		2.47	135	6.8	85.9	9.3	29.46	29.18
12	100000	184.37	61.79962955						62	20.4		2.47	135	6.8	85.9	9.3	29.46	29.18
Mean			48.4	49.7	49.0	49.0	49.0	49.7		21.0		2.5	135.0	6.8	85.9	9.3	29.46	29.18

$p = \text{molar mass} \cdot \text{absolute pressure}$
 $K = \text{gas temperature}$

Mean density 1.201 kg/m³

$V_{\text{velocity}} = K \cdot \sqrt{\frac{2 \cdot \Delta p}{\rho}}$

Mean velocity	7.45 m/sec
Standard uncertainty of velocity	1.4 % of value
Expanded uncertainty in velocity	2.7 % of value

Circular duct	
Flow rate	46135 m ³ /hour
Volume flow rate expanded uncertainty	2086 m ³ /hour
Volume flow rate expanded uncertainty	4.5 % of value

SCV 2A Uncertainty Calculations

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Uncertainty calculation for Gaseous Measurement NOx EN14792 - PG-250 (AS0450)- SRM Version
v3.6
SCV 2A
Feb-15

Limit value	107 mg/m ³ (corrected) NOx	Cal gas conc	305.51 mg/m ³ (NO _x)
Measured concentration	33.43 mg/m ³ (101.3kPa, 273K)	Full Scale	513.25 mg/m ³ (NO _x)
Measured concentration	49.37 mg/m ³ (Corrected)	Gas	NO
NO/NO ₂ ratio	100.00	Full Scale	250 ppm
		Cal gas conc	149.3 ppm
		Conversion	2.053

Correction for reference conditions				
	ref	measured	Moisture, %	Temperature, K
			3.00	101.30
			8.81	101.30
			0.00	273.00
Factors			0.46	0.00
Uncertainty in factor			1.48	1.00
Correction Factor			0.06	0.00
			1.48	0.06

Performance characteristics	Value	specification
Response time	165 seconds	≤200
Logger sampling interval	60 seconds	
Measurement period	61 minutes	
Number of readings in measurement	61	
Repeatability at zero	0.03 % full scale	≤±1 % range
Repeatability at span level	0.09 % full scale	≤±2 % range
Deviation from linearity(lack of fit)	0.56 % full scale	≤±2 % range
Zero drift	0.21 mg/m ³	≤5% span value
Span drift	2.88 mg/m ³	≤5% span value
volume or pressure flow dependence	0.36 % full scale	
atmospheric pressure dependence	-0.16 % full scale/2kPa	≤±3 % range/2kPa
ambient temperature dependence	-2.43 % full scale/10K	≤±3 % range/10K
Cross sensitivity	0.5 % full scale	≤±4 % range
dependence on voltage losses in the line (leak)	-0.1 % full scale/10V	≤±2% range/10V
Converter efficiency	96.1 % value	≥95% of value
Uncertainty of calibration gas	1 % value	≤±2% of value

Effect of drift	0.52 mg/m ³
-----------------	------------------------

ranges	min	max	value at cellb
flow	0.4	0.4	0.4 l/min
pressure	101.0	102	101.2 kPa
temp	281	286	281 K
Voltage	110	110	110 V

Performance characteristic	Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero	U ₀	use rep at span
Standard deviation of repeatability at span level	U _s	for mean
Lack of fit	U _{li}	for mean
Drift	U _{dr}	1.66
volume or pressure flow dependence	U _{vol}	0.30
atmospheric pressure dependence	U _{atm}	0.00
ambient temperature dependence	U _{atm}	-0.17
Cross sensitivity	U _{amp}	-3.60
Dependence on voltage losses in the line (leak)	U _{line}	1.48
Uncertainty of calibration gas	U _{cal}	0.00
converter efficiency	U _{conv}	0.19
Uncertainty in factor	U _f	0.75
		1.88

Measurement uncertainty		
Combined uncertainty	k =	4.32 mg/m ³
Expanded uncertainty	2	8.63 mg/m ³
Uncertainty corrected to std conds		13.30 mg/m ³
Expanded uncertainty	expressed with a level of confidence of 95%	13.30 mg/m ³
Expanded uncertainty	expressed with a level of confidence of 95%	26.94 % value
Expanded uncertainty	expressed with a level of confidence of 95%	8.07 % ELV

Requirement in standard is for uncertainty to be < 10% of ELV (dry before O₂ correction)

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Uncertainty calculation for Gaseous Measurement CO EN15058 - PG-250 AS0450 - SRM Version
v4.1
SCV 2A
Feb-15

Limit value	N/A	mg/m ³ (corrected)	Cal gas conc	191.38 mg.m ⁻³
Measured concentration	86.22	mg/m ³ (101.3kPa, 273K)	Full Scale	250.00 mg/m ³
Measured concentration	127.36	mg/m ³ (Corrected)	Gas	CO
			Full Scale	200 ppm
			Cal gas conc	153.1 ppm
			Conversion	1.25

Correction for reference conditions				
ref	O ₂ , %	Moisture, %	Pressure, kPa	Temperature, K
measured	0.00	0.00	101.30	273.00
Uncert	0.46	1.00	0.00	0.00
Factors	1.48	1.00	1.00	1.00
Uncertainty in factor	0.06	0.01	0.00	0.00
Correction Factor	1.48	uf	0.06	0.00

Performance characteristics	Value	specification
Response time	76	seconds
Logger sampling interval	60	seconds
Measurement period	61	minutes
Number of readings in measurement	61	
Repeatability at zero	0.09	% full scale
Repeatability at span level	0.08	% full scale
Deviation from linearity(lack of fit)	0.71	% full scale
Zero drift	0	mg/m ³
Span drift	0	mg/m ³
volume or pressure flow dependence	-0.27	% full scale
atmospheric pressure dependence	0.09	% full scale/2kPa
ambient temperature dependence	1.48	% full scale/10K
Cross sensitivity	0.1	% full scale
dependence on voltage	0.33	% full scale/10V
losses in the line (leak)	0	% value
Uncertainty of calibration gas	1	% value

Effect of drift	0.00	mg/m ³
-----------------	------	-------------------

ranges	min	max	value at calib
flow	0.4	0.4	0.4 l/min
pressure	101.0	102	101.2 kPa
temp	281	286	281 K
Voltage	110	110	110 V

Performance characteristic	Uncertainty	Value of uncertainty quantity
Standard deviation of repeatability at zero	U ₀	mg/m ³
Standard deviation of repeatability at span level	U _s	use rep at span
Lack of fit	U _{li}	for mean
Drift	U _{dr}	for mean
volume or pressure flow dependence	U _{vol}	0.03
atmospheric pressure dependence	U _{atm}	1.02
ambient temperature dependence	U _{atemp}	0.00
Cross sensitivity	U _{cross}	0.00
Dependence on voltage	U _{volt}	0.05
losses in the line (leak)	U _{leak}	1.07
Uncertainty of calibration gas	U _{cal}	0.12
		0.00
		0.00
		0.50
Uncertainty in factor	uf	4.85

Measurement uncertainty		
Combined uncertainty	k =	1.57
Expanded uncertainty		3.13
Uncertainty corrected to std conds		10.77
Expanded uncertainty expressed with a level of confidence of 95%		21.54
Expanded uncertainty expressed with a level of confidence of 95%		8.45 % value

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Uncertainty calculation for Gaseous Measurement O2 EN14789 - PG-250 AS0450 SRM
v3.0
SCV 2A

Limit value	N/A	Cal gas conc	14.93 % vol
Measured concentration	8.81 % vol	Full Scale	23 % vol

Performance characteristics	Value	specification
Response time	85 seconds	≤200
Logger sampling interval	60 seconds	
Measurement period	61 minutes	
Number of readings in measurement	61	
Repeatability at zero	0.03 % range	±0.2 % range
Repeatability at span level	0.03 % range	±0.4 % range
Deviation from linearity(lack of fit)	-0.05 % vol	<0.3 % vol
Zero drift	0.12 % vol	<5% span value
Span drift	0.26 % vol	<5% span value
volume or pressure flow dependence	0.08 % range	<1% range
atmospheric pressure dependence	0.91 % full scale /kPa	±1.5 % range/kPa
ambient temperature dependence	0.19 % vol/10K	±0.3 % vol/10K
Cross Sensitivity	0.2 % vol	±0.2 % vol
dependence on voltage losses in the line (leak)	0.05 % vol/10V	±0.1% vol/10V
losses in the line (leak)	0 % value	± 2% of value
Uncertainty of calibration gas	1 % value	± 2% of value

Effect of drift	0.27 % vol
-----------------	------------

ranges	min	max	value at calib
flow	0.4	0.4	0.4 l/min
pressure	101.0	102	101.2 kPa
temp	281	286	281 K
Voltage	110	110	110 V

Performance characteristic	Uncertainty	Value of uncertainty quantity	% vol
Standard deviation of repeatability at zero	U_{r0}	for mean	use rep at span
Standard deviation of repeatability at span level	U_{rs}	for mean	0.00
Lack of fit	U_{rlf}		-0.03
Drift	U_{dr}		0.16
volume or pressure flow dependence	U_{vps}		0.00
atmospheric pressure dependence	U_{aps}		0.09
ambient temperature dependence	U_{atemp}		0.05
Cross Sensitivity	U_{cst}		0.11
Dependence on voltage losses in the line (leak)	U_{vll}		0.00
Uncertainty of calibration gas	U_{cal}		0.00
			0.05

Measurement uncertainty	Value	% vol
Combined uncertainty	0.23	% vol
Expanded uncertainty	0.46	% vol
Expanded uncertainty expressed with a level of confidence of 95%	0.46 % vol	
Expanded uncertainty expressed with a level of confidence of 95%	5.19 % value	

Requirement in standard is for uncertainty to be < 6% of value (dry)

2.2.7 - Calculations Used in Reporting Results

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Nozzle Selection

For isokinetic sampling, the pressure difference of the orifice meter must equal the pressure difference of the Pitot tube pressure multiplied by the K-factor. Where:

$$K = \text{Constant} \times C_p^2 \times D_n^4 \times DH_{@} \times \left(\frac{M_d}{M_s} \right) \left(\frac{1 - B_{wm}}{1 - B_{ws}} \right)^2 \left(\frac{T_m + 273}{T_s + 273} \right) \left(\frac{P_s}{P_m} \right)$$

$$DH = K \times D_p$$

Where:-

Constant: is a constant dependent on the units used to measure the nozzle (8.038×10^{-5} for mm)

D_n the nozzle diameter mm

$DH_{@}$ a constant dependent on the sampler control box orifice and gas meter

B_{ws} the percent water vapour in the emission as a fraction i.e. 12% = 0.12

B_{wm} the percentage water vapour in the air around the meter box often assumed to be zero

C_p Pitot tube coefficient dependent on the Pitot tube type

T_m the meter temperature in °C

T_s the stack temperature in °C

P_s the stack pressure

P_m the meter pressure

M_d dry gas molecular weight

M_s apparent stack gas molecular weight

DH pressure drop across the orifice (mm water)

DP differential Pitot pressure (mm water)

From this the correct nozzle size can be determined.

$$D_n = \sqrt{\left(\frac{\text{Constant} \cdot Q_m \cdot P_m}{(T_m + 273) C_p} \right) \left(\frac{1 - B_{wm}}{1 - B_{ws}} \right) \sqrt{\frac{(T_s + 273) M_s}{(P_s \cdot (\Delta P)_{avg})}}$$

Where the Constant = 0.6071 Metric

Q_m = Orifice flow rate normally 21.2 actual lmin^{-1}

$$= K_m \sqrt{\frac{(T_m + 273) \Delta H}{P_m M_m}}$$

Where K_m = Orifice meter coefficient

$$K_m = Q_m \sqrt{\frac{P_m M_m}{\Delta H (T_m + 273)}} = \text{Const} \sqrt{\frac{1}{\Delta H_{@}}}$$

Where Const = 183.7 metric

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

Moisture Determination Calculations

These calculations are based at 273K and 101.325kPa

To calculate moisture the following equation is used:

$$B_{ws} = \frac{0.001245 \times W_I \times 100}{(0.001245 \times W_I) + 0.359V_m \left(\frac{P_b + \frac{\Delta H_{avg}}{13.6}}{(T_m + 273)} \right)}$$

Particulate Concentration C_s in stack Gases

At 273K and 101.325kPa and dry gas

$$C_s = \frac{W_t}{V_m} \times \frac{T_m + 273}{273} \times \frac{760}{\left(P_b + \frac{\Delta H_{avg}}{13.6} \right)} \times 1000 \quad \text{mg/Nm}^3$$

Oxygen Concentration Correction C_{oxy} to Particulate concentration

$$C_{oxy} = C \times \frac{(20.9 - \%O_2 \text{ ref})}{(20.9 - \%O_2 \text{ Meas})} \quad \text{mg/Nm}^3$$

Dry Molecular Weight of gases

$$M_D = 0.44(\%CO_2) + 0.32(\%O_2) + 0.28(\%CO + \%N)$$

Stack Molecular Weight of gases

$$M_s = 0.18(B_{ws}) + \frac{M_d}{100}(100 - B_{ws})$$

Stack Gas Velocity

$$(V_s)_{avg} = 34.96 \times C_p \times \sqrt{(\Delta P)_{avg}} \sqrt{\frac{T_s + 273}{P_s M_s}} \quad \text{m/s}$$

Mass Emission Rate M_R

$$M_R = \frac{C_m \times (V_s)_{avg} \times A \times 3600}{10^6} \quad \text{kg/hr}$$

NATIONAL PHYSICAL LABORATORY
Continuation Sheet

IsoKinicity

$$I = \frac{2.12 \times 10^8 \times V_m \times Y \times \left(P_b + \left(\frac{\Delta H_{avg}}{13.6} \right) \right) \left(\frac{273 + T_s}{273 + T_m} \right)}{\Theta P_s \pi D_n^2 (Vs)_{avg} (100 - B_{ws})} \%$$

- W_I = the weight change of the impingers during sampling in g
- V_m = volume of dry gas sample in litres at temperature of the meter box
- B_{ws} = the percent water vapour in the emission
- Q = length of time sampling in minutes
- Y = Gas Meter Calibration correction factor
- V_s = Velocity of stack gas m/s
- C_M = measured concentration of particulate matter (mg/m³)
- T_m = average temperature at dry gas meter (°C)
- P_b = atmospheric pressure (mmHg)
- $\%O_{2ref}$ = % oxygen at standard temperature & pressure
- $\%O_{2Meas}$ = % oxygen measured on site
- C_P = Pitot tube coefficient
- DP = mean differential Pitot pressure drop (mm H₂O)
- DH = mean orifice pressure drop (mm H₂O)
- D_s = diameter of stack (m)
- D_n = Nozzle diameter (mm)
- T_s = stack temperature (°C)
- M_d = molecular weight of dry stack gas
- B_w = moisture fraction
- P_s = stack pressure (mmHg)
- A = duct c.s.a. (m²)
- M_s = molecular weight of wet stack gas
- M_d = molecular weight of dry stack gas
- W_t = total weight of particulate matter (g)