

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | | |

Input values

| Component | Certification range | | Emissions limit value | | Confidence interval |
|-----------|---------------------|-------|-----------------------|---------|---------------------|
| CO | 300.00 | mg/m³ | 1,000.00 | mg/m³ | 10 % |
| CO2 | 25.00 | Vol% | 25.00 | Vol% * | 20 % * |
| NO | 400.00 | mg/m³ | 650.00** | mg/m³ | 20 % |
| NO2 | 100.00 | mg/m³ | 100.00 | mg/m³ | 20 % |
| N2O | 50.00 | mg/m³ | 50.00 | mg/m³ * | 20 % * |
| SO2 | 300.00 | mg/m³ | 400.00 | mg/m³ | 20 % |
| HCl | 90.00 | mg/m³ | 100.00 | mg/m³ | 40 % |
| HF | 10.00 | mg/m³ | 10.00 | mg/m³ | 40 % |
| NH3 | 50.00 | mg/m³ | 60.00 | mg/m³ * | 40 % * |
| H2O | 40.00 | Vol% | 40.00 | Vol% * | 40 % * |
| CH4 | 50.00 | mg/m³ | | mg/m³ * | 20 % * |
| Corg | 50.00 | mg/m³ | 100.00 | mg/m³ | 30 % |
| O2 | 21.00 | Vol% | 25.00 | Vol% * | 20 % * |

* For this measuring component no emission limit values and confidence intervals are defined: Therefore full scale values and exemplary confidence intervals are used here.

** The emissions limit value for NOx is given as NO2-concentration, therefore the value as NO-concentration is decreased by the factor 1.53.

| Interferent | Concentration | | Interferent | Concentration | |
|------------------------|---------------|-------|-------------------------|---------------|-------|
| Oxygen (O2) | 3.00 | Vol% | Ammonia (NH3) | 20.00 | mg/m³ |
| Oxygen (O2) | 21.00 | Vol% | Sulfur dioxide (SO2) | 200.00 | mg/m³ |
| Water (H2O) | 30.00 | Vol% | Sulfur dioxide (SO2) | 1,000.00 | mg/m³ |
| Carbon monoxide (CO) | 300.00 | mg/m³ | Hydrogen chloride (HCl) | 50.00 | mg/m³ |
| Carbon dioxide (CO2) | 15.00 | Vol% | Hydrogen chloride (HCl) | 200.00 | mg/m³ |
| Methane (CH4) | 50.00 | mg/m³ | | | |
| Dinitrogen oxide (N2O) | 20.00 | mg/m³ | | | |
| Dinitrogen oxide (N2O) | 100.00 | mg/m³ | | | |
| Nitrogen monoxide (NO) | 300.00 | mg/m³ | | | |
| Nitrogen dioxide (NO2) | 30.00 | mg/m³ | | | |

Required quality of the measurement

| | | | | |
|-----------------------------------|----|-----|----|---|
| Requirement to response time | 25 | % | ** | Requirement of the legislation, the customer or authority |
| Averaging time of measured values | 30 | min | | |

** Possible values are 25% for dynamic (standard) or 10 % for highly dynamic processes (EN ISO 14956, 7.2)

Summary of the results

| Component | Response time | s(AMS) values | | Quality of the measurement |
|-----------|------------------------|---------------|------------|----------------------------|
| | | Zero point | Span point | |
| CO | Requirements fulfilled | 6.5727 | 9.0515 | Requirements fulfilled |
| CO2 | Requirements fulfilled | 0.6360 | 0.7807 | Requirements fulfilled |
| NO | Requirements fulfilled | 10.0161 | 13.9600 | Requirements fulfilled |
| NO2 | Requirements fulfilled | 3.4784 | 3.6244 | Requirements fulfilled |
| N2O | Requirements fulfilled | 1.0292 | 1.3577 | Requirements fulfilled |
| SO2 | Requirements fulfilled | 8.1319 | 11.7655 | Requirements fulfilled |
| HCl | Requirements fulfilled | 3.3212 | 3.6177 | Requirements fulfilled |
| HF | Requirements fulfilled | 0.4901 | 0.4225 | Requirements fulfilled |
| NH3 | Requirements fulfilled | 1.5732 | 1.4275 | Requirements fulfilled |
| H2O | Requirements fulfilled | 0.9910 | 1.0581 | Requirements fulfilled |
| CH4 | | | | |
| Corg | Requirements fulfilled | 2.3356 | 2.2500 | Requirements fulfilled |
| O2 | Requirements fulfilled | 0.2816 | 0.3226 | Requirements fulfilled |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|-------------------------|----------------------------|------------------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | CO |

Input values

| | | | | | |
|-----------------------|-------|-------------------|-----------------------------------|----|-----|
| Certification range | 300 | mg/m ³ | Requirement to response time | 25 | % |
| Emissions limit value | 1,000 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 10 | % | | | |

General information

| | | | | | |
|-----------------------------|---|--------|------------------------|------|-------------------|
| Maintenance interval | 6 | months | Detection limit | 0.32 | mg/m ³ |
|-----------------------------|---|--------|------------------------|------|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 2.97 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|-------------------------|--------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 Vol% Water (H ₂ O) | 0.00 mg/m ³ | 3.00 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | | |
| 15 Vol% Carbon dioxide (CO ₂) | 4.80 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | -5.40 mg/m ³ | -10.50 mg/m ³ |
| 300 mg/m ³ Nitrogen monoxide (NO) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Ammonia (NH ₃) | 0.00 mg/m ³ | 2.10 mg/m ³ |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|-------|-------------------|
| 4.80 | mg/m ³ |
| -5.40 | mg/m ³ |

| | |
|--------|-------------------|
| 5.10 | mg/m ³ |
| -10.50 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | CO |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

Largest difference according to type approval

| | Zero point | Span point |
|--|-------------------------|--------------------------|
| Lack-of-fit (Linearity) | 6.00 mg/m ³ | 6.00 mg/m ³ |
| Zero drift from the field test | 4.50 mg/m ³ | 0.00 mg/m ³ |
| Span drift from the field test | 0.00 mg/m ³ | -5.40 mg/m ³ |
| Influence of ambient temperature at span point | 0.60 mg/m ³ | -5.10 mg/m ³ |
| Influence of sample gas pressure | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Influence of sample gas flow | 0.00 mg/m ³ | -0.30 mg/m ³ |
| Influence of voltage | 0.60 mg/m ³ | 0.90 mg/m ³ |
| Cross-sensitivity | -5.40 mg/m ³ | -10.50 mg/m ³ |
| Repeatability at span point | 0.16 mg/m ³ | 0.22 mg/m ³ |
| Standard deviation from paired measurements under field conditions | 2.73 mg/m ³ | 2.73 mg/m ³ |
| Uncertainty of provided reference material | 6.00 mg/m ³ | 6.00 mg/m ³ |
| Misalignment | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Changes of response factors | 0.00 mg/m ³ | 0.00 mg/m ³ |

Process characteristics

Standard uncertainty

| | | Zero point | Span point |
|--|-------------|---------------------------|---------------------------|
| Lack-of-fit (Linearity) | U_{lof} = | 3.4641 mg/m ³ | 3.4641 mg/m ³ |
| Zero drift from the field test | $U_{d,z}$ = | 2.5981 mg/m ³ | 0.0000 mg/m ³ |
| Span drift from the field test | $U_{d,s}$ = | 0.0000 mg/m ³ | -3.1177 mg/m ³ |
| Influence of ambient temperature at span point | U_t = | 0.3464 mg/m ³ | -2.9445 mg/m ³ |
| Influence of sample gas pressure | U_p = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Influence of sample gas flow | U_f = | 0.0000 mg/m ³ | -0.1732 mg/m ³ |
| Influence of voltage | U_v = | 0.3464 mg/m ³ | 0.5196 mg/m ³ |
| Cross-sensitivity | U_i = | -3.1177 mg/m ³ | -6.0622 mg/m ³ |
| Repeatability at span point | U_r = | 0.0924 mg/m ³ | 0.1270 mg/m ³ |
| Standard deviation from paired measurements under field conditions | U_D = | 1.5780 mg/m ³ | 1.5780 mg/m ³ |
| Uncertainty of provided reference material | U_{rm} = | 3.4641 mg/m ³ | 3.4641 mg/m ³ |
| Misalignment | U_{mb} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | U_{ce} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Changes of response factors | U_{rf} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |

Calculation of the combined standard uncertainties

| | | Zero point | Span point |
|-------------------------------|---------------|--------------------------|--------------------------|
| Combined standard uncertainty | s(AMS) values | 6.5727 mg/m ³ | 9.0515 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|--------|-------------------|--|
| Combined standard uncertainty | 9.42 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 18.46 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 1.85 | % | of the emissions limit value of 1000 mg/m ³ |
| Allowed expanded uncertainty | 10.00 | % | of the emissions limit value of 1000 mg/m ³ |
| Allowed expanded uncertainty | 100.00 | mg/m ³ | |

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | CO2 |

Input values

| | | | | | |
|---------------------|----|------|-----------------------------------|----|-----|
| Certification range | 25 | Vol% | Requirement to response time | 25 | % |
| Measuring range | 25 | Vol% | Averaging time of measured values | 30 | min |
| Confidence interval | 20 | % | | | * |

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

General information

| | | | | | |
|----------------------|---|--------|-----------------|------|------|
| Maintenance interval | 3 | months | Detection limit | 0.06 | Vol% |
|----------------------|---|--------|-----------------|------|------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 3.03 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | | Span point | |
|-----------------------------------|------------|------|------------|------|
| 3 Vol% Oxygen (O2) | 0.00 | Vol% | 0.00 | Vol% |
| 21 Vol% Oxygen (O2) | 0.00 | Vol% | 0.00 | Vol% |
| 30 Vol% Water (H2O) | -0.33 | Vol% | 0.48 | Vol% |
| 300 mg/m³ Carbon monoxide (CO) | 0.00 | Vol% | 0.00 | Vol% |
| 15 Vol% Carbon dioxide (CO2) | | | | |
| 50 mg/m³ Methane (CH4) | 0.00 | Vol% | 0.00 | Vol% |
| 20 mg/m³ Dinitrogen oxide (N2O) | 0.00 | Vol% | 0.00 | Vol% |
| 100 mg/m³ Dinitrogen oxide (N2O) | -0.40 | Vol% | -0.35 | Vol% |
| 300 mg/m³ Nitrogen monoxide (NO) | 0.00 | Vol% | 0.00 | Vol% |
| 30 mg/m³ Nitrogen dioxide (NO2) | 0.00 | Vol% | 0.00 | Vol% |
| 20 mg/m³ Ammonia (NH3) | 0.00 | Vol% | -0.35 | Vol% |
| 200 mg/m³ Sulfur dioxide (SO2) | 0.00 | Vol% | 0.00 | Vol% |
| 1000 mg/m³ Sulfur dioxide (SO2) | 0.00 | Vol% | 0.33 | Vol% |
| 50 mg/m³ Hydrogen chloride (HCl) | 0.00 | Vol% | 0.00 | Vol% |
| 200 mg/m³ Hydrogen chloride (HCl) | 0.00 | Vol% | -0.13 | Vol% |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|-------|------|
| 0.00 | Vol% |
| -0.73 | Vol% |

| | |
|-------|------|
| 0.80 | Vol% |
| -0.83 | Vol% |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | CO2 |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

Largest difference according to type approval

| | Zero point | Span point |
|--|------------|------------|
| Lack-of-fit (Linearity) | 0.18 Vol% | 0.18 Vol% |
| Zero drift from the field test | 0.53 Vol% | 0.00 Vol% |
| Span drift from the field test | 0.00 Vol% | 0.68 Vol% |
| Influence of ambient temperature at span point | 0.08 Vol% | 0.53 Vol% |
| Influence of sample gas pressure | 0.00 Vol% | 0.00 Vol% |
| Influence of sample gas flow | 0.00 Vol% | -0.03 Vol% |
| Influence of voltage | -0.03 Vol% | 0.10 Vol% |
| Cross-sensitivity | -0.73 Vol% | -0.83 Vol% |
| Repeatability at span point | 0.03 Vol% | 0.05 Vol% |
| Standard deviation from paired measurements under field conditions | 0.35 Vol% | 0.35 Vol% |
| Uncertainty of provided reference material | 0.50 Vol% | 0.50 Vol% |
| Misalignment | 0.00 Vol% | 0.00 Vol% |
| Conversion rate of AMS for measurement of NOx | 0.00 Vol% | 0.00 Vol% |
| Changes of response factors | 0.00 Vol% | 0.00 Vol% |

Process characteristics

Standard uncertainty

| | Zero point | Span point |
|--|--------------|--------------|
| Lack-of-fit (Linearity) | 0.1010 Vol% | 0.1010 Vol% |
| Zero drift from the field test | 0.3031 Vol% | 0.0000 Vol% |
| Span drift from the field test | 0.0000 Vol% | 0.3897 Vol% |
| Influence of ambient temperature at span point | 0.0433 Vol% | 0.3031 Vol% |
| Influence of sample gas pressure | 0.0000 Vol% | 0.0000 Vol% |
| Influence of sample gas flow | 0.0000 Vol% | -0.0144 Vol% |
| Influence of voltage | -0.0144 Vol% | 0.0577 Vol% |
| Cross-sensitivity | -0.4186 Vol% | -0.4763 Vol% |
| Repeatability at span point | 0.0173 Vol% | 0.0289 Vol% |
| Standard deviation from paired measurements under field conditions | 0.2046 Vol% | 0.2046 Vol% |
| Uncertainty of provided reference material | 0.2887 Vol% | 0.2887 Vol% |
| Misalignment | 0.0000 Vol% | 0.0000 Vol% |
| Conversion rate of AMS for measurement of NOx | 0.0000 Vol% | 0.0000 Vol% |
| Changes of response factors | 0.0000 Vol% | 0.0000 Vol% |

Calculation of the combined standard uncertainties

| | s(AMS) values | Zero point | Span point |
|-------------------------------|---------------|-------------|-------------|
| Combined standard uncertainty | | 0.6360 Vol% | 0.7807 Vol% |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|------|-----------------------------------|
| Combined standard uncertainty | 0.84 | Vol% | according to EN 15267-3 |
| Expanded uncertainty | 1.64 | Vol% | according to EN 15267-3 |
| Relative expanded uncertainty | 6.57 | % | of the measuring range of 25 Vol% |
| Allowed expanded uncertainty | 20.00 | % | of the measuring range of 25 Vol% |
| Allowed expanded uncertainty | 5.00 | Vol% | |

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | NO |

Input values

| | | | | | |
|-----------------------|-----|-------------------|-----------------------------------|----|-----|
| Certification range | 400 | mg/m ³ | Requirement to response time | 25 | % |
| Emissions limit value | 650 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 20 | % | | | |

General information

| | | | | | |
|----------------------|---|--------|-----------------|------|-------------------|
| Maintenance interval | 6 | months | Detection limit | 0.76 | mg/m ³ |
|----------------------|---|--------|-----------------|------|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 2.93 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|--------------------------|-------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 Vol% Water (H ₂ O) | -10.40 mg/m ³ | 6.40 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 15 Vol% Carbon dioxide (CO ₂) | 0.00 mg/m ³ | -9.60 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 4.00 mg/m ³ |
| 300 mg/m ³ Nitrogen monoxide (NO) | | |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Ammonia (NH ₃) | 2.80 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 2.40 mg/m ³ |
| 50 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 3.20 mg/m ³ |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|--------|-------------------|
| 2.80 | mg/m ³ |
| -10.40 | mg/m ³ |

| | |
|-------|-------------------|
| 16.00 | mg/m ³ |
| -9.60 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | NO |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

Largest difference according to type approval

| | Zero point | Span point |
|--|--------------------------|--------------------------|
| Lack-of-fit (Linearity) | 6.00 mg/m ³ | 6.00 mg/m ³ |
| Zero drift from the field test | 7.20 mg/m ³ | 0.00 mg/m ³ |
| Span drift from the field test | 0.00 mg/m ³ | -12.00 mg/m ³ |
| Influence of ambient temperature at span point | -1.60 mg/m ³ | -6.00 mg/m ³ |
| Influence of sample gas pressure | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Influence of sample gas flow | 0.00 mg/m ³ | -0.40 mg/m ³ |
| Influence of voltage | 0.40 mg/m ³ | -3.20 mg/m ³ |
| Cross-sensitivity | -10.40 mg/m ³ | 16.00 mg/m ³ |
| Repeatability at span point | 0.38 mg/m ³ | 0.78 mg/m ³ |
| Standard deviation from paired measurements under field conditions | 6.18 mg/m ³ | 6.18 mg/m ³ |
| Uncertainty of provided reference material | 8.00 mg/m ³ | 8.00 mg/m ³ |
| Misalignment | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Changes of response factors | 0.00 mg/m ³ | 0.00 mg/m ³ |

Process characteristics

Standard uncertainty

| | | Zero point | Span point |
|--|-------------|---------------------------|---------------------------|
| Lack-of-fit (Linearity) | U_{lof} = | 3.4641 mg/m ³ | 3.4641 mg/m ³ |
| Zero drift from the field test | $U_{d,z}$ = | 4.1569 mg/m ³ | 0.0000 mg/m ³ |
| Span drift from the field test | $U_{d,s}$ = | 0.0000 mg/m ³ | -6.9282 mg/m ³ |
| Influence of ambient temperature at span point | U_t = | -0.9238 mg/m ³ | -3.4641 mg/m ³ |
| Influence of sample gas pressure | U_p = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Influence of sample gas flow | U_f = | 0.0000 mg/m ³ | -0.2309 mg/m ³ |
| Influence of voltage | U_v = | 0.2309 mg/m ³ | -1.8475 mg/m ³ |
| Cross-sensitivity | U_i = | -6.0044 mg/m ³ | 9.2376 mg/m ³ |
| Repeatability at span point | U_r = | 0.2194 mg/m ³ | 0.4503 mg/m ³ |
| Standard deviation from paired measurements under field conditions | U_D = | 3.5705 mg/m ³ | 3.5705 mg/m ³ |
| Uncertainty of provided reference material | U_{rm} = | 4.6188 mg/m ³ | 4.6188 mg/m ³ |
| Misalignment | U_{mb} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | U_{ce} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Changes of response factors | U_{rf} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |

Calculation of the combined standard uncertainties

| | | Zero point | Span point |
|-------------------------------|---------------|---------------------------|---------------------------|
| Combined standard uncertainty | s(AMS) values | 10.0161 mg/m ³ | 13.9600 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|--------|-------------------|---|
| Combined standard uncertainty | 14.57 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 28.55 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 4.39 | % | of the emissions limit value of 650 mg/m ³ |
| Allowed expanded uncertainty | 20.00 | % | of the emissions limit value of 650 mg/m ³ |
| Allowed expanded uncertainty | 130.00 | mg/m ³ | |

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|-------------------------|----------------------------|------------------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | NO2 |

Input values

| | | | | | |
|-----------------------|-----|-------------------|-----------------------------------|----|-----|
| Certification range | 100 | mg/m ³ | Requirement to response time | 25 | % |
| Emissions limit value | 100 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 20 | % | | | |

General information

| | | | | | |
|-----------------------------|---|--------|------------------------|------|-------------------|
| Maintenance interval | 6 | months | Detection limit | 0.38 | mg/m ³ |
|-----------------------------|---|--------|------------------------|------|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 3.30 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|-------------------------|-------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 Vol% Water (H ₂ O) | -1.50 mg/m ³ | 1.00 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 15 Vol% Carbon dioxide (CO ₂) | 0.50 mg/m ³ | -1.10 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | -0.90 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | 1.00 mg/m ³ | 0.00 mg/m ³ |
| 300 mg/m ³ Nitrogen monoxide (NO) | 3.20 mg/m ³ | 3.00 mg/m ³ |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | | |
| 20 mg/m ³ Ammonia (NH ₃) | 0.00 mg/m ³ | -1.00 mg/m ³ |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | -2.60 mg/m ³ |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|-------|-------------------|
| 4.70 | mg/m ³ |
| -2.40 | mg/m ³ |

| | |
|-------|-------------------|
| 4.00 | mg/m ³ |
| -4.70 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | NO2 |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

Largest difference according to type approval

| | Zero point | Span point |
|--|-------------------------|-------------------------|
| Lack-of-fit (Linearity) | 1.40 mg/m ³ | 1.40 mg/m ³ |
| Zero drift from the field test | -2.30 mg/m ³ | 0.00 mg/m ³ |
| Span drift from the field test | 0.00 mg/m ³ | 3.00 mg/m ³ |
| Influence of ambient temperature at span point | 1.30 mg/m ³ | 0.90 mg/m ³ |
| Influence of sample gas pressure | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Influence of sample gas flow | 0.00 mg/m ³ | -0.10 mg/m ³ |
| Influence of voltage | 0.30 mg/m ³ | -0.60 mg/m ³ |
| Cross-sensitivity | 4.70 mg/m ³ | -4.70 mg/m ³ |
| Repeatability at span point | 0.19 mg/m ³ | 0.47 mg/m ³ |
| Standard deviation from paired measurements under field conditions | 1.09 mg/m ³ | 1.09 mg/m ³ |
| Uncertainty of provided reference material | 2.00 mg/m ³ | 2.00 mg/m ³ |
| Misalignment | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Changes of response factors | 0.00 mg/m ³ | 0.00 mg/m ³ |

Process characteristics

Standard uncertainty

| | | Zero point | Span point |
|--|-------------|---------------------------|---------------------------|
| Lack-of-fit (Linearity) | U_{lof} = | 0.8083 mg/m ³ | 0.8083 mg/m ³ |
| Zero drift from the field test | $U_{d,z}$ = | -1.3279 mg/m ³ | 0.0000 mg/m ³ |
| Span drift from the field test | $U_{d,s}$ = | 0.0000 mg/m ³ | 1.7321 mg/m ³ |
| Influence of ambient temperature at span point | U_t = | 0.7506 mg/m ³ | 0.5196 mg/m ³ |
| Influence of sample gas pressure | U_p = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Influence of sample gas flow | U_f = | 0.0000 mg/m ³ | -0.0577 mg/m ³ |
| Influence of voltage | U_v = | 0.1732 mg/m ³ | -0.3464 mg/m ³ |
| Cross-sensitivity | U_i = | 2.7135 mg/m ³ | -2.7135 mg/m ³ |
| Repeatability at span point | U_r = | 0.1097 mg/m ³ | 0.2714 mg/m ³ |
| Standard deviation from paired measurements under field conditions | U_D = | 0.6267 mg/m ³ | 0.6267 mg/m ³ |
| Uncertainty of provided reference material | U_{rm} = | 1.1547 mg/m ³ | 1.1547 mg/m ³ |
| Misalignment | U_{mb} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | U_{ce} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Changes of response factors | U_{rf} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |

Calculation of the combined standard uncertainties

| | | Zero point | Span point |
|-------------------------------|---------------|--------------------------|--------------------------|
| Combined standard uncertainty | s(AMS) values | 3.4784 mg/m ³ | 3.6244 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|-------------------|---|
| Combined standard uncertainty | 3.86 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 7.57 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 7.57 | % | of the emissions limit value of 100 mg/m ³ |
| Allowed expanded uncertainty | 20.00 | % | of the emissions limit value of 100 mg/m ³ |
| Allowed expanded uncertainty | 20.00 | mg/m ³ | |

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|-------------------------|----------------------------|------------------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | N2O |

Input values

| | | | | | |
|---------------------|----|-------------------|-----------------------------------|----|-----|
| Certification range | 50 | mg/m ³ | Requirement to response time | 25 | % |
| Measuring range | 50 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 20 | % | | | * |

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

General information

| | | | | | |
|-----------------------------|---|--------|------------------------|------|-------------------|
| Maintenance interval | 6 | months | Detection limit | 0.08 | mg/m ³ |
|-----------------------------|---|--------|------------------------|------|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 2.92 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|-------------------------|-------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.40 mg/m ³ | 0.80 mg/m ³ |
| 30 Vol% Water (H ₂ O) | -0.50 mg/m ³ | -0.80 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | 0.00 mg/m ³ | -0.30 mg/m ³ |
| 15 Vol% Carbon dioxide (CO ₂) | -0.70 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | | |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | | |
| 300 mg/m ³ Nitrogen monoxide (NO) | 0.95 mg/m ³ | 0.95 mg/m ³ |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Ammonia (NH ₃) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|-------|-------------------|
| 1.35 | mg/m ³ |
| -1.20 | mg/m ³ |

| | |
|-------|-------------------|
| 1.75 | mg/m ³ |
| -1.10 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | N2O |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

Largest difference according to type approval

| | Zero point | Span point |
|--|------------------------|-------------------------|
| Lack-of-fit (Linearity) | 0.50 mg/m ³ | 0.50 mg/m ³ |
| Zero drift from the field test | 0.25 mg/m ³ | 0.00 mg/m ³ |
| Span drift from the field test | 0.00 mg/m ³ | -0.90 mg/m ³ |
| Influence of ambient temperature at span point | 0.10 mg/m ³ | -0.55 mg/m ³ |
| Influence of sample gas pressure | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Influence of sample gas flow | 0.00 mg/m ³ | -0.05 mg/m ³ |
| Influence of voltage | 0.05 mg/m ³ | 0.20 mg/m ³ |
| Cross-sensitivity | 1.35 mg/m ³ | 1.75 mg/m ³ |
| Repeatability at span point | 0.04 mg/m ³ | 0.25 mg/m ³ |
| Standard deviation from paired measurements under field conditions | 0.17 mg/m ³ | 0.17 mg/m ³ |
| Uncertainty of provided reference material | 1.00 mg/m ³ | 1.00 mg/m ³ |
| Misalignment | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Changes of response factors | 0.00 mg/m ³ | 0.00 mg/m ³ |

Process characteristics

Standard uncertainty

| | | Zero point | Span point |
|--|-------------|--------------------------|---------------------------|
| Lack-of-fit (Linearity) | U_{lof} = | 0.2887 mg/m ³ | 0.2887 mg/m ³ |
| Zero drift from the field test | $U_{d,z}$ = | 0.1443 mg/m ³ | 0.0000 mg/m ³ |
| Span drift from the field test | $U_{d,s}$ = | 0.0000 mg/m ³ | -0.5196 mg/m ³ |
| Influence of ambient temperature at span point | U_t = | 0.0577 mg/m ³ | -0.3175 mg/m ³ |
| Influence of sample gas pressure | U_p = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Influence of sample gas flow | U_f = | 0.0000 mg/m ³ | -0.0289 mg/m ³ |
| Influence of voltage | U_v = | 0.0289 mg/m ³ | 0.1155 mg/m ³ |
| Cross-sensitivity | U_i = | 0.7794 mg/m ³ | 1.0104 mg/m ³ |
| Repeatability at span point | U_r = | 0.0231 mg/m ³ | 0.1443 mg/m ³ |
| Standard deviation from paired measurements under field conditions | U_D = | 0.1009 mg/m ³ | 0.1009 mg/m ³ |
| Uncertainty of provided reference material | U_{rm} = | 0.5774 mg/m ³ | 0.5774 mg/m ³ |
| Misalignment | U_{mb} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | U_{ce} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Changes of response factors | U_{rf} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |

Calculation of the combined standard uncertainties

| | | Zero point | Span point |
|-------------------------------|---------------|--------------------------|--------------------------|
| Combined standard uncertainty | s(AMS) values | 1.0292 mg/m ³ | 1.3577 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|-------------------|--|
| Combined standard uncertainty | 1.37 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 2.68 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 5.35 | % | of the measuring range of 50 mg/m ³ |
| Allowed expanded uncertainty | 20.00 | % | of the measuring range of 50 mg/m ³ |
| Allowed expanded uncertainty | 10.00 | mg/m ³ | |

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|-------------------------|----------------------------|------------------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | SO2 |

Input values

| | | | | | |
|-----------------------|-----|-------------------|-----------------------------------|----|-----|
| Certification range | 300 | mg/m ³ | Requirement to response time | 25 | % |
| Emissions limit value | 400 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 20 | % | | | |

General information

| | | | | | |
|-----------------------------|---|--------|------------------------|------|-------------------|
| Maintenance interval | 6 | months | Detection limit | 0.24 | mg/m ³ |
|-----------------------------|---|--------|------------------------|------|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 3.05 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|------------------------|------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 Vol% Water (H ₂ O) | 5.10 mg/m ³ | 6.60 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | 0.00 mg/m ³ | 3.30 mg/m ³ |
| 15 Vol% Carbon dioxide (CO ₂) | 3.00 mg/m ³ | 1.50 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | 0.00 mg/m ³ | 1.50 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 1.50 mg/m ³ |
| 300 mg/m ³ Nitrogen monoxide (NO) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Ammonia (NH ₃) | 0.00 mg/m ³ | 1.50 mg/m ³ |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | | |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | | |
| 50 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|------|-------------------|
| 8.10 | mg/m ³ |
| 0.00 | mg/m ³ |

| | |
|-------|-------------------|
| 15.90 | mg/m ³ |
| 0.00 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | SO2 |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

Largest difference according to type approval

| | Zero point | Span point |
|--|-------------------------|-------------------------|
| Lack-of-fit (Linearity) | 3.30 mg/m ³ | 3.30 mg/m ³ |
| Zero drift from the field test | -7.50 mg/m ³ | 0.00 mg/m ³ |
| Span drift from the field test | 0.00 mg/m ³ | 9.30 mg/m ³ |
| Influence of ambient temperature at span point | 3.90 mg/m ³ | -4.50 mg/m ³ |
| Influence of sample gas pressure | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Influence of sample gas flow | 0.00 mg/m ³ | -0.30 mg/m ³ |
| Influence of voltage | -2.40 mg/m ³ | 0.30 mg/m ³ |
| Cross-sensitivity | 8.10 mg/m ³ | 15.90 mg/m ³ |
| Repeatability at span point | 0.12 mg/m ³ | 0.16 mg/m ³ |
| Standard deviation from paired measurements under field conditions | 2.94 mg/m ³ | 2.94 mg/m ³ |
| Uncertainty of provided reference material | 6.00 mg/m ³ | 6.00 mg/m ³ |
| Misalignment | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Changes of response factors | 0.00 mg/m ³ | 0.00 mg/m ³ |

Process characteristics

Standard uncertainty

| | Zero point | Span point |
|--|---------------------------|---------------------------|
| Lack-of-fit (Linearity) | 1.9053 mg/m ³ | 1.9053 mg/m ³ |
| Zero drift from the field test | -4.3301 mg/m ³ | 0.0000 mg/m ³ |
| Span drift from the field test | 0.0000 mg/m ³ | 5.3694 mg/m ³ |
| Influence of ambient temperature at span point | 2.2517 mg/m ³ | -2.5981 mg/m ³ |
| Influence of sample gas pressure | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Influence of sample gas flow | 0.0000 mg/m ³ | -0.1732 mg/m ³ |
| Influence of voltage | -1.3856 mg/m ³ | 0.1732 mg/m ³ |
| Cross-sensitivity | 4.6765 mg/m ³ | 9.1799 mg/m ³ |
| Repeatability at span point | 0.0693 mg/m ³ | 0.0924 mg/m ³ |
| Standard deviation from paired measurements under field conditions | 1.6994 mg/m ³ | 1.6994 mg/m ³ |
| Uncertainty of provided reference material | 3.4641 mg/m ³ | 3.4641 mg/m ³ |
| Misalignment | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Changes of response factors | 0.0000 mg/m ³ | 0.0000 mg/m ³ |

Calculation of the combined standard uncertainties

| | Zero point | Span point |
|-------------------------------|--------------------------|---------------------------|
| Combined standard uncertainty | 8.1319 mg/m ³ | 11.7655 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|-------------------|---|
| Combined standard uncertainty | 12.61 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 24.72 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 6.18 | % | of the emissions limit value of 400 mg/m ³ |
| Allowed expanded uncertainty | 20.00 | % | of the emissions limit value of 400 mg/m ³ |
| Allowed expanded uncertainty | 80.00 | mg/m ³ | |

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|-------------------------|----------------------------|------------------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | HCI |

Input values

| | | | | | |
|-----------------------|-----|-------------------|-----------------------------------|----|-----|
| Certification range | 90 | mg/m ³ | Requirement to response time | 25 | % |
| Emissions limit value | 100 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 40 | % | | | |

General information

| | | | | | |
|-----------------------------|---|--------|------------------------|------|-------------------|
| Maintenance interval | 6 | months | Detection limit | 0.08 | mg/m ³ |
|-----------------------------|---|--------|------------------------|------|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 3.18 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|------------------------|------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 Vol% Water (H ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 15 Vol% Carbon dioxide (CO ₂) | 0.54 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | 0.81 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | 1.35 mg/m ³ | 1.71 mg/m ³ |
| 300 mg/m ³ Nitrogen monoxide (NO) | 0.54 mg/m ³ | 0.00 mg/m ³ |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | 0.00 mg/m ³ | 2.07 mg/m ³ |
| 20 mg/m ³ Ammonia (NH ₃) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | 1.08 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Hydrogen chloride (HCl) | | |
| 200 mg/m ³ Hydrogen chloride (HCl) | | |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|------|-------------------|
| 4.32 | mg/m ³ |
| 0.00 | mg/m ³ |

| | |
|------|-------------------|
| 3.78 | mg/m ³ |
| 0.00 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | HCI |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

| | Largest difference according to type approval | | | |
|--|---|-------------------|------------|-------------------|
| | Zero point | | Span point | |
| Lack-of-fit (Linearity) | 1.80 | mg/m ³ | 1.80 | mg/m ³ |
| Zero drift from the field test | -2.52 | mg/m ³ | 0.00 | mg/m ³ |
| Span drift from the field test | 0.00 | mg/m ³ | 2.70 | mg/m ³ |
| Influence of ambient temperature at span point | 0.72 | mg/m ³ | -3.15 | mg/m ³ |
| Influence of sample gas pressure | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Influence of sample gas flow | 0.00 | mg/m ³ | -0.09 | mg/m ³ |
| Influence of voltage | -0.45 | mg/m ³ | 0.63 | mg/m ³ |
| Cross-sensitivity | 4.32 | mg/m ³ | 3.78 | mg/m ³ |
| Repeatability at span point | 0.04 | mg/m ³ | 0.15 | mg/m ³ |
| Standard deviation from paired measurements under field conditions | 0.94 | mg/m ³ | 0.94 | mg/m ³ |
| Uncertainty of provided reference material | 1.80 | mg/m ³ | 1.80 | mg/m ³ |
| Misalignment | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Changes of response factors | 0.00 | mg/m ³ | 0.00 | mg/m ³ |

Process characteristics

| | | Standard uncertainty | | | |
|--|-------------|----------------------|-------------------|------------|-------------------|
| | | Zero point | | Span point | |
| Lack-of-fit (Linearity) | u_{lof} = | 1.0392 | mg/m ³ | 1.0392 | mg/m ³ |
| Zero drift from the field test | $u_{d,z}$ = | -1.4549 | mg/m ³ | 0.0000 | mg/m ³ |
| Span drift from the field test | $u_{d,s}$ = | 0.0000 | mg/m ³ | 1.5588 | mg/m ³ |
| Influence of ambient temperature at span point | u_t = | 0.4157 | mg/m ³ | -1.8187 | mg/m ³ |
| Influence of sample gas pressure | u_p = | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Influence of sample gas flow | u_f = | 0.0000 | mg/m ³ | -0.0520 | mg/m ³ |
| Influence of voltage | u_v = | -0.2598 | mg/m ³ | 0.3637 | mg/m ³ |
| Cross-sensitivity | u_i = | 2.4942 | mg/m ³ | 2.1824 | mg/m ³ |
| Repeatability at span point | u_r = | 0.0231 | mg/m ³ | 0.0866 | mg/m ³ |
| Standard deviation from paired measurements under field conditions | u_D = | 0.5410 | mg/m ³ | 0.5410 | mg/m ³ |
| Uncertainty of provided reference material | u_{rm} = | 1.0392 | mg/m ³ | 1.0392 | mg/m ³ |
| Misalignment | u_{mb} = | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Conversion rate of AMS for measurement of NOx | u_{ce} = | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Changes of response factors | u_{rf} = | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |

Calculation of the combined standard uncertainties

| | | Zero point | Span point |
|-------------------------------|---------------|--------------------------|--------------------------|
| Combined standard uncertainty | s(AMS) values | 3.3212 mg/m ³ | 3.6177 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|-------------------|---|
| Combined standard uncertainty | 4.08 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 8.00 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 8.00 | % | of the emissions limit value of 100 mg/m ³ |
| Allowed expanded uncertainty | 40.00 | % | of the emissions limit value of 100 mg/m ³ |
| Allowed expanded uncertainty | 40.00 | mg/m ³ | |

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | HF |

Input values

| | | | | | |
|-----------------------|----|-------------------|-----------------------------------|----|-----|
| Certification range | 10 | mg/m ³ | Requirement to response time | 25 | % |
| Emissions limit value | 10 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 40 | % | | | |

General information

| | | | | | |
|----------------------|---|--------|-----------------|------|-------------------|
| Maintenance interval | 3 | months | Detection limit | 0.08 | mg/m ³ |
|----------------------|---|--------|-----------------|------|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 3.30 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|--------------------------------|--------------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 Vol% Water (H ₂ O) | 0.15 mg/m ³ | 0.00 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | -0.28 mg/m ³ | -0.25 mg/m ³ |
| 15 Vol% Carbon dioxide (CO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | 0.23 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 300 mg/m ³ Nitrogen monoxide (NO) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | 0.00 mg/m ³ | 0.15 mg/m ³ |
| 20 mg/m ³ Ammonia (NH ₃) | 0.10 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Hydrogen chloride (HCl) | 0.10 mg/m ³ | -0.10 mg/m ³ |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|-------|-------------------|
| 0.58 | mg/m ³ |
| -0.28 | mg/m ³ |

| | |
|-------|-------------------|
| 0.15 | mg/m ³ |
| -0.35 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | HF |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

| Process characteristics | Largest difference according to type approval | | | |
|--|---|-------------------|------------|-------------------|
| | Zero point | | Span point | |
| Lack-of-fit (Linearity) | 0.17 | mg/m ³ | 0.17 | mg/m ³ |
| Zero drift from the field test | -0.39 | mg/m ³ | 0.00 | mg/m ³ |
| Span drift from the field test | 0.00 | mg/m ³ | -0.30 | mg/m ³ |
| Influence of ambient temperature at span point | -0.37 | mg/m ³ | 0.47 | mg/m ³ |
| Influence of sample gas pressure | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Influence of sample gas flow | 0.00 | mg/m ³ | -0.01 | mg/m ³ |
| Influence of voltage | 0.10 | mg/m ³ | 0.13 | mg/m ³ |
| Cross-sensitivity | 0.58 | mg/m ³ | -0.35 | mg/m ³ |
| Repeatability at span point | 0.04 | mg/m ³ | 0.05 | mg/m ³ |
| Standard deviation from paired measurements under field conditions | 0.13 | mg/m ³ | 0.13 | mg/m ³ |
| Uncertainty of provided reference material | 0.20 | mg/m ³ | 0.20 | mg/m ³ |
| Misalignment | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Changes of response factors | 0.00 | mg/m ³ | 0.00 | mg/m ³ |

Process characteristics

| | | Standard uncertainty | | | |
|--|-------------|----------------------|-------------------|------------|-------------------|
| | | Zero point | | Span point | |
| Lack-of-fit (Linearity) | $u_{lof} =$ | 0.0981 | mg/m ³ | 0.0981 | mg/m ³ |
| Zero drift from the field test | $u_{d,z} =$ | -0.2252 | mg/m ³ | 0.0000 | mg/m ³ |
| Span drift from the field test | $u_{d,s} =$ | 0.0000 | mg/m ³ | -0.1732 | mg/m ³ |
| Influence of ambient temperature at span point | $u_t =$ | -0.2136 | mg/m ³ | 0.2714 | mg/m ³ |
| Influence of sample gas pressure | $u_p =$ | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Influence of sample gas flow | $u_f =$ | 0.0000 | mg/m ³ | -0.0058 | mg/m ³ |
| Influence of voltage | $u_v =$ | 0.0577 | mg/m ³ | 0.0751 | mg/m ³ |
| Cross-sensitivity | $u_i =$ | 0.3349 | mg/m ³ | -0.2021 | mg/m ³ |
| Repeatability at span point | $u_r =$ | 0.0231 | mg/m ³ | 0.0289 | mg/m ³ |
| Standard deviation from paired measurements under field conditions | $u_D =$ | 0.0736 | mg/m ³ | 0.0736 | mg/m ³ |
| Uncertainty of provided reference material | $u_{rm} =$ | 0.1155 | mg/m ³ | 0.1155 | mg/m ³ |
| Misalignment | $u_{mb} =$ | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Conversion rate of AMS for measurement of NOx | $u_{ce} =$ | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Changes of response factors | $u_{rf} =$ | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |

Calculation of the combined standard uncertainties

| Combined standard uncertainty | s(AMS) values | Zero point | Span point |
|-------------------------------|---------------|--------------------------|--------------------------|
| | | 0.4901 mg/m ³ | 0.4225 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|-------------------|--|
| Combined standard uncertainty | 0.55 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 1.07 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 10.74 | % | of the emissions limit value of 10 mg/m ³ |
| Allowed expanded uncertainty | 40.00 | % | of the emissions limit value of 10 mg/m ³ |
| Allowed expanded uncertainty | 4.00 | mg/m ³ | |

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|-------------------------|----------------------------|------------------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | NH3 |

Input values

| | | | | | |
|---------------------|----|-------------------|-----------------------------------|----|-----|
| Certification range | 50 | mg/m ³ | Requirement to response time | 25 | % |
| Measuring range | 60 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 40 | % | | | * |

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

General information

| | | | | | |
|-----------------------------|---|--------|------------------------|------|-------------------|
| Maintenance interval | 3 | months | Detection limit | 0.05 | mg/m ³ |
|-----------------------------|---|--------|------------------------|------|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 3.32 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|------------------------|-------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 Vol% Water (H ₂ O) | 0.40 mg/m ³ | 0.00 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | 0.35 mg/m ³ | 0.00 mg/m ³ |
| 15 Vol% Carbon dioxide (CO ₂) | 0.45 mg/m ³ | -0.50 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | 0.60 mg/m ³ | -0.20 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.20 mg/m ³ |
| 300 mg/m ³ Nitrogen monoxide (NO) | 0.65 mg/m ³ | -0.50 mg/m ³ |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | 0.00 mg/m ³ | -0.25 mg/m ³ |
| 20 mg/m ³ Ammonia (NH ₃) | | |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 50 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|------|-------------------|
| 2.45 | mg/m ³ |
| 0.00 | mg/m ³ |

| | |
|-------|-------------------|
| 0.20 | mg/m ³ |
| -1.45 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | NH3 |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

| | Largest difference according to type approval | | | |
|--|---|-------------------|------------|-------------------|
| | Zero point | | Span point | |
| Lack-of-fit (Linearity) | -0.30 | mg/m ³ | -0.30 | mg/m ³ |
| Zero drift from the field test | 0.25 | mg/m ³ | 0.00 | mg/m ³ |
| Span drift from the field test | 0.00 | mg/m ³ | 1.47 | mg/m ³ |
| Influence of ambient temperature at span point | -0.25 | mg/m ³ | -0.70 | mg/m ³ |
| Influence of sample gas pressure | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Influence of sample gas flow | 0.00 | mg/m ³ | -0.05 | mg/m ³ |
| Influence of voltage | -0.25 | mg/m ³ | 0.35 | mg/m ³ |
| Cross-sensitivity | 2.45 | mg/m ³ | -1.45 | mg/m ³ |
| Repeatability at span point | 0.02 | mg/m ³ | 0.07 | mg/m ³ |
| Standard deviation from paired measurements under field conditions | 0.38 | mg/m ³ | 0.38 | mg/m ³ |
| Uncertainty of provided reference material | 1.00 | mg/m ³ | 1.00 | mg/m ³ |
| Misalignment | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 | mg/m ³ | 0.00 | mg/m ³ |
| Changes of response factors | 0.00 | mg/m ³ | 0.00 | mg/m ³ |

Process characteristics

| | | Standard uncertainty | | | |
|--|-------------|----------------------|-------------------|------------|-------------------|
| | | Zero point | | Span point | |
| Lack-of-fit (Linearity) | U_{lof} = | -0.1732 | mg/m ³ | -0.1732 | mg/m ³ |
| Zero drift from the field test | $U_{d,z}$ = | 0.1443 | mg/m ³ | 0.0000 | mg/m ³ |
| Span drift from the field test | $U_{d,s}$ = | 0.0000 | mg/m ³ | 0.8487 | mg/m ³ |
| Influence of ambient temperature at span point | U_t = | -0.1443 | mg/m ³ | -0.4041 | mg/m ³ |
| Influence of sample gas pressure | U_p = | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Influence of sample gas flow | U_f = | 0.0000 | mg/m ³ | -0.0289 | mg/m ³ |
| Influence of voltage | U_v = | -0.1443 | mg/m ³ | 0.2021 | mg/m ³ |
| Cross-sensitivity | U_i = | 1.4145 | mg/m ³ | -0.8372 | mg/m ³ |
| Repeatability at span point | U_r = | 0.0115 | mg/m ³ | 0.0404 | mg/m ³ |
| Standard deviation from paired measurements under field conditions | U_D = | 0.2198 | mg/m ³ | 0.2198 | mg/m ³ |
| Uncertainty of provided reference material | U_{rm} = | 0.5774 | mg/m ³ | 0.5774 | mg/m ³ |
| Misalignment | U_{mb} = | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Conversion rate of AMS for measurement of NOx | U_{ce} = | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |
| Changes of response factors | U_{rf} = | 0.0000 | mg/m ³ | 0.0000 | mg/m ³ |

Calculation of the combined standard uncertainties

| | | Zero point | Span point |
|-------------------------------|---------------|--------------------------|--------------------------|
| Combined standard uncertainty | s(AMS) values | 1.5732 mg/m ³ | 1.4275 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|-------------------|--|
| Combined standard uncertainty | 1.83 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 3.59 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 5.99 | % | of the measuring range of 60 mg/m ³ |
| Allowed expanded uncertainty | 40.00 | % | of the measuring range of 60 mg/m ³ |
| Allowed expanded uncertainty | 24.00 | mg/m ³ | |

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | H2O |

Input values

| | | | | | |
|---------------------|----|------|-----------------------------------|----|-----|
| Certification range | 40 | Vol% | Requirement to response time | 25 | % |
| Measuring range | 40 | Vol% | Averaging time of measured values | 30 | min |
| Confidence interval | 40 | % | | | * |

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

General information

| | | | | | |
|----------------------|---|--------|-----------------|------|------|
| Maintenance interval | 6 | months | Detection limit | 0.04 | Vol% |
|----------------------|---|--------|-----------------|------|------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 2.93 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|-----------------------------------|-------------------|-------------------|
| 3 Vol% Oxygen (O2) | 0.00 Vol% | 0.00 Vol% |
| 21 Vol% Oxygen (O2) | 0.00 Vol% | 0.00 Vol% |
| 30 Vol% Water (H2O) | | |
| 300 mg/m³ Carbon monoxide (CO) | 0.80 Vol% | 0.76 Vol% |
| 15 Vol% Carbon dioxide (CO2) | 0.00 Vol% | 0.00 Vol% |
| 50 mg/m³ Methane (CH4) | -0.20 Vol% | -0.36 Vol% |
| 20 mg/m³ Dinitrogen oxide (N2O) | 0.00 Vol% | 0.00 Vol% |
| 100 mg/m³ Dinitrogen oxide (N2O) | 0.00 Vol% | 0.00 Vol% |
| 300 mg/m³ Nitrogen monoxide (NO) | 0.00 Vol% | 0.00 Vol% |
| 30 mg/m³ Nitrogen dioxide (NO2) | 0.00 Vol% | 0.00 Vol% |
| 20 mg/m³ Ammonia (NH3) | 0.00 Vol% | -0.20 Vol% |
| 200 mg/m³ Sulfur dioxide (SO2) | 0.00 Vol% | 0.00 Vol% |
| 1000 mg/m³ Sulfur dioxide (SO2) | 0.00 Vol% | 0.00 Vol% |
| 50 mg/m³ Hydrogen chloride (HCl) | 0.00 Vol% | 0.00 Vol% |
| 200 mg/m³ Hydrogen chloride (HCl) | 0.00 Vol% | -0.20 Vol% |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|-------|------|
| 0.80 | Vol% |
| -0.20 | Vol% |

| | |
|-------|------|
| 0.76 | Vol% |
| -0.76 | Vol% |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | H2O |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

| | Largest difference according to type approval | | | |
|--|---|------|------------|------|
| | Zero point | | Span point | |
| Lack-of-fit (Linearity) | 0.64 | Vol% | 0.64 | Vol% |
| Zero drift from the field test | -1.04 | Vol% | 0.00 | Vol% |
| Span drift from the field test | 0.00 | Vol% | 1.16 | Vol% |
| Influence of ambient temperature at span point | 0.12 | Vol% | 0.48 | Vol% |
| Influence of sample gas pressure | 0.00 | Vol% | 0.00 | Vol% |
| Influence of sample gas flow | 0.00 | Vol% | -0.04 | Vol% |
| Influence of voltage | 0.08 | Vol% | 0.00 | Vol% |
| Cross-sensitivity | 0.80 | Vol% | 0.76 | Vol% |
| Repeatability at span point | 0.02 | Vol% | 0.06 | Vol% |
| Standard deviation from paired measurements under field conditions | 0.39 | Vol% | 0.39 | Vol% |
| Uncertainty of provided reference material | 0.80 | Vol% | 0.80 | Vol% |
| Misalignment | 0.00 | Vol% | 0.00 | Vol% |
| Conversion rate of AMS for measurement of NOx | 0.00 | Vol% | 0.00 | Vol% |
| Changes of response factors | 0.00 | Vol% | 0.00 | Vol% |

Process characteristics

| | | Standard uncertainty | | | |
|--|-----------|----------------------|--------------|------------|------|
| | | Zero point | | Span point | |
| Lack-of-fit (Linearity) | u_{lof} | = | 0.3695 Vol% | 0.3695 | Vol% |
| Zero drift from the field test | $u_{d,z}$ | = | -0.6004 Vol% | 0.0000 | Vol% |
| Span drift from the field test | $u_{d,s}$ | = | 0.0000 Vol% | 0.6697 | Vol% |
| Influence of ambient temperature at span point | u_t | = | 0.0693 Vol% | 0.2771 | Vol% |
| Influence of sample gas pressure | u_p | = | 0.0000 Vol% | 0.0000 | Vol% |
| Influence of sample gas flow | u_f | = | 0.0000 Vol% | -0.0231 | Vol% |
| Influence of voltage | u_v | = | 0.0462 Vol% | 0.0000 | Vol% |
| Cross-sensitivity | u_i | = | 0.4619 Vol% | 0.4388 | Vol% |
| Repeatability at span point | u_r | = | 0.0115 Vol% | 0.0346 | Vol% |
| Standard deviation from paired measurements under field conditions | u_D | = | 0.2266 Vol% | 0.2266 | Vol% |
| Uncertainty of provided reference material | u_{rm} | = | 0.4619 Vol% | 0.4619 | Vol% |
| Misalignment | u_{mb} | = | 0.0000 Vol% | 0.0000 | Vol% |
| Conversion rate of AMS for measurement of NOx | u_{ce} | = | 0.0000 Vol% | 0.0000 | Vol% |
| Changes of response factors | u_{rf} | = | 0.0000 Vol% | 0.0000 | Vol% |

Calculation of the combined standard uncertainties

| Combined standard uncertainty | s(AMS) values | Zero point | | Span point | |
|-------------------------------|---------------|------------|------|------------|------|
| | | 0.9910 | Vol% | 1.0581 | Vol% |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|------|-----------------------------------|
| Combined standard uncertainty | 1.23 | Vol% | according to EN 15267-3 |
| Expanded uncertainty | 2.40 | Vol% | according to EN 15267-3 |
| Relative expanded uncertainty | 6.01 | % | of the measuring range of 40 Vol% |
| Allowed expanded uncertainty | 40.00 | % | of the measuring range of 40 Vol% |
| Allowed expanded uncertainty | 16.00 | Vol% | |

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|-------------------------|----------------------------|------------------|-------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | Corg |

Input values

| | | | | | |
|-----------------------|-----|-------------------|-----------------------------------|----|-----|
| Certification range | 50 | mg/m ³ | Requirement to response time | 25 | % |
| Emissions limit value | 100 | mg/m ³ | Averaging time of measured values | 30 | min |
| Confidence interval | 30 | % | | | |

General information

| | | | | | |
|-----------------------------|---|--------|------------------------|---|-------------------|
| Maintenance interval | 2 | months | Detection limit | 0 | mg/m ³ |
|-----------------------------|---|--------|------------------------|---|-------------------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 0.82 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | Span point |
|---|-------------------------------|--------------------------------|
| 3 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 21 Vol% Oxygen (O ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 30 Vol% Water (H ₂ O) | 0.57 mg/m ³ | 0.60 mg/m ³ |
| 300 mg/m ³ Carbon monoxide (CO) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 15 Vol% Carbon dioxide (CO ₂) | 0.44 mg/m ³ | -0.50 mg/m ³ |
| 50 mg/m ³ Methane (CH ₄) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 100 mg/m ³ Dinitrogen oxide (N ₂ O) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 300 mg/m ³ Nitrogen monoxide (NO) | 0.27 mg/m ³ | 0.00 mg/m ³ |
| 30 mg/m ³ Nitrogen dioxide (NO ₂) | 0.27 mg/m ³ | 0.00 mg/m ³ |
| 20 mg/m ³ Ammonia (NH ₃) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 1000 mg/m ³ Sulfur dioxide (SO ₂) | 0.00 mg/m ³ | -0.27 mg/m ³ |
| 50 mg/m ³ Hydrogen chloride (HCl) | 0.00 mg/m ³ | 0.00 mg/m ³ |
| 200 mg/m ³ Hydrogen chloride (HCl) | 0.27 mg/m ³ | 0.30 mg/m ³ |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|------|-------------------|
| 1.80 | mg/m ³ |
| 0.00 | mg/m ³ |

| | |
|-------|-------------------|
| 0.90 | mg/m ³ |
| -0.77 | mg/m ³ |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | Corg |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

Largest difference according to type approval

| | Zero point | Span point |
|--|-------------------------|-------------------------|
| Lack-of-fit (Linearity) | 0.34 mg/m ³ | 0.34 mg/m ³ |
| Zero drift from the field test | 0.88 mg/m ³ | 0.00 mg/m ³ |
| Span drift from the field test | 0.00 mg/m ³ | -1.41 mg/m ³ |
| Influence of ambient temperature at span point | 0.70 mg/m ³ | -0.65 mg/m ³ |
| Influence of sample gas pressure | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Influence of sample gas flow | 0.30 mg/m ³ | -0.35 mg/m ³ |
| Influence of voltage | -0.05 mg/m ³ | 0.25 mg/m ³ |
| Cross-sensitivity | 1.80 mg/m ³ | 0.90 mg/m ³ |
| Repeatability at span point | 0.00 mg/m ³ | 0.01 mg/m ³ |
| Standard deviation from paired measurements under field conditions | 0.15 mg/m ³ | 0.15 mg/m ³ |
| Uncertainty of provided reference material | 1.00 mg/m ³ | 1.00 mg/m ³ |
| Misalignment | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | 0.00 mg/m ³ | 0.00 mg/m ³ |
| Changes of response factors | 3.27 mg/m ³ | 3.27 mg/m ³ |

Process characteristics

Standard uncertainty

| | | Zero point | Span point |
|--|-------------|---------------------------|---------------------------|
| Lack-of-fit (Linearity) | u_{lof} = | 0.1934 mg/m ³ | 0.1934 mg/m ³ |
| Zero drift from the field test | $u_{d,z}$ = | 0.5052 mg/m ³ | 0.0000 mg/m ³ |
| Span drift from the field test | $u_{d,s}$ = | 0.0000 mg/m ³ | -0.8112 mg/m ³ |
| Influence of ambient temperature at span point | u_t = | 0.4041 mg/m ³ | -0.3753 mg/m ³ |
| Influence of sample gas pressure | u_p = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Influence of sample gas flow | u_f = | 0.1732 mg/m ³ | -0.2021 mg/m ³ |
| Influence of voltage | u_v = | -0.0289 mg/m ³ | 0.1443 mg/m ³ |
| Cross-sensitivity | u_i = | 1.0363 mg/m ³ | 0.5196 mg/m ³ |
| Repeatability at span point | u_r = | 0.0000 mg/m ³ | 0.0058 mg/m ³ |
| Standard deviation from paired measurements under field conditions | u_D = | 0.0877 mg/m ³ | 0.0877 mg/m ³ |
| Uncertainty of provided reference material | u_{rm} = | 0.5774 mg/m ³ | 0.5774 mg/m ³ |
| Misalignment | u_{mb} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Conversion rate of AMS for measurement of NOx | u_{ce} = | 0.0000 mg/m ³ | 0.0000 mg/m ³ |
| Changes of response factors | u_{rf} = | 1.8850 mg/m ³ | 1.8850 mg/m ³ |

Calculation of the combined standard uncertainties

| | | Zero point | Span point |
|-------------------------------|---------------|--------------------------|--------------------------|
| Combined standard uncertainty | s(AMS) values | 2.3356 mg/m ³ | 2.2500 mg/m ³ |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|-------------------|---|
| Combined standard uncertainty | 2.47 | mg/m ³ | according to EN 15267-3 |
| Expanded uncertainty | 4.85 | mg/m ³ | according to EN 15267-3 |
| Relative expanded uncertainty | 4.85 | % | of the emissions limit value of 100 mg/m ³ |
| Allowed expanded uncertainty | 30.00 | % | of the emissions limit value of 100 mg/m ³ |
| Allowed expanded uncertainty | 30.00 | mg/m ³ | |

Result

Requirements fulfilled

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|-------------------------|----------------------------|------------------|------------|
| Customer | DGtek / Volund GMAB Margam | | |
| Identification | ZTA-4943685 | | |
| Serial number | 1628 0511 | Date | 2016-09-14 |
| Measuring system | MCS100FT | Component | O2 |

Input values

| | | | | | |
|---------------------|----|------|-----------------------------------|----|-----|
| Certification range | 21 | Vol% | Requirement to response time | 25 | % |
| Measuring range | 25 | Vol% | Averaging time of measured values | 30 | min |
| Confidence interval | 20 | % | | | * |

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.

General information

| | | | | | |
|-----------------------------|---|-------|------------------------|------|------|
| Maintenance interval | 4 | weeks | Detection limit | 0.03 | Vol% |
|-----------------------------|---|-------|------------------------|------|------|

Required performance regarding dynamic operating conditions

| | | | | |
|------------------------------|------|-----|-------------------------------------|--|
| Measured response time | 2.27 | min | | |
| Requirement to response time | 7.50 | min | 25% of the averaging time of 30 min | |

Result

Requirements fulfilled

Calculation of the expanded uncertainty

| Interferent | Zero point | | Span point | |
|-----------------------------------|------------|------|------------|------|
| 3 Vol% Oxygen (O2) | 0.00 | Vol% | 0.00 | Vol% |
| 21 Vol% Oxygen (O2) | 0.00 | Vol% | 0.00 | Vol% |
| 30 Vol% Water (H2O) | 0.00 | Vol% | 0.00 | Vol% |
| 300 mg/m³ Carbon monoxide (CO) | 0.00 | Vol% | 0.00 | Vol% |
| 15 Vol% Carbon dioxide (CO2) | 0.00 | Vol% | 0.00 | Vol% |
| 50 mg/m³ Methane (CH4) | 0.00 | Vol% | 0.00 | Vol% |
| 20 mg/m³ Dinitrogen oxide (N2O) | 0.00 | Vol% | 0.00 | Vol% |
| 100 mg/m³ Dinitrogen oxide (N2O) | 0.00 | Vol% | 0.00 | Vol% |
| 300 mg/m³ Nitrogen monoxide (NO) | 0.00 | Vol% | 0.00 | Vol% |
| 30 mg/m³ Nitrogen dioxide (NO2) | 0.00 | Vol% | 0.00 | Vol% |
| 20 mg/m³ Ammonia (NH3) | 0.00 | Vol% | 0.00 | Vol% |
| 200 mg/m³ Sulfur dioxide (SO2) | 0.00 | Vol% | 0.00 | Vol% |
| 1000 mg/m³ Sulfur dioxide (SO2) | 0.00 | Vol% | 0.00 | Vol% |
| 50 mg/m³ Hydrogen chloride (HCl) | 0.00 | Vol% | 0.00 | Vol% |
| 200 mg/m³ Hydrogen chloride (HCl) | 0.00 | Vol% | 0.00 | Vol% |

Sum of the positive cross-sensitivities
Sum of the negative cross-sensitivities

| | |
|------|------|
| 0.00 | Vol% |
| 0.00 | Vol% |

| | |
|------|------|
| 0.00 | Vol% |
| 0.00 | Vol% |

Calculation of measurement uncertainty

according to EN ISO 14956, EN 14181 and EN 15267-3

Version 5.2

Device data

| | | | |
|------------------|----------------------------|-----------|------------|
| Customer | DGtek / Volund GMAB Margam | Date | 2016-09-14 |
| Identification | ZTA-4943685 | Component | O2 |
| Serial number | 1628 0511 | | |
| Measuring system | MCS100FT | | |

Influences of the process characteristics

Process characteristics

| | Largest difference according to type approval | | | |
|--|---|------|------------|------|
| | Zero point | | Span point | |
| Lack-of-fit (Linearity) | -0.14 | Vol% | -0.14 | Vol% |
| Zero drift from the field test | 0.18 | Vol% | 0.00 | Vol% |
| Span drift from the field test | 0.00 | Vol% | -0.20 | Vol% |
| Influence of ambient temperature at span point | 0.02 | Vol% | 0.24 | Vol% |
| Influence of sample gas pressure | 0.00 | Vol% | 0.00 | Vol% |
| Influence of sample gas flow | -0.02 | Vol% | 0.01 | Vol% |
| Influence of voltage | 0.01 | Vol% | -0.10 | Vol% |
| Cross-sensitivity | 0.00 | Vol% | 0.00 | Vol% |
| Repeatability at span point | 0.01 | Vol% | 0.01 | Vol% |
| Standard deviation from paired measurements under field conditions | 0.09 | Vol% | 0.09 | Vol% |
| Uncertainty of provided reference material | 0.42 | Vol% | 0.42 | Vol% |
| Misalignment | 0.00 | Vol% | 0.00 | Vol% |
| Conversion rate of AMS for measurement of NOx | 0.00 | Vol% | 0.00 | Vol% |
| Changes of response factors | 0.00 | Vol% | 0.00 | Vol% |

Process characteristics

| | | Standard uncertainty | | | |
|--|-------------|----------------------|------|------------|------|
| | | Zero point | | Span point | |
| Lack-of-fit (Linearity) | u_{lof} = | -0.0808 | Vol% | -0.0808 | Vol% |
| Zero drift from the field test | $u_{d,z}$ = | 0.1039 | Vol% | 0.0000 | Vol% |
| Span drift from the field test | $u_{d,s}$ = | 0.0000 | Vol% | -0.1155 | Vol% |
| Influence of ambient temperature at span point | u_t = | 0.0115 | Vol% | 0.1386 | Vol% |
| Influence of sample gas pressure | u_p = | 0.0000 | Vol% | 0.0000 | Vol% |
| Influence of sample gas flow | u_f = | -0.0115 | Vol% | 0.0058 | Vol% |
| Influence of voltage | u_v = | 0.0058 | Vol% | -0.0577 | Vol% |
| Cross-sensitivity | u_i = | 0.0000 | Vol% | 0.0000 | Vol% |
| Repeatability at span point | u_r = | 0.0058 | Vol% | 0.0058 | Vol% |
| Standard deviation from paired measurements under field conditions | u_D = | 0.0533 | Vol% | 0.0533 | Vol% |
| Uncertainty of provided reference material | u_{rm} = | 0.2425 | Vol% | 0.2425 | Vol% |
| Misalignment | u_{mb} = | 0.0000 | Vol% | 0.0000 | Vol% |
| Conversion rate of AMS for measurement of NOx | u_{ce} = | 0.0000 | Vol% | 0.0000 | Vol% |
| Changes of response factors | u_{rf} = | 0.0000 | Vol% | 0.0000 | Vol% |

Calculation of the combined standard uncertainties

| Combined standard uncertainty | s(AMS) values | Zero point | | Span point | |
|-------------------------------|---------------|------------|------|------------|------|
| | | 0.2816 | Vol% | 0.3226 | Vol% |

Verification of compliance with the requirements

| | | | |
|-------------------------------|-------|------|-----------------------------------|
| Combined standard uncertainty | 0.34 | Vol% | according to EN 15267-3 |
| Expanded uncertainty | 0.66 | Vol% | according to EN 15267-3 |
| Relative expanded uncertainty | 2.66 | % | of the measuring range of 25 Vol% |
| Allowed expanded uncertainty | 20.00 | % | of the measuring range of 25 Vol% |
| Allowed expanded uncertainty | 5.00 | Vol% | |

Result

Requirements fulfilled

Attention: The 2001/80/EC and 2000/76/EC gives no requirements for these components.