



## Schedule 6 - Notification

Part B - to be submitted as soon as practicable

Permit Number	BR9383ID	Notification Reference	EP_EX_142
Name of operator	Knauf Insulation Ltd		
Location of Facility	Chemistry Lane, Queensferry, Deeside, Flintshire, CH5 2DA		
Time and date of the detection	12/9/14		

Any more accurate information on the matters for notification under Part A.	Detailed information provided in the Part A Notification
Measures taken, or intended to be taken, to prevent a recurrence of the incident	Enviro Technology who are contracted to maintain the equipment came to site on the 12/9/14 to determine why the CEMS equipment was not fully functioning. They discovered that there were no readings from the AR 650 s/N 650-1-192 and that flow measurements were dropping to nil- generally after 17:00hrs even though gases were flowing through the system.
Measures taken, or intended to be taken, to rectify, limit or prevent any pollution of the environment which has been or may be caused by the emission	<p>The engineers report of his finding are as follows :-</p> <ol style="list-style-type: none"> <li>1. On arrival the analyser not measuring as the PC was continually rebooting</li> <li>2. Laser found to be not working.</li> <li>3. Replaced the laser and optimized values by adjustment.</li> <li>4. On completion the system was put back into measure mode and normal operation resumed.</li> <li>5. Flow problem: Durag probes DR-X 250T and DR-X 250D were checked for any obvious fault conditions – none were found.</li> </ol> <p>Data loss from DRX seems to be prevalent after 17:00hrs . System will be checked remotely by Enviro Technology for anything in the OPSIS panel that may be timing out</p>
The dates of any unauthorised emissions from the facility in the preceding 24 months.	10/08/12, 04/09/12, 19/11/12, 21/11/12, 28/11/12, 30/01/13, 07/02/13, 26/02/13, 11/03/13, 21/03/13, 22/03/13, 04/04/13, 07/04/13, 18/04/13, 26/04/13, 09/05/13, 21/05/13, 31/05/13, 25/06/13,27/09/13, 04/11/13, 14/11/13, 20/11/13, 29/11/13,04/12/13, 06/12/13, 20/01/14, 05/02/14, 12/02/14, 13/03/14, 08/04/14, 214/05/14, 29/06/14, 02/07/14, 11/7/14,19/7/14, 20/7/14, 21/7/14.

Name*	Nigel Heyes
Post	Systems Manager
Signature	
Date	7 <sup>th</sup> October 2014

\* authorised to sign on behalf of Knauf Insulation Ltd

