


Schedule 6 - Notification

Part B - to be submitted as soon as practicable

Permit Number	BR9383ID	Notification Reference	EP_EX_183 EP_EX_184 EP_EX_185
Name of operator	Knauf Insulation Ltd		
Location of Facility	Chemistry Lane, Queensferry, Deeside, Flintshire, CH5 2DA		
Time and date of the detection	EP_EX_183: Day of 2 nd October 2017 EP_EX_184: Day of 3 rd October 2017 EP_EX_185: Day of 4 th October 2017		

Any more accurate information on the matters for notification under Part A.	The validated Daily Averages were as follows: EP_EX_183: Daily Average SO2 emission of 1458.24mg/Nm3 EP_EX_184: Daily Average SO2 emission of 1565.28mg/Nm3 EP_EX_185: Daily Average SO2 emission of 1501.49mg/Nm3
Measures taken, or intended to be taken, to prevent a recurrence of the incident	On investigation by the CEMs equipment specialist, the oxygen probe was failing to consistently Span check. Therefore, the oxygen reading during this period was not being corrected for oxygen resulting in CEMs data recording SO2 values in excess of the Emission Limit Value. The oxygen probe was removed and the spare probe installed.
Measures taken, or intended to be taken, to rectify, limit or prevent any pollution of the environment which has been or may be caused by the emission	It is believed that no permanent harm will result from the incident.
The dates of any unauthorised emissions from the facility in the preceding 24 months.	5-7/10/15, 27/10/15, 02/02/16, 23/02/16, 15/06/16, 27/06/16, 28/06/16, 10/07/16, 15/08/16, 16/08/16, 13/09/16, 14/09/16, 25/04/2017, 24/05/17, 07/07/17, 04/08/17, 14-15/08/17, 06/09/17, 14/09/17, 16-20/09/17

Name*	C. Keouski
Post	HSE Manager
Signature	
Date	4 th December 2017

* authorised to sign on behalf of Knauf Insulation Ltd