

STACK EMISSIONS MONITORING REPORT



Unit 5 Crown Industrial Estate
Kenwood Road
Stockport
SK5 6PH
Tel: 0161 443 0980
Fax: 0161 443 0989

| Your contact at ESG |
|---|
| <p>Mark Woodruff Business Manager - North Tel: 0161 443 0982 Email: mark.woodruff@esg.co.uk</p> |

| Operator & Address: |
|--|
| <p>Eni Liverpool Bay Operating Company Ltd Point Of Ayr Terminal Station Road Holywell Clwyd CH8 9RD</p> |

| Permit: |
|----------------------|
| EPR Permit: ZP3331LM |

| Release Point: |
|------------------|
| Thermal Oxidiser |

| Sampling Date(s): |
|-------------------|
| 19th May 2014 |

| | |
|-------------------------|------------------------------|
| ESG Job Number: | LNO 11828 |
| Report Date: | 20th May 2014 |
| Version: | 1 |
| Report By: | Keith Bird |
| MCERTS Number: | MM 07 825 |
| MCERTS Level: | MCERTS Level 2 - Team Leader |
| Technical Endorsements: | 1, 2, 3 & 4 |
| Report Approved By: | Dave Armitage |
| MCERTS Number: | MM 04 516 |
| Business Title: | MCERTS Level 2 - Team Leader |
| Technical Endorsements: | 1, 2, 3 & 4 |
| Signature: | |



1015

CONTENTS

EXECUTIVE SUMMARY

Stack Emissions Monitoring Objectives

- Plant
- Operator
- Stack Emissions Monitoring Test House

Emissions Summary

Monitoring Times

Process Details

Monitoring Methods

Analytical Methods

- Sampling Methods with Subsequent Analysis
- On-Site Testing

Sampling Location

- Sampling Plane Validation Criteria
- Duct Characteristics
- Sampling Lines & Sample Points
- Sampling Platform
- Sampling Location / Platform Improvement Recommendations

Sampling and Analytical Method Deviations

APPENDICES

APPENDIX 1 - Monitoring Schedule, Calibration Checklist & Monitoring Team

APPENDIX 2 - Summaries, Calculations, Raw Data and Charts

APPENDIX 3 - Measurement Uncertainty Budget Calculations

EXECUTIVE SUMMARY

MONITORING OBJECTIVES

Eni Liverpool Bay Operating Company Ltd operates a gas enrichment process at Point Of Ayr Terminal which is subject to EPR Permit ZP3331LM, under the Environmental Permitting Regulations 2010.

Environmental Scientifics Group Limited were commissioned by Eni Liverpool Bay Operating Company Ltd to carry out stack emissions monitoring to determine the release of prescribed pollutants from the following Plant under normal operating conditions.

The results of these tests shall be used to demonstrate compliance with a set of emission limit values for prescribed pollutants as specified in the Plant's EPR Permit, ZP3331LM.

Plant

Thermal Oxidiser

Operator

Eni Liverpool Bay Operating Company Ltd
Point Of Ayr Terminal
Station Road
Holywell
Clwyd
CH8 9RD

EPR Permit: ZP3331LM

Stack Emissions Monitoring Test House

Environmental Scientifics Group Limited - Stockport Laboratory
Unit 5 Crown Industrial Estate
Kenwood Road
Stockport
SK5 6PH
UKAS and MCERTS Accreditation Number: 1015

Opinions and interpretations expressed herein are outside the scope of UKAS accreditation.
MCERTS accredited results will only be claimed where both the sampling and analytical stages are UKAS accredited.
This test report shall not be reproduced, except in full, without written approval of Environmental Scientifics Group Limited.

EXECUTIVE SUMMARY

*
↓

| EMISSIONS SUMMARY | | | | | |
|--|-------------------|--------|-------------------------------|-------|--------------------------|
| Parameter | Units | Result | Calculated Uncertainty +/- | Limit | MCERTS accredited result |
| Oxides of Nitrogen (as NO ₂) | mg/m ³ | 40 | 3.5 | 120 | ✓ |
| Oxides of Nitrogen (as NO ₂) Emission Rate | g/hr | - | - | - | ✓ |
| Sulphur Dioxide | mg/m ³ | 0.91 | 4.7 | 190 | ✓ |
| Sulphur Dioxide Emission Rate | g/hr | - | - | - | ✓ |
| Carbon Monoxide | mg/m ³ | 15 | 2.1 | - | ✓ |
| Carbon Monoxide Emission Rate | g/hr | - | - | - | ✓ |
| Carbon Dioxide | % v/v | 7.3 | 0.30 | - | ✓ |
| Oxygen | % v/v | 9.9 | 0.32 | - | ✓ |

ND = None Detected,

Results at or below the limit of detection are highlighted by bold italic text.

The above volumetric flow rate is calculated using data from the preliminary survey. Mass emissions for non isokinetic tests are calculated using these values. For all isokinetic testing the mass emission is calculated using test specific flow data and not the above values.

Reference conditions are 273K, 101.3kPa, dry gas .

EXECUTIVE SUMMARY

| MONITORING TIMES | | | |
|------------------|------------------|----------------|-------------------|
| Parameter | Sampling Date(s) | Sampling Times | Sampling Duration |
| Combustion Gases | 19 May 2014 | 13:15 - 14:15 | 60 minutes |

EXECUTIVE SUMMARY

| PROCESS DETAILS | |
|---|-------------------|
| Parameter | Process Details |
| Description of process | Gas enrichment |
| Continuous or batch | Continuous |
| Product Details | - |
| Part of batch to be monitored (if applicable) | Stable Conditions |
| Normal load, throughput or continuous rating | Normal load |
| Fuel used during monitoring | Natural Gas |
| Abatement | None |
| Plume Appearance | None Visible |

EXECUTIVE SUMMARY

Monitoring Methods

The selection of standard reference / alternative methods employed by Environmental Scientifics Group Limited is determined, wherever possible by the hierarchy of method selection outlined in Environment Agency Technical Guidance Note (Monitoring) M2. i.e. CEN, ISO, BS, US EPA etc.

| MONITORING METHODS | | | | | | |
|--------------------|---|-------------------------------|--------------------|--------------------------------|--------------------------------|---------------------------|
| Species | Method Standard Reference Method / Alternative Method | ESG Technical Procedure | UKAS Lab Number | MCERTS Accredited Method | Limit of Detection (LOD) | Calculated MU +/- % |
| NO _x | SRM - BS EN 14792 | AE 102 | 1015 | Yes | 0.41 mg/m ³ | 8.7% |
| SO ₂ | AM - M21 | AE 102 | 1015 | Yes | 0.62 mg/m ³ | 521.3% |
| CO | SRM - BS EN 15058 | AE 102 | 1015 | Yes | 0.35 mg/m ³ | 13.8% |
| CO ₂ | SRM - ISO 12039 | AE 102 | 1015 | Yes | 0.003 % | 4.1% |
| O ₂ | AM - BS EN 14789 | AE 102 | 1015 | Yes | 0.01% | 3.2% |

Permit NO_x
SO₂

ISO 10849 : 1996
ISO 6069 - 4.4

6 monthly
continuous

EXECUTIVE SUMMARY

Analytical Methods

The following tables list the analytical methods employed together with the custody and archiving details:

| SAMPLING METHODS WITH SUBSEQUENT ANALYSIS | | | | | | | |
|---|----------------------|----------------------|-----------------|------------------------------|-----------------------------------|-------------------------|----------------|
| Species | Analytical Technique | Analytical Procedure | UKAS Lab Number | UKAS Accredited Lab Analysis | Analysis Lab (ESG or Subcontract) | Sample Archive Location | Archive Period |
| - | - | - | - | - | - | - | - |

| ON-SITE TESTING | | | | | | | |
|-----------------|--------------------------|----------------------|-----------------|----------------------------|---------------|-----------------------|----------------|
| Species | Analytical Technique | Analytical Procedure | UKAS Lab Number | MCERTS Accredited Analysis | Laboratory | Data Archive Location | Archive Period |
| NO _x | Chemiluminescence | AE 102 | 1015 | Yes | ESG Stockport | ESG Stockport | 5 years |
| SO ₂ | Non Dispersive Infra Red | AE 102 | 1015 | Yes | ESG Stockport | ESG Stockport | 5 years |
| CO | Non Dispersive Infra Red | AE 102 | 1015 | Yes | ESG Stockport | ESG Stockport | 5 years |
| CO ₂ | Non Dispersive Infra Red | AE 102 | 1015 | Yes | ESG Stockport | ESG Stockport | 5 years |
| O ₂ | Zirconia Cell | AE 102 | 1015 | Yes | ESG Stockport | ESG Stockport | 5 years |

EXECUTIVE SUMMARY

| SAMPLING LOCATION | | | | | |
|--|-------|-------|--------------|-----------|---------------|
| Sampling Plane Validation Criteria | Value | Units | Requirement | Compliant | Method |
| Lowest Differential Pressure | - | Pa | ≥ 5 Pa | - | BS EN 13284-1 |
| Lowest Gas Velocity | - | m/s | - | - | - |
| Highest Gas Velocity | - | m/s | - | - | - |
| Ratio of Gas Velocities | - | : 1 | $< 3 : 1$ | - | BS EN 13284-1 |
| Mean Velocity | - | m/s | - | - | - |
| Maximum angle of flow with regard to duct axis | - | ° | $< 15^\circ$ | - | BS EN 13284-1 |
| No local negative flow | - | - | - | - | BS EN 13284-1 |

| DUCT CHARACTERISTICS | | |
|----------------------|----------|----------------|
| | Value | Units |
| Shape | Circular | - |
| Depth | - | m |
| Width | - | m |
| Area | - | m ² |
| Port Depth | 40 | mm |

| SAMPLING LINES & POINTS | | | |
|-------------------------|-----------------------------|-----------------------------|--------------------|
| | Isokinetic (CEN Methods) | Isokinetic (ISO Methods) | Non-Iso & Gases |
| Sample port size | - | - | 20mm Hole |
| Number of lines used | - | - | 1 |
| Number of points / line | - | - | 1 |
| Duct orientation | - | - | Vertical |

| SAMPLING PLATFORM | |
|--|-----------|
| General Platform Information | |
| Permanent / Temporary Platform / Ground level / Floor Level / Roof | Permanent |
| Inside / Outside | Outside |

| M1 Platform requirements | |
|---|-----|
| Is there a sufficient working area so work can be performed in a compliant manner | Yes |
| Platform has 2 levels of handrails (approximately 0.5 m & 1.0 m high) | Yes |
| Platform has vertical base boards (approximately 0.25 m high) | Yes |
| Platform has removable chains / self closing gates at the top of ladders | Yes |
| Handrail / obstructions do not hamper insertion of sampling equipment | Yes |
| Depth of Platform = >Stack depth / diameter + wall and port thickness + 1.5m | No |

Sampling Platform Improvement Recommendations (if applicable)

Although small the platform is adequate for the method of sampling carried out. The sample port however is not of sufficient size to allow any stack dimension or flow measurements.

EXECUTIVE SUMMARY

Sampling & Analytical Method Deviations

Stack Dimension and Flow Measurements

* It is not possible to calculate a mass emission due to the inability to fit the measurement devices into the stack. *

Required?

APPENDICES

CONTENTS

APPENDIX 1 - Monitoring Schedule, Calibration Checklist & Monitoring Team

APPENDIX 2 - Summaries, Calculations, Raw Data and Charts

APPENDIX 3 - Measurement Uncertainty Budget Calculations

APPENDIX 1 - Monitoring Schedule, Calibration Checklist & Monitoring Team

| MONITORING SCHEDULE | | | | | |
|---------------------|---|-------------------------------|--------------------|--------------------------------|----------------------|
| Species | Method Standard Reference Method / Alternative Method | ESG Technical Procedure | UKAS Lab Number | MCERTS Accredited Method | Number of Samples |
| NO _x | SRM - BS EN 14792 | AE 102 | 1015 | Yes | 1 |
| SO ₂ | AM - M21 | AE 102 | 1015 | Yes | 1 |
| CO | SRM - BS EN 15058 | AE 102 | 1015 | Yes | 1 |
| CO ₂ | SRM - ISO 12039 | AE 102 | 1015 | Yes | 1 |
| O ₂ | AM - BS EN 14789 | AE 102 | 1015 | Yes | 1 |

APPENDIX 1 - Monitoring Schedule, Calibration Checklist & Monitoring Team

| CALIBRATEABLE EQUIPMENT CHECKLIST | | | | | |
|-----------------------------------|----------------|---------------------------------|----------------|---------------------------|----------------|
| Extractive Sampling | | Instrumental Analyser/s | | Miscellaneous | |
| Equipment | Equipment I.D. | Equipment | Equipment I.D. | Equipment | Equipment I.D. |
| Control Box DGM | - | Horiba PG-250 Analyser | LNO 21-13 | Laboratory Balance | - |
| Box Thermocouples | - | FT-IR | - | Tape Measure | - |
| Meter In Thermocouple | - | FT-IR Oven Box | - | Stopwatch | - |
| Meter Out Thermocouple | - | Bernath 3006 FID | - | Protractor | - |
| Control Box Timer | - | Signal 3030 FID | - | Barometer | LNO 08-KB |
| Oven Box | - | Servomex | - | Digital Micromanometer | - |
| Probe | - | JCT Heated Head Filter | - | Digital Temperature Meter | - |
| Probe Thermocouple | - | Thermo FID | - | Stack Thermocouple | - |
| Probe | - | Stackmaster | - | Mass Flow Controller | - |
| Probe Thermocouple | - | FTIR Heater Box for Heated Line | - | MFC Display module | - |
| S-Pitot | - | Anemometer | - | 1m Heated Line (1) | - |
| L-Pitot | - | Ecophysics NOx Analyser | - | 1m Heated Line (2) | - |
| Site Balance | - | Chiller (JCT/MAK 10) | LNO 21-38 | 1m Heated Line (3) | - |
| Last Impinger Arm | - | Heated Line Controller (1) | LNO 03-86 | 5m Heated Line (1) | - |
| Dioxins Cond. Thermocouple | - | Heated Line Controller (2) | - | 10m Heated Line (1) | - |
| Callipers | - | Site temperature Logger | - | 10m Heated Line (2) | - |
| Small DGM | - | | - | 15m Heated Line (1) | - |
| Heater Controller | - | | - | 20m Heated Line (1) | LNO 18-86 |
| Inclinometer (Swirl Device) | - | | - | 20m Heated Line (2) | - |

NOTE: If the equipment I.D. is represented by a dash (-), then this piece of equipment has not been used for this test.

| CALIBRATION GASES | | | | | |
|------------------------------|----------------------|----------|-------|------|----------------------------|
| Gas (traceable to ISO 17025) | Cylinder I.D. Number | Supplier | ppm | % | Analytical Tolerance +/- % |
| Nitric Oxide | HPC 787 | BOC | 200 | - | 2 |
| Sulphur Dioxide | HPC 755 | BOC | 158.9 | - | 2 |
| Carbon Monoxide | HPC 773 | BOC | 167 | - | 2 |
| Carbon Dioxide | HPC 752 | BOC | - | 8.11 | 2 |

APPENDIX 1 - Monitoring Schedule, Calibration Checklist & Monitoring Team

STACK EMISSIONS MONITORING TEAM

Team Leader

Keith Bird
MCERTS Level 2, Technical Endorsements 1, 2, 3 & 4
MM 07 825
MCERTS Expiry Date - Dec 2018
H&S Expiry Date - Apr 2017

Technician

Mark Derbyshire
MCERTS Level 1, Technical Endorsement 1
MM 07 824
MCERTS Expiry Date - Dec 2017
H&S Expiry Date - Apr 2017

APPENDIX 2 - Summaries, Calculations, Raw Data and Charts

COMBUSTION GASES SUMMARY

| Test | Sampling Time and Date | Concentration mg/m ³ | LOD mg/m ³ | Limit mg/m ³ | Emission Rate g/hr |
|-----------------|------------------------------|------------------------------------|--------------------------|----------------------------|-----------------------|
| NO _x | 13:15 - 14:15 19 May 2014 | 40 | 0.41 | 120 | - |
| SO ₂ | 13:15 - 14:15 19 May 2014 | 0.91 | 0.62 | 190 | - |
| CO | 13:15 - 14:15 19 May 2014 | 15.2 | 0.35 | - | - |

| Test | Sampling Time and Date | Concentration % | LOD % |
|-----------------|------------------------------|--------------------|----------|
| CO ₂ | 13:15 - 14:15 19 May 2014 | 7.3 | 0.003 |
| O ₂ | 13:15 - 14:15 19 May 2014 | 9.9 | 0.01 |

Reference conditions are 273K, 101.3kPa, dry gas

PRE-SAMPLING CALIBRATION DATA

| | |
|------------|-------------|
| Date | 19 May 2014 |
| Start Time | 11:00 |
| End Time | 11:30 |

| | |
|--------------------------|-------|
| Chiller Temperature (°C) | 2.2 |
| Requirement | < 4°C |
| Compliant | Yes |

| Gas | Range (ppm / %) | Zero Reading at analyser | Span Reading at analyser | Zero Check at analyser | Zero Check down line | Span Check down line | Response Time (Secs) | Leak Rate % |
|-----------------|--------------------|-----------------------------|-----------------------------|---------------------------|-------------------------|-------------------------|-------------------------|----------------|
| NO | 200 | 0 | 200 | -0.1 | 0.1 | 198 | 55 | 1.00 |
| SO ₂ | 200 | 0 | 158.9 | 0.1 | 0.2 | 157.7 | 65 | 0.76 |
| CO | 200 | 0 | 167 | 0.2 | 0.3 | 166.5 | 45 | 0.30 |
| CO ₂ | 25 | 0 | 8.11 | 0.04 | 0.06 | 8.15 | 40 | -0.49 |
| O ₂ | 25 | 0 | 9.9 | 0.03 | 0.07 | 9.98 | 40 | -0.81 |

POST-SAMPLING CALIBRATION DATA


| | |
|------------|-------------|
| Date | 19 May 2014 |
| Start Time | 14:18 |
| End Time | 14:30 |

| | |
|--------------------------|-------|
| Chiller Temperature (°C) | 2.2 |
| Requirement | < 4°C |
| Compliant | Yes |


| Gas | Zero Check down line | Span Check down line | Zero Drift (%) | Span Drift (%) |
|-----------------|-------------------------|-------------------------|-------------------|-------------------|
| NO | 0.2 | 199 | 0.05 | 0.45 |
| SO ₂ | 0.1 | 159 | -0.05 | 0.70 |
| CO | 0.2 | 164 | -0.05 | -1.20 |
| CO ₂ | 0.06 | 8.12 | 0.00 | -0.12 |
| O ₂ | 0.08 | 9.99 | 0.04 | 0.00 |

APPENDIX 2 - Summaries, Calculations, Raw Data and Charts


OXIDES OF NITROGEN (as NO_x) EMISSIONS CHART




SULPHUR DIOXIDE EMISSIONS CHART




CARBON MONOXIDE EMISSIONS CHART



CARBON DIOXIDE EMISSIONS CHART








STACK DIAGRAM


| Non-Isokinetic/Gases Sampling | | | |
|-------------------------------|-----------------------|---------------------|-------|
| Sampling Point | Distance (% of Depth) | Distance into Stack | Units |
| - | - | - | - |



-  Isokinetic sampling point
 Isokinetic sampling points not used
 Non Isokinetic/Gases sampling point

[illegible]

SAMPLING LOCATION



APPENDIX 3 - Measurement Uncertainty Budget Calculations

MEASUREMENT UNCERTAINTY BUDGET - OXIDES OF NITROGEN

| | | |
|--------------------------------|------|-------------------|
| Limit value | 120 | mg/m ³ |
| Concentration @ Ref conditions | 39.6 | mg/m ³ |
| Cal gas conc | 410 | mg/m ³ |
| Analyser Full Scale | 410 | mg/m ³ |

| Performance characteristics | Value | Units | specification | MU Met? |
|------------------------------------|-------|-----------------------|--------------------|---------|
| Response time | 55 | seconds | 180 | Yes |
| Logger sampling interval | 60 | seconds | - | - |
| Measurement period | 60 | minutes | - | - |
| Number of readings in measurement | 60 | - | - | - |
| Repeatability at zero | 0.25 | % full scale | <1 % range | Yes |
| Repeatability at span level | 0.15 | % full scale | <2 % range | Yes |
| Deviation from linearity | 0.7 | % of value | <2 % range | Yes |
| Zero drift | 0.10 | % full scale | <2% range / 24hr | Yes |
| Span drift | 0.90 | % full scale | <2% range/24hr | Yes |
| volume or pressure flow dependence | 0.02 | % of full scale/3 kPa | <2 % / 3 kPa | Yes |
| atmospheric pressure dependence | 0.8 | % of full scale/2 kPa | <3% / 2 kPa | Yes |
| ambient temperature dependence | 0.01 | % full scale/10K | <3% range / 10 K | Yes |
| dependence on voltage | 0.1 | % full scale/10V | < 0.1%vol /10 volt | Yes |
| losses in the line (leak) | 1.00 | % of value | < 2% of value | Yes |
| Uncertainty of calibration gas | 1 | % of value | < 2% of value | Yes |

| Performance characteristic | Uncertainty | Value of uncertainty quantity |
|---|-------------|-------------------------------|
| Standard deviation of repeatability at zero | ur0 | 0.02 |
| Standard deviation of repeatability at span level | urs | 0.02 |
| Lack of fit | ufit | 1.66 |
| Drift | u0dr | 0.11 |
| volume or pressure flow dependence | uspres | 0.002 |
| atmospheric pressure dependence | uapres | 0.10 |
| ambient temperature dependence | utemp | 0.00001 |
| Dependence on voltage | uvolt | 0.35 |
| losses in the line (leak) | uleak | 0.23 |
| Uncertainty of calibration gas | ucalib | 0.23 |
| Uncertainty in factor | uf | 0.00 |

| | | |
|--|------|-------------------|
| Measurement uncertainty (Concentration Measured) | 39.6 | mg/m ³ |
| Combined uncertainty | 1.7 | mg/m ³ |
| Expanded at a 95% confidence interval | 3.5 | mg/m ³ |

| | | |
|--|-----|-------|
| Expanded uncertainty expressed with a level of confidence of 95% | 2.9 | % ELV |
|--|-----|-------|

| | | |
|--|-----|-------------------|
| Expanded uncertainty expressed with a level of confidence of 95% | 3.5 | mg/m ³ |
|--|-----|-------------------|

| | | |
|--|-----|---------|
| Expanded uncertainty expressed with a level of confidence of 95% | 8.7 | % value |
|--|-----|---------|

Developed for the STA by R Robinson, NPL

APPENDIX 3 - Measurement Uncertainty Budget Calculations

MEASUREMENT UNCERTAINTY BUDGET - SULPHUR DIOXIDE

| | | |
|--------------------------------|---------|-------------------|
| Limit value | 190 | mg/m ³ |
| Concentration @ Ref conditions | 0.9 | mg/m ³ |
| Cal gas conc | 451.276 | mg/m ³ |
| Analyser Full Scale | 572 | mg/m ³ |

| Performance characteristics | Value | Units | specification | MU Met? |
|------------------------------------|-------|-----------------------|---------------------|---------|
| Response time | 65 | seconds | 180 | Yes |
| Logger sampling interval | 60 | seconds | - | - |
| Measurement period | 60 | minutes | - | - |
| Number of readings in measurement | 60 | - | - | - |
| Repeatability at zero | 0.25 | % full scale | <1 % range | Yes |
| Repeatability at span level | 0.15 | % full scale | <2 % range | Yes |
| Deviation from linearity | 0.7 | % of value | <2 % range | Yes |
| Zero drift | -0.10 | % full scale | <2% range / 24hr | Yes |
| Span drift | 1.40 | % full scale | <2% range/24hr | Yes |
| volume or pressure flow dependence | 0.02 | % of full scale/3 kPa | <2 % / 3 kPa | Yes |
| atmospheric pressure dependence | 0.8 | % of full scale/2 kPa | <3% / 2 kPa | Yes |
| ambient temperature dependence | 0.01 | % full scale/10K | <3% range / 10 K | Yes |
| dependence on voltage | 0.1 | % full scale/10V | < 0.1%vol / 10 volt | Yes |
| losses in the line (leak) | 0.76 | % of value | < 2% of value | Yes |
| Uncertainty of calibration gas | 1 | % of value | < 2% of value | Yes |

| Performance characteristic | Uncertainty | Value of uncertainty quantity |
|---|-------------|-------------------------------|
| Standard deviation of repeatability at zero | ur0 | 0.019 |
| Standard deviation of repeatability at span level | urs | 0.02 |
| Lack of fit | ufit | 2.31 |
| Drift | u0dr | -0.06 |
| volume or pressure flow dependence | uspres | 0.00 |
| atmospheric pressure dependence | uapres | 0.14 |
| ambient temperature dependence | utemp | 0.00 |
| Dependence on voltage | uvolt | 0.49 |
| losses in the line (leak) | uleak | 0.00 |
| Uncertainty of calibration gas | ucalib | 0.01 |
| Uncertainty in factor | uf | 0.00 |

| | | |
|--|-----|-------------------|
| Measurement uncertainty (Concentration Measured) | 0.9 | mg/m ³ |
| Combined uncertainty | 2.4 | mg/m ³ |
| Expanded uncertainty | 4.7 | mg/m ³ |

| | | |
|--|-------|-------------------|
| Expanded uncertainty expressed with a level of confidence of 95% | 2.5 | % ELV |
| Expanded uncertainty expressed with a level of confidence of 95% | 4.7 | mg/m ³ |
| Expanded uncertainty expressed with a level of confidence of 95% | 521.3 | % value |

Developed for the STA by R Robinson, NPL

APPENDIX 3 - Measurement Uncertainty Budget Calculations

MEASUREMENT UNCERTAINTY BUDGET - CARBON MONOXIDE

| | | |
|--------------------------------|--------|-------------------|
| Limit value | - | mg/m ³ |
| Concentration @ Ref conditions | 15.2 | mg/m ³ |
| Cal gas conc | 208.75 | mg/m ³ |
| Analyser Full Scale | 250 | mg/m ³ |

| Performance characteristics | Value | Units | specification | MU Met? |
|------------------------------------|-------|-----------------------|---------------------|---------|
| Response time | 45 | seconds | 180 | Yes |
| Logger sampling interval | 60 | seconds | - | - |
| Measurement period | 60 | minutes | - | - |
| Number of readings in measurement | 60 | - | - | - |
| Repeatability at zero | 0.25 | % full scale | <1 % range | Yes |
| Repeatability at span level | 0.15 | % full scale | <2 % range | Yes |
| Deviation from linearity | 0.7 | % of value | <2 % range | Yes |
| Zero drift | -0.10 | % full scale | <2% range / 24hr | Yes |
| Span drift | -2.40 | % full scale | <2% range/24hr | Yes |
| volume or pressure flow dependence | 0.02 | % of full scale/3 kPa | <2 % / 3 kPa | Yes |
| atmospheric pressure dependence | 0.80 | % of full scale/2 kPa | <3% / 2 kPa | Yes |
| ambient temperature dependence | 0.01 | % full scale/10K | <3% range / 10 K | Yes |
| dependence on voltage | 0.10 | % full scale/10V | < 0.1%vol / 10 volt | Yes |
| losses in the line (leak) | 0.30 | % of value | < 2% of value | Yes |
| Uncertainty of calibration gas | 1 | % of value | < 2% of value | Yes |

| Performance characteristic | Uncertainty | Value of uncertainty quantity |
|---|-------------|-------------------------------|
| Standard deviation of repeatability at zero | ur0 | 0.02 |
| Standard deviation of repeatability at span level | urs | 0.02 |
| Lack of fit | ufit | 1.01 |
| Drift | u0dr | -0.16 |
| volume or pressure flow dependence | uspres | 0.001 |
| atmospheric pressure dependence | uapres | 0.06 |
| ambient temperature dependence | utemp | 0.00 |
| Dependence on voltage | uvolt | 0.22 |
| losses in the line (leak) | uleak | 0.03 |
| Uncertainty of calibration gas | ucalib | 0.09 |
| Uncertainty in factor | uf | 0.00 |

| | | |
|--|------|-------------------|
| Measurement uncertainty (Concentration Measured) | 15.2 | mg/m ³ |
| Combined uncertainty | 1.1 | mg/m ³ |
| Expanded uncertainty | 2.1 | mg/m ³ |

| | | |
|--|------|-------------------|
| Expanded uncertainty expressed with a level of confidence of 95% | - | % ELV |
| Expanded uncertainty expressed with a level of confidence of 95% | 2.1 | mg/m ³ |
| Expanded uncertainty expressed with a level of confidence of 95% | 13.8 | % value |

Developed for the STA by R Robinson, NPL

APPENDIX 3 - Measurement Uncertainty Budget Calculations

MEASUREMENT UNCERTAINTY BUDGET - CARBON DIOXIDE

| | | |
|------------------------|------|------|
| Limit value | - | %vol |
| Reported Concentration | 7.31 | %vol |
| Calibration gas | 8.11 | %vol |
| Analyser Full Scale | 25 | %vol |

| Performance characteristics | Value | Units | specification | MU Met? |
|--|-------|---------------------|----------------------|---------|
| Response time | 40 | seconds | < 200 s | Yes |
| Logger sampling interval | 60 | seconds | - | - |
| Measurement period | 60 | minutes | - | - |
| Number of readings in measurement | 60 | - | - | - |
| Repeatability at zero | 0.015 | % by volume | <0.2 % range | Yes |
| Repeatability at span level | 0.014 | % by volume | <0.4 % range | Yes |
| Deviation from linearity | 0.13 | % vol | <0.3 % volume | Yes |
| Zero drift (during measurement period) | 0.00 | % vol at zero level | <2% of volume / 24hr | Yes |
| Span drift (during measurement period) | -0.03 | % vol at span level | <2% volume/24hr | Yes |
| volume or pressure flow dependence | 0.02 | % of fs / 10l/h | <1% range | Yes |
| atmospheric pressure dependence | 0.8 | % of fs/kPa | < 1.5 % range | Yes |
| ambient temperature dependence | 0.01 | % by volume /10K | <0.3% volume 10 K | Yes |
| Combined interference | 0.56 | % range | <2% range | Yes |
| Dependence on voltage | 0.1 | % by volume /10V | < 0.1%vol /10 volt | Yes |
| Losses in the line (leak) | -0.49 | % of value | < 2% of value | Yes |
| Uncertainty of calibration gas | 1 | % of value | < 2% of value | Yes |

| Performance characteristic | Uncertainty | Value of uncertainty quantity |
|---|-------------|-------------------------------|
| Standard deviation of repeatability at zero | ur0 | - |
| Standard deviation of repeatability at span level | urs | 0.00 |
| Lack of fit | ufit | 0.08 |
| Drift | u0dr | -0.02 |
| volume or pressure flow dependence | uspres | 0.00003 |
| atmospheric pressure dependence | uapres | 0.01 |
| ambient temperature dependence | utemp | 0.001 |
| Combined interference (from mcerts) | - | 0.08 |
| dependence on voltage | uvolt | 0.086 |
| losses in the line (leak) | uleak | -0.02 |
| Uncertainty of calibration gas | ucalib | 0.04 |

| | | |
|--|------|------|
| Measurement uncertainty (Concentration Measured) | 7.31 | %vol |
| Combined uncertainty | 0.15 | %vol |
| % of value | 2.04 | % |

| | | |
|--|------|------------|
| Expanded uncertainty expressed with a level of confidence of 95% | 4.08 | % of value |
|--|------|------------|

| | | |
|--|------|-------|
| Expanded uncertainty expressed with a level of confidence of 95% | 0.30 | % vol |
|--|------|-------|

Developed for the STA by R Robinson, NPL

APPENDIX 3 - Measurement Uncertainty Budget Calculations

MEASUREMENT UNCERTAINTY BUDGET - OXYGEN

| | | |
|------------------------|------|------|
| Reference | N/A | %vol |
| Reported Concentration | 9.86 | %vol |
| Calibration gas | 9.9 | %vol |
| Analyser Full Scale | 25 | %vol |

| Performance characteristics | Value | Units | specification | MU Met? |
|--|-------|---------------------|----------------------|---------|
| Response time | 40 | seconds | < 200 s | Yes |
| Logger sampling interval | 60 | seconds | - | - |
| Measurement period | 60 | minutes | - | - |
| Number of readings in measurement | 60 | - | - | - |
| Repeatability at zero | 0.015 | % by volume | <0.2 % range | Yes |
| Repeatability at span level | 0.014 | % by volume | <0.4 % range | Yes |
| Deviation from linearity | 0.13 | % vol | <0.3 % volume | Yes |
| Zero drift (during measurement period) | 0.01 | % vol at zero level | <2% of volume / 24hr | Yes |
| Span drift (during measurement period) | 0.00 | % vol at span level | <2% volume/24hr | Yes |
| volume or pressure flow dependence | 0.02 | % of fs / 10l/h | <1% range | Yes |
| atmospheric pressure dependence | 0.80 | % of fs/kPa | < 1.5 % range | Yes |
| ambient temperature dependence | 0.01 | % by volume /10K | <0.3% volume 10 K | Yes |
| Combined interference | 0.14 | % range | <2% range | Yes |
| Dependence on voltage | 0.10 | % by volume /10V | < 0.1%vol /10 volt | Yes |
| Losses in the line (leak) | -0.81 | % of value | < 2% of value | Yes |
| Uncertainty of calibration gas | 1.00 | % of value | < 2% of value | Yes |

| Performance characteristic | Uncertainty | Value of uncertainty quantity |
|---|-------------|-------------------------------|
| Standard deviation of repeatability at zero | ur0 | - |
| Standard deviation of repeatability at span level | urs | 0.0018 |
| Lack of fit | ufit | 0.0751 |
| Drift | u0dr | 0.0058 |
| volume or pressure flow dependence | uspres | 0.00003 |
| atmospheric pressure dependence | uapres | 0.0122 |
| ambient temperature dependence | utemp | 0.0005 |
| Combined interference (from mcerts) | - | 0.0808 |
| dependence on voltage | uvolt | 0.0862 |
| losses in the line (leak) | uleak | -0.0460 |
| Uncertainty of calibration gas | ucalib | 0.0569 |

| | | |
|--|------|------|
| Measurement uncertainty (Concentration Measured) | 9.86 | %vol |
| Combined uncertainty | 0.16 | %vol |
| % of value | 1.61 | % |

| | | |
|--|-----|------------|
| Expanded uncertainty expressed with a level of confidence of 95% | 3.2 | % of value |
|--|-----|------------|

| | | |
|--|-------|-------|
| Expanded uncertainty expressed with a level of confidence of 95% | 0.317 | % vol |
|--|-------|-------|

Developed for the STA by R Robinson, NPL

END OF REPORT