

PIPETTE CALIBRATION REPORT

Product code: 4641120
Description: Finnpiquette F1 1-10ml
Pipette sn: KH20930
Volume: 1000 - 10000 µl
Nr.of channels: 1
Tip: Finntip 10 ml

Manufacturer: Thermo Fisher Scientific, Joensuu, Finland
Tested by: 4913

Date:	8.3.2013	Time:	6:24:47
Temperature:	21.3 °C	Air Pressure:	1002,2 hPa
Z-factor:	1,0032	Humidity:	Moisture trap in use
Balance:	Sartorius ME215S		
Balance sn:	12902083		

Volume	1000,0 µl	10000,0 µl
1	1000,7	10003,6
2	1006,2	10030,2
3	1011,0	10011,4
4	1007,8	10013,4
5	1004,0	10013,0

Mean	1005,9 µl	10014,3 µl
Acc±%	0,59	0,14
CV%	0,39	0,10
	PASSED	PASSED

Manufacturer specifications:

Volume:	1000 µl	Acc±% 2,00	CV% 0,80
Volume:	10000 µl	Acc±% 0,50	CV% 0,20

Manufacturer specifications meet the requirements of ISO 8655