

O. Smetes.
DRAINED 24/06/19.

Sample Results Print

Thermo
SCIENTIFIC

This report contains details of one or more Sample records. This report was generated using the Print Sample menu option and the 'Sample Results Print' report layout.

Id Numeric : 1980144
Sample : SM-17-JUN-2019-361045
Status : C

Sample Results List

Analysis	Component Name	Type	Result	Units	Status
0176\$OIL/1	Appearance (Pass = 1, Fail = 0)	N	1		E
0843\$ORG/2	Methanol	N	<3	ppm_wt_v	E
0843\$ORG/2	TMS	N	<0.1	ppm_wt_v	E
0281\$D/1	Final weight of BOTTOM Filter	N	0.1944	g	E
0281\$D/1	Final weight of TOP Filter	N	0.1919	g	E
0281\$D/1	Initial weight of BOTTOM Filter	N	0.1942	g	E
0281\$D/1	Initial weight of TOP Filter	N	0.1902	g	E
0281\$D/1	Suspended solids	K	3.2	ppm_wt_v	E
0281\$D/1	Volume of water filtered	N	470	ml	E
0007\$A/1	pH	N	9.26		E