

BS EN 14181 QAL2 Calibration and Validation of Continuous Emissions
Monitoring Systems Report

Permit Number: EPR/BL7108IM

Operator: Tata Steel Strip UK

Works: Port Talbot

Plant: Power Plant

Location: A62 - Boiler 3

Monitoring dates: 14th - 16th May 2019



Monitoring Organisation: Tata Steel
 Environmental Technology, Group Environment

Address: Swinden Technology Centre
 Moorgate
 Rotherham
 S60 3AR

Date of Report: 15/07/2019

Section 1- Executive Summary

Section 1A

Monitoring Objectives

A Continuous Emission Monitoring System (CEMS) is installed on Boiler 3 stack at Tata Steel Strip UK, Port Talbot. Tata Steel, Environmental Technology (ET), Group Environment were commissioned to perform the calibration of the CEMS in accordance with BS EN 14181. A QAL2 has been carried out to determine a Calibration Function and verify that the CEM meets the uncertainty requirements of the directive.

Operator: Tata Steel Strip UK

Site Contact: Sean Langstone

Address Port Talbot
South Wales
SA13 2NG

Type of Process: Multi fuel boiler using works arising gases to produce steam for the 11 bar steam network for site.

Plant: Power Plant

Installation Name: A62 - Boiler 3

Permit Number: EPR/BL7108IM

Monitoring dates: 14th - 16th May 2019

Report Version and Date Version 1 15/07/2019

Monitoring Organisation: Tata Steel, Environmental Technology, Group Environment

Address: Swinden Technology Centre
Moorgate
Rotherham
S60 3AR

UKAS Lab No.: 0312

Report Prepared By: Kriss Wells


MCERTS Registration Number: MM 06 701
Function: MCerts Level 2 (TE1)
Tel: 0114 2043057
Email: kriss.wells@tatasteeleurope.com

Report Approved By: Simon Johnston


MCERTS Registration Number: MM 02 114
Function: MCerts Level 2 (TE1 TE2 TE3 TE4)
Tel: 0114 2043019
Email: simon.johnston@tatasteeleurope.com