

Customer:
Blazer Fuels Limited
Brickfield Lane
Ruthin, Denbighshire
LL15 2TN

Service Report



Contact: Steve Davies / Rebecca Taylor **Telephone:** 01824 708928 **Email:** Steve.Davies@blazersfuels.co.uk / rebecca.taylor@advantagebiogas.com
Job Number: 220260087 **Engineer:** Paul Bell(501)

Model	Equipment	Serial No	Description	Complete	Removed	Addition	Certs Issued
MA35M-000230V1	MA35M-000230V1 Infrared Moisture Analyze	33712275	Maintenance B/Standard, Cert with nat. accreditation pre/post	Y			
MA35M-230N	MA35M-230N	37506360	Maintenance B/Standard, Cert with nat. accreditation pre/post	Y			
MA35M-230N	MA35M-230N Infrared Moisture Analyzer, t	37008595	Maintenance B/Standard, Cert with nat. accreditation pre/post	Y			

Task	Complete	N/A	Part Code	Qty	Charge	Fitted To (S/N)
Check the mains or adaptor leads for any damage	✓		No parts added			
Check Display Segment Test	✓					
Equipment is level (If applicable)	✓					
Check under the pan carrier and pan carrier pivot for any foreign matter.	✓					
Check the interior mechanism for foreign matter, damage to parts, wear on parts or loose parts	✓					
Reproducibility Checked	✓					
Check eccentric error	✓					
Linearity checked	✓					
Heating Element working (Moisture Products only) Measure temperature at 2 points IR Dryers	✓					
Check the analysis using a 10% saline solution (Using filter pads)	✓					
Other - Please state below (Free Text) UKAS certs to be sent to the customer	✓					

Complete Date: 17/05/2021
Travel Time: 3 hours
On Site Time: 5.5 hours
Arrival Time:
Departure Time:

Engineer Signature:

PBM

Customer Signature:

PB

Date:

17/05/2021