

## BS EN 14181 Report

Permit Number: **RP3133LD**  
Operator: **RWE Generation UK plc.**  
Installation: **Aberthaw Power Station**  
Installation Type: **Coal-Fired Power Station**  
Emission Point: **Unit 7**  
Monitoring Dates: **5<sup>th</sup> – 10<sup>th</sup> March 2015**



1709



Contract Reference: FTBS 29312  
Operator: RWE Generation UK plc.  
Address: Aberthaw Power Plant  
The Leys  
Aberthaw, Nr Barry  
South Glamorgan  
CF62 4ZW  
Client Contact: Richard Kadim  
Monitoring Organisation: RPS Consultants  
Address: Noble House, Capital Drive, Linford Wood,  
Milton Keynes. MK14 6QP.  
Report Date: 17<sup>th</sup> September 2015  
Report Author: Glyn Harrison  
Report Approved By: Richard Harvey  
Position: Principal Consultant  
MCERTS Qualifications: Level 2, Technical Endorsements 1, 2, 3 & 4  
MCERTS Registration No.: MM 02 020  
Signature: 

RPS Consultants has produced this report within the term of the contract with the client and taking account of the resources devoted to it by agreement with the client.

We disclaim any responsibility to the client and others in respect of any matters outside the scope of the above. This report is confidential to the client and we accept no responsibility of whatsoever nature to third parties to whom this report, or any part thereof, is made known. Any such party relies on the report at their own risk.

## CONTENTS

SECTION 1 – Executive Summary	3
SECTION 2 – Information about the regulated installation	6
SECTION 3 – Information about the monitoring campaign	12
SECTION 4A1 – Data & calculations – QAL2 SICK GM32	15
SECTION 4B1 – Data & calculations – AST Procal 1	46
SECTION 4B2 – Data & calculations – AST Procal 2	65
SECTION 4B3 – Data & calculations – AST SICK OMD 41	84
SECTION 5 – Functional Test Results	92
APPENDIX 1 – SRM on site calibrations	145
APPENDIX 2 – Accreditation Schedule	147

## Section 1 – Executive Summary

### 1B.1 Result Summary – AST

EN 14181 Test Type		AST						
Stack designation		Unit 7						
Measurand	Correlation coefficient of parallel data ( $R^2$ )	Derived Calibration function ( $y_i$ )		Established Calibrated Range <sup>2</sup>	AST Calibrated Range	Extrapolated Calibrated Range	Variability Test	Test of Calibration Function
		$y =$	$a + bx_i$					
Nitric Oxide as total NOx (Procal 1)	0.9834	-3.578 <sup>1</sup>	1.022 <sup>1</sup>	0 – 1634.6 mg/m <sup>3 2</sup>	0 – 1172 mg/m <sup>3 2</sup>	-	Pass	Pass
Sulphur Dioxide (Procal 1 – Low Range)	0.9014	-6.1 <sup>1</sup>	1.082 <sup>1</sup>	0 – 749.3 mg/m <sup>3 2</sup>	0 – 275 mg/m <sup>3 2</sup>	-	Pass	Pass
Sulphur Dioxide (Procal 1 – High Range)	0.9027	7.158 <sup>1</sup>	0.952 <sup>1</sup>	0 – 2302.4 mg/m <sup>3 2</sup>	0 – 345 mg/m <sup>3 2</sup>	-	Pass	Pass
Nitric Oxide as total NOx (Procal 2)	0.9821	-1.5 <sup>1</sup>	0.979 <sup>1</sup>	0 – 1320 mg/m <sup>3 2</sup>	0 – 1092 mg/m <sup>3 2</sup>	-	Pass	Pass
Sulphur Dioxide (Procal 2 – Low Range)	0.9640	-6.072 <sup>1</sup>	0.885 <sup>1</sup>	0 – 566 mg/m <sup>3 2</sup>	0 – 281 mg/m <sup>3 2</sup>	-	Pass	Pass
Sulphur Dioxide (Procal 2 – High Range)	0.9623	-5.884 <sup>1</sup>	0.757 <sup>1</sup>	0 – 599 mg/m <sup>3 2</sup>	0 – 261 mg/m <sup>3 2</sup>	-	Pass	Pass
Particulate Matter (Erwin SICK)	0.4722	0.6 <sup>1</sup>	1.433 <sup>1</sup>	0 – 55 mg/m <sup>3 3</sup>	0 – 19.8 mg/m <sup>3 3</sup>	Not applicable	Pass	Pass

## 1B.2 Result Summary – QAL2

EN 14181 Test Type		QAL2				
Stack designation		Unit 7				
Measurand	Correlation coefficient of parallel data ( $R^2$ )	Derived Calibration function ( $y_i$ )		Calibrated Range	Extrapolated Calibrated Range	Variability Test
		$y_i =$	$a + bx_i$			
Oxides of Nitrogen (GM32)	0.9936	21.795 <sup>1</sup>	0.889 <sup>1</sup>	0 - 1098 mg/m <sup>3</sup>	0 – 2158 mg/m <sup>3</sup>	Pass
Sulphur Dioxide (GM32 – Low Range)	0.9883	0.83 <sup>1</sup>	1.066 <sup>1</sup>	0 – 310.7 mg/m <sup>3</sup>	0 – 1596 mg/m <sup>3</sup>	Pass
Sulphur Dioxide (GM32 – High Range)	0.9882	0.659 <sup>1</sup>	1.064 <sup>1</sup>	0 – 310.6 mg/m <sup>3</sup>	0 – 1596 mg/m <sup>3</sup>	Pass

Notes (Both Tables):

- 1 – Calibration function derived using Method A.
- 2 – Calibrated range derived using AST parallel test data extended 10%.
- 3 - Calibrated range derived using QAL2 parallel test data extended 100%.
- 4 – Extrapolated calibrated range from QAL 2 - derived using reference materials

**Note: The calibration functions, once applied, only remain valid as long as the QAL 3 data remains within control limits, and there are no manual adjustments made to the CEMS other than those allowed to bring the settings back within the QAL 3 control limits.**

## 1C Deviations

<b>SRM deviations</b>	None
<b>Reason for deviation</b>	None
<b>EN 14181 deviations</b>	None
<b>Reason for deviation</b>	N/A
<b>Impact on results</b>	N/A
<b>Further actions required</b>	None

## Section 2 - Information about the Regulated Installation

### 2.1 Regulatory Information

<b>Name of operator</b>	RWE Generation UK plc.
<b>Name of Installation</b>	Aberthaw Power Station
<b>Address of installation</b>	The Leys Aberthaw, Nr Barry South Glamorgan CF62 4ZW
<b>Sector</b>	LCPD
<b>Permit Number</b>	RP3133LD
<b>Date of last QAL 2/AST</b>	March 2014 AST – Procal 1 & 2, SICK OMD 41 & SICK GM32

### Regulated Determinands

Determinand	Emission Point	48-Hour Mean	Calendar Monthly	Uncertainty Requirement
Oxides of Nitrogen	Unit 7	1210 mg/m <sup>3</sup> (NO <sub>x</sub> as NO <sub>2</sub> )	1100 mg/m <sup>3</sup> (NO <sub>x</sub> as NO <sub>2</sub> )	20% at the ELV
Sulphur Dioxide	Unit 7	440 mg/m <sup>3</sup>	400 mg/m <sup>3</sup>	20% at the ELV
Total particulate Matter	Unit 7	55 mg/m <sup>3</sup>	25 mg/m <sup>3</sup>	30% at the ELV

Note: ELVs at reference conditions 273K, 101.3kPa, 6% oxygen, dry gas


### 2.2 Operational Information and site monitoring provisions

#### 2.2.1 Process type and emissions variations

<b>Process Type</b>	Continuous - Coal-Fired Power Station
<b>Process Variations</b>	Once operational at maximum load variation is minimal
<b>Expected emissions variations</b>	
1 Oxides of nitrogen	850 - 1050 mg/m <sup>3</sup> as NO <sub>2</sub> – variable
2 Sulphur Dioxide	150 - 300 mg/m <sup>3</sup> – variable
3 Oxygen	6 – 8 % – constant
<b>Possible low level emissions</b>	Sulphur Dioxide
<b>Provision to deal with low level emissions</b>	Use values from linearity checks to derive calibrated range
<b>Other factors affecting monitoring results</b>	None
<b>Fuel type</b>	Coal
<b>Abatement</b>	Flue Gas Desulphurisation & Combustion control

## 2.3 Monitoring Provisions at the installation

### 2.3.1 Stack & sampling ports



### Photograph of Stack Photo of Monitoring Platform



Access was limited to ports due to the stack layout inside the windshield. The port shown is one of four.



### 2.3.2 Monitoring platform and site provisions

Requirement	Compliant	Notes
<ul style="list-style-type: none"> <li>A safe and clean working environment with sufficient space and weather protections.</li> </ul>	CEMs cabinet – Monitoring probe -	Both cabinet and probe are located inside the windshield and are thus in a spacious, clean and weatherproof environment.
<ul style="list-style-type: none"> <li>Easy and safe access to the CEM.</li> </ul>	CEMS cabinet – Monitoring probe -	Stairways to the relevant levels.
<ul style="list-style-type: none"> <li>Adequate supplies of reference materials, tools and spare parts.</li> </ul>	Yes	
<ul style="list-style-type: none"> <li>Facilities to introduce the reference materials for gaseous-monitoring systems, both at the inlet of the sampling line (where present), and at the inlet of the CEM.</li> </ul>	N/A	There is no facility to introduce reference materials to the sample line or CEM. The CEMs are cross-duct IR (NO, NO <sub>2</sub> , SO <sub>2</sub> ) and cross-duct forward scatter (particulate).
<ul style="list-style-type: none"> <li>Compliance with TGN M1</li> </ul>	No	Unable to access sample port B
<ul style="list-style-type: none"> <li>Compliance with EN 15259 –</li> </ul> <p>flow stability criteria (if applicable),</p> <p>Stack gas homogeneity.</p>	<p>Yes</p> <p>Yes</p>	<p>Stack gas homogeneity carried out previously by Atkins.</p>

## Temperature and Velocity Profile

Company Name: RWE Date: 06/03/15  
Site Name: Aberthaw Run: 1  
Sampling Point Ref: Unit 7 Barometric Press. Mbar: 1024  
Project Reference: FTBS 29312  
Stack Diamter (m): 6.00  
Stack Static press.mm H<sub>2</sub>O: -24 Stack Area (m<sup>2</sup>): 28.274

Traverse Point No.	Port A			Port C			Port D		
	Δ p, mm H <sub>2</sub> O	Root Δ p	Stack Temp °C	Δ p, mm H <sub>2</sub> O	Root Δ p	Stack Temp °C	Δ p, mm H <sub>2</sub> O	Root Δ p	Stack Temp °C
1	30	5.477	60						
2	32	5.657	61						
3	32	5.657	61						
4									
5									
6									
7									
8				30	5.477	60	30	5.477	60
9				32	5.657	61	32	5.657	61
10				32	5.657	61	34	5.831	61
Minimum	30.0	5.477	60	30.0	5.477	60	30.0	5.477	60
Maximum	32.0	5.657	61	32.0	5.657	61	34.0	5.831	61
Mean	31.3	5.597	60.7	31.3	5.597	60.7	32.0	5.655	60.7
Sum	94	16.791	182	94	16.791	182	94	16.791	182
Total Sum							282	50.373	546

Max. pitot press. = 34.0  
Min. pitot press. = 30.0  
Ratio Max:Min = 1.1 :1

Notes : Port B inaccessible due to wall  
Traverse points shown are those reachable with usable length probe in windshell

### Gas Data

Oxygen %	7.0
CO <sub>2</sub> %	12.00

### Oxygen Correction

Required Correction Value	0
Actual Oxygen Factor	1
Enter 0 if correction is not required	

BS EN 13284-1 & M1 Sample Point Requirements	Requirement Met?
Duct gas Flow: angle with regard to duct access <15°?	Y
Duct Gas Flow Negative Velocity: Not Permitted	Y
Duct Gas Flow: Ratio of max to min velocity <3:1?	Y
Working Area > 5m <sup>2</sup> ?	N
Handrails with removable chains / self closing gates across the top of the ladder?	Y
Handrails (approx 0,5 and 1,0 m high) and vertical baseboards (approx 0,25m high)?	Y
Scaffold Built to 'Heavy Duty' Scafftag Rating or at least 2.5kN/m2 loading	N/A
Handrails not restricting access to ports?	N
Room opposite sampling port equal or greater than the length of the sampling probe plus 1 metre?	N
Sufficient Power (Waterproof 110V BS4343 Standard) close or on the platform?	Y

## 2.4.1 Continuous Emissions Monitoring Systems at the installation

	Procal 1 & 2	SICK	SICK
Determinand	NO <sub>x</sub> & SO <sub>2</sub>	NO <sub>x</sub> & SO <sub>2</sub>	Particulate
Type	In Situ IR	In-situ Probe – UV-DOAS	Cross Duct Forward Scatter
Make	Procal	SICK Maihak GmbH	Erwin Sick
Model	Pulsi 200 series	GM32 – In-situ	OMD41
MCERTS Certificate	MC990006/07	MC100163/01	MC040042/01
Certification ranges	NO = 0 – 1000 ppm SO <sub>2</sub> low = 0 – 250 ppm SO <sub>2</sub> high = 0 – 1000 ppm	Measuring path length 1.25m: NO 0 to 70 mg/m <sup>3</sup> & 0 to 700 mg/m <sup>3</sup> SO <sub>2</sub> 0 to 75mg/m <sup>3</sup> & 0 to 1000mg/m <sup>3</sup>  Measuring path length 1.00m NO 0 to 87.5 mg/m <sup>3</sup> & 0 to 875 mg/m <sup>3</sup> SO <sub>2</sub> 0 to 93.8 mg/m <sup>3</sup> and 0 to 1250 mg/m <sup>3</sup>	0 – 0.3 extinction
Operational ranges	As Above	SO <sub>2</sub> Low: 0 – 250 ppm SO <sub>2</sub> High: 0 – 1000ppm NO: 0 – 1000ppm NO <sub>2</sub> : 0 - 100ppm	0 – 50 mg/m <sup>3</sup>
Principle	Dual wavelength infra-red	UV DOAS	Opacity
Raw data units	ppm	ppm	Extinction
Reference condition of raw data	wet gas, no oxygen correction	wet gas, no oxygen correction	wet gas, no oxygen, temp or pressure correction
Signal output	4-20 mA	4 – 20mA	Fibre optics
Provision for logging of 14181 data	Data logged by DCS		
Location of sample	Measurement taken at monitoring platform. SRM sample obtained from monitoring platform.		
Moisture – Measurement or calculated	Measured Measured	Measured	Measured

### Section 3 – Information about the Monitoring campaign

**Table 3.1 - Monitoring Organisation Staff Details**

Project Manager	Position	MCERTS Level	Technical Endorsements	Expiry Dates	MCERTS Registration Number
Glyn Harrison	Operations Manager (Stack Emissions)	2	1	10/19	MM 03 228
			2	06/17	
			3	12/16	

Site Team	Position	MCERTS Level	Technical Endorsements	Expiry Dates	MCERTS Registration Number
Ian Baggley	Consultant	2	1	06/18	MM 05 642
			2	03/19	
			3	12/17	
			4	03/17	
Chris Davies	Consultant	2	1	12/19	MM 03 252
			2	12/15	
			3	12/17	
			4	06/16	
Will Doward	Technician	1	-	-	MM 13 1249

Report Author	Position	MCERTS Level	Technical Endorsements	Expiry Dates	MCERTS Registration Number
Glyn Harrison	Operations Manager (Stack Emissions)	2	1	10/19	MM 03 228
			2	06/17	
			3	12/16	

Report Reviewer	Position	MCERTS Level	Technical Endorsements	Expiry Dates	MCERTS Registration Number
Richard Harvey	Principal Consultant	2	1	11/17	MM 02 020
			2	03/19	
			3	03/16	
			4	12/15	

### 3.2 - Monitoring Organisation Method Details

Emission Parameter	Standard Method	Monitoring Procedure No.	Monitoring Accreditation Status	Analysis Technique	Expected Uncertainty (%)	Analysis Procedure No.	Analytical Laboratory	Analysis Accreditation Status
Oxides of Nitrogen (as NO <sub>2</sub> )	BS EN 14792:2005	RPSCE/1/21f	MCERTS	Chemiluminescence	6	N/A	N/A	N/A
Oxygen	BS EN 14789:2005	RPSCE/1/21g	MCERTS	Zirconia Cell	5	N/A	N/A	N/A
Sulphur Dioxide	TGN M22	RPSCE/1/24	MCERTS	FTIR	10	N/A	N/A	N/A
Total particulate Matter	BS EN 13284:2002	RPSCE/1/7c	MCERTS	Gravimetric	10 - 30	D9	RPS	UKAS

## Equipment details

Emission Parameter	Analysis Technique	Analyser	Analyser Certification Status	Certified Ranges	Operational Ranges	Operating Principle
Oxygen	Zirconia Cell	Horiba PG 250 SRM	MCERTs certificate No MC110186/03	0 – 25%	0-25%	Extractive, multicomponent dry gas analyser. Sample extracted through sample probe and 5metre heated sample line (with integral heated filter) – line temperature 180°C. Sample line connected directly to a gas conditioner (peltier cooler) set at 3°C. Cold dry sample then passes to analyser. Sample is drawn through system by integral pump built into analyser.
Oxides of Nitrogen	Chemiluminescence	Horiba PG 250 SRM	MCERTs certificate No MC110186/03	0 – 130 mg/m <sup>3</sup> as NO 0 – 200 mg/m <sup>3</sup> as NO <sub>2</sub>	0 – 1000ppm	
Sulphur Dioxide	FTIR	Gasmet DX4000	MCERTs certificate No MC30014/05	0 – 75mg/m <sup>3</sup>	0 – 500 ppm	Extractive wet gas analyser. Sample obtained non-isokinetically. Sample extracted through sample probe and filtered before passing through 5metre heated sample line (with integral heated filter) – line temperature 180°C. Sample line connected directly to a heated sample pump which in turn was connected to the FTIR. Hot, wet sample then passes to analyser.
Stack Gas Moisture	FTIR	Gasmet DX4000	MCERTs certificate No MC30014/05	0-40%	0-40%	
Total Particulate Matter	Multipoint isokinetic sampling with in stack filtration	N/A	N/A	0 – 50mg/m <sup>3</sup>	-	Extractive manual test. Sample obtained isokinetically through sharp edged nozzle. Sample gas passed through a pre weighed, pre blown filter. Filter holder mounted in-stack.

## **Section 4A1: Data & calculations – QAL2 – Unit 7, SICK GM32**

## Section 4A – Data and calculations – QAL2 SICK GM32

### 4A1.1 Table 4.1.1 – Raw monitoring Data – Oxides of Nitrogen

Test No	Test Date	Test Start Time	Test End Time	CEMS Raw Value (Wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw Value (dry)	SRM Oxygen (Dry)	SRM Moisture	SRM at CEMs Raw conditions (wet)
		hr:min		NO (ppm)	(%)	(%)	NOx (ppm)	(%)	(%)	NOx (ppm)
1	05/03/2015	15:10	16:10	450.8	6.21	2.31	425.97	6.7	2.1	417.0
2	05/03/2015	17:12	18:12	438.6	6.19	2.31	414.98	6.6	2.1	406.2
3	05/03/2015	19:14	20:14	455.9	6.70	2.23	432.12	6.8	2.1	423.2
4	05/03/2015	21:16	22:16	442.4	6.47	2.25	421.10	6.6	2.1	412.3
5	06/03/2015	01:02	02:02	389.1	6.83	2.30	369.56	7.0	2.1	361.7
6	06/03/2015	03:04	04:04	433.4	7.77	2.13	413.27	7.7	2.0	405.2
7	06/03/2015	05:06	06:06	367.3	7.90	1.97	353.81	7.8	1.8	347.5
8	06/03/2015	07:08	08:08	454.3	6.35	2.21	427.81	6.5	2.0	419.2
9	06/03/2015	09:09	10:09	468.4	6.53	2.25	442.43	6.6	2.1	432.9
10	06/03/2015	11:11	12:11	472.6	6.85	2.22	444.16	6.9	2.2	434.3
11	06/03/2015	13:13	14:13	510.0	5.97	2.45	479.21	6.5	2.4	467.7
12	09/03/2015	12:45	13:45	293.3	4.95	1.97	290.85	5.8	2.0	284.9
13	09/03/2015	14:47	15:47	359.8	6.42	2.10	369.90	6.5	2.1	362.1
14	09/03/2015	16:49	17:49	359.5	6.00	1.98	368.93	6.2	2.0	361.7
15	09/03/2015	18:51	19:51	298.6	5.77	1.92	299.71	6.4	1.9	294.0
16	09/03/2015	20:53	21:53	248.1	6.33	1.89	250.28	6.6	1.9	245.6
17	09/03/2015	22:55	23:55	244.8	6.72	1.88	246.92	6.9	1.8	242.4
18	10/03/2015	01:01	02:01	233.1	6.21	1.94	236.06	6.6	1.9	231.6
19	10/03/2015	03:03	04:03	238.2	6.58	2.04	240.94	7.0	2.0	236.1
20	10/03/2015	05:05	06:05	248.2	6.97	1.92	250.26	7.2	1.9	245.6
21	10/03/2015	07:07	08:07	248.1	6.67	1.88	251.34	7.0	1.8	246.8
22	10/03/2015	09:09	10:09	243.6	6.66	1.90	243.34	6.9	1.9	238.8
23	10/03/2015	13:13	14:13	301.0	6.43	1.88	297.92	6.6	1.9	292.1

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa



#### 4A1.2 Table 4.2.1 - Standardised monitoring Data – Oxides of Nitrogen


Test No	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
	hr:min		NO as NO <sub>2</sub> (mg/m <sup>3</sup> )	NOx as NO <sub>2</sub> (mg/m <sup>3</sup> )	(mg/m <sup>3</sup> )
1	15:10	16:10	961.2	918.4	30.4
2	17:12	18:12	934.0	887.4	29.4
3	19:14	20:14	1004.8	936.3	31.0
4	21:16	22:16	959.6	902.8	29.9
5	1:02	2:02	865.9	815.2	27.0
6	3:04	4:04	1032.4	960.4	31.8
7	5:06	6:06	882.1	825.6	27.4
8	7:08	8:08	976.8	911.6	30.2
9	9:09	10:09	1020.6	948.5	34.0
10	11:11	12:11	1052.5	971.0	32.2
11	13:13	14:13	1071.9	1015.2	33.6
12	12:45	13:45	574.1	590.9	19.7
13	14:47	15:47	776.4	785.8	26.1
14	16:49	17:49	753.4	765.7	25.4
15	18:51	19:51	615.5	632.5	21.1
16	20:53	21:53	531.1	534.3	17.9
17	22:55	23:55	538.3	540.0	18.1
18	1:01	2:01	495.1	503.8	16.9
19	3:03	4:03	519.6	531.5	17.4
20	5:05	6:05	555.8	558.3	18.7
21	7:07	8:07	543.7	553.7	18.5
22	9:09	10:09	533.5	532.3	17.8
23	13:13	14:13	648.7	638.3	21.3

Note:


Emission concentrations expressed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4A1.3 – Plot 1.1 - Time Series of Standardised CEM versus Standardised SRM data – Oxides of Nitrogen, (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A1.4 – Elimination of Outliers – Oxides of Nitrogen



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.

Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9936, an outlier test has not been undertaken.

#### 4A1.5.1 Determination of Method A or Method B - Oxides of Nitrogen

Test No	Test Date	Test Start Time	Test End Time	SRM measured value (y)	SRM O2	SRM Standardised
		hr:min		(ppm)	(%)	(mg/m3)
1	05-Mar-15	15:10	16:10	426.0	6.7	918.4
2	05-Mar-15	17:12	18:12	415.0	6.6	887.4
3	05-Mar-15	19:14	20:14	432.1	6.8	936.3
4	05-Mar-15	21:16	22:16	421.1	6.6	902.8
5	06-Mar-15	1:02	2:02	369.6	7.0	815.2
6	06-Mar-15	3:04	4:04	413.3	7.7	960.4
7	06-Mar-15	5:06	6:06	353.8	7.8	825.6
8	06-Mar-15	7:08	8:08	427.8	6.5	911.6
9	06-Mar-15	9:09	10:09	442.4	6.6	948.5
10	06-Mar-15	11:11	12:11	444.2	6.9	971.0
11	06-Mar-15	13:13	14:13	479.2	6.5	1015.2
12	09-Mar-15	12:45	13:45	290.9	5.8	590.9
13	09-Mar-15	14:47	15:47	369.9	6.5	785.8
14	09-Mar-15	16:49	17:49	368.9	6.2	765.7
15	09-Mar-15	20:53	21:53	299.7	6.6	639.8
16	09-Mar-15	22:55	23:55	250.3	6.9	547.4
17	09-Mar-15	1:01	2:01	246.9	6.6	527.0
18	10-Mar-15	3:03	4:03	236.1	7.0	520.7
19	10-Mar-15	5:05	6:05	240.9	7.2	537.5
20	10-Mar-15	7:07	8:07	250.3	7.0	551.4
21	10-Mar-15	9:09	10:09	251.3	6.9	549.8
22	10-Mar-15	13:13	14:13	243.3	6.6	521.3
23	10-Mar-15	13:13	14:13	297.9	6.6	638.3
Sum				7970.88		
Emission Limit Value (ELV) =				1210 mg/Nm <sup>3</sup>	Y <sub>max</sub>	1015.22
15% of the ELV =				181.5 mg/Nm <sup>3</sup>	Y <sub>min</sub>	520.71
Therefore Ymax - Ymin > 15% of the ELV				<b>Method A</b>	Y <sub>max</sub> - Ymin	494.51

#### 4A1.6.1 Table 4.3.1 - Data used to derive calibration function - Oxides of Nitrogen,

Test No	Test Date	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)	Yi	Xi	Xi * Yi	Xi <sup>2</sup>	b
		hr:min		NOx (ppm)	NO (ppm)	1	2	3	4	
1	09-May-14		Reference	0	0.33	-325.3751019	-341.3130754	111054.7767	116494.6155	
2	05-Mar-15	15:10	16:10	417.0	450.8	91.66	109.17	10006.54	11917.65	
3	05-Mar-15	17:12	18:12	406.2	438.6	80.80	96.93	7831.75	9394.95	
4	05-Mar-15	19:14	20:14	423.2	455.9	97.86	114.30	11185.28	13063.77	
5	05-Mar-15	21:16	22:16	412.3	442.4	86.92	100.73	8754.80	10146.01	
6	06-Mar-15	1:02	2:02	361.7	389.1	36.31	47.49	1724.70	2255.65	
7	06-Mar-15	3:04	4:04	405.2	433.4	79.83	91.78	7327.15	8424.02	
8	06-Mar-15	5:06	6:06	347.5	367.3	22.09	25.64	566.51	657.59	
9	06-Mar-15	7:08	8:08	419.2	454.3	93.83	112.64	10568.73	12686.86	
10	06-Mar-15	9:09	10:09	432.9	468.4	107.57	126.74	13633.47	16063.16	
11	06-Mar-15	11:11	12:11	434.3	472.6	108.96	130.96	14269.47	17149.92	
12	06-Mar-15	13:13	14:13	467.7	510.0	142.31	168.40	23964.11	28357.60	
13	09-Mar-15	12:45	13:45	284.9	293.3	-40.48	-48.38	1958.38	2340.59	
14	09-Mar-15	14:47	15:47	362.1	359.8	36.77	18.12	666.52	328.51	
15	09-Mar-15	16:49	17:49	361.7	359.5	36.30	17.89	649.32	319.92	
16	09-Mar-15	18:51	19:51	294.0	298.6	-31.40	-43.07	1352.42	1855.45	
17	09-Mar-15	20:53	21:53	245.6	248.1	-79.78	-93.55	7463.76	8751.65	
18	09-Mar-15	22:55	23:55	242.4	244.8	-83.00	-96.82	8036.00	9373.14	
19	10-Mar-15	1:01	2:01	231.6	233.1	-93.80	-108.51	10178.01	11774.04	
20	10-Mar-15	3:03	4:03	236.1	238.2	-89.23	-103.47	9232.70	10705.85	
21	10-Mar-15	5:05	6:05	245.6	248.2	-79.76	-93.41	7449.97	8725.16	
22	10-Mar-15	7:07	8:07	246.8	248.1	-78.58	-93.58	7353.82	8757.39	
23	10-Mar-15	9:09	10:09	238.8	243.6	-86.56	-98.05	8487.70	9613.85	
24	10-Mar-15	13:13	14:13	292.1	301.0	-33.25	-40.63	1351.01	1650.95	
Sum				7809.00	8199.43	0.00	0.00	285066.91	320808.29	0.89

#### 4A1.7.1 Determination of Calibration Function - Oxides of Nitrogen

##### Method A

If Ymax-Ymin >15% of the ELV, the following formulae are used:

$$b = \frac{\sum_{i=1}^N (x_i - \bar{x})(y_i - \bar{y})}{\sum_{i=1}^N (x_i - \bar{x})^2} \quad \text{where} \quad \bar{x} = \frac{1}{N} \sum_{i=1}^N x_i \quad \bar{y} = \frac{1}{N} \sum_{i=1}^N y_i$$


$$a = y - bx \quad a = 325.38 - 341.65 * 0.889 \quad a = 21.795$$

The calibration is function  $y_i = a + b x_i$  or  $y_i = 21.795 + 0.889 * x_i$

#### 4A1.8.1 Table 4.4.1 - Calculation of calibrated CEMS values - Oxides of Nitrogen

Test No	Test Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS Moisture	CEMS wet Oxygen	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		NO (ppm)	NO (ppm)	(%)	(%)	NO as NO2 (mg/m3)	NOx as NO2 (mg/m3)
1	09-May-14	Reference Gas		0.3	22.1			22.1	0.0
2	05-Mar-15	15:10	16:10	450.8	422.4	2.3	6.2	900.6	918.4
3	05-Mar-15	17:12	18:12	438.6	411.5	2.3	6.2	876.3	887.4
4	05-Mar-15	19:14	20:14	455.9	426.9	2.2	6.7	940.9	936.3
5	05-Mar-15	21:16	22:16	442.4	414.9	2.3	6.5	899.9	902.8
6	06-Mar-15	01:02	02:02	389.1	367.6	2.3	6.8	817.9	815.2
7	06-Mar-15	03:04	04:04	433.4	406.9	2.1	7.8	969.3	960.4
8	06-Mar-15	05:06	06:06	367.3	348.2	2.0	7.9	836.2	825.6
9	06-Mar-15	07:08	08:08	454.3	425.5	2.2	6.3	914.8	911.6
10	06-Mar-15	09:09	10:09	468.4	438.0	2.3	6.5	954.4	948.5
11	06-Mar-15	11:11	12:11	472.6	441.7	2.2	6.8	983.8	971.0
12	06-Mar-15	13:13	14:13	510.0	475.0	2.5	6.0	998.3	1015.2
13	09-Mar-15	12:45	13:45	293.3	282.4	2.0	5.0	552.8	590.9
14	09-Mar-15	14:47	15:47	359.8	341.5	2.1	6.4	737.0	785.8
15	09-Mar-15	16:49	17:49	359.5	341.3	2.0	6.0	715.1	765.7
16	09-Mar-15	18:51	19:51	298.6	287.1	1.9	5.8	591.8	632.5
17	09-Mar-15	20:53	21:53	248.1	242.2	1.9	6.3	518.6	534.3
18	09-Mar-15	22:55	23:55	244.8	239.3	1.9	6.7	526.3	540.0
19	10-Mar-15	01:01	02:01	233.1	229.0	1.9	6.2	486.2	503.8
20	10-Mar-15	03:03	04:03	238.2	233.4	2.0	6.6	509.2	531.5
21	10-Mar-15	05:05	06:05	248.2	242.4	1.9	7.0	542.7	558.3
22	10-Mar-15	07:07	08:07	248.1	242.2	1.9	6.7	530.9	553.7
23	10-Mar-15	09:09	10:09	243.6	238.2	1.9	6.7	521.8	532.3
24	10-Mar-15	13:13	14:13	301.0	289.3	1.9	6.4	623.4	638.3
Sum								16970.3	
Emission Limit Value (ELV) =				1210	mg/Nm <sup>3</sup>				
				Reference Oxygen		6 %			

**4A1.9.1 Plot 2.1 CEM versus SRM Parallel Test Data at CEM measurement conditions –NOx ppm, wet gas.**



#### 4A1.10.1 Table 4.5.1 – Data used for the Variability Test – Oxides of Nitrogen


Test No	Test Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1		Reference Gas		22.1	0.0	-22.09	-34.14	1165.32
2	05-Mar-15	15:10	16:10	900.6	918.4	17.82	5.78	33.36
3	05-Mar-15	17:12	18:12	876.3	887.4	11.04	-1.01	1.02
4	05-Mar-15	19:14	20:14	940.9	936.3	-4.63	-16.68	278.18
5	05-Mar-15	21:16	22:16	899.9	902.8	2.81	-9.24	85.42
6	06-Mar-15	01:02	02:02	817.9	815.2	-2.77	-14.82	219.60
7	06-Mar-15	03:04	04:04	969.3	960.4	-8.84	-20.89	436.38
8	06-Mar-15	05:06	06:06	836.2	825.6	-10.58	-22.63	512.14
9	06-Mar-15	07:08	08:08	914.8	911.6	-3.21	-15.26	232.77
10	06-Mar-15	09:09	10:09	954.4	948.5	-5.83	-17.88	319.76
11	06-Mar-15	11:11	12:11	983.8	971.0	-12.74	-24.79	614.38
12	06-Mar-15	13:13	14:13	998.3	1015.2	16.91	4.86	23.65
13	09-Mar-15	12:45	13:45	552.8	590.9	38.08	26.03	677.50
14	09-Mar-15	14:47	15:47	737.0	785.8	48.85	36.80	1354.30
15	09-Mar-15	16:49	17:49	715.1	765.7	50.54	38.49	1481.59
16	09-Mar-15	18:51	19:51	591.8	632.5	40.71	28.66	821.67
17	09-Mar-15	20:53	21:53	518.6	534.3	15.69	3.64	13.24
18	09-Mar-15	22:55	23:55	526.3	540.0	13.76	1.71	2.92
19	10-Mar-15	01:01	02:01	486.2	503.8	17.63	5.58	31.11
20	10-Mar-15	03:03	04:03	509.2	531.5	22.27	10.23	104.56
21	10-Mar-15	05:05	06:05	542.7	558.3	15.63	3.58	12.84
22	10-Mar-15	07:07	08:07	530.9	553.7	22.85	10.80	116.70
23	10-Mar-15	09:09	10:09	521.8	532.3	10.42	-1.63	2.65
24	10-Mar-15	13:13	14:13	623.4	638.3	14.84	2.80	7.81
24	Tests		Mean			12.05		
	Sum							8548.87

#### 4A1.11.1 - Variability Test Calculation – Oxides of Nitrogen

SD=	Root(1-Number).Integral(D1-D) <sup>2</sup>	19.28	mg/m3(s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2*1210 mg/m3 (s,d,6%O <sub>2</sub> )/1.96	123.47	mg/m3(s,d),6%O <sub>2</sub>
For 25 tests, k <sub>v</sub> =	0.9861		
Therefore variability=	19.28 <= 123.47 * 0.9861		
	or 19.28 <=	121.75	
Which is TRUE therefore the CEMS passes the test			



**4A1.12.1 Plot 3.1 –Standardised CEM data versus standardised SRM - Oxides of Nitrogen (as NO<sub>2</sub>) – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen**



#### 4A1.1.2 Table 4.1.2 – Raw monitoring Data – Sulphur Dioxide, Low Range

Test No	Test Date	Test Start Time	Test End Time	CEMS Raw Value (Wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw Value (Wet)	SRM Oxygen (Dry)	SRM Moisture	SRM at CEMs Raw conditions (wet)
		hr:min		SO2 (ppm)	(%)	(%)	SO2 (ppm)	(%)	(%)	SO2 (ppm)
1	05/03/2015	15:10	16:10	63.4	6.21	2.31	67.99	6.7	2.1	68.0
2	05/03/2015	17:12	18:12	51.7	6.19	2.31	55.77	6.6	2.1	55.8
3	05/03/2015	19:14	20:14	54.9	6.70	2.23	54.98	6.8	2.1	55.0
4	05/03/2015	21:16	22:16	77.8	6.47	2.25	83.10	6.6	2.1	83.1
5	06/03/2015	01:02	02:02	84.1	6.83	2.30	89.82	7.0	2.1	89.8
6	06/03/2015	03:04	04:04	79.2	7.77	2.13	85.68	7.7	2.0	85.7
7	06/03/2015	05:06	06:06	77.1	7.90	1.97	84.21	7.8	1.8	84.2
8	06/03/2015	07:08	08:08	87.1	6.35	2.21	92.40	6.5	2.0	92.4
9	06/03/2015	09:09	10:09	79.3	6.53	2.25	84.79	6.6	2.1	84.8
10	06/03/2015	11:11	12:11	72.6	6.85	2.22	76.61	6.9	2.2	76.6
11	06/03/2015	13:13	14:13	70.7	5.97	2.45	74.34	6.5	2.4	74.3
12	09/03/2015	12:45	13:45	54.6	4.95	1.97	59.06	5.8	2.0	59.1
13	09/03/2015	14:47	15:47	45.6	6.42	2.10	49.42	6.5	2.1	49.4
14	09/03/2015	16:49	17:49	49.0	6.00	1.98	53.39	6.2	2.0	53.4
15	09/03/2015	18:51	19:51	55.1	5.77	1.92	59.36	6.4	1.9	59.4
16	09/03/2015	20:53	21:53	55.9	6.33	1.89	61.15	6.6	1.9	61.2
17	09/03/2015	22:55	23:55	53.6	6.72	1.88	59.25	6.9	1.8	59.3
18	10/03/2015	01:01	02:01	59.0	6.21	1.94	65.53	6.6	1.9	65.5
19	10/03/2015	03:03	04:03	53.6	6.58	2.04	59.42	7.0	2.0	59.4
20	10/03/2015	05:05	06:05	51.5	6.97	1.92	57.07	7.2	1.9	57.1
21	10/03/2015	07:07	08:07	62.6	6.67	1.88	68.99	7.0	1.8	69.0
22	10/03/2015	09:09	10:09	70.4	6.66	1.90	77.28	6.9	1.9	77.3
23	10/03/2015	13:13	14:13	65.1	6.43	1.88	71.51	6.6	1.9	71.5

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

#### 4A1.2.2 Table 4.2.2 -Standardised monitoring Data – Sulphur Dioxide, Low Range


Test No	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
	hr:min		SO <sub>2</sub> (mg/m3)	SO <sub>2</sub> (mg/m3)	(mg/m3)
1	15:10	16:10	188.2	208.2	31.9
2	17:12	18:12	153.1	169.6	31.7
3	19:14	20:14	168.2	169.6	32.2
4	21:16	22:16	234.9	252.7	32.2
5	1:02	2:02	260.4	281.0	32.4
6	3:04	4:04	262.4	282.0	32.4
7	5:06	6:06	257.4	278.7	32.4
8	7:08	8:08	260.5	278.7	32.4
9	9:09	10:09	240.4	257.8	32.2
10	11:11	12:11	225.0	238.2	32.1
11	13:13	14:13	206.8	225.2	31.4
12	12:45	13:45	148.6	169.9	31.7
13	14:47	15:47	136.9	149.2	31.7
14	16:49	17:49	142.9	157.8	31.7
15	18:51	19:51	158.0	177.7	31.8
16	20:53	21:53	166.4	185.6	31.8
17	22:55	23:55	163.9	183.5	31.8
18	1:01	2:01	174.3	198.9	31.9
19	3:03	4:03	162.7	185.7	31.1
20	5:05	6:05	160.5	180.8	31.1
21	7:07	8:07	190.9	215.2	32.0
22	9:09	10:09	214.6	239.5	32.1
23	13:13	14:13	195.2	217.0	32.0

Note:


Emission concentrations expressed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4A1.3.2 – Plot 1.2 - Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide, Low Range (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A1.4.2 – Elimination of Outliers – Sulphur Dioxide, Low Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.

Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9883, an outlier test has not been undertaken.

#### 4A1.5.2 Determination of Method A or Method B - Sulphur Dioxide, Low Range.

Test No	Test Date	Test Start Time	Test End Time	SRM measured value (y)	SRM Moisture	SRM O2	SRM Standardised
		hr:min		(ppm)	(%)	(%)	(mg/m3)
1	05-Mar-15	15:10	16:10	68.0	2.10	6.7	208.2
2	05-Mar-15	17:12	18:12	55.8	2.10	6.6	169.6
3	05-Mar-15	19:14	20:14	55.0	2.10	6.8	169.6
4	05-Mar-15	21:16	22:16	83.1	2.10	6.6	252.7
5	06-Mar-15	1:02	2:02	89.8	2.10	7.0	281.0
6	06-Mar-15	3:04	4:04	85.7	2.00	7.7	282.0
7	06-Mar-15	5:06	6:06	84.2	1.80	7.8	278.7
8	06-Mar-15	7:08	8:08	92.4	2.00	6.5	278.7
9	06-Mar-15	9:09	10:09	84.8	2.10	6.6	257.8
10	06-Mar-15	11:11	12:11	76.6	2.20	6.9	238.2
11	06-Mar-15	13:13	14:13	74.3	2.40	6.5	225.2
12	09-Mar-15	12:45	13:45	59.1	2.00	5.8	169.9
13	09-Mar-15	14:47	15:47	49.4	2.10	6.5	149.2
14	09-Mar-15	16:49	17:49	53.4	2.00	6.2	157.8
15	09-Mar-15	18:51	19:51	59.4	1.90	6.4	177.7
16	09-Mar-15	20:53	21:53	61.2	1.90	6.6	185.6
17	09-Mar-15	22:55	23:55	59.3	1.80	6.9	183.5
18	10-Mar-15	1:01	2:01	65.5	1.90	6.6	198.9
19	10-Mar-15	3:03	4:03	59.4	2.00	7.0	185.7
20	10-Mar-15	5:05	6:05	57.1	1.90	7.2	180.8
21	10-Mar-15	7:07	8:07	69.0	1.80	7.0	215.2
22	10-Mar-15	9:09	10:09	77.3	1.90	6.9	239.5
23	10-Mar-15	13:13	14:13	71.5	1.90	6.6	217.0
Sum				1591.12			
Emission Limit Value (ELV) =			440	mg/Nm <sup>3</sup>		Y <sub>max</sub>	281.97
15% of the ELV =			66	mg/Nm <sup>3</sup>		Y <sub>min</sub>	149.24
Therefore Ymax - Ymin > 15% of the ELV			Method A			Y <sub>max</sub> - Ymin	132.73

#### 4A1.6.2 Table 4.3.2 - Data used to derive calibration function - Sulphur Dioxide, Low Range.

Test No	Test Date	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)	Yi	Xi	Xi * Yi	Xi <sup>2</sup>	b
		hr:min		SO2 (ppm)	SO2 (ppm)	1	2	3	4	
1	17-Feb-15	Reference Gas		0	0.37	-66.29651389	-61.05410566	4047.674364	3727.603818	
2	05-Mar-15	15:10	16:10	68.0	63.4	1.69	2.02	3.42	4.08	
3	05-Mar-15	17:12	18:12	55.8	51.7	-10.53	-9.75	102.59	94.99	
4	05-Mar-15	19:14	20:14	55.0	54.9	-11.32	-6.56	74.23	43.02	
5	05-Mar-15	21:16	22:16	83.1	77.8	16.80	16.40	275.54	268.88	
6	06-Mar-15	1:02	2:02	89.8	84.1	23.52	22.67	533.32	514.01	
7	06-Mar-15	3:04	4:04	85.7	79.2	19.38	17.78	344.72	316.27	
8	06-Mar-15	5:06	6:06	84.2	77.1	17.91	15.64	280.12	244.54	
9	06-Mar-15	7:08	8:08	92.4	87.1	26.10	25.63	668.97	656.96	
10	06-Mar-15	9:09	10:09	84.8	79.3	18.49	17.89	330.79	319.94	
11	06-Mar-15	11:11	12:11	76.6	72.6	10.31	11.19	115.40	125.19	
12	06-Mar-15	13:13	14:13	74.3	70.7	8.04	9.32	74.99	86.93	
13	09-Mar-15	12:45	13:45	59.1	54.6	-7.24	-6.85	49.56	46.90	
14	09-Mar-15	14:47	15:47	49.4	45.6	-16.88	-15.84	267.33	250.92	
15	09-Mar-15	16:49	17:49	53.4	49.0	-12.91	-12.41	160.12	153.91	
16	09-Mar-15	18:51	19:51	59.4	55.1	-6.94	-6.36	44.10	40.42	
17	09-Mar-15	20:53	21:53	61.2	55.9	-5.15	-5.55	28.57	30.81	
18	09-Mar-15	22:55	23:55	59.3	53.6	-7.05	-7.84	55.27	61.53	
19	10-Mar-15	1:01	2:01	65.5	59.0	-0.77	-2.44	1.87	5.93	
20	10-Mar-15	3:03	4:03	59.4	53.6	-6.88	-7.80	53.65	60.87	
21	10-Mar-15	5:05	6:05	57.1	51.5	-9.23	-9.93	91.59	98.55	
22	10-Mar-15	7:07	8:07	69.0	62.6	2.69	1.19	3.20	1.41	
23	10-Mar-15	9:09	10:09	77.3	70.4	10.98	8.98	98.66	80.68	
24	10-Mar-15	13:13	14:13	71.5	65.1	5.21	3.66	19.09	13.41	
Sum				1591.12	1474.18	0.00	0.00	7724.76	7247.74	1.07

#### 4A1.7.2 Determination of Calibration Function - Sulphur Dioxide, Low Range.

##### Method A

If Ymax-Ymin >15% of the ELV, the following formulae are used:

$$b = \frac{\sum_{i=1}^N (x_i - \bar{x})(y_i - \bar{y})}{\sum_{i=1}^N (x_i - \bar{x})^2} \quad \text{where} \quad \bar{x} = \frac{1}{N} \sum_{i=1}^N x_i \quad \bar{y} = \frac{1}{N} \sum_{i=1}^N y_i$$

$$a = y - bx \quad a = 66.3 - 61.43 * 1.065$$


The calibration is function  $y_i = a + b x_i$  or  $y_i = 0.83 + 1.066 * x_i$

#### 4A1.8.2 Table A.4.4.2 - Calculation of calibrated CEMS values - Sulphur Dioxide, Low Range

Test No	Test Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS Moisture	CEMS dry Oxygen	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		SO2 (ppm)	SO2 (ppm)	(%)	(%)	SO2 (mg/m3)	SO2 (mg/m3)
1	17-Feb-15	Reference Gas		0.4	1.2			3.5	0.0
2	05-Mar-15	15:10	16:10	63.4	68.4	2.3	6.2	203.1	208.2
3	05-Mar-15	17:12	18:12	51.7	55.9	2.3	6.2	165.6	169.6
4	05-Mar-15	19:14	20:14	54.9	59.3	2.2	6.7	181.9	169.6
5	05-Mar-15	21:16	22:16	77.8	83.8	2.3	6.5	252.8	252.7
6	06-Mar-15	01:02	02:02	84.1	90.5	2.3	6.8	280.1	281.0
7	06-Mar-15	03:04	04:04	79.2	85.3	2.1	7.8	282.4	282.0
8	06-Mar-15	05:06	06:06	77.1	83.0	2.0	7.9	277.1	278.7
9	06-Mar-15	07:08	08:08	87.1	93.6	2.2	6.4	280.1	278.7
10	06-Mar-15	09:09	10:09	79.3	85.4	2.3	6.5	258.7	257.8
11	06-Mar-15	11:11	12:11	72.6	78.2	2.2	6.9	242.4	238.2
12	06-Mar-15	13:13	14:13	70.7	76.2	2.5	6.0	222.8	225.2
13	09-Mar-15	12:45	13:45	54.6	59.0	2.0	5.0	160.6	169.9
14	09-Mar-15	14:47	15:47	45.6	49.4	2.1	6.4	148.4	149.2
15	09-Mar-15	16:49	17:49	49.0	53.1	2.0	6.0	154.7	157.8
16	09-Mar-15	18:51	19:51	55.1	59.5	1.9	5.8	170.8	177.7
17	09-Mar-15	20:53	21:53	55.9	60.4	1.9	6.3	179.8	185.6
18	09-Mar-15	22:55	23:55	53.6	57.9	1.9	6.7	177.3	183.5
19	10-Mar-15	01:01	02:01	59.0	63.7	1.9	6.2	188.3	198.9
20	10-Mar-15	03:03	04:03	53.6	58.0	2.0	6.6	176.0	185.7
21	10-Mar-15	05:05	06:05	51.5	55.7	1.9	7.0	173.6	180.8
22	10-Mar-15	07:07	08:07	62.6	67.6	1.9	6.7	206.0	215.2
23	10-Mar-15	09:09	10:09	70.4	75.9	1.9	6.7	231.2	239.5
24	10-Mar-15	13:13	14:13	65.1	70.2	1.9	6.4	210.5	217.0
Sum								4827.7	
Emission Limit Value (ELV) =				440	mg/Nm <sup>3</sup>				
				Reference Oxygen		6 %			



#### 4A1.9.2 Plot 2.2 CEM versus SRM Parallel Test Data at CEM measurement conditions –Sulphur dioxide, wet gas




#### 4A1.10.2 Table 4.5.2 – Data used for the Variability Test – Sulphur Dioxide, Low Range.

Test No	Test Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1	17-Feb-15	Reference Gas		3.5	0.0	-3.50	-6.61	43.64
2	05-Mar-15	15:10	16:10	203.1	208.2	5.15	2.04	4.18
3	05-Mar-15	17:12	18:12	165.6	169.6	3.96	0.85	0.73
4	05-Mar-15	19:14	20:14	181.9	169.6	-12.29	-15.40	237.25
5	05-Mar-15	21:16	22:16	252.8	252.7	-0.14	-3.25	10.55
6	06-Mar-15	01:02	02:02	280.1	281.0	0.84	-2.27	5.13
7	06-Mar-15	03:04	04:04	282.4	282.0	-0.46	-3.57	12.72
8	06-Mar-15	05:06	06:06	277.1	278.7	1.53	-1.57	2.48
9	06-Mar-15	07:08	08:08	280.1	278.7	-1.36	-4.47	20.01
10	06-Mar-15	09:09	10:09	258.7	257.8	-0.87	-3.98	15.81
11	06-Mar-15	11:11	12:11	242.4	238.2	-4.20	-7.30	53.36
12	06-Mar-15	13:13	14:13	222.8	225.2	2.35	-0.76	0.58
13	09-Mar-15	12:45	13:45	160.6	169.9	9.27	6.17	38.02
14	09-Mar-15	14:47	15:47	148.4	149.2	0.84	-2.26	5.13
15	09-Mar-15	16:49	17:49	154.7	157.8	3.07	-0.04	0.00
16	09-Mar-15	18:51	19:51	170.8	177.7	6.90	3.79	14.39
17	09-Mar-15	20:53	21:53	179.8	185.6	5.75	2.64	6.97
18	09-Mar-15	22:55	23:55	177.3	183.5	6.20	3.09	9.56
19	10-Mar-15	01:01	02:01	188.3	198.9	10.61	7.50	56.24
20	10-Mar-15	03:03	04:03	176.0	185.7	9.74	6.63	43.96
21	10-Mar-15	05:05	06:05	173.6	180.8	7.17	4.06	16.47
22	10-Mar-15	07:07	08:07	206.0	215.2	9.17	6.07	36.80
23	10-Mar-15	09:09	10:09	231.2	239.5	8.33	5.23	27.31
24	10-Mar-15	13:13	14:13	210.5	217.0	6.52	3.41	11.66
24	Tests		Mean			3.11		
Sum								672.96

#### 4A1.11.2 - Variability Test Calculation – Sulphur Dioxide, Low Range.

SD=	Root(1-Number).Integral(D1-D) <sup>2</sup>	5.41	mg/m3(s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2*440 mg/m3 (s,d,6%O <sub>2</sub> )/1.96	44.90	mg/m3(s,d),6%O <sub>2</sub>
For 25 tests, k <sub>v</sub> =	0.9861		
Therefore variability=	5.41 <= 44.9 * 0.9861		
	or 5.41 <=	44.27	
Which is TRUE therefore the CEMS passes the test			

**4A1.12.2 Plot 3.2 –Standardised CEM data versus standardised SRM - Sulphur dioxide, Low Range – Reference conditions 273K, 101.3kPa 6% oxygen, dry gas.**



#### 4A1.1.3 Table 4.1.3– Raw monitoring Data – Sulphur Dioxide, High Range

Test No	Test Date	Test Start Time	Test End Time	CEMS Raw Value (Wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw Value (Wet)	SRM Oxygen (Dry)	SRM Moisture	SRM at CEMS Raw conditions (wet)
		hr:min		SO2 (ppm)	(%)	(%)	SO2 (ppm)	(%)	(%)	SO2 (ppm)
1	05/03/2015	15:10	16:10	63.7	6.21	2.31	67.99	6.7	2.1	68.0
2	05/03/2015	17:12	18:12	52.0	6.19	2.31	55.77	6.6	2.1	55.8
3	05/03/2015	19:14	20:14	55.1	6.70	2.23	54.98	6.8	2.1	55.0
4	05/03/2015	21:16	22:16	78.1	6.47	2.25	83.10	6.6	2.1	83.1
5	06/03/2015	01:02	02:02	84.4	6.83	2.30	89.82	7.0	2.1	89.8
6	06/03/2015	03:04	04:04	79.5	7.77	2.13	85.68	7.7	2.0	85.7
7	06/03/2015	05:06	06:06	77.3	7.90	1.97	84.21	7.8	1.8	84.2
8	06/03/2015	07:08	08:08	87.3	6.35	2.21	92.40	6.5	2.0	92.4
9	06/03/2015	09:09	10:09	79.6	6.53	2.25	84.79	6.6	2.1	84.8
10	06/03/2015	11:11	12:11	72.9	6.85	2.22	76.61	6.9	2.2	76.6
11	06/03/2015	13:13	14:13	71.1	5.97	2.45	74.34	6.5	2.4	74.3
12	09/03/2015	12:45	13:45	54.9	4.95	1.97	59.06	5.8	2.0	59.1
13	09/03/2015	14:47	15:47	45.9	6.42	2.10	49.42	6.5	2.1	49.4
14	09/03/2015	16:49	17:49	49.4	6.00	1.98	53.39	6.2	2.0	53.4
15	09/03/2015	18:51	19:51	55.4	5.77	1.92	59.36	6.4	1.9	59.4
16	09/03/2015	20:53	21:53	56.2	6.33	1.89	61.15	6.6	1.9	61.2
17	09/03/2015	22:55	23:55	53.9	6.72	1.88	59.25	6.9	1.8	59.3
18	10/03/2015	01:01	02:01	59.3	6.21	1.94	65.53	6.6	1.9	65.5
19	10/03/2015	03:03	04:03	53.9	6.58	2.04	59.42	7.0	2.0	59.4
20	10/03/2015	05:05	06:05	51.8	6.97	1.92	57.07	7.2	1.9	57.1
21	10/03/2015	07:07	08:07	62.9	6.67	1.88	68.99	7.0	1.8	69.0
22	10/03/2015	09:09	10:09	70.6	6.66	1.90	77.28	6.9	1.9	77.3
23	10/03/2015	13:13	14:13	65.4	6.43	1.88	71.51	6.6	1.9	71.5

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

#### 4A1.2.3 Table 4.2.3 -Standardised monitoring Data – Sulphur Dioxide, High Range


Test No	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
	hr:min		SO <sub>2</sub> (mg/m3)	SO <sub>2</sub> (mg/m3)	(mg/m3)
1	15:10	16:10	189.1	208.2	31.9
2	17:12	18:12	154.0	169.6	31.7
3	19:14	20:14	169.1	169.6	32.2
4	21:16	22:16	235.7	252.7	32.2
5	1:02	2:02	261.4	281.0	32.4
6	3:04	4:04	263.4	282.0	32.4
7	5:06	6:06	258.4	278.7	32.4
8	7:08	8:08	261.3	278.7	32.4
9	9:09	10:09	241.3	257.8	32.2
10	11:11	12:11	226.0	238.2	32.1
11	13:13	14:13	207.8	225.2	31.4
12	12:45	13:45	149.5	169.9	31.7
13	14:47	15:47	138.0	149.2	31.7
14	16:49	17:49	144.0	157.8	31.7
15	18:51	19:51	159.0	177.7	31.8
16	20:53	21:53	167.3	185.6	31.8
17	22:55	23:55	164.8	183.5	31.8
18	1:01	2:01	175.2	198.9	31.9
19	3:03	4:03	163.6	185.7	31.1
20	5:05	6:05	161.4	180.8	31.1
21	7:07	8:07	191.8	215.2	32.0
22	9:09	10:09	215.2	239.5	32.1
23	13:13	14:13	196.0	217.0	32.0

Note:


Emission concentrations expressed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4A1.3.3 – Plot 1.3 - Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide, High Range (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A1.4.3 – Elimination of Outliers – Sulphur Dioxide, High Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.

Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9882, an outlier test has not been undertaken.

#### 4A1.5.3 Determination of Method A or Method B - Sulphur Dioxide, High Range.

Test No	Test Date	Test Start Time	Test End Time	SRM measured value (y)	SRM Moisture	SRM O2	SRM Standardised
		hr:min		(ppm)	(%)	(%)	(mg/m3)
1	05-Mar-15	15:10	16:10	68.0	2.10	6.7	208.2
2	05-Mar-15	17:12	18:12	55.8	2.10	6.6	169.6
3	05-Mar-15	19:14	20:14	55.0	2.10	6.8	169.6
4	05-Mar-15	21:16	22:16	83.1	2.10	6.6	252.7
5	06-Mar-15	1:02	2:02	89.8	2.10	7.0	281.0
6	06-Mar-15	3:04	4:04	85.7	2.00	7.7	282.0
7	06-Mar-15	5:06	6:06	84.2	1.80	7.8	278.7
8	06-Mar-15	7:08	8:08	92.4	2.00	6.5	278.7
9	06-Mar-15	9:09	10:09	84.8	2.10	6.6	257.8
10	06-Mar-15	11:11	12:11	76.6	2.20	6.9	238.2
11	06-Mar-15	13:13	14:13	74.3	2.40	6.5	225.2
12	09-Mar-15	12:45	13:45	59.1	2.00	5.8	169.9
13	09-Mar-15	14:47	15:47	49.4	2.10	6.5	149.2
14	09-Mar-15	16:49	17:49	53.4	2.00	6.2	157.8
15	09-Mar-15	18:51	19:51	59.4	1.90	6.4	177.7
16	09-Mar-15	20:53	21:53	61.2	1.90	6.6	185.6
17	09-Mar-15	22:55	23:55	59.3	1.80	6.9	183.5
18	10-Mar-15	1:01	2:01	65.5	1.90	6.6	198.9
19	10-Mar-15	3:03	4:03	59.4	2.00	7.0	185.7
20	10-Mar-15	5:05	6:05	57.1	1.90	7.2	180.8
21	10-Mar-15	7:07	8:07	69.0	1.80	7.0	215.2
22	10-Mar-15	9:09	10:09	77.3	1.90	6.9	239.5
23	10-Mar-15	13:13	14:13	71.5	1.90	6.6	217.0
Sum				1591.12			
Emission Limit Value (ELV) =				440	mg/Nm <sup>3</sup>		
15% of the ELV =				66	mg/Nm <sup>3</sup>		
Therefore Ymax - Ymin > 15% of the ELV				Method A			
						Y <sub>max</sub>	281.97
						Y <sub>min</sub>	149.24
						Y <sub>max</sub> - Ymin	132.73



#### 4A1.6.3 Table 4.3.3 - Data used to derive calibration function - Sulphur Dioxide, High Range.

Test No	Test Date	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)	Yi	Xi	Xi * Yi	Xi <sup>2</sup>	b
		hr:min		SO2 (ppm)	SO2 (ppm)					
1	17-Feb-15	Reference Gas		0	0.37	-66.29651389	-61.3427173	4066.808309	3762.928965	
2	05-Mar-15	15:10	16:10	68.0	63.7	1.69	2.02	3.41	4.07	
3	05-Mar-15	17:12	18:12	55.8	52.0	-10.53	-9.74	102.54	94.89	
4	05-Mar-15	19:14	20:14	55.0	55.1	-11.32	-6.57	74.31	43.11	
5	05-Mar-15	21:16	22:16	83.1	78.1	16.80	16.39	275.36	268.54	
6	06-Mar-15	1:02	2:02	89.8	84.4	23.52	22.68	533.58	514.51	
7	06-Mar-15	3:04	4:04	85.7	79.5	19.38	17.80	345.04	316.87	
8	06-Mar-15	5:06	6:06	84.2	77.3	17.91	15.63	280.03	244.37	
9	06-Mar-15	7:08	8:08	92.4	87.3	26.10	25.63	669.02	657.05	
10	06-Mar-15	9:09	10:09	84.8	79.6	18.49	17.89	330.85	320.05	
11	06-Mar-15	11:11	12:11	76.6	72.9	10.31	11.20	115.55	125.53	
12	06-Mar-15	13:13	14:13	74.3	71.1	8.04	9.37	75.40	87.88	
13	09-Mar-15	12:45	13:45	59.1	54.9	-7.24	-6.79	49.12	46.07	
14	09-Mar-15	14:47	15:47	49.4	45.9	-16.88	-15.77	266.15	248.70	
15	09-Mar-15	16:49	17:49	53.4	49.4	-12.91	-12.33	159.08	151.92	
16	09-Mar-15	18:51	19:51	59.4	55.4	-6.94	-6.30	43.72	39.73	
17	09-Mar-15	20:53	21:53	61.2	56.2	-5.15	-5.54	28.50	30.66	
18	09-Mar-15	22:55	23:55	59.3	53.9	-7.05	-7.85	55.32	61.63	
19	10-Mar-15	1:01	2:01	65.5	59.3	-0.77	-2.42	1.86	5.87	
20	10-Mar-15	3:03	4:03	59.4	53.9	-6.88	-7.81	53.71	61.00	
21	10-Mar-15	5:05	6:05	57.1	51.8	-9.23	-9.93	91.59	98.55	
22	10-Mar-15	7:07	8:07	69.0	62.9	2.69	1.18	3.18	1.39	
23	10-Mar-15	9:09	10:09	77.3	70.6	10.98	8.92	97.95	79.53	
24	10-Mar-15	13:13	14:13	71.5	65.4	5.21	3.67	19.12	13.45	
Sum				1591.12	1481.11	0.00	0.00	7741.19	7278.30	1.06

#### 4A1.7.3 Determination of Calibration Function - Sulphur Dioxide, High Range.

##### Method A

If  $Y_{max} - Y_{min} > 15\%$  of the ELV, the following formulae are used:

$$b = \frac{\sum_{i=1}^N (x_i - \bar{x})(y_i - \bar{y})}{\sum_{i=1}^N (x_i - \bar{x})^2} \quad \text{where} \quad \bar{x} = \frac{1}{N} \sum_{i=1}^N x_i \quad \bar{y} = \frac{1}{N} \sum_{i=1}^N y_i$$


$$a = \bar{y} - b\bar{x} \quad a = 66.3 - 61.72 * 1.063$$

The calibration is function  $y_i = a + b x_i$  or  $y_i = 0.659 + 1.064 * x_i$

#### 4A1.8.3 Table A.4.4.3 - Calculation of calibrated CEMS values - Sulphur Dioxide, High Range.

Test No	Test Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS Moisture	CEMS dry Oxygen	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		SO2 (ppm)	SO2 (ppm)	(%)	(%)	SO2 (mg/m3)	SO2 (mg/m3)
1	17-Feb-15	Reference Gas		0.4	1.1			3.0	0.0
2	05-Mar-15	15:10	16:10	63.7	68.4	2.3	6.2	203.0	208.2
3	05-Mar-15	17:12	18:12	52.0	55.9	2.3	6.2	165.7	169.6
4	05-Mar-15	19:14	20:14	55.1	59.3	2.2	6.7	181.9	169.6
5	05-Mar-15	21:16	22:16	78.1	83.7	2.3	6.5	252.7	252.7
6	06-Mar-15	01:02	02:02	84.4	90.4	2.3	6.8	280.0	281.0
7	06-Mar-15	03:04	04:04	79.5	85.2	2.1	7.8	282.4	282.0
8	06-Mar-15	05:06	06:06	77.3	82.9	2.0	7.9	277.0	278.7
9	06-Mar-15	07:08	08:08	87.3	93.6	2.2	6.4	279.9	278.7
10	06-Mar-15	09:09	10:09	79.6	85.3	2.3	6.5	258.6	257.8
11	06-Mar-15	11:11	12:11	72.9	78.2	2.2	6.9	242.4	238.2
12	06-Mar-15	13:13	14:13	71.1	76.3	2.5	6.0	222.9	225.2
13	09-Mar-15	12:45	13:45	54.9	59.1	2.0	5.0	160.8	169.9
14	09-Mar-15	14:47	15:47	45.9	49.5	2.1	6.4	148.7	149.2
15	09-Mar-15	16:49	17:49	49.4	53.2	2.0	6.0	155.0	157.8
16	09-Mar-15	18:51	19:51	55.4	59.6	1.9	5.8	171.0	177.7
17	09-Mar-15	20:53	21:53	56.2	60.4	1.9	6.3	179.9	185.6
18	09-Mar-15	22:55	23:55	53.9	57.9	1.9	6.7	177.3	183.5
19	10-Mar-15	01:01	02:01	59.3	63.7	1.9	6.2	188.3	198.9
20	10-Mar-15	03:03	04:03	53.9	58.0	2.0	6.6	176.0	185.7
21	10-Mar-15	05:05	06:05	51.8	55.7	1.9	7.0	173.7	180.8
22	10-Mar-15	07:07	08:07	62.9	67.6	1.9	6.7	206.0	215.2
23	10-Mar-15	09:09	10:09	70.6	75.8	1.9	6.7	230.9	239.5
24	10-Mar-15	13:13	14:13	65.4	70.2	1.9	6.4	210.5	217.0
Sum								4827.6	
Emission Limit Value (ELV) =				440	mg/Nm <sup>3</sup>				
				Reference Oxygen		6 %			

**4A1.9.3 Plot 2.3 CEM versus SRM Parallel Test Data at CEM measurement conditions –Sulphur dioxide, wet gas.**




#### 4A1.10.3 Table 4.5.3 – Data used for the Variability Test – Sulphur Dioxide, High Range.

Test No	Test Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1	17-Feb-15	Reference Gas		3.0	0.0	-3.01	-6.12	37.43
2	05-Mar-15	15:10	16:10	203.0	208.2	5.17	2.06	4.26
3	05-Mar-15	17:12	18:12	165.7	169.6	3.88	0.77	0.59
4	05-Mar-15	19:14	20:14	181.9	169.6	-12.32	-15.43	237.97
5	05-Mar-15	21:16	22:16	252.7	252.7	0.00	-3.11	9.66
6	06-Mar-15	01:02	02:02	280.0	281.0	0.96	-2.15	4.61
7	06-Mar-15	03:04	04:04	282.4	282.0	-0.39	-3.50	12.23
8	06-Mar-15	05:06	06:06	277.0	278.7	1.67	-1.44	2.08
9	06-Mar-15	07:08	08:08	279.9	278.7	-1.20	-4.31	18.58
10	06-Mar-15	09:09	10:09	258.6	257.8	-0.76	-3.87	14.96
11	06-Mar-15	11:11	12:11	242.4	238.2	-4.17	-7.28	52.98
12	06-Mar-15	13:13	14:13	222.9	225.2	2.25	-0.86	0.74
13	09-Mar-15	12:45	13:45	160.8	169.9	9.06	5.95	35.36
14	09-Mar-15	14:47	15:47	148.7	149.2	0.51	-2.60	6.74
15	09-Mar-15	16:49	17:49	155.0	157.8	2.74	-0.37	0.14
16	09-Mar-15	18:51	19:51	171.0	177.7	6.69	3.58	12.85
17	09-Mar-15	20:53	21:53	179.9	185.6	5.67	2.56	6.55
18	09-Mar-15	22:55	23:55	177.3	183.5	6.17	3.06	9.35
19	10-Mar-15	01:01	02:01	188.3	198.9	10.55	7.44	55.40
20	10-Mar-15	03:03	04:03	176.0	185.7	9.71	6.60	43.61
21	10-Mar-15	05:05	06:05	173.7	180.8	7.10	3.99	15.91
22	10-Mar-15	07:07	08:07	206.0	215.2	9.21	6.10	37.17
23	10-Mar-15	09:09	10:09	230.9	239.5	8.60	5.49	30.18
24	10-Mar-15	13:13	14:13	210.5	217.0	6.53	3.42	11.68
24	Tests		Mean			3.11		
Sum								661.02

#### 4A1.11.3 - Variability Test Calculation – Sulphur Dioxide, High Range.

SD=	Root(1-Number).Integral(D1-D) <sup>2</sup>	5.36	mg/m3(s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2*440 mg/m3 (s,d,6%O <sub>2</sub> )/1.96	44.90	mg/m3(s,d),6%O <sub>2</sub>
For 30 tests, k <sub>v</sub> =	0.9885		
Therefore variability=	5.36 <= 44.9 * 0.9885		
	or 5.36 <=	44.38	
Which is TRUE therefore the CEMS passes the test			

**4A1.12.3 Plot 3.3 –Standardised CEM data versus standardised SRM - Sulphur dioxide, High Range – Reference conditions 273K, 101.3kPa., 6% oxygen, dry gas.**



## **Section 4B1: Data & Calculations – AST – Unit 7, Procal 1**

## Section 4B1 – Data and calculations – AST Procal 1

**4B1 4.1.1 Table 4.1.1 – Raw monitoring Data – Oxides of Nitrogen**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (dry)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMs Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	05-Mar-15	15:10	16:10	438.13	6.21	2.31	425.97	6.71	2.10	417.04
2	05-Mar-15	17:12	18:12	428.47	6.19	2.31	414.98	6.59	2.12	406.18
3	05-Mar-15	19:14	20:14	451.51	6.70	2.23	432.12	6.78	2.06	423.24
4	05-Mar-15	21:16	22:16	438.19	6.47	2.25	421.10	6.63	2.09	412.29
5	06-Mar-15	1:02	2:02	383.85	6.83	2.30	369.56	7.03	2.13	361.69
6	06-Mar-15	3:04	4:04	437.98	7.77	2.13	413.27	7.73	1.95	405.21
7	06-Mar-15	5:06	6:06	377.60	7.90	1.97	353.81	7.79	1.79	347.47
8	06-Mar-15	7:08	8:08	448.41	6.35	2.21	427.81	6.54	2.01	419.21
9	06-Mar-15	9:09	10:09	465.80	6.53	2.25	442.43	6.63	2.14	432.95
10	06-Mar-15	11:11	12:11	469.95	6.85	2.22	444.16	6.90	2.21	434.34
11	06-Mar-15	13:13	14:13	499.71	5.97	2.45	479.21	6.46	2.41	467.68

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4B1 4.2.1 Table 4.2.1 – Standardised monitoring Data – Oxides of Nitrogen**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(mg/m3)	(mg/m3)	(mg/m3)
1	05-Mar-15	15:10	16:10	934.1	918.4	30.4
2	05-Mar-15	17:12	18:12	912.5	887.4	29.4
3	05-Mar-15	19:14	20:14	995.1	936.3	31.0
4	05-Mar-15	21:16	22:16	950.5	902.8	29.9
5	06-Mar-15	1:02	2:02	854.1	815.2	27.0
6	06-Mar-15	3:04	4:04	1043.2	960.4	31.8
7	06-Mar-15	5:06	6:06	906.9	825.6	27.4
8	06-Mar-15	7:08	8:08	964.2	911.6	30.2
9	06-Mar-15	9:09	10:09	1015.0	948.5	34.0
10	06-Mar-15	11:11	12:11	1046.6	971.0	32.2
11	06-Mar-15	13:13	14:13	1050.2	1015.2	33.6

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa


6 % Oxygen, dry gas

**4B1 4.3.1 Plot 1.1 –Time Series of Standardised CEM versus Standardised SRM data – Oxides of Nitrogen. Reference conditions: 273K, 101.3kPa, dry gas, 6% oxygen.**





#### 4B1 4.4.1 –Elimination of Outliers – Oxides of Nitrogen.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.


Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9834, an outlier test has not been undertaken.

**4B1 4.5.1 Table 4.3.1 - Data used to Calculate Calibrated Values – Oxides of Nitrogen**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS dry Oxygen	CEMS Moisture	CEMS Standardised Value (dry)	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		(ppm)	(ppm)	%	%	(mg/Nm <sup>3</sup> )	(mg/Nm <sup>3</sup> )	(mg/m3)
1	18-Feb-15	Reference		0.3	-3.2	N/A	N/A	0.7	-6.6	0.0
2	05-Mar-15	15:10	16:10	438.1	444.2	6.21	2.31	934.1	947.1	918.4
3	05-Mar-15	17:12	18:12	428.5	434.3	6.19	2.31	912.5	924.9	887.4
4	05-Mar-15	19:14	20:14	451.5	457.9	6.70	2.23	995.1	1009.1	936.3
5	05-Mar-15	21:16	22:16	438.2	444.3	6.47	2.25	950.5	963.7	902.8
6	06-Mar-15	01:02	02:02	383.9	388.7	6.83	2.30	854.1	865.0	815.2
7	06-Mar-15	03:04	04:04	438.0	444.0	7.77	2.13	1043.2	1057.7	960.4
8	06-Mar-15	05:06	06:06	377.6	382.3	7.90	1.97	906.9	918.2	825.6
9	06-Mar-15	07:08	08:08	448.4	454.7	6.35	2.21	964.2	977.7	911.6
10	06-Mar-15	09:09	10:09	465.8	472.5	6.53	2.25	1015.0	1029.5	948.5
11	06-Mar-15	11:11	12:11	470.0	476.7	6.85	2.22	1046.6	1061.7	971.0
12	06-Mar-15	13:13	14:13	499.7	507.1	5.97	2.45	1050.2	1065.8	1015.2
Sum								10673.12		
Emission Limit Value (ELV) =				1210	mg/Nm <sup>3</sup>					
Reference Oxygen						6	%			
Established Calibration Function y <sub>i</sub> =				-3.578 + 1.022x <sub>i</sub>						

**Plot of CEM versus SRM Parallel Test Data at CEM measurement conditions – wet gas**



**4B1 4.6.1 Table 4.4.1 – Data used for the Variability Test – Oxides of Nitrogen**

Test No	Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1	18-Feb-15	Reference		-6.6	0.0	6.6	66.74	4454.75
2	05-Mar-15	15:10	16:10	947.1	918.4	-28.67	31.43	987.70
3	05-Mar-15	17:12	18:12	924.9	887.4	-37.55	22.55	508.61
4	05-Mar-15	19:14	20:14	1009.1	936.3	-72.79	-12.69	161.07
5	05-Mar-15	21:16	22:16	963.7	902.8	-60.91	-0.81	0.66
6	06-Mar-15	01:02	02:02	865.0	815.2	-49.82	10.28	105.69
7	06-Mar-15	03:04	04:04	1057.7	960.4	-97.23	-37.13	1378.73
8	06-Mar-15	05:06	06:06	918.2	825.6	-92.65	-32.55	1059.44
9	06-Mar-15	07:08	08:08	977.7	911.6	-66.07	-5.97	35.66
10	06-Mar-15	09:09	10:09	1029.5	948.5	-80.95	-20.85	434.76
11	06-Mar-15	11:11	12:11	1061.7	971.0	-90.62	-30.52	931.44
12	06-Mar-15	13:13	14:13	1065.8	1015.2	-50.57	9.52	90.71
Mean						-60.10		
Sum								10149.22

**4B1 4.7.1 - Variability Test and Test of Calibration Function Calculations – Oxides of Nitrogen**


**Variability Test**

SD=	$\text{Root}(1-\text{Number}).\text{Integral}(D1-D)^2$	30.38	mg/m3(s,d),6%O2
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. $O_0$ is therefore calculated as:-			
$O_0$ =	$0.2 \times 1210 \text{ mg/m3 (s,d,6\%O2)}/1.96$	123.47	mg/m3(s,d),6%O2
For 12 tests, kv =	0.9695		
Therefore variability=	$30.38 \leq 123.47 \times 0.9695 \times 1.5$		
	or $30.38 \leq 179.56$		
Which is TRUE therefore the CEMS passes the test			

**Test of Calibration Function**

The calibration of the AMS is accepted if:		$D1 \leq t_{0.95}(N-1) \times (s,d/\text{root } N) + O_0$
D1=	60.10	
For 12 tests, $t_{0.95}(N-1)$ :	1.796	
Therefore test of calibration =	$60.1 \leq 1.796 \times (30.38/\text{root } 12) + 123.47$	
or	$60.10 \leq 139.22$	
Which is TRUE therefore the calibration function is VALID		

**4B1 4.8.1 Plot 2.1 –Standardised CEM data versus standardised SRM - Oxides of Nitrogen (as NO<sub>2</sub>) – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen.**



**4B1 4.1.2 Table 4.1.2 – Raw monitoring Data – Sulphur Dioxide Low Range**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (wet)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMS Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	05-Mar-15	15:10	16:10	68.49	6.21	2.31	67.99	6.71	2.10	67.99
2	05-Mar-15	17:12	18:12	55.34	6.19	2.31	55.77	6.59	2.12	55.77
3	05-Mar-15	19:14	20:14	56.80	6.70	2.23	54.98	6.78	2.06	54.98
4	05-Mar-15	21:16	22:16	73.96	6.47	2.25	83.10	6.63	2.09	83.10
5	06-Mar-15	1:02	2:02	78.03	6.83	2.30	89.82	7.03	2.13	89.82
6	06-Mar-15	3:04	4:04	75.23	7.77	2.13	85.68	7.73	1.95	85.68
7	06-Mar-15	5:06	6:06	72.81	7.90	1.97	84.21	7.79	1.79	84.21
8	06-Mar-15	7:08	8:08	82.19	6.35	2.21	92.40	6.54	2.01	92.40
9	06-Mar-15	9:09	10:09	75.79	6.53	2.25	84.79	6.63	2.14	84.79
10	06-Mar-15	11:11	12:11	68.77	6.85	2.22	76.61	6.90	2.21	76.61
11	06-Mar-15	13:13	14:13	61.58	5.97	2.45	74.34	6.46	2.41	74.34

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**B1 4.2.2 Table 4.2.2 – Standardised monitoring Data – Sulphur Dioxide - Low Range**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(mg/m3)	(mg/m3)	(mg/m3)
1	05-Mar-15	15:10	16:10	203.2	208.3	31.9
2	05-Mar-15	17:12	18:12	164.0	169.5	31.7
3	05-Mar-15	19:14	20:14	174.2	169.2	32.2
4	05-Mar-15	21:16	22:16	223.2	253.2	32.2
5	06-Mar-15	1:02	2:02	241.6	281.6	32.4
6	06-Mar-15	3:04	4:04	249.3	282.6	32.4
7	06-Mar-15	5:06	6:06	243.3	278.4	32.4
8	06-Mar-15	7:08	8:08	245.9	279.6	32.4
9	06-Mar-15	9:09	10:09	229.8	258.5	32.2
10	06-Mar-15	11:11	12:11	213.1	238.3	32.1
11	06-Mar-15	13:13	14:13	180.1	224.5	31.4

Note:


Emission concentrations expressed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4B1 4.3.2 Plot 1.2 –Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide - Low Range. Reference conditions: 273K, 101.3kPa, dry gas, 6% oxygen.**



#### 4B1 4.4.2 – Elimination of Outliers – Sulphur Dioxide - Low Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.


Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9014, an outlier test has not been undertaken

**4B1 4.5.2 Table 4.3.2 - Data used to Calculate Calibrated Values – Sulphur Dioxide - Low Range.**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS dry Oxygen	CEMS Moisture	CEMS Standardised Value (dry)	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		(ppm)	(ppm)	%	%	(mg/Nm <sup>3</sup> )	(mg/Nm <sup>3</sup> )	(mg/m3)
1	18-Feb-15	Reference		-0.5	-6.6	N/A	N/A	-0.5	-19.0	0.0
2	05-Mar-15	15:10	16:10	68.5	68.0	6.21	2.31	203.2	201.8	208.3
3	05-Mar-15	17:12	18:12	55.3	53.8	6.19	2.31	164.0	159.3	169.5
4	05-Mar-15	19:14	20:14	56.8	55.4	6.70	2.23	174.2	169.7	169.2
5	05-Mar-15	21:16	22:16	74.0	73.9	6.47	2.25	223.2	223.1	253.2
6	06-Mar-15	01:02	02:02	78.0	78.3	6.83	2.30	241.6	242.5	281.6
7	06-Mar-15	03:04	04:04	75.2	75.3	7.77	2.13	249.3	249.5	282.6
8	06-Mar-15	05:06	06:06	72.8	72.7	7.90	1.97	243.3	242.8	278.4
9	06-Mar-15	07:08	08:08	82.2	82.8	6.35	2.21	245.9	247.8	279.6
10	06-Mar-15	09:09	10:09	75.8	75.9	6.53	2.25	229.8	230.1	258.5
11	06-Mar-15	11:11	12:11	68.8	68.3	6.85	2.22	213.1	211.7	238.3
12	06-Mar-15	13:13	14:13	61.6	60.5	5.97	2.45	180.1	177.0	224.5
Sum								2366.99		
Emission Limit Value (ELV) =				440	mg/Nm <sup>3</sup>					
Reference Oxygen				6						
Established Calibration Function y <sub>i</sub> =				-6.1 + 1.082x <sub>i</sub>						
				%						

**Plot of CEM versus SRM Data at CEM measurement conditions– Sulphur Dioxide - Low Range. wet gas**





#### 4B1 4.6.2 Table 4.4.2 – Data used for the Variability Test – Sulphur Dioxide - Low Range

Test No	Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1	18-Feb-15	Reference		-19.0	0.0	19.0	-6.63	43.92
2	05-Mar-15	15:10	16:10	201.8	208.3	6.57	-19.04	362.38
3	05-Mar-15	17:12	18:12	159.3	169.5	10.18	-15.42	237.74
4	05-Mar-15	19:14	20:14	169.7	169.2	-0.53	-26.13	682.68
5	05-Mar-15	21:16	22:16	223.1	253.2	30.06	4.46	19.90
6	06-Mar-15	01:02	02:02	242.5	281.6	39.14	13.54	183.25
7	06-Mar-15	03:04	04:04	249.5	282.6	33.02	7.42	55.02
8	06-Mar-15	05:06	06:06	242.8	278.4	35.53	9.93	98.64
9	06-Mar-15	07:08	08:08	247.8	279.6	31.76	6.16	37.90
10	06-Mar-15	09:09	10:09	230.1	258.5	28.35	2.75	7.54
11	06-Mar-15	11:11	12:11	211.7	238.3	26.63	1.02	1.05
12	06-Mar-15	13:13	14:13	177.0	224.5	47.54	21.94	481.19
Mean						25.60		
Sum								2211.21

#### 4B1 4.7.2 - Variability Test and Test of Calibration Function Calculations – Sulphur Dioxide - Low Range.


##### Variability Test

SD=  $\text{Root}(1-\text{Number}).\text{Integral}(\text{D1}-\text{D})^2$  14.18 mg/m3(s,d),6%O2  
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval.  $O_0$  is therefore calculated as:-  
 $O_0 = 0.2 \times 440 \text{ mg/m3 (s,d,6\%O2)}/1.96$  44.90 mg/m3(s,d),6%O2  
For 12 tests, kv = 0.9695  
Therefore variability= 14.18 <= 44.9 \* 0.9695\*1.5  
or 14.18 <= 65.29  
Which is TRUE therefore the CEMS passes the test

##### Test of Calibration Function

The calibration of the AMS is accepted if:  $\text{D1} \leq t_{0.95}(N-1) \times (\text{s,d}/\text{root } N) + O_0$   
 $\text{D1} = 25.60$   
For 12 tests,  $t_{0.95}(N-1) : 1.796$   
Therefore test of calibration =  $25.6 \leq 1.796 \times (14.18/\text{root } 12) + 44.9$   
or  $25.60 \leq 52.25$   
Which is TRUE therefore the calibration function is VALID

**4B1 4.8.2 Plot 2.2 – Standardised CEM data versus standardised SRM - Sulphur Dioxide - Low Range – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen.**



**4B1 4.1.3 Table 4.1.3 – Raw monitoring Data – Sulphur Dioxide High Range**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (wet)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMS Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	05-Mar-15	15:10	16:10	84.46	6.21	2.31	67.99	6.71	2.10	67.99
2	05-Mar-15	17:12	18:12	69.44	6.19	2.31	55.77	6.59	2.12	55.77
3	05-Mar-15	19:14	20:14	70.94	6.70	2.23	54.98	6.78	2.06	54.98
4	05-Mar-15	21:16	22:16	90.53	6.47	2.25	83.10	6.63	2.09	83.10
5	06-Mar-15	1:02	2:02	95.26	6.83	2.30	89.82	7.03	2.13	89.82
6	06-Mar-15	3:04	4:04	91.74	7.77	2.13	85.68	7.73	1.95	85.68
7	06-Mar-15	5:06	6:06	88.65	7.90	1.97	84.21	7.79	1.79	84.21
8	06-Mar-15	7:08	8:08	99.82	6.35	2.21	92.40	6.54	2.01	92.40
9	06-Mar-15	9:09	10:09	92.64	6.53	2.25	84.79	6.63	2.14	84.79
10	06-Mar-15	11:11	12:11	84.80	6.85	2.22	76.61	6.90	2.21	76.61
11	06-Mar-15	13:13	14:13	76.85	5.97	2.45	74.34	6.46	2.41	74.34

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4B1 4.2.3 Table 4.2.3 – Standardised monitoring Data – Sulphur Dioxide - High Range**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(mg/m3)	(mg/m3)	(mg/m3)
1	05-Mar-15	15:10	16:10	250.6	208.3	31.9
2	05-Mar-15	17:12	18:12	205.8	169.5	31.7
3	05-Mar-15	19:14	20:14	217.5	169.2	32.2
4	05-Mar-15	21:16	22:16	273.2	253.2	32.2
5	06-Mar-15	1:02	2:02	294.9	281.6	32.4
6	06-Mar-15	3:04	4:04	304.0	282.6	32.4
7	06-Mar-15	5:06	6:06	296.2	278.4	32.4
8	06-Mar-15	7:08	8:08	298.6	279.6	32.4
9	06-Mar-15	9:09	10:09	280.8	258.5	32.2
10	06-Mar-15	11:11	12:11	262.7	238.3	32.1
11	06-Mar-15	13:13	14:13	224.7	224.5	31.4

Note:


Emission concentrations expressed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4B1 4.3.3 Plot 1.3 –Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide - High Range. Reference conditions: 273K, 101.3kPa, dry gas, 6% oxygen.**



#### 4B1 4.4.3 – Elimination of Outliers – Sulphur Dioxide - High Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.


Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9014, an outlier test has not been undertaken

#### 4B1 4.5.3 Table 4.3.3 - Data used to Calculate Calibrated Values – Sulphur Dioxide - High Range.

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS dry Oxygen	CEMS Moisture	CEMS Standardised Value (dry)	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		(ppm)	(ppm)	%	%	(mg/Nm <sup>3</sup> )	(mg/Nm <sup>3</sup> )	(mg/m3)
1	18-Feb-15	Reference		-0.5	6.7	N/A	N/A	-0.5	19.1	0.0
2	05-Mar-15	15:10	16:10	84.5	87.6	6.21	2.31	250.6	259.8	208.3
3	05-Mar-15	17:12	18:12	69.4	73.3	6.19	2.31	205.8	217.1	169.5
4	05-Mar-15	19:14	20:14	70.9	74.7	6.70	2.23	217.5	229.0	169.2
5	05-Mar-15	21:16	22:16	90.5	93.3	6.47	2.25	273.2	281.7	253.2
6	06-Mar-15	01:02	02:02	95.3	97.8	6.83	2.30	294.9	302.9	281.6
7	06-Mar-15	03:04	04:04	91.7	94.5	7.77	2.13	304.0	313.1	282.6
8	06-Mar-15	05:06	06:06	88.7	91.6	7.90	1.97	296.2	305.9	278.4
9	06-Mar-15	07:08	08:08	99.8	102.2	6.35	2.21	298.6	305.7	279.6
10	06-Mar-15	09:09	10:09	92.6	95.3	6.53	2.25	280.8	289.1	258.5
11	06-Mar-15	11:11	12:11	84.8	87.9	6.85	2.22	262.7	272.3	238.3
12	06-Mar-15	13:13	14:13	76.9	80.3	5.97	2.45	224.7	234.9	224.5
Sum								2908.64		
Emission Limit Value (ELV) =				440	mg/Nm <sup>3</sup>					
Reference Oxygen										
Established Calibration Function y <sub>i</sub> =				7.158 + 0.952x <sub>i</sub>						
				6 %						

#### Plot of CEM versus SRM Data at CEM measurement conditions– Sulphur Dioxide - High Range. wet gas



**4B1 4.6.3 Table 4.4.3 – Data used for the Variability Test – Sulphur Dioxide - High Range.**

Test No	Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1	18-Feb-15	Reference		19.1	0.0	-19.1	13.16	173.12
2	05-Mar-15	15:10	16:10	259.8	208.3	-51.45	-19.20	368.61
3	05-Mar-15	17:12	18:12	217.1	169.5	-47.57	-15.32	234.79
4	05-Mar-15	19:14	20:14	229.0	169.2	-59.81	-27.56	759.69
5	05-Mar-15	21:16	22:16	281.7	253.2	-28.55	3.70	13.69
6	06-Mar-15	01:02	02:02	302.9	281.6	-21.28	10.97	120.27
7	06-Mar-15	03:04	04:04	313.1	282.6	-30.59	1.65	2.74
8	06-Mar-15	05:06	06:06	305.9	278.4	-27.55	4.70	22.05
9	06-Mar-15	07:08	08:08	305.7	279.6	-26.14	6.11	37.32
10	06-Mar-15	09:09	10:09	289.1	258.5	-30.60	1.65	2.73
11	06-Mar-15	11:11	12:11	272.3	238.3	-34.01	-1.76	3.11
12	06-Mar-15	13:13	14:13	234.9	224.5	-10.34	21.91	480.12
Mean						-32.25		
Sum								2218.25

**4B1 4.7.3 - Variability Test and Test of Calibration Function Calculations – Sulphur Dioxide - High Range.**


Variability Test

SD=  $\text{Root}(1-\text{Number}).\text{Integral}(D1-D)^2$  14.20 mg/m3(s,d),6%O2  
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval.  $O_0$  is therefore calculated as:-  
 $O_0 = 0.2 * 440 \text{ mg/m3 (s,d,6\%O2)}/1.96$  44.90 mg/m3(s,d),6%O2  
For 12 tests, kv = 0.9695  
Therefore variability=  $14.2 \leq 44.9 * 0.9695 * 1.5$   
or  $14.20 \leq 65.29$   
Which is TRUE therefore the CEMS passes the test

Test of Calibration Function

The calibration of the AMS is accepted if:  $D1 \leq t_{0.95}(N-1)*(s,d/\text{root } N)+O_0$   
 $D1 = 32.25$   
For 12 tests,  $t_{0.95}(N-1) : 1.796$   
Therefore test of calibration =  $32.25 \leq 1.796*(14.2/\text{root } 12)+44.9$   
or  $32.25 \leq 52.26$   
Which is TRUE therefore the calibration function is VALID

**4B1 4.8.3 Plot 2.3 – Standardised CEM data versus standardised SRM - Sulphur Dioxide - High Range – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen.**





## **Section 4B2: Data & calculations – AST – Unit 7, Procal 2**

## Section 4B2 – Data and calculations – AST Procal 2

**4B2 4.1.1 Table 4.1.1 – Raw monitoring Data – Oxides of Nitrogen**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (dry)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMS Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	05-Mar-15	15:10	16:10	426.32	6.21	1.85	425.97	6.71	2.10	417.04
2	05-Mar-15	17:12	18:12	414.67	6.19	1.86	414.98	6.59	2.12	406.18
3	05-Mar-15	19:14	20:14	421.59	6.70	1.88	432.12	6.78	2.06	423.24
4	05-Mar-15	21:16	22:16	416.34	6.47	1.85	421.10	6.63	2.09	412.29
5	06-Mar-15	1:02	2:02	371.90	6.83	2.02	369.56	7.03	2.13	361.69
6	06-Mar-15	3:04	4:04	406.82	7.77	1.95	413.27	7.73	1.95	405.21
7	06-Mar-15	5:06	6:06	347.75	7.90	1.86	353.81	7.79	1.79	347.47
8	06-Mar-15	7:08	8:08	428.66	6.35	1.90	427.81	6.54	2.01	419.21
9	06-Mar-15	9:09	10:09	439.96	6.53	1.95	442.43	6.63	2.14	432.95
10	06-Mar-15	11:11	12:11	443.13	6.85	2.00	444.16	6.90	2.21	434.34
11	06-Mar-15	13:13	14:13	485.68	5.97	2.12	479.21	6.46	2.41	467.68

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4B2 4.2.1 Table 4.2.1 – Standardised monitoring Data – Oxides of Nitrogen**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(mg/m3)	(mg/m3)	(mg/m3)
1	05-Mar-15	15:10	16:10	904.7	918.4	30.4
2	05-Mar-15	17:12	18:12	879.1	887.4	29.4
3	05-Mar-15	19:14	20:14	925.9	936.3	31.0
4	05-Mar-15	21:16	22:16	899.4	902.8	29.9
5	06-Mar-15	1:02	2:02	825.2	815.2	27.0
6	06-Mar-15	3:04	4:04	967.2	960.4	31.8
7	06-Mar-15	5:06	6:06	834.2	825.6	27.4
8	06-Mar-15	7:08	8:08	918.8	911.6	30.2
9	06-Mar-15	9:09	10:09	955.7	948.5	34.0
10	06-Mar-15	11:11	12:11	984.7	971.0	32.2
11	06-Mar-15	13:13	14:13	1017.2	1015.2	33.6

Note:


Emission concentrations expressed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4B2 4.3.1 Plot 1.1 –Time Series of Standardised CEM versus Standardised SRM data – Oxides of Nitrogen. Reference conditions: 273K, 101.3kPa, dry gas, 6% oxygen.**



#### 4B2 4.4.1 –Elimination of Outliers – Oxides of Nitrogen.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.


Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9821, an outlier test has not been undertaken.

**4B2 4.5.1 Table 4.3.1 - Data used to Calculate Calibrated Values – Oxides of Nitrogen**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS dry Oxygen	CEMS Moisture	CEMS Standardised Value (dry)	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		(ppm)	(ppm)	%	%	(mg/Nm <sup>3</sup> )	(mg/Nm <sup>3</sup> )	(mg/m3)
1	18-Feb-15	Reference		-1.3	-2.8	N/A	N/A	-2.7	-5.8	0.0
2	05-Mar-15	15:10	16:10	426.3	415.9	6.21	1.85	904.7	882.5	918.4
3	05-Mar-15	17:12	18:12	414.7	404.5	6.19	1.86	879.1	857.4	887.4
4	05-Mar-15	19:14	20:14	421.6	411.2	6.70	1.88	925.9	903.1	936.3
5	05-Mar-15	21:16	22:16	416.3	406.1	6.47	1.85	899.4	877.3	902.8
6	06-Mar-15	01:02	02:02	371.9	362.6	6.83	2.02	825.2	804.5	815.2
7	06-Mar-15	03:04	04:04	406.8	396.8	7.77	1.95	967.2	943.3	960.4
8	06-Mar-15	05:06	06:06	347.8	338.9	7.90	1.86	834.2	813.1	825.6
9	06-Mar-15	07:08	08:08	428.7	418.2	6.35	1.90	918.8	896.3	911.6
10	06-Mar-15	09:09	10:09	440.0	429.2	6.53	1.95	955.7	932.4	948.5
11	06-Mar-15	11:11	12:11	443.1	432.3	6.85	2.00	984.7	960.7	971.0
12	06-Mar-15	13:13	14:13	485.7	474.0	5.97	2.12	1017.2	992.7	1015.2
Sum								10109.43		
Emission Limit Value (ELV) =				1210	mg/Nm <sup>3</sup>					
Reference Oxygen					6		%			
Established Calibration Function y=				-1.5 + 0.979xi						

**Plot of CEM versus SRM Parallel Test Data at CEM measurement conditions – wet gas**



**4B2 4.6.1 Table 4.4.1 – Data used for the Variability Test – Oxides of Nitrogen**

Test No	Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1	18-Feb-15	Reference		-5.8	0.0	5.8	-13.80	190.51
2	05-Mar-15	15:10	16:10	882.5	918.4	35.86	16.30	265.77
3	05-Mar-15	17:12	18:12	857.4	887.4	29.96	10.39	108.05
4	05-Mar-15	19:14	20:14	903.1	936.3	33.17	13.61	185.27
5	05-Mar-15	21:16	22:16	877.3	902.8	25.47	5.91	34.94
6	06-Mar-15	01:02	02:02	804.5	815.2	10.65	-8.91	79.45
7	06-Mar-15	03:04	04:04	943.3	960.4	17.11	-2.45	6.03
8	06-Mar-15	05:06	06:06	813.1	825.6	12.47	-7.09	50.27
9	06-Mar-15	07:08	08:08	896.3	911.6	15.34	-4.22	17.84
10	06-Mar-15	09:09	10:09	932.4	948.5	16.15	-3.41	11.64
11	06-Mar-15	11:11	12:11	960.7	971.0	10.32	-9.24	85.39
12	06-Mar-15	13:13	14:13	992.7	1015.2	22.48	2.92	8.51
Mean						19.56		
Sum								1043.66

#### 4B2 4.7.1 - Variability Test and Test of Calibration Function Calculations – Oxides of Nitrogen


##### Variability Test

SD=  $\text{Root}(1-\text{Number}).\text{Integral}(D1-D)^2$  9.74 mg/m3(s,d),6%O2  
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval.  $O_0$  is therefore calculated as:-  
 $O_0 = 0.2 * 1210 \text{ mg/m3 (s,d,6\%O2)}/1.96$  123.47 mg/m3(s,d),6%O2  
For 12 tests, kv = 0.9695  
Therefore variability=  $9.74 \leq 123.47 * 0.9695 * 1.5$   
or  $9.74 \leq 179.56$   
Which is TRUE therefore the CEMS passes the test

##### Test of Calibration Function

The calibration of the AMS is accepted if:  $D1 \leq t_{0.95}(N-1) * (s,d/\text{root } N) + O_0$   
 $D1 = 19.56$   
For 12 tests,  $t_{0.95}(N-1) : 1.796$   
Therefore test of calibration =  $19.56 \leq 1.796 * (9.74/\text{root } 12) + 123.47$   
or  $19.56 \leq 128.52$   
Which is TRUE therefore the calibration function is VALID

**4B2 4.8.1 Plot 2.1 –Standardised CEM data versus standardised SRM - Oxides of Nitrogen (as NO<sub>2</sub>) – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen.**



**4B2 4.1.2 Table 4.1.2 – Raw monitoring Data – Sulphur Dioxide Low Range**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (wet)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMS Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	05-Mar-15	15:10	16:10	84.48	6.21	1.85	67.99	6.71	2.10	67.99
2	05-Mar-15	17:12	18:12	71.10	6.19	1.86	55.77	6.59	2.12	55.77
3	05-Mar-15	19:14	20:14	72.78	6.70	1.88	54.98	6.78	2.06	54.98
4	05-Mar-15	21:16	22:16	92.86	6.47	1.85	83.10	6.63	2.09	83.10
5	06-Mar-15	1:02	2:02	99.93	6.83	2.02	89.82	7.03	2.13	89.82
6	06-Mar-15	3:04	4:04	93.89	7.77	1.95	85.68	7.73	1.95	85.68
7	06-Mar-15	5:06	6:06	91.37	7.90	1.86	84.21	7.79	1.79	84.21
8	06-Mar-15	7:08	8:08	103.69	6.35	1.90	92.40	6.54	2.01	92.40
9	06-Mar-15	9:09	10:09	95.48	6.53	1.95	84.79	6.63	2.14	84.79
10	06-Mar-15	11:11	12:11	87.89	6.85	2.00	76.61	6.90	2.21	76.61
11	06-Mar-15	13:13	14:13	84.54	5.97	2.12	74.34	6.46	2.41	74.34

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4B2 4.2.2 Table 4.2.2 – Standardised monitoring Data – Sulphur Dioxide - Low Range**

Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(mg/m3)	(mg/m3)	(mg/m3)
1	05-Mar-15	15:10	16:10	249.4	208.3	31.9
2	05-Mar-15	17:12	18:12	209.7	169.5	31.7
3	05-Mar-15	19:14	20:14	222.4	169.2	32.2
4	05-Mar-15	21:16	22:16	279.1	253.2	32.2
5	06-Mar-15	1:02	2:02	308.5	281.6	32.4
6	06-Mar-15	3:04	4:04	310.6	282.6	32.4
7	06-Mar-15	5:06	6:06	305.0	278.4	32.4
8	06-Mar-15	7:08	8:08	309.2	279.6	32.4
9	06-Mar-15	9:09	10:09	288.6	258.5	32.2
10	06-Mar-15	11:11	12:11	271.7	238.3	32.1
11	06-Mar-15	13:13	14:13	246.3	224.5	31.4


Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa


6 % Oxygen, dry gas



**4B2 4.3.2 Plot 1.2 –Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide - Low Range. Reference conditions: 273K, 101.3kPa, dry gas, 6% oxygen.**



#### 4B2 4.4.2 – Elimination of Outliers – Sulphur Dioxide - Low Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the R2 value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.


Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the R2 value for this determinand was 0.9640, an outlier test has not been undertaken.

**4B2 4.5.2 Table 4.3.2 - Data used to Calculate Calibrated Values – Sulphur Dioxide - Low Range.**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS dry Oxygen	CEMS Moisture	CEMS Standardised Value (dry)	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		(ppm)	(ppm)	%	%	(mg/Nm <sup>3</sup> )	(mg/Nm <sup>3</sup> )	(mg/m3)
1	18-Feb-15	Reference		-0.5	-6.5	N/A	N/A	-0.5	-18.6	0.0
2	05-Mar-15	15:10	16:10	84.5	68.7	6.21	1.85	249.4	202.8	208.3
3	05-Mar-15	17:12	18:12	71.1	56.8	6.19	1.86	209.7	167.7	169.5
4	05-Mar-15	19:14	20:14	72.8	58.3	6.70	1.88	222.4	178.3	169.2
5	05-Mar-15	21:16	22:16	92.9	76.1	6.47	1.85	279.1	228.7	253.2
6	06-Mar-15	01:02	02:02	99.9	82.4	6.83	2.02	308.5	254.3	281.6
7	06-Mar-15	03:04	04:04	93.9	77.0	7.77	1.95	310.6	254.8	282.6
8	06-Mar-15	05:06	06:06	91.4	74.8	7.90	1.86	305.0	249.6	278.4
9	06-Mar-15	07:08	08:08	103.7	85.7	6.35	1.90	309.2	255.5	279.6
10	06-Mar-15	09:09	10:09	95.5	78.4	6.53	1.95	288.6	237.0	258.5
11	06-Mar-15	11:11	12:11	87.9	71.7	6.85	2.00	271.7	221.7	238.3
12	06-Mar-15	13:13	14:13	84.5	68.7	5.97	2.12	246.3	200.3	224.5
Sum								2999.99		
Emission Limit Value (ELV) =				440	mg/Nm <sup>3</sup>					
Reference Oxygen				6%						
Established Calibration Function y <sub>i</sub> =				-6.072 + 0.885x <sub>i</sub>						

**Plot of CEM versus SRM Data at CEM measurement conditions– Sulphur Dioxide - Low Range. wet gas**



**4B2 4.6.2 Table 4.4.2 – Data used for the Variability Test – Sulphur Dioxide - Low Range**

Test No	Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1	18-Feb-15	Reference		-18.6	0.0	18.6	0.99	0.98
2	05-Mar-15	15:10	16:10	202.8	208.3	5.50	-12.12	146.92
3	05-Mar-15	17:12	18:12	167.7	169.5	1.84	-15.78	248.98
4	05-Mar-15	19:14	20:14	178.3	169.2	-9.03	-26.65	710.24
5	05-Mar-15	21:16	22:16	228.7	253.2	24.41	6.79	46.07
6	06-Mar-15	01:02	02:02	254.3	281.6	27.37	9.75	94.98
7	06-Mar-15	03:04	04:04	254.8	282.6	27.78	10.16	103.16
8	06-Mar-15	05:06	06:06	249.6	278.4	28.77	11.14	124.18
9	06-Mar-15	07:08	08:08	255.5	279.6	24.01	6.39	40.77
10	06-Mar-15	09:09	10:09	237.0	258.5	21.42	3.80	14.43
11	06-Mar-15	11:11	12:11	221.7	238.3	16.59	-1.03	1.06
12	06-Mar-15	13:13	14:13	200.3	224.5	24.19	6.57	43.20
Mean						17.62		
Sum								1574.98

#### 4B2 4.7.2 - Variability Test and Test of Calibration Function Calculations – Sulphur Dioxide - Low Range.


##### Variability Test

SD=  $\text{Root}(1-\text{Number}).\text{Integral}(\text{D1}-\text{D})^2$  11.97 mg/m3(s,d),6%O2  
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval.  $O_0$  is therefore calculated as:-  
 $O_0 = 0.2 * 440 \text{ mg/m3 (s,d,6\%O2)}/1.96$  44.90 mg/m3(s,d),6%O2  
For 12 tests, kv = 0.9695  
Therefore variability=  $11.97 \leq 44.9 * 0.9695 * 1.5$   
or  $11.97 \leq 65.29$   
Which is TRUE therefore the CEMS passes the test

##### Test of Calibration Function

The calibration of the AMS is accepted if:  $\text{D1} \leq t_{0.95}(N-1) * (\text{s,d}/\text{root } N) + O_0$   
 $\text{D1} = 17.62$   
For 12 tests,  $t_{0.95}(N-1) : 1.796$   
Therefore test of calibration =  $17.62 \leq 1.796 * (11.97/\text{root } 12) + 44.9$   
or  $17.62 \leq 51.10$   
Which is TRUE therefore the calibration function is VALID

**4B2 4.8.2 Plot 2.2 – Standardised CEM data versus standardised SRM - Sulphur Dioxide - Low Range – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen.**



**4B2 4.1.3 Table 4.1.3 – Raw monitoring Data – Sulphur Dioxide High Range**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (wet)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMS Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	05-Mar-15	15:10	16:10	92.76	6.21	1.85	67.99	6.71	2.10	67.99
2	05-Mar-15	17:12	18:12	78.58	6.19	1.86	55.77	6.59	2.12	55.77
3	05-Mar-15	19:14	20:14	80.30	6.70	1.88	54.98	6.78	2.06	54.98
4	05-Mar-15	21:16	22:16	101.55	6.47	1.85	83.10	6.63	2.09	83.10
5	06-Mar-15	1:02	2:02	108.99	6.83	2.02	89.82	7.03	2.13	89.82
6	06-Mar-15	3:04	4:04	102.50	7.77	1.95	85.68	7.73	1.95	85.68
7	06-Mar-15	5:06	6:06	99.84	7.90	1.86	84.21	7.79	1.79	84.21
8	06-Mar-15	7:08	8:08	112.97	6.35	1.90	92.40	6.54	2.01	92.40
9	06-Mar-15	9:09	10:09	104.32	6.53	1.95	84.79	6.63	2.14	84.79
10	06-Mar-15	11:11	12:11	96.21	6.85	2.00	76.61	6.90	2.21	76.61
11	06-Mar-15	13:13	14:13	92.65	5.97	2.12	74.34	6.46	2.41	74.34

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4B2 4.2.3 Table 4.2.3 – Standardised monitoring Data – Sulphur Dioxide - High Range**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(mg/m3)	(mg/m3)	(mg/m3)
1	05-Mar-15	15:10	16:10	273.9	208.3	31.9
2	05-Mar-15	17:12	18:12	231.8	169.5	31.7
3	05-Mar-15	19:14	20:14	245.3	169.2	32.2
4	05-Mar-15	21:16	22:16	305.2	253.2	32.2
5	06-Mar-15	1:02	2:02	336.4	281.6	32.4
6	06-Mar-15	3:04	4:04	339.0	282.6	32.4
7	06-Mar-15	5:06	6:06	333.2	278.4	32.4
8	06-Mar-15	7:08	8:08	336.9	279.6	32.4
9	06-Mar-15	9:09	10:09	315.3	258.5	32.2
10	06-Mar-15	11:11	12:11	297.5	238.3	32.1
11	06-Mar-15	13:13	14:13	270.0	224.5	31.4

Note:


Emission concentrations expressed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4B2 4.3.3 Plot 1.3 –Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide - High Range. Reference conditions: 273K, 101.3kPa, dry gas, 6% oxygen.**



#### 4B2 4.4.3 – Elimination of Outliers – Sulphur Dioxide - High Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.

Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'


As the  $R^2$  value for this determinand was 0.9623, an outlier test has not been undertaken.



**4B2 4.5.3 Table 4.3.3 - Data used to Calculate Calibrated Values – Sulphur Dioxide - High Range.**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (x)	CEMS Calibrated signal	CEMS dry Oxygen	CEMS Moisture	CEMS Standardised Value (dry)	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		(ppm)	(ppm)	%	%	(mg/Nm <sup>3</sup> )	(mg/Nm <sup>3</sup> )	(mg/m3)
1	18-Feb-15	Reference		-0.5	-6.3	N/A	N/A	-0.5	-17.9	0.0
2	05-Mar-15	15:10	16:10	92.8	64.3	6.21	1.85	273.9	190.0	208.3
3	05-Mar-15	17:12	18:12	78.6	53.6	6.19	1.86	231.8	158.1	169.5
4	05-Mar-15	19:14	20:14	80.3	54.9	6.70	1.88	245.3	167.8	169.2
5	05-Mar-15	21:16	22:16	101.6	71.0	6.47	1.85	305.2	213.4	253.2
6	06-Mar-15	01:02	02:02	109.0	76.6	6.83	2.02	336.4	236.5	281.6
7	06-Mar-15	03:04	04:04	102.5	71.7	7.77	1.95	339.0	237.2	282.6
8	06-Mar-15	05:06	06:06	99.8	69.7	7.90	1.86	333.2	232.6	278.4
9	06-Mar-15	07:08	08:08	113.0	79.6	6.35	1.90	336.9	237.5	279.6
10	06-Mar-15	09:09	10:09	104.3	73.1	6.53	1.95	315.3	220.9	258.5
11	06-Mar-15	11:11	12:11	96.2	66.9	6.85	2.00	297.5	207.0	238.3
12	06-Mar-15	13:13	14:13	92.6	64.3	5.97	2.12	270.0	187.2	224.5
Sum								3284.08		
Emission Limit Value (ELV) =				440 mg/Nm <sup>3</sup>						
				Reference Oxygen						
				6 %						
Established Calibration Function y <sub>i</sub> =				-5.884 + 0.757x <sub>i</sub>						

**Plot of CEM versus SRM Data at CEM measurement conditions– Sulphur Dioxide - High Range. wet gas**



**4B2 4.6.3 Table 4.4.3 – Data used for the Variability Test – Sulphur Dioxide - High Range.**

Test No	Date	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
		hr:min		mg/m3	mg/m3			
1	18-Feb-15	Reference		-17.9	0.0	17.9	-13.22	174.89
2	05-Mar-15	15:10	16:10	190.0	208.3	18.36	-12.76	162.87
3	05-Mar-15	17:12	18:12	158.1	169.5	11.42	-19.70	388.00
4	05-Mar-15	19:14	20:14	167.8	169.2	1.47	-29.65	878.92
5	05-Mar-15	21:16	22:16	213.4	253.2	39.78	8.67	75.08
6	06-Mar-15	01:02	02:02	236.5	281.6	45.12	14.01	196.18
7	06-Mar-15	03:04	04:04	237.2	282.6	45.36	14.25	202.96
8	06-Mar-15	05:06	06:06	232.6	278.4	45.75	14.63	214.17
9	06-Mar-15	07:08	08:08	237.5	279.6	42.08	10.96	120.07
10	06-Mar-15	09:09	10:09	220.9	258.5	37.58	6.46	41.74
11	06-Mar-15	11:11	12:11	207.0	238.3	31.31	0.19	0.04
12	06-Mar-15	13:13	14:13	187.2	224.5	37.29	6.17	38.09
Mean						31.12		
Sum								2493.02

**4B2 4.7.3 - Variability Test and Test of Calibration Function Calculations – Sulphur Dioxide - High Range.**


Variability Test

SD=  $\text{Root}(1-\text{Number}).\text{Integral}(\text{D1}-\text{D})^2$  15.05 mg/m3(s,d),6%O2  
 The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval.  $O_0$  is therefore calculated as:-  
 $O_0 = 0.2 \times 440 \text{ mg/m3 (s,d,6\%O2)} / 1.96$  44.90 mg/m3(s,d),6%O2  
 For 12 tests,  $k_v = 0.9695$   
 Therefore variability=  $15.05 \leq 44.9 \times 0.9695 \times 1.5$   
 or  $15.05 \leq 65.29$   
 Which is TRUE therefore the CEMS passes the test

Test of Calibration Function

The calibration of the AMS is accepted if:  $\text{D1} \leq t_{0.95}(N-1) \times (\text{s,d}/\text{root } N) + O_0$   
 $\text{D1} = 31.12$   
 For 12 tests,  $t_{0.95}(N-1) : 1.796$   
 Therefore test of calibration =  $31.12 \leq 1.796 \times (15.05/\text{root } 12) + 44.9$   
 or  $31.12 \leq 52.70$   
 Which is TRUE therefore the calibration function is VALID

**4B2 4.8.3 Plot 2.3 – Standardised CEM data versus standardised SRM - Sulphur Dioxide - High Range – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen.**



## **Section 4B3: Data & Calculations – AST – Unit 7, SICK OMD 41**

**4B3 4.1 Table 4.1 – Raw monitoring Data – Total Particulate Matter**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (Extinction)	CEMS Oxygen (Wet)	CEMS Moisture (Wet)	CEM Stack Temp	CEM Stack Press	SRM Raw value (dry)	SRM Moisture (Wet FT)	SRM Oxygen (Dry)	SRM Stack Temp	SRM Stack Press	SRM at CEMs Raw conditions
		hr:min		%	(%)	(%)	C	kpa	(mg/m3)	(%)	(%)	C	kpa	(mg/m3)
1	06-Mar-15	9:05	9:55	3.3	6.4	2.25	63.5	100.2	6.8	2.1	6.6	60.9	102.2	5.5
2	06-Mar-15	10:10	11:00	4.2	5.8	2.26	64.0	100.2	3.6	2.1	6.6	61.9	102.1	2.9
3	06-Mar-15	11:10	12:00	4.7	6.8	2.31	66.2	100.2	11.2	2.2	6.9	63.0	102.1	9.0
4	06-Mar-15	12:10	13:00	4.3	6.3	1.97	66.1	100.2	8.3	2.2	6.7	63.8	102.1	6.7
5	09-Mar-15	13:00	13:55	3.2	4.9	1.99	56.4	100.1	8.1	2.0	5.8	54.3	101.2	6.6
6	09-Mar-15	14:10	15:00	2.6	5.6	2.03	57.3	100.1	0.6	2.0	6.0	55.0	101.2	0.5
7	10-Mar-15	10:07	10:58	1.9	6.4	1.93	56.5	100.2	2.7	1.9	6.8	54.3	102.3	2.2

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4B3 4.2 Table 4.2 – Standardised monitoring Data – Total Particulate Matter**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised Value	SRM Standardised Value	SRM Uncertainty
		hr:min		(mg/m3)	mg/m3	(mg/m3)
1	06-Mar-15	9:05	9:55	4.3	7.1	0.31
2	06-Mar-15	10:10	11:00	5.3	3.8	0.31
3	06-Mar-15	11:10	12:00	6.4	11.9	1.31
4	06-Mar-15	12:10	13:00	5.6	8.8	0.32
5	09-Mar-15	13:00	13:55	3.8	8.0	0.34
6	09-Mar-15	14:10	15:00	3.2	0.6	0.47
7	10-Mar-15	10:07	10:58	2.4	2.8	0.47

Note:


Emission concentrations expressed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4B3 4.3 Plot 1 –Time Series of Standardised CEM versus Standardised SRM data – Total Particulate Matter. Reference conditions: 273K, 101.3kPa, dry gas, 6% oxygen.**



#### 4B3 4.4 – Elimination of Outliers – Total Particulate Matter.




Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (Extinction)	SRM Value at CEMS Raw conditions	Difference Di	Difference Di - Di	Is Result an Outlier - Di - Di > 2SD
		hr:min		%	(mg/m3)			
1	06-Mar-15	09:05	09:55	3.3	5.5	2.26	0.94	No
2	06-Mar-15	10:10	11:00	4.2	2.9	-1.32	-2.64	No
3	06-Mar-15	11:10	12:00	4.7	9.0	4.28	2.96	No
4	06-Mar-15	12:10	13:00	4.3	6.7	2.40	1.07	No
5	09-Mar-15	13:00	13:55	3.2	6.6	3.35	2.02	No
6	09-Mar-15	14:10	15:00	2.6	0.5	-2.08	-3.40	No
7	10-Mar-15	10:07	10:58	1.9	2.2	0.37	-0.96	No
					Average Di	1.32		
					Standard Deviation	2.39		
					Standard Deviation x2	4.78		

**4B3 4.5 Table 4.3 - Data used to Calculate Calibrated Values – Total Particulate Matter.**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (Extinction)	CEMS Calibrated signal	CEMS Moisture	CEMS Temp	CEMS Pressure	CEMS Dry Oxygen	CEMS Standardised Value (dry)	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		%	(mg/m3)	(%)	(°C)	(kPa)	(%)	(mg/Nm³)	(mg/Nm³)	(mg/m3)
1	06-Mar-15	Reference		0.0	0.6	N/A	N/A	N/A	0.0	0.0	0.4	0.0
2	06-Mar-15	09:05	09:55	3.3	5.3	2.3	63.5	100.2	6.4	4.3	6.9	7.1
3	06-Mar-15	10:10	11:00	4.2	6.7	2.3	64.0	100.2	5.8	5.3	8.4	3.8
4	06-Mar-15	11:10	12:00	4.7	7.3	2.3	66.2	100.2	6.8	6.4	9.9	11.9
5	06-Mar-15	12:10	13:00	4.3	6.7	2.0	66.1	100.2	6.3	5.6	8.8	8.8
6	09-Mar-15	13:00	13:55	3.2	5.2	2.0	56.4	100.1	4.9	3.8	6.1	8.0
7	09-Mar-15	14:10	15:00	2.6	4.3	2.0	57.3	100.1	5.6	3.2	5.3	0.6
8	10-Mar-15	10:07	10:58	1.9	3.3	1.9	56.5	100.2	6.4	3.2	4.2	0.6
Sum										31.65		
Emission Limit Value (ELV) =					55 mg/Nm³							
Reference Oxygen										6 %		
Established Calibration Function y <sub>i</sub> =					0.6 + 1.433x <sub>i</sub>							



**Plot of CEM versus SRM Data at CEM measurement conditions– Total Particulate Matter, at stack temperature, pressure and moisture conditions.**



**4B3 4.6 Table 4.4 – Data used for the Variability Test – Total Particulate Matter.**

Test No	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
	hr:min		mg/m3	mg/m3			
1	Reference		0.4	0.0	-0.43	0.72	0.51
2	09:05	09:55	6.9	7.1	0.25	1.39	1.94
3	10:10	11:00	8.4	3.8	-4.63	-3.48	12.12
4	11:10	12:00	9.9	11.9	1.99	3.14	9.84
5	12:10	13:00	8.8	8.8	-0.06	1.08	1.17
6	13:00	13:55	6.1	8.0	1.89	3.04	9.23
7	14:10	15:00	5.3	0.6	-4.63	-3.49	12.19
8	10:07	10:58	4.2	0.6	-3.54	-2.39	5.73
8 Tests		Mean			-1.14		
Sum							52.72

**4B3 4.7 - Variability Test and Test of Calibration Function Calculations – Total Particulate Matter.**


Variability Test

SD=	$\text{Root}(1-\text{Number}).\text{Integral}(D1-D)^2$	2.74	mg/m3(s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 30% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	$0.3*55 \text{ mg/m3 (s,d,6\%O}_2\text{)}/1.96$	8.42	mg/m3(s,d),6%O <sub>2</sub>
For 8 tests, k <sub>v</sub> =	0.9521		
Therefore variability=	$2.74 \leq 8.42 * 0.9521 * 1.5$		
or	$2.74 \leq 12.02$		
Which is TRUE therefore the CEMS passes the test			

Test of Calibration Function

The calibration of the AMS is accepted if:		$D1 \leq t_{0.95}(N-1)*(s,d/\text{root } N)+O_0$
	D1=	1.14
For 8 tests, t <sub>0.95</sub> (N-1) =	1.895	
Therefore test of calibration =	$1.14 \leq 1.895*(2.74/\text{root } 8)+8.42$	
or	$1.14 \leq 10.26$	
Which is TRUE therefore the calibration function is VALID		

**4B3 4.8 Plot 2 – Standardised CEM data versus standardised SRM - Total Particulate Matter – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen.**



## Section 5 – Results of Functional tests

Table 5.1 - Audit of functional tests

<b>Operator</b>	RWE Generation UK plc.	
<b>Site</b>	Aberthaw Power Plant	
<b>Stack</b>	Unit 7	
<b>Process Sector</b>	LCPD	
<b>Analyser A - Make Model MCERTs Certificate Number</b>	Procal 1 - Pulsi 2000, In Situ IR MC990006/07	
<b>Analyser B - Make Model MCERTs Certificate Number</b>	Procal 2 - Pulsi 2000, In Situ IR MC990006/07	
<b>Analyser C - Make Model MCERTs Certificate Number</b>	Erwin SICK OMD41, Cross Duct Forward Scatter MC040042/01	
<b>Analyser D - Make Model MCERTs Certificate Number</b>	SICK Maihak GmbH GM32 – In-situ MC100163/01	
<b>Parameters Tested</b>	<b>Daily ELV</b>	<b>Certified range</b>
<b>NO &amp; NO<sub>2</sub></b>	1210 mg/m <sup>3</sup> (NO <sub>x</sub> as NO <sub>2</sub> )	PROCAL: NO = 0 – 1000 ppm  GM 32: Measuring path length 1.25m: NO 0 to 70 mg/m <sup>3</sup> & 0 to 700 mg/m <sup>3</sup>  GM 32: Measuring path length 1.00m NO 0 to 87.5 mg/m <sup>3</sup> & 0 to 875 mg/m <sup>3</sup>
<b>Total Particulate Matter</b>	55 mg/m <sup>3</sup>	Erwin SICK: TPM = 0 – 200 mg/m <sup>3</sup>
<b>SO<sub>2</sub></b>	440 mg/m <sup>3</sup>	PROCAL = SO <sub>2</sub> low = 0 – 250 ppm SO <sub>2</sub> high = 0 – 1000 ppm  GM 32: Measuring path length 1.25m: SO <sub>2</sub> 0 to 75mg/m <sup>3</sup> & 0 to 1000mg/m <sup>3</sup>  GM 32: Measuring path length 1.00m SO <sub>2</sub> 0 to 93.8 mg/m <sup>3</sup> and 0 to 1250 mg/m <sup>3</sup>

Analysers A & B	
Organisation carrying out tests -	Parker/Procal
Status of organisation – CEMS manufacturer/operator/service contractor	CEMS OEM
Test engineer	M Findley
Date of tests	17 <sup>th</sup> – 19 <sup>th</sup> February 2015
Analyser C & D	
Organisation carrying out tests -	SICK
Status of organisation – CEMS manufacturer/operator/service contractor	CEMS OEM
Test engineer	Rhodri Jones
Date of tests	25 <sup>th</sup> February 2015

## Functional Test compliance with EN 14181

Requirement	Compliance Y/N	Notes
<p>1 – Alignment and cleanliness All checks specified in MID 14181 carried out ?</p> <p>– Sampling System</p> <p>A visual inspection of the sampling system shall be performed, noting the condition of the following components, when fitted:</p> <ul style="list-style-type: none"> <li>- sampling probe;</li> <li>- gas conditioning systems;</li> <li>- pumps;</li> <li>- all connections;</li> <li>- sample lines;</li> <li>- power lines;</li> <li>- filters.</li> </ul> <p>The sampling system shall be in good condition and free of any visible faults, which may decrease the quality of the testing.</p>	<p>Yes</p>	<p><b>Procal</b> – Yes <b>SICK (OMD)</b> – Yes – Optics cleaned <b>SICK (GM32)</b> – Yes</p>
<p>2 - Leak Test</p> <p>Leak testing shall be performed according to the AMS manuals. The test shall cover the entire sampling system.</p>	<p>Yes</p>	
<p>Results of leak check compliant with requirements of relevant standards</p>	<p>N/A</p>	<p><b>Procal</b> – N/A although differential pressure noted as OK <b>SICK (OMD &amp; GM32)</b> – N/A</p>
<p>3 - Zero and Span Check Analyser</p> <p>Reference zero and span materials shall be used to verify the corresponding readings of the AMS.</p>	<p>N/A</p>	
<p>Results compliant with requirements of relevant standards</p> <p>Parameter:</p>	<p>Yes</p>	<p><b>Procal</b> – Yes <b>SICK (OMD 41)</b> – Yes – although based on the use of filters <b>SICK (GM32)</b> – Yes, although based on the use of filters</p>
<p>NO</p> <p>SO<sub>2</sub></p> <p>TPM</p>	<p>Yes (Procal)</p> <p>Yes (Procal)</p> <p>Yes</p>	<p><b>SICK (GM32)</b> – Span value of filters not stated, though analyser response recorded <b>SICK (OMD41)</b> - Span values recorded as mA signals</p>

Requirement	Compliance Y/N	Notes
4 - Zero and Span Check Full System  Reference zero and span materials shall be used to verify the corresponding readings of the AMS.	N/A	Procal, & SICK GM 32 analysers are in situ type and therefore there is as such, no 'full system'.
5 – NOx converter efficiency check	N/A	
6 - Linearity  The linearity of the analysers shall be checked using five different reference materials, including zero concentration.	Yes – Procal, SICK OMD 41 No – SICK GM32	<b>Procal</b> – Yes <b>SICK (OMD)</b> – Yes – although based on the use of filters
The reference material with zero concentration, as well as the reference materials with four different concentrations, shall have a verifiable quantity and quality.	Yes	<b>Procal</b> – Yes <b>SICK</b> – Yes – although based on the use of filters <b>SICK (GM32)</b> - N/A
The reference material concentrations shall be selected such that the measured values are approximately 20%, 40%, 60% and 80% of two times the emission limit.	See note	<b>Procal</b> – Yes <b>SICK</b> – Used 8 points using filters <b>SICK (GM32)</b> - N/A
The dry test reference material shall be applied to the inlet of the AMS.  Reference materials can be introduced directly into the analyser as long as the integrity of the sample system has been proved	N/A	N/A

Requirement	Compliance Y/N	Notes
<p>6 – Linearity (continued)</p> <p>After each change in concentration, the first instrument readings shall be taken after a time period equal to at least three times the response time of the AMS. At each reference material concentration, at least three readings shall be made. The time period between the start of each of the three readings shall be separated by at least four times the response time.</p> <p>A risk based approach may be adopted in order to reduce the time for the linearity tests</p>	<p>No</p> <p>N/A</p>	<p><b>Procal</b> – No times stated in report</p> <p><b>SICK (OMD &amp; GM32)</b> – No times stated in report</p>
Linearity Test Pass		<p><b>Procal</b> – Yes</p> <p><b>SICK (OMD)</b> – Yes</p>
Parameter		
NO, SO <sub>2</sub> & TPM	Yes	
6 – Interferences (only required in the event of a failure of the QAL 2/AST)	N/A	
7 – Zero and Span Drift (Audit)	Yes	
<p>8- Response time</p> <p>The response of the AMS shall be checked. This can be performed, if appropriate, by feeding the reference material at the end of the sampling probe. The response time shall not exceed the measured value as identified during QAL 1.</p>	See Note	<p><b>Procal</b> – Yes</p> <p><b>SICK (OMD 41)</b> – Yes</p> <p><b>SICK (GM 32)</b> – No</p>



Requirement	Compliance Y/N	Notes
<b>9a – Service Report - PROCAL</b>		
• Document reference	Yes	
• Instrument manufacturer	Yes	Procal
• Instrument Type	Yes	In-situ
• Instrument model	Yes	PULSI 200
• Instrument Serial No's	Yes	
• Operating principal	Yes	
• Operating range	Yes	
• Certification details	Yes	
• Compliance with MCERTS	Yes	
• Location	Yes	
• Date and time of work	Yes	Date only
• Equipment used - Type serial no's etc	Yes	Gas divider and regulator nos
• Gases used – certificate numbers, expiry dates, type	No	Copy of gas certificates offered to customer
• NOx converter efficiency check	N/A	
• Calibration and linearity data	Yes	
• Logged data for period of calibration/linearity	No	
• Name & signature of test engineer	Yes	

Requirement	Compliance Y/N	Notes
<b>9c – Service Report – SICK OMD</b>		
• Document reference	Yes	
• Instrument manufacturer	Yes	
• Instrument Type	Yes	
• Instrument model	Yes	
• Instrument Serial No's	Yes	
• Operating principal	Yes	
• Operating range	Yes	
• Certification details	Yes	
• Compliance with MCERTS	Yes	
• Location	Yes	
• Date and time of work	Yes	Date Only
• Equipment used - Type serial no's etc	Yes	On Linearity Sheet
• Gases used – certificate numbers, expiry dates, type	N/A	Linearity carried out using filters. Filter numbers stated (F1 to F8)
• NOx converter efficiency check	N/A	
• Calibration and linearity data	Yes	
• Logged data for period of calibration/linearity	Yes	Data supplied separately by client
• Name & signature of test engineer	Yes	

Requirement	Compliance Y/N	Notes
<b>9c – Service Report – SICK GM32</b>		
• Document reference	Yes	
• Instrument manufacturer	Yes	
• Instrument Type	Yes	
• Instrument model	Yes	
• Instrument Serial No's	Yes	
• Operating principal	Yes	
• Operating range	Yes	
• Certification details	Yes	
• Compliance with MCERTS	Yes	
• Location	Yes	
• Date and time of work	Yes	
• Equipment used - Type serial no's etc	No	
• Gases used – certificate numbers, expiry dates, type	Yes	
• NOx converter efficiency check	N/A	
• Calibration and linearity data	Yes	
• Logged data for period of calibration/linearity	Yes	Analyser readout checked against DCS values
• Name & signature of test engineer	Yes	



## CALIBRATION CERTIFICATE

Site	Aberthaw		
Date	18/02/2015		
Instrument	Unit 7 Primary		
Instrument Ref	8800623	8500442M	
Engineer	M.Findley		


### Results:

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
CO	PPM	500	333	Site Gas	310	331	-0.4
NO	PPM	1000	667	Site Gas	679	676	0.9
SO2-H	PPM	1000	800	5501912	811	811	1.1
SO2-L	PPM	250	200	5503378	211	211	4.4

### Cross sensitivity tests:

Test/ Gas conc	Response			% Error FSD		
	CO	NO	SO2-H	CO	NO	SO2-H
CO2 15%	10	14	-1	2.0	1.4	-0.1
H2O 2.5%	10	-3	12	2.0	-0.3	1.2
CO2 15% & H2O 2.5%	11	5	10	2.2	0.5	1.0

Test/ Gas conc	Response			% Error FSD		
	SO2-L			SO2-L		
CO2 15%	-1			-0.4		
H2O 2.5%	5			2.0		
CO2 15% & H2O 2.5%	4			1.6		

Signature:   
M.Findley  
Procal Customer Support Engineer

Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK  
Tel: +44 (0) 17336 232495 Email: procalsales@parker.com Website: www.procal.com



## Functionality Test Results

### DECLARATION OF CONFORMITY

We,

Kittiwake Procal Ltd  
5 Maxwell Road  
Woodston  
Peterborough  
Cambridgeshire  
PE2 7HU  
United Kingdom

Tel: (+44) (0) 1733 232495

Fax: (+44) (0) 1733 235255

e-mail: [post@procal.com](mailto:post@procal.com)

Web site: [www.procal.com](http://www.procal.com)

Declare that the product: **Type PULSI 200 Analyser** installed at

Customer	Aberthaw
Control Unit Serial Number	8800623
P200 Serial Number	8500442M
Site Identification	Unit 7 Primary

To which this declaration relates is in conformance with the MCERTs Performance Standard for Continuous Emission Monitoring Systems: Version 2. Revision 1 (April 2003).

#### MCERTS Product Conformance Certificates:

Licence	Certificate Number	Variation	Dated
ACU	SIRA MC 990006/	8	02/10/09
ACWn	SIRA MC 050060/	6	02/10/09

#### MCERTs Certification Body:

Sira Certification Service	
12 Acorn Industrial Park	Tel: (+44) (0) 1322 520500
Crayford Road	
Crayford	Fax: (+44) (0) 1322 520501
Dartford	
Kent	
DA1 4AL	e-mail: <a href="mailto:info@siraenvironmental.com">info@siraenvironmental.com</a>
United Kingdom	Web site: <a href="http://www.siracertification.com">www.siracertification.com</a>

#### Quality System:

ISO 9001:2008	Certificate Nr: 062043	Dated: 11/06/09
EN 15267-2	Dated: 2009	Technical Standards Used:
EN 15267-3	Dated: 2007	

All tests carried out to Kittiwake Procal's working procedures and those within Annex A of EN14181, Version 2.3 June 2010, relating to functionality testing by CEMs' manufacturers.

Signed:  Date: 18/2/2015

Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, United Kingdom  
Tel: +44 (0) 1733 232495 E-mail: [procal@parker.com](mailto:procal@parker.com) Web: [www.procal.com](http://www.procal.com)

F:/Drawings & Forms/Forms & procedures/Forms/Current/7-7880-01 Functionality Test Results inc App1  
Change Note: 7008144 Date: 24/09/10

Page 1 of 6





## 1.0 Alignment and Cleanliness.

### 1.1 Internal Check of Analyser

Analyser under test	Status	Notes
Optical path clear	√	
Optics clean	√	Process Mirror replaced
Internal wiring ok	√	
Air supply on	√	
Purge flow >300 ltrs/hr	√	Ok
Pressure differential <80mBar	√	38

## 2.0 Zero and Span Checks.

### 2.1 Zero Check

Component:	H2O	CO2	CO	NO	SO2-H	SO2-L	Ta	Ts	Ps
Units:	%	mau	PPM	PPM	PPM	PPM			
Process conc	2.4	145	673	459	47	36	26.8	130	1006
Zero conc	0	0	-5	-2	-1	-1	26.8	129	1044
% error	<2	<2	<2	<2	<2	<2			
Pass/Fail	Pass	Pass	Pass	Pass	Pass	Pass			
T90 recovery									

### 2.2 Span Check

Component:	H2O	CO2	CO	NO	SO2-H	SO2-L	Ta	Ts	Ps
Units:	%	mau	PPM	PPM	PPM	PPM			
Full range	10		500	1000	1000	250			
Zero Conc	0	0	-5	-2	-1	-1	26.8	129	1044
Test Gas Conc			333	667	800	200			
Cert number			Site Gas	Site Gas	040005 501912	040005 503378			
Response Conc			331	676	867	187	26.8	132	1035
% Error			<2	<2	>2	>2			
Pass/Fail			Pass	Pass	Fail	Fail			
T90 time Seconds				75	70	75			

### 2.3 Filter Level Checks

All on zero air	F6	F2	F7	F5G	F3	F1G	F4	F8
Factory	1890	3979	4090	2306	3471	2218	539	3481
Last visit	1270	2479	2701	1538	2409	1589	413	2422
This visit	1691	3364	3697	2111	3166	2090	447	2873



## 2.4 Interference Checks

Component: Units:	H2O %	CO2 mau	CO PPM	NO PPM	SO2-H PPM	SO2-L PPM
ZERO	0	0	1	1	1	1
CO2 15%	0	148	10	14	-1	-1
H2O 2.5 %	2.6	2	10	-3	12	5
H2O 2.5% & CO2 15%	2.7	146	11	5	10	4

## 3.0 Data Acquisition.

Displayed \ Component		H2O %	CO PPM	NO PPM	SO2-H PPM	SO2-L PPM
Zero	Analyser	0	-5	-2	-1	-1
	DCS					
	4-20mA					
Span	Analyser	2.6	331	676	867	187
	DCS					
	4-20mA					

## 4.0 Linear coefficient record.

Coefficient \ Component		H2O	CO2	CO	NO	SO2-H	SO2-L
Last test	Linear	2.160e-2	1	7.612e+0	1.379e+1	4.932e+0	4.423e+0
	Squared	4.210e-5	0	1.122e-1	4.429e-2	-5.139e-4	1.431e-3
	Cubed						
This test	Linear	2.16e-2	1	8.152e0	1.379e1	5.105e0	4.423e0
	Squared	4.12e-5	0	1.202e-1	4.429e-2	-5.319e-4	1.431e-3
	Cubed						

## 5.0 Test Information

5.1 Procal Engineer: Michael Findley

5.2 Test House:

5.3 Test House Engineer:





## Appendix 1. Pro-forma for assessing and reporting the results of the functional tests

Requirement	Notes	
<b>1. Alignment and Cleanliness</b>		
A visual inspection, with reference to the CEMs manuals, shall be carried out on the following when applicable:		
• Internal check of the CEM	√	All Functions ok
• Cleanliness of the optical components	√	Process Mirror replaced
• Flushing air supply	√	Ok
• Obstructions in the optical path	√	None
• After re-assembly at the measurement location at least the following shall be checked	√	No re-assembly required
• Alignment of the measuring system	√	
• Contamination control (internal check of optical surfaces)	√	Ok
<b>2. Sampling Systems</b>		
A visual inspection of the sampling system shall be performed, noting the condition of the following components, when fitted:		
• Sampling probe	n/a	
• Gas conditioning systems	n/a	
• Pumps	n/a	
• All connections	√	Ok
• Sample lines	n/a	
• Power supplies	√	Ok
• Filters	√	Ok
• NOx converters - if the sampling system contains a NOx converter, then the test laboratory shall record when the last efficiency-test was performed, and the result of this test.	n/a	
• The sampling system shall be in good condition and free of any visible faults, which may decrease the quality of data	√	Ok
<b>3. Leak testing</b>		
• Leak testing shall be performed according to the CEMs manuals. The tests shall cover the entire sampling system.	n/a	
<b>4. Zero and Span check</b>		
• Reference zero and span materials shall be used to verify the corresponding readings of the CEM.	√	Ok
• In case of non-extractive CEM, zero and span checks shall be performed using a reference-path free of flue gas before and after readjustment and after re-assembly of the CEM at the measurement location.	√	Ok
<b>5. Linearity</b>		
• During the calibration/linearity tests the applied concentrations should be logged onto the DCS to prove the complete system i.e. concentration applied to the instrument is represented by the instrument output and identical to the value logged on the DCS.	√	Data collected on ACU and customer DCS during testing





Requirement		Notes
DCS logged values should be included in the instrument service report.	√	Data collected on ACU and customer DCS during testing
<ul style="list-style-type: none"> <li>The linearity of the CEM's response shall be checked using five different reference materials, including a zero concentration.</li> </ul>	√	See linearity Cert
<ul style="list-style-type: none"> <li>The reference material with zero concentration, as well as the reference materials with four different concentrations, shall have a verifiable quantity and quality.</li> </ul>	√	Recorded on Test Sheet
<ul style="list-style-type: none"> <li>In case of gaseous reference materials, these four reference materials can be obtained from different gas cylinders or can be prepared by means of a calibrated dilution system from one single gas concentration.</li> </ul>	√	Gas points derived from blending full gas range using a Procal P900, a water generator/gas blender
<ul style="list-style-type: none"> <li>The reference material concentrations shall be selected such that the measured values are at approximately 20 %, 40 %, 60 % and 80 % of the range of two times the emission limit. It is necessary to know the values of the ratios of their concentrations precisely enough so that an incorrect failure of the linearity test does not occur. The dry test reference material shall be applied to the inlet of the CEM.</li> </ul>	√	See linearity Cert
The individual CEMs are tested using the following concentrations applied in a randomized sequence:		
<ul style="list-style-type: none"> <li>Reference material with zero concentration;</li> </ul>	√	See linearity Cert
Reference material concentration approximately 20 % of 2 times the emission limit;	√	See linearity Cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 40 % of 2 times the emission limit;</li> </ul>	√	See linearity Cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 60 % of 2 times the emission limit;</li> </ul>	√	See linearity Cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 80 % of 2 times the emission limit;</li> </ul>	√	See linearity Cert
<ul style="list-style-type: none"> <li>Reference material with zero concentration;</li> </ul>	√	See linearity Cert
<p>After each change in concentration, the first instrument reading shall ordinarily be taken after a time period equal to at least three times the response time of the CEM. At each reference material concentration, at least three readings shall be made, six readings shall be taken at zero. The time period between the start of each of the three readings shall be separated by at least four times the response time.</p> <p>The test-laboratory may apply a risk-based approach to linearity testing in order to reduce the time for the tests. For example, the readings may be taken after less than 3x the response time; however, if the CEM fails the linearity test, then the test laboratory shall repeat the linearity test and wait at least 3x the response time as stated above. Alternatively, the number of repetitions of the test may be reduced if the CEM passes the required performance criteria by a factor of at least 2 (i.e. half the allowable residual). Increasing the waiting time to 5x the response time, for example, may be a means of meeting this requirement. Where no other method is possible, the linearity can also be performed with the aid of reference materials such as grating filters or gas filters.</p>		



Requirement	Notes
The linearity shall be calculated and tested using the procedure as given in EN 14181 annex B. If the CEM does not pass this test; then the problem shall be identified and rectified.	
<b>6. Interferences</b>	
<ul style="list-style-type: none"> <li>A test shall be undertaken if the process gases to be monitored contain components that are known interferences, as identified during QAL1 and there is a failure of the QAL2 or AST which could be due to interferences.</li> </ul>	<ul style="list-style-type: none"> <li>√ Interference checks made with water vapour, CO2 and each gas species. See Calcert</li> </ul>
<b>7. Zero and Span drift (Audit)</b>	
<ul style="list-style-type: none"> <li>The test laboratory shall assess whether the operator has a QAL3 procedure in place, and whether the operator has applied this procedure. The evidence would comprise (i) a documented procedure, (ii) zero and span data, (iii) control charts.</li> </ul>	<ul style="list-style-type: none"> <li>√ Customer has procedure in place and documented evidence</li> </ul>
<b>Response Time</b>	
<ul style="list-style-type: none"> <li>The response time of the CEM shall be checked. This can be performed, if appropriate, by feeding of the reference material at the end of the sampling probe. The response time shall not exceed the performance requirement applied during the QAL1 tests.</li> </ul>	<ul style="list-style-type: none"> <li>√ Response time recorded on test sheet and this report</li> </ul>
<b>9. Service Report</b>	
As a minimum requirement the service report should include the following:	
<ul style="list-style-type: none"> <li>Document reference for work instruction for the type of work being undertaken</li> </ul>	<ul style="list-style-type: none"> <li>√ Parker Procal Method Statement</li> </ul>
<ul style="list-style-type: none"> <li>Instrument manufacturer</li> </ul>	<ul style="list-style-type: none"> <li>√ Parker Procal (UK)</li> </ul>
<ul style="list-style-type: none"> <li>Instrument type</li> </ul>	<ul style="list-style-type: none"> <li>√ Pulsi 2000</li> </ul>
<ul style="list-style-type: none"> <li>Instrument model</li> </ul>	<ul style="list-style-type: none"> <li>√ P2000</li> </ul>
<ul style="list-style-type: none"> <li>Instrument Serial No</li> </ul>	<ul style="list-style-type: none"> <li>√ 8500442M</li> </ul>
<ul style="list-style-type: none"> <li>Operating principle</li> </ul>	<ul style="list-style-type: none"> <li>√ GFCIR/NIR</li> </ul>
<ul style="list-style-type: none"> <li>Operating range</li> </ul>	<ul style="list-style-type: none"> <li>√ Recorded on this report</li> </ul>
<ul style="list-style-type: none"> <li>Certification details</li> </ul>	<ul style="list-style-type: none"> <li>√ mcerted</li> </ul>
<ul style="list-style-type: none"> <li>Compliance with MCERTS (including certificate no.)</li> </ul>	<ul style="list-style-type: none"> <li>√ SIRA MC 990006/08</li> </ul>
<ul style="list-style-type: none"> <li>Location</li> </ul>	<ul style="list-style-type: none"> <li>√ Aberthaw Power Station, Wales</li> </ul>
<ul style="list-style-type: none"> <li>Date and time work was undertaken</li> </ul>	<ul style="list-style-type: none"> <li>√ 18/02/2015</li> </ul>
<ul style="list-style-type: none"> <li>Equipment used - type, serial no's, calibration dates</li> </ul>	<ul style="list-style-type: none"> <li>√ Procal Water Generator/Gas divider Pcal #2</li> </ul>
<ul style="list-style-type: none"> <li>Gases used - certificate numbers, expiry dates, binary / mix</li> </ul>	<ul style="list-style-type: none"> <li>√ Copy of Gas certs can be provided</li> </ul>
<ul style="list-style-type: none"> <li>NOx converter efficiency test, if applicable</li> </ul>	<ul style="list-style-type: none"> <li>n/a</li> </ul>
<ul style="list-style-type: none"> <li>Calibration and linearity data as required by EN14181</li> </ul>	<ul style="list-style-type: none"> <li>√ See Linearity &amp; Calcert data</li> </ul>
<ul style="list-style-type: none"> <li><b>Logged</b> data for period of calibration and linearity. Note there may be gaps in the data, in such cases, the test laboratory shall state why there are gaps in the data</li> </ul>	<ul style="list-style-type: none"> <li>√ See Linearity &amp; Calcert data</li> </ul>






## LINEARISATION RESULTS

Site	Aberthaw	
Date	18/02/2015	
Instrument	Unit 7 Primary	
Instrument Ref	8800623	8500442M
Engineer	Michael Findley	


The above analyser has undergone a linearisation check and calibration, if required, in accordance with the company's quality system. The following results were obtained.

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
NO	PPM	1000	800	40005501912	811	811	1.1
					814	814	1.4
					815	815	1.5
NO	PPM	1000	600	40005501912	609	609	0.9
					612	612	1.2
					613	613	1.3
NO	PPM	1000	400	40005501912	407	407	0.7
					408	408	0.8
					405	405	0.5
NO	PPM	1000	200	40005501912	206	206	0.6
					205	205	0.5
					204	204	0.4



Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK  
Tel: +44 (0) 17336 232495 Email: [procalsales@parker.com](mailto:procalsales@parker.com) Website: [www.procal.com](http://www.procal.com)

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
SO2-H	PPM	1000	800	40005503378	867	867	6.7
					866	866	6.6
					866	866	6.6
SO2-H	PPM	1000	600	40005503378	647	647	4.7
					650	650	5.0
					650	650	5.0
SO2-H	PPM	1000	400	40005503378	432	432	3.2
					431	431	3.1
					432	432	3.2
SO2-H	PPM	1000	200	40005503378	211	211	1.1
					211	211	1.1
					212	212	1.2




Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK TEL:+44(0)1733 23495 FAX:+44(0)1733 235255  
E-mail: asampson@procal.com Website: www.procal.com

7-7862-04 Linearisation Certificate

Change note: 7009251

Date: 30/07/12

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
SO2-L	PPM	250	200	40005503378	187	187	-5.2
					187	187	-5.2
					187	187	-5.2
SO2-L	PPM	250	150	40005503378	140	140	-4.0
					139	139	-4.4
					139	139	-4.4
SO2-L	PPM	250	100	40005503378	91	91	-3.6
					92	92	-3.2
					92	92	-3.2
SO2-L	PPM	250	50	40005503378	44	44	-2.4
					44	44	-2.4
					44	44	-2.4



Zero	H2O	CO2	CO	NO	SO2 - H	SO2 - L
#1	0	0	1	1	1	1
#2	0	1	2	5	0	0
#3	0	0	-5	-2	-1	-1
#4	0	0	-6	4	3	3
#5	0	0	5	0	-2	-2
#6	0	-1	1	-6	-4	-4

Signature:

Procal Customer Support Engineer

Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK TEL:+44(0)1733 23495 FAX:+44(0)1733 235255

E-mail: [asampson@procal.com](mailto:asampson@procal.com) Website: [www.procal.com](http://www.procal.com)

7-7862-04 Linearisation Certificate

Change note: 7009251

Date: 30/07/12





## Functionality Test Results

### DECLARATION OF CONFORMITY

We,

Kittiwake Procal Ltd  
5 Maxwell Road  
Woodston  
Peterborough  
Cambridgeshire  
PE2 7HU  
United Kingdom

Tel: (+44) (0) 1733 232495

Fax: (+44) (0) 1733 235255

e-mail: [post@procal.com](mailto:post@procal.com)

Web site: [www.procal.com](http://www.procal.com)

declare that the product: **Type PULSI 200 Analyser** installed at

Customer	RWE Aberthaw
Control Unit Serial Number	8800624
P200 Serial Number	8500705
Site Identification	Unit 7 Secondary

to which this declaration relates is in conformance with the **MCERTs Performance Standard for Continuous Emission Monitoring Systems: Version 2. Revision 1 (April 2003)**.

#### MCERTS Product Conformance Certificates:

Licence	Certificate Number	Variation	Dated
ACU	SIRA MC 990006/	8	02/10/09
ACWn	SIRA MC 050060/	6	02/10/09

#### MCERTs Certification Body:

Sira Certification Service	
12 Acorn Industrial Park	Tel: (+44) (0) 1322 520500
Crayford Road	
Crayford	Fax: (+44) (0) 1322 520501
Dartford	
Kent	
DA1 4AL	e-mail: <a href="mailto:info@siraenvironmental.com">info@siraenvironmental.com</a>
United Kingdom	Web site: <a href="http://www.siracertification.com">www.siracertification.com</a>

#### Quality System:

ISO 9001:2008	Certificate Nr: 062043	Dated: 11/06/09
EN 15267-2	Dated: 2009	
EN 15267-3	Dated: 2007	

#### Technical Standards Used:

**All tests carried out to Kittiwake Procal's working procedures and those within Annex A of EN14181, Version 2.3 June 2010, relating to functionality testing by CEMs' manufacturers.**

Signed:.....

Date:..18/02/2015.....

Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, United Kingdom

Tel: +44 (0) 1733 232495 E-mail: [procal@parker.com](mailto:procal@parker.com) Web: [www.procal.com](http://www.procal.com)

F:/Drawings & Forms/Forms & procedures/Forms/Current/7-7880-01 Functionality Test Results inc App1

Change Note: 7008144

Date: 24/09/10

Page 1 of 6



## 1.0 Alignment and Cleanliness.

### 1.1 Internal Check of Analyser

Analyser under test	Status	Notes
Optical path clear	Yes	OK
Optics clean	Yes	OK
Internal wiring ok	Yes	
Air supply on	Yes	
Purge flow >300 ltrs/hr	Yes	
Pressure differential <80mBar	Yes	16 mbar

## 2.0 Zero and Span Checks.

### 2.1 Zero Check

Component: Units:	H2O %	CO2 mAU	CO ppm	NO ppm	SO2-H ppm	SO2-L ppm	Ta	Ts	Ps
Process conc	1.9	218	1000	423	74	67	25.5	129.7	1027.5
Zero conc	-0.1	-4	1	-4	1	1	25.5	129.2	1043.2
% error	<2	<2	<2	<2	<2	<2			
Pass/Fail	Pass	Pass	Pass	Pass	Pass	Pass			
T90 recovery									

### 2.2 Span Check

Component: Units:	H2O %	CO2 mAU	CO ppm	NO ppm	SO2-H ppm	SO2-L ppm	Ta	Ts	Ps
Full range	10		500	1000	1000	250			
Zero Conc	-0.1	-6	5	-2	-2	-2	25.2	130.2	1036.9
Test Gas Conc			306	800	800	200			
Cert number			5682016	5501912	040005503377				
Response Conc			316	820	820	196			
% Error			2	2	2	<2			
Pass/Fail			Pass	Pass	Pass	Pass			
T90 time				70	70	70			

### 2.3 Filter Level Checks

All on zero air	F2	F4	F6	F1G	F7	F5G	F8	F3	
Factory	2093	3650	3904	2426	3776	2399	463	2281	
Last visit	1995	2473	3808	2437	3805	2672	491	2381	
This visit	1996	3466	3773	2426	3783	2677	494	2358	





## 2.4 Interference Checks

Component: Units:	H2O %	CO2 mAU	CO ppm	NO ppm	SO2-H ppm	SO2-L ppm
ZERO	0	-2	-3	-4	0	0
CO2 15%	0.2	222	1	-4	3	1
H2O 2.5%	2	3	-1	-15	9	6
H2O + CO2	2.3	227	1	-13	12	9

## 3.0 Data Acquisition.

Displayed \ Component						
Zero	Analyser					
	DCS					
	4-20mA					
Span	Analyser					
	DCS					
	4-20mA					

## 4.0 Linear coefficient record.

Coefficient \ Component		H2O	CO2	CO	NO	SO2-H	SO2-L
Last test	Linear	1.650e-2		7.400e+0	1.200e+1	6.373e+0	5.810e+0
	Squared	4.400e-5		1.120e-1	4.150e-2	-1.570e-3	5.790e-3
	Cubed						
This test	Linear	1.650e-2		7.400e+0	1.200e+1	6.373e+0	5.810e+0
	Squared	4.400e-5		1.120e-1	4.150e-2	-1.570e-3	5.790e-3
	Cubed						

## 5.0 Test Information

5.1 Procal Engineer: Panteleimon Kompeas-Charidis

5.2 Test House:

5.3 Test House Engineer:





## Appendix 1. Pro-forma for assessing and reporting the results of the functional tests

Requirement	Notes	
<b>1. Alignment and Cleanliness</b>		
A visual inspection, with reference to the CEMs manuals, shall be carried out on the following when applicable:		
• Internal check of the CEM	✓	All functions ok
• Cleanliness of the optical components	✓	
• Flushing air supply	✓	Ok
• Obstructions in the optical path	✓	None
• After re-assembly at the measurement location at least the following shall be checked	✓	No re-assembly required
• Alignment of the measuring system	✓	No alignment required
• Contamination control (internal check of optical surfaces)	✓	Filter levels ok
<b>2. Sampling Systems</b>		
A visual inspection of the sampling system shall be performed, noting the condition of the following components, when fitted:		
• Sampling probe	n/a	
• Gas conditioning systems	n/a	
• Pumps	n/a	
• All connections	✓	Ok
• Sample lines	n/a	
• Power supplies	✓	Ok
• Filters	✓	Ok
• NOx converters - if the sampling system contains a NOx converter, then the test laboratory shall record when the last efficiency-test was performed, and the result of this test.	n/a	
• The sampling system shall be in good condition and free of any visible faults, which may decrease the quality of data	✓	Ok
<b>3. Leak testing</b>		
• Leak testing shall be performed according to the CEMs manuals. The tests shall cover the entire sampling system.	n/a	
<b>4. Zero and Span check</b>		
• Reference zero and span materials shall be used to verify the corresponding readings of the CEM.	✓	Ok
• In case of non-extractive CEM, zero and span checks shall be performed using a reference-path free of flue gas before and after readjustment and after re-assembly of the CEM at the measurement location.	✓	Ok
<b>5. Linearity</b>		
• During the calibration/linearity tests the applied concentrations should be logged onto the DCS to prove the complete system i.e. concentration applied to the instrument is represented by the instrument output and identical to the value logged on the DCS.	✓	Data collected on ACU MK4 & customer DCS during testing



Requirement		Notes
DCS logged values should be included in the instrument service report.	√	Data collected on ACU MK4 & customer DCS during testing
• The linearity of the CEM's response shall be checked using five different reference materials, including a zero concentration.	√	See linearity Cert
• The reference material with zero concentration, as well as the reference materials with four different concentrations, shall have a verifiable quantity and quality.	√	See gas certs
• In case of gaseous reference materials, these four reference materials can be obtained from different gas cylinders or can be prepared by means of a calibrated dilution system from one single gas concentration.	√	Gas points derived from blending full gas range using a Procal P9000, a water generator/gas blender
• The reference material concentrations shall be selected such that the measured values are at approximately 20 %, 40 %, 60 % and 80 % of the range of two times the emission limit. It is necessary to know the values of the ratios of their concentrations precisely enough so that an incorrect failure of the linearity test does not occur. The dry test reference material shall be applied to the inlet of the CEM.	√	See linearity cert
The individual CEMs are tested using the following concentrations applied in a randomised sequence:		
• Reference material with zero concentration;	√	See linearity cert
Reference material concentration approximately 20 % of 2 times the emission limit;	√	See linearity cert
• Reference material concentration approximately 40 % of 2 times the emission limit;	√	See linearity cert
• Reference material concentration approximately 60 % of 2 times the emission limit;	√	See linearity cert
• Reference material concentration approximately 80 % of 2 times the emission limit;	√	See linearity cert
• Reference material with zero concentration;	√	See linearity cert
After each change in concentration, the first instrument reading shall ordinarily be taken after a time period equal to at least three times the response time of the CEM. At each reference material concentration, at least three readings shall be made, six readings shall be taken at zero. The time period between the start of each of the three readings shall be separated by at least four times the response time.		
The test-laboratory may apply a risk-based approach to linearity testing in order to reduce the time for the tests. For example, the readings may be taken after less than 3x the response time; however, if the CEM fails the linearity test, then the test laboratory shall repeat the linearity test and wait at least 3x the response time as stated above. Alternatively, the number of repetitions of the test may be reduced if the CEM passes the required performance criteria by a factor of at least 2 (i.e. half the allowable residual). Increasing the waiting time to 5x the response time, for example, may be a means of meeting this requirement. Where no other method is possible, the linearity can also be performed with the aid of reference materials such as grating filters or gas filters.		





Requirement	Notes
The linearity shall be calculated and tested using the procedure as given in EN 14181 annex B. If the CEM does not pass this test; then the problem shall be identified and rectified.	
<b>6. Interferences</b>	
<ul style="list-style-type: none"> <li>A test shall be undertaken if the process gases to be monitored contain components that are known interferences, as identified during QAL1 and there is a failure of the QAL2 or AST which could be due to interferences.</li> </ul>	<ul style="list-style-type: none"> <li>✓ Interference checks made with Water vapour, CO2 and each gas species. See also Calcert</li> </ul>
<b>7. Zero and Span drift (Audit)</b>	
<ul style="list-style-type: none"> <li>The test laboratory shall assess whether the operator has a QAL3 procedure in place, and whether the operator has applied this procedure. The evidence would comprise (i) a documented procedure, (ii) zero and span data, (iii) control charts.</li> </ul>	<ul style="list-style-type: none"> <li>✓</li> </ul>
<b>Response Time</b>	
<ul style="list-style-type: none"> <li>The response time of the CEM shall be checked. This can be performed, if appropriate, by feeding of the reference material at the end of the sampling probe. The response time shall not exceed the performance requirement applied during the QAL1 tests.</li> </ul>	<ul style="list-style-type: none"> <li>✓ Response time recorded on test sheet.</li> </ul>
<b>9. Service Report</b>	
As a minimum requirement the service report should include the following:	
<ul style="list-style-type: none"> <li>Document reference for work instruction for the type of work being undertaken</li> </ul>	<ul style="list-style-type: none"> <li>✓ Procal Method Statement</li> </ul>
<ul style="list-style-type: none"> <li>Instrument manufacturer</li> </ul>	<ul style="list-style-type: none"> <li>✓ Procal</li> </ul>
<ul style="list-style-type: none"> <li>Instrument type</li> </ul>	<ul style="list-style-type: none"> <li>✓ Pulsi 2000</li> </ul>
<ul style="list-style-type: none"> <li>Instrument model</li> </ul>	<ul style="list-style-type: none"> <li>✓ P2000</li> </ul>
<ul style="list-style-type: none"> <li>Instrument Serial No</li> </ul>	<ul style="list-style-type: none"> <li>✓ 8500705</li> </ul>
<ul style="list-style-type: none"> <li>Operating principle</li> </ul>	<ul style="list-style-type: none"> <li>✓ GFCIR &amp; NIR</li> </ul>
<ul style="list-style-type: none"> <li>Operating range</li> </ul>	<ul style="list-style-type: none"> <li>✓ Recorded on this report</li> </ul>
<ul style="list-style-type: none"> <li>Certification details</li> </ul>	<ul style="list-style-type: none"> <li>✓ Mcerted</li> </ul>
<ul style="list-style-type: none"> <li>Compliance with MCERTS (including certificate no.)</li> </ul>	<ul style="list-style-type: none"> <li>✓ SIRA MC 990006/08</li> </ul>
<ul style="list-style-type: none"> <li>Location</li> </ul>	<ul style="list-style-type: none"> <li>✓ RWE Aberthaw</li> </ul>
<ul style="list-style-type: none"> <li>Date and time work was undertaken</li> </ul>	<ul style="list-style-type: none"> <li>✓ 18/02/2015</li> </ul>
<ul style="list-style-type: none"> <li>Equipment used - type, serial no's, calibration dates</li> </ul>	<ul style="list-style-type: none"> <li>✓ Procal Water Generator/Gas divider Pcal #6</li> </ul>
<ul style="list-style-type: none"> <li>Gases used - certificate numbers, expiry dates, binary / mix</li> </ul>	<ul style="list-style-type: none"> <li>✓ Copy of Gas certs can be provided</li> </ul>
<ul style="list-style-type: none"> <li>NOx converter efficiency test, if applicable</li> </ul>	<ul style="list-style-type: none"> <li>n/a</li> </ul>
<ul style="list-style-type: none"> <li>Calibration and linearity data as required by EN14181</li> </ul>	<ul style="list-style-type: none"> <li>✓ See Linearity &amp; Calcert Data</li> </ul>
<ul style="list-style-type: none"> <li>Logged data for period of calibration and linearity. Note there may be gaps in the data, in such cases, the test laboratory shall state why there are gaps in the data</li> </ul>	<ul style="list-style-type: none"> <li>✓ See Linearity &amp; Calcert Data</li> </ul>



Visit Ref:0000548



## Service Engineers Report

<b>Client</b> Aberthaw Power Station RWE INNOGY PLC THE LEYS (NEAR BARRY) SOUTH GLAMORGAN  <b>Contact</b> James Beveridge  <b>Tel No.</b> 01446 750271	<b>Job Category</b> Chargeable UK  <b>Activity</b> Functionality inc. Linearisation  <b>Date of Visit</b> 16/02/2015  <b>Analyser/System No.</b> U8 Secondary U9 Secondary U7 Secondary
---	---

Qty	Description	Part No.

Travel Time	9 Hours	Mileage driven	498
Site Time	3 Days	Expenses incurred	£358.94
Engineer	Pandeli Kompeas-Charidis	For Client	

### Notes for Call

8800624: Linearisation completed on all Secondary Units 7, 8 & 9

Unit7 Secondary 8500705: Dry gases no coefficient changes made, NO reads quite OK, SO2-H reads slightly high, SO2-L reads rather low and CO reads a bit high but within spec

Unit8 Secondary 8500726: Dry gases, all coefficients had to be changed; NO, SO2 & CO gases now all within spec, SO2-L & SO2-H readings now in close match which was achieved also by reverting back to older X-product coefficients

Unit9 Secondary 8500707: Dry gases no coefficient changes made, NO reads rather low, SO2-H reads rather high, SO2-L and CO read quite OK

Please refer to certs for more details

Kittiwake Procal Limited - 5 Maxwell Road, Woodston, Peterborough, PE2 7HU  
Telephone: +44 01733 232495 Fax: +44 01733 235255 Web: www.procal.com




## LINEARISATION RESULTS

Site	RWE Aberthaw	
Date	18/02/2015	
Instrument	Unit 7 Secondary	
Instrument Ref	8800624	8500705
Engineer	PKC	

The above analyser has undergone a linearisation check and calibration, if required, in accordance with the company's quality system. The following results were obtained.


Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
NO	ppm	1000	800	40005501912	800	800	0.0
					801	801	0.1
					803	803	0.3
NO	ppm	1000	200	40005501912	206	206	0.6
					205	205	0.5
					202	202	0.2
NO	ppm	1000	600	40005501912	601	601	0.1
					598	598	-0.2
					600	600	0.0
NO	ppm	1000	400	40005501912	401	401	0.1
					403	403	0.3
					395	395	-0.5



Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK  
Tel: +44 (0) 17336 232495 Email: [procal@parker.com](mailto:procal@parker.com) Website: [www.procal.com](http://www.procal.com)



Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
SO2_H	ppm	1000	800	40005503377	820	820	2.0
					823	823	2.3
					819	819	1.9
SO2_H	ppm	1000	200	40005503377	206	206	0.6
					206	206	0.6
					207	207	0.7
SO2_H	ppm	1000	600	40005503377	617	617	1.7
					618	618	1.8
					619	619	1.9
SO2_H	ppm	1000	400	40005503377	410	410	1.0
					412	412	1.2
					411	411	1.1




Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK TEL:+44(0)1733 23495 FAX:+44(0)1733 235255  
E-mail: [asampson@procal.com](mailto:asampson@procal.com) Website: [www.procal.com](http://www.procal.com)

7-7862-04 Linearisation Certificate


Change note: 7009251

Date: 30/07/12

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
SO2_L	ppm	250	200	40005503377	195	195	-2.0
					196	196	-1.6
					197	197	-1.2
SO2_L	ppm	250	50	40005503373	49	49	-0.4
					48	48	-0.8
					50	50	0.0
SO2_L	ppm	250	150	40005503373	141	141	-3.6
					142	142	-3.2
					144	144	-2.4
SO2_L	ppm	250	100	40005503373	94	94	-2.4
					91	91	-3.6
					94	94	-2.4



Zero	H2O	CO2	CO	NO	SO2 - H	SO2 - L
#1	0	0	0	0	-1	-1
#2	0	-1	2	-1	1	1
#3	-0.1	-4	1	-4	1	1
#4	-0.1	-6	5	-2	-2	-2
#5	-0.1	-2	1	-1	1	1
#6	0	-2	0	0	-1	-1

Signature:   
 PKC  
 Procal Customer Support Engineer

Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK TEL:+44(0)1733 23495 FAX:+44(0)1733 235255  
 E-mail: [asampson@procal.com](mailto:asampson@procal.com) Website: [www.procal.com](http://www.procal.com)



Visit Ref:0000601

## Service Engineers Report

<b>Client</b> Aberthaw Power Station RWE INNOGY PLC THE LEYS (NEAR BARRY) SOUTH GLAMORGAN  <b>Contact</b> James Beveridge  <b>Tel No.</b> 01446 750271	<b>Job Category</b> Contract  <b>Activity</b> Functionality inc. Linearisation  <b>Date of Visit</b> 16/03/2015  <b>Analyser/System No.</b> U8 Secondary U7 Primary U8 Primary U9 Primary
---	--

Qty	Description	Part No.

Travel Time	10hrs	Mileage driven	510
Site Time	4 Days	Expenses incurred	£455.55
Engineer	Michael Findley	For Client	

### Notes for Call

#### Unit 7 Primary 8500443M

Dry gas test carried out on functionality test, NO, CO, within spec. SO2-H & SO2-L out of spec but no adjustment made.  
Linearization tests completed on NO, SO2-H & SO2-L channels. SO2-H & SO2-L greater than 2% but reading on the high scale for SO2-H and the low scale for SO2-L. No adjustments made to SO2-H or SO2-L channel. NO channel within spec.

#### Unit 8 Primary 8500916M

Dry gas test carried out on functionality test, NO, CO, within spec. SO2-H & SO2-L out of spec but no adjustment made.  
Linearization tests completed on NO, SO2-H & SO2-L channels. SO2-H & SO2-L greater than 2% but reading on the high scale for SO2-H & SO2-L channel. No adjustments made to SO2-H or SO2-L channel. NO channel within spec.

#### Unit 9 Primary 8500442M

Dry gas test carried out on functionality test, CO & SO2-L, within spec. NO & SO2-L slightly out of spec but no adjustment made.  
Linearization tests completed on NO, SO2-H & SO2-L channels. SO2-L within spec and SO2-H & NO channels greater than 2% no adjustments made to SO2-H or NO channel.

Kittiwake Procal Limited - 5 Maxwell Road, Woodston, Peterborough, PE2 7HU  
Telephone: +44 01733 232495 Fax: +44 01733 235255 Web: www.procal.com




**SICK (UK) LTD**  
Waldkirch House,  
39 Hedley Rd  
St Albans,  
Herts,  
AL1 5BN

# SICK

Tel : 0044 (0)1727 831121  
Fax : 0044 (0)1727 855332  
E – Mail : info@sick.co.uk

## U7 - OMD 41 Particulate Monitor

BS EN 14181 – Functional Test &  
Maintenance Report

Customer	RWE nPower
Site	Aberthaw
Contact	James Beveridge
Service Ref No.	SVON017322
Survey Dates	12.02.2015
Device Type	OMD41
Receiver S/N	9807 8000
Reflector S/N	9807 8000
Evaluation Unit S/N	9807 8000
Device Location	Unit 7
MCERTS Cert No	MC 040042/01
Certification Details	See Certificate
Measurement Principle	Opacity
SICK Engineer	Rhodri Jones
Signature	

Dust reading before check			
Opacity or Extinction		Extinction	
Dust in mg / m <sup>3</sup>	mg/m <sup>3</sup>	0.0	
Contamination	mA	0.000%	
Dust reading, measured values	mA	4.0	
Control cycle, Zero point values	mA	4.0	
Control cycle, Span value	mA	16.8	

1. Alignment & Cleanliness			
1.1 Visual Inspection of the analyser	<input checked="" type="checkbox"/>	Analyser dusty but OK	
1.2 Internal check of the analyser	<input checked="" type="checkbox"/>	Internal optics cleaned	
1.3 Cleanliness of optical components	<input checked="" type="checkbox"/>	Cleaned using optic cloth	
1.4 Flushing of air supply	<input checked="" type="checkbox"/>	Purge air OK	
1.5 Obstructions in the optical path	<input checked="" type="checkbox"/>	Cross duct - no obstructions	

2. Sampling System			
2.1 System in good condition and free from faults	<input checked="" type="checkbox"/>	No faults or warnings	
2.2 Visual Inspection of sampling probe	<input checked="" type="checkbox"/>	N/A	
2.3 Check of the Gas Conditioning System	<input checked="" type="checkbox"/>	N/A	
2.4 Check of the sample pump	<input checked="" type="checkbox"/>	N/A	
2.5 Check of all hoses, gas lines and connections	<input checked="" type="checkbox"/>	NA	
2.6 Check of power supplies	<input checked="" type="checkbox"/>	Power supplies OK	
2.7 Check of all filters	<input checked="" type="checkbox"/>	Purge air filter OK	
2.8 Check of heated lines are operating correctly	<input checked="" type="checkbox"/>	N/A	
2.9 Efficiency of NOx converter > 95% (if applicable)	<input checked="" type="checkbox"/>	N/A	
2.10 Has the response time been checked by applying gas to probe	<input checked="" type="checkbox"/>	N/A	

3. Serviceability			
3.1 Safe & Clean working environment for servicing	<input checked="" type="checkbox"/>	Workshop available	
3.2 Adequate supply of reference gases	<input checked="" type="checkbox"/>	NA	
3.3 Adequate supply of spare parts	<input checked="" type="checkbox"/>	Whole spare analysers available	
3.4 Ability to apply gas at probe	<input checked="" type="checkbox"/>	NA	
3.5 Output signals checked	<input checked="" type="checkbox"/>	Analogue outputs spot on	

<b>4. Zero Span &amp; Leak Checks</b>		
4.1 Leak check of whole system carried out	<input checked="" type="checkbox"/>	N/A
4.2 Are the QAL3 checks (control cycle) being done	<input checked="" type="checkbox"/>	Yes - 3 x Daily
4.3 Are the QAL3 checks being logged and plotted on control charts	<input checked="" type="checkbox"/>	Yes - OK
4.4 Are the QAL3 charts exhibiting signs of drift	<input checked="" type="checkbox"/>	No - OK

<b>5. Linearity</b>		
5.1 Checks at 0, 20%, 40%, 60% & 80% of 2 X Daily ELV's	<input checked="" type="checkbox"/>	N/A – % of Analyser Range
5.2 Different concentrations applied in a random sequence	<input checked="" type="checkbox"/>	Refer to linearity report
5.3 Has the response time for each component been recorded (T90)	<input checked="" type="checkbox"/>	Refer to linearity report
5.4 Have the first readings been taken at 3 x T90	<input checked="" type="checkbox"/>	Refer to linearity report
5.5 Have three readings been taken at least 4 x T90 apart	<input checked="" type="checkbox"/>	Refer to linearity report
5.6 Does the filter set have a valid calibration	<input checked="" type="checkbox"/>	Refer to linearity report

<b>6. Visual Inspection</b>		
6.1 Check if measured values are plausible in the control room	<input checked="" type="checkbox"/>	Yes
6.2 Check if control cycle (zero/span) is valid	<input checked="" type="checkbox"/>	Recorded & stored
6.3 Check if there is any error message	<input checked="" type="checkbox"/>	No errors
6.4 Check for mechanical or electrical connection damage	<input checked="" type="checkbox"/>	No damage
6.5 Check mechanical mounting	<input checked="" type="checkbox"/>	Sound

<b>7. Purge Air Unit</b>		
7.1 Check the hoses and ring nozzles	<input checked="" type="checkbox"/>	All OK
7.2 Clean the pre-filter	<input checked="" type="checkbox"/>	Replaced
7.3 Replace the filter insert	<input checked="" type="checkbox"/>	Replaced
7.4 Check function of filter monitor purge switch.	<input checked="" type="checkbox"/>	OK

<b>8. Fail Safe Shutter (FSS Option)</b>		
8.1 Check function of fail-safe shutter	<input checked="" type="checkbox"/>	Not fitted on stack



<b>9. Optical unit</b>		
9.1	Clean housing	<input checked="" type="checkbox"/> Housing cleaned
9.2	Check and, if necessary readjust optical axis	<input checked="" type="checkbox"/> Not required
9.3	Clean optical boundary surfaces for sender & receiver	<input checked="" type="checkbox"/> Optical surfaces cleaned
9.4	Check and replace desiccant cartridge, part no. 2008475	<input checked="" type="checkbox"/> Not required
9.5	Check the transmitter diode	<input checked="" type="checkbox"/> OK
9.6	Check the stepper motor	<input checked="" type="checkbox"/> Stepper motor functioning well
9.7	Check zero point. Re-calibrate zero point if required	<input checked="" type="checkbox"/> Reading Zero in workshop
9.8	Check function of control cycle	<input checked="" type="checkbox"/> Good
9.9	Carry out internal linearity check, record results below	<input checked="" type="checkbox"/> See below

Linearity Measurement	Set value	Actual Value	Deviation in %	< 2% OK?
Zero	4.00 mA	4.00 mA	0.0	OK
Span 2 (approx. 30%)	14.1 mA	14.1 mA	0.0	OK
Span 1 (approx. 70%)	16.8 mA	16.8 mA	0.0	OK
Span 3 (100%)	20.00 mA	20.00 mA	0.0	OK

<b>10. Connection unit</b>		
10.1	Check the analogue and digital input and output signals	<input checked="" type="checkbox"/> Reading 4.0 @ 4.0mA

<b>11. Final check</b>		
11.1	Check function of complete measurement system	<input checked="" type="checkbox"/> System functioning well
11.2	Check the availability of the system	<input checked="" type="checkbox"/> 100%
11.3	Write device data into the report file	<input checked="" type="checkbox"/> Device data saved
11.4	Complete Service Report and Measurement Log-file	<input checked="" type="checkbox"/> Report complete

<b>12. Additional details:</b>	
Quality of maintenance work by plant personnel: Good	
Cleanliness of measuring station and devices: OK	
Maintenance work done (in addition to Annex 1): NA	
Problems found, solution applied: No problems	
Recommendation for repair, changes: NA	
Availability of the unit:	100 %
Proposal how to reach higher availability and better emission data quality: NA	
Training provided to plant personnel, proposal for settlement: Training recently provided	
Training requirements detected: NA	
Others:NA	

Dust reading after check			
Opacity or Extinction		Extinction	
Dust in mg / m <sup>3</sup>		mg/m <sup>3</sup>	0.0
Contamination		mA	0.00%
Dust reading, measured values		mA	0.0mg/m3
Control cycle, Zero point values		mA	4.0
Control cycle, Span value		mA	16.8

<b>Remarks</b> Analyser functioning correctly
Approved by customer Name: <div style="text-align: center;">Position:</div>

Sick (UK) Ltd  
Waldkirch House  
39 Hedley Road  
St.Albans  
Hertfordshire  
AL1 5BN




Tel: 0044 (0) 1727 831121  
Fax:0044 (0) 1727 855332  
E-Mail: info@sick.co.uk

Linearity Results Sheet			
Service Ref: SVON017322		Date: 12.02.2015	
		Customer Order No.: 0	
Customer:		Status	Chargable
		Device-type:	OMD41-2
		Serial Number :	98078000
Contact: James Beveridge		Filter Box Serial Number :	95278033

Location	Unit 7
----------	--------

#### 4 Results:

Component:		Particulate	Range			1.60	Extinction
	Expected Extinction	Expected %range	Measured Extinction	Measured %range	% Error of Range	Pass/Fail	
Filter 1	0.022	1.363%	0.021	1.333%	-0.03%	PASS	
Filter 2	0.049	3.056%	0.049	3.063%	0.01%	PASS	
Filter 3	0.083	5.181%	0.082	5.146%	-0.04%	PASS	
Filter 4	0.175	10.919%	0.174	10.875%	-0.04%	PASS	
Filter 5	0.270	16.856%	0.229	14.333%	-2.52%	PASS	
Filter 6	0.375	23.431%	0.373	23.333%	-0.10%	PASS	
Filter 7	0.809	50.550%	0.805	50.333%	-0.22%	PASS	
Filter 8	1.276	79.763%	1.251	78.208%	-1.55%	PASS	



Engineer's Comments : LINEARITY TEST WAS SUCCESSFUL



<b>SICK (UK)LTD</b> Waldkirch House, 39 Hedley Rd St Albans, Herts, AL1 5BN	<h1 style="color: blue; margin: 0;">SICK</h1>	Tel : 0044 (0)1727 831121 Fax : 0044 (0)1727 855332 E – Mail : info@sick.co.uk
<b>Maintenance Report OMD41</b>		
<input checked="" type="checkbox"/> Maintenance <span style="float: right;"><input type="checkbox"/> Repair</span>		
Customer: <u>RWE nPower</u>	Customer no: <u>FGD U7</u>	
Country: <u>UK</u>	Location: <u>Aberthaw, Barry</u>	
Receiver: <u>M221</u>	Serial no.: <u>98078003</u>	
Reflector: <u>R22</u>	Serial no.: <u>98078003</u>	
Eval.Unit: <u>A3</u>	Serial no.: <u>98078000</u>	
Plant operating status: <b>Instrument in workshop</b>		

	done	Remarks:
Check weekly/monthly/ quarterly log files: Log files ok?	<input checked="" type="checkbox"/>	NA
Training recommended:	<input checked="" type="checkbox"/>	Training provided

#### Dust reading before check

Opacity or Extinction	Extinction
Dust in mg / m <sup>3</sup>	mg/m <sup>3</sup> 0.0 mg/m <sup>3</sup>
Contamination	mA 0.0%
Dust reading, measured values	mA 4.0
Control cycle, Zero point values	mA 4.0
Control cycle, Span value	mA 16.80

#### 1. Visual Inspection

1.1 Check if measured values are plausible in the control room	<input checked="" type="checkbox"/>	Yes
1.2 Check if control cycle is (zero/span) valid (control room or recorder or evaluation unit)	<input checked="" type="checkbox"/>	QAL3 charts consistent
1.3 Check if there is any error message	<input checked="" type="checkbox"/>	No errors or warnings
1.4 Check for mechanical or electrical connection damage	<input checked="" type="checkbox"/>	No mechanical damage
1.5 Check mechanical mounting	<input checked="" type="checkbox"/>	All sound

#### 2. Purge Air Unit

2.1 Check the hoses and ring nozzles	<input checked="" type="checkbox"/>	All sound
2.2 Clean the pre-filter	<input checked="" type="checkbox"/>	NA
2.3 Replace the filter insert, part no. 5306091 for OMD41 SN > 97xxxxxxx, or part no.6700149 for OMD41 SN <96xxxxxxx	<input checked="" type="checkbox"/>	Purge air maintained internal
2.4 Check function of filter monitor purge switch.	<input checked="" type="checkbox"/>	Not fitted

#### 3. Fail Safe Shutter (FSS Optional)

3.1 <input type="checkbox"/> Check function of fail-safe shutter	<input checked="" type="checkbox"/>	Not fitted on stack
--	-------------------------------------	---------------------

#### 4. Optical unit

4.1 Clean housing	<input checked="" type="checkbox"/>	Housing cleaned
4.2 Check and, if necessary readjust optical axis	<input checked="" type="checkbox"/>	Not required
4.3 Clean optical boundary surfaces for sender/receiver and reflector, 2 x part no. 4003353	<input checked="" type="checkbox"/>	Both optical surfaces cleaned
4.4 Check and replace desiccant cartridge, part no. 2008475	<input checked="" type="checkbox"/>	Blue - OK

4.5.	Check the transmitter diode	<input checked="" type="checkbox"/>	Course = 132 Good
4.6	Check the stepper motor	<input checked="" type="checkbox"/>	All motors functioning correct
4.7	Check zero point, re-calibrate at smoke-free measurement point if it deviates	<input checked="" type="checkbox"/>	Not required
4.8	Check function of control cycle	<input checked="" type="checkbox"/>	Good
4.9	Linearity check	<input checked="" type="checkbox"/>	See below

Linearity Measurement	Set value	Actual Value	Deviation in %	< 2% OK?
Zero	4.00 mA	4.00 mA	0.0	OK
Span 2 (aprox. 30%)	14.11 mA	14.11 mA	0.0	OK
Span 1 (aprox. 70%)	16.8 mA	16.8 mA	0.0	OK
Span 3 (100%)	20.00 mA	20.00 mA	0.0	OK

## 5. Connection unit

Check the analogue and digital input and output signals	<input checked="" type="checkbox"/>	Analogue outputs +/- 0.1 mA
---	-------------------------------------	-----------------------------

## 6. Final check

6.1	Check function of complete measurement system	<input checked="" type="checkbox"/>	System functioning Well
6.2	Check the availability of the system	<input checked="" type="checkbox"/>	100%
6.3	Write device data into the report file	<input checked="" type="checkbox"/>	Device data saved
6.4	Complete Service Report and Measurement Log-file	<input checked="" type="checkbox"/>	Report complete

## 7. Report in writing to the plant manager as follows:

Quality of maintenance work by plant personnel: Good
Cleanliness of measuring station and devices: OK
Maintenance work done(in addition to Annex 1): NA
Problems found, solution applied: No problems
Recommendation for repair, changes: NA
Availability of the unit: 100 %
Proposal how to reach higher availability and better emission data quality: NA
Training provided to plant personnel, proposal for settlement: Training already provided
Training requirements detected: NA
Others:

## Dust reading after check

Opacity or Extinction	Extinction
Dust in mg / m <sup>3</sup>	mg/m <sup>3</sup> 0.0 mg/m <sup>3</sup>
Contamination	mA 0.000%
Dust reading, measured values	mA 0.0 mg/m <sup>3</sup>
Control cycle, Zero point values	mA 4.00
Control cycle, Span value	mA 16.8

## Remarks :

System functioning perfectly	
Date : 09.02.2015	Plant person: _____ Technician: Rhodri Jones



## Unit 7 - GM32 Commissioning

### Customer data

Customer: RWE N Power	Customer no: UNIT 7
Country: UK - Wales	City: Cardiff
Plant: Aberthaw	Location: U8

### 1 Device data

Device type: GM32	Device no: 1210606
Serial no: 1406 8012	
Process optic: GMP <input checked="" type="checkbox"/> GPP <input type="checkbox"/> CD <input type="checkbox"/>	
Serial no: 1406 8012	Device no: Type: In Situ

### 2 Plant data

Location:	Outside <input type="checkbox"/>	Under cover <input type="checkbox"/>	Inside <input checked="" type="checkbox"/>
Tag number			
Orientation of the stack	Horizontal <input type="checkbox"/>	Vertical <input checked="" type="checkbox"/>	Angle of 90 °
Orientation of the GM32	Horizontal <input checked="" type="checkbox"/>	Vertical <input type="checkbox"/>	Angle of 0 °
Length of probe / Flange-Flange (CD)	1500 mm	Active measurement distance / Aperture	500 mm
Zero path	NA mm	Differential pressure	1024 hpa
Ambient temperature	8 °C	Gas temperature	68 °C
I/O Modules on site	<input checked="" type="checkbox"/>	I/O Modules relocated	<input type="checkbox"/>
Plant operating status	Online - Base Load		

### 3 Prerequisite

	Y	N	Remarks
3.1 Documentation + Delivery complete	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Delivery Complete
3.2 Platform at measurement spot has suitable dimension?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Plenty of room
3.3 If this measurement location is under legal regulation, has it been acknowledged by an official body?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Yes, QAL2 to be carried out
3.4 Customer specific data for parameterization available?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Yes
3.5 Purge air unit installed and electrically connected?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Yes
3.6 Connection unit installed and electrically connected?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Yes
3.7 Zero point stands / tube available (only CD)?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	NA

4 Preliminary work		Y	N	Remarks
4.1	Mounting of flanges like described in the Operating Instruction?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Yes
4.2	Check for damage	<input checked="" type="checkbox"/>	<input type="checkbox"/>	No Damage
4.3	Check ambient conditions (ref. ch. 2)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Clean, warm & Dry
4.4	Check mounting conditions (ref. ch. 2)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	OK
4.5	Check mounting	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Good
4.6	Check cables / wires for correct installation	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Correct
4.7	Check main power supply voltage	<input checked="" type="checkbox"/>	<input type="checkbox"/>	240V

5 Purge Air Unit (only GMP / Cross-Duct)		Y	N	Remarks
5.1	Purge air unit type	X	X	SIEMENS
5.2	Check the rotation direction	<input checked="" type="checkbox"/>	<input type="checkbox"/>	+ve Pressure
5.3	Check hoses for correct installation	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Installed correctly
5.4	Purge air heating installed?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Leister air heater
5.5	Differential pressure monitor installed?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Yes

6 Sender / receiver unit		Y	N	Remarks
6.1	Clean all optical surfaces	<input checked="" type="checkbox"/>	<input type="checkbox"/>	All clean
6.2	Dessicant cartridges exchanged?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	Yes x 2
6.3	Check output of power supply voltages	<input checked="" type="checkbox"/>	<input type="checkbox"/>	See Below
6.3.1	26V±0,3V: 25.72V	<input checked="" type="checkbox"/>	<input type="checkbox"/>	OK
6.3.2	115V±1V: 113.2	<input checked="" type="checkbox"/>	<input type="checkbox"/>	OK
6.4	Check emission free zero point	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Reading zero for all components
6.5	Adjust purge air heating (option)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Value: position 9
6.6	Check and adjust differential pressure monitor (option)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	OK
6.7	Install the GM32 at the measurement spot	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Installed
6.8	Adjust optical alignment	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Adjusted to centre
6.9	Check signals	<input checked="" type="checkbox"/>	<input type="checkbox"/>	See Below
A: 275      B: 216				OK
C: 291      D: 217				OK
6.10	Note lamp data	<input checked="" type="checkbox"/>	<input type="checkbox"/>	100% (Zero hours)
Maximum lamp intensity: 18750				Exposure: 259 ms
				Lamp pulse: 120 mA
6.11	Note software revision	<input checked="" type="checkbox"/>	<input type="checkbox"/>	See below
Device Process: 9171698_X938_130708				Display: 9104869
Process optic: /				



6.12	Check stepper motor for proper function	<input checked="" type="checkbox"/>	<input type="checkbox"/>	All cycles perform well
6.13	Check parameterization	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Configured to site requirements
6.14	Parameterize the I/O Modules	<input checked="" type="checkbox"/>	<input type="checkbox"/>	OK
6.15	Perform loop test	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Analogue outputs & Ranging checked
6.16	Measured values are plausible (ref. Ch. 8)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Yes
6.17	Save device data	<input checked="" type="checkbox"/>	<input type="checkbox"/>	SOPAS project & Device File saved
6.18	Complete Commissioning Sign-Off Sheet	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Report complete
6.19	Instruct the operator personnel Hand over the maintenance manual and check lists - Measurement reading - Perform customer maintenance - Read messages	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Basic training provided

## 7 Input / Output / Ranges

### 7.1 Analog output / Display

Live zero: 4.0 mA			Output of control values: Y <input checked="" type="checkbox"/> / N <input type="checkbox"/>				
Analog output	Comp.	Source	Unit	Range		Control value	Remark
				Start	Ende		
1	SO2	MV01	ppm	0	1200	<input checked="" type="checkbox"/>	Checked to DCS
2	SO2	MV01	ppm	0	250	<input checked="" type="checkbox"/>	Checked to DCS
3	NO	MV02	ppm	0	1000	<input checked="" type="checkbox"/>	Checked to DCS
4	NO2	MV03	ppm	0	100	<input checked="" type="checkbox"/>	Checked to DCS
5	Temp	MV05	C	0	200	<input type="checkbox"/>	
6	Press.	MV06	hPa	900	1200	<input type="checkbox"/>	

### 7.2 Analog input

Analog input	Source	Unit	Live zero	Start	End
1	NA				
2					

### 7.3 Digital output

Digital output	Signal	Inv.
1	Failure	<input checked="" type="checkbox"/>
2	Maintenance Request	<input type="checkbox"/>
3	Not Measuring	<input type="checkbox"/>
4	Output Control Values	<input type="checkbox"/>
5	Uncertain	<input type="checkbox"/>
6	Extended	<input type="checkbox"/>
7	Purge air failure	<input type="checkbox"/>
8	No Function	<input type="checkbox"/>

### 7.4 Digital input


Digital input	Source	Inv.
1	Check Cycle	<input type="checkbox"/>
2	Maintenance	<input type="checkbox"/>
3	Output Control Values	<input type="checkbox"/>
4	Disable Check Cycle	<input type="checkbox"/>

8 Measurement values after commissioning					
	Unit	Range	Reading	Zero point	Span point
SO2	ppm	1200	18	0	840
NO	ppm	1000	399	0	700
SO2 low	ppm	250	18.3	0	175
NO2	ppm	100	3.8	0	70
Temp.			70	- X -	- X -
Press.			1013	- X -	- X -

#### Remarks

Analyser functioning correctly - Ok for live service

Date : 20 <sup>th</sup> February 2015		Name	
		Plant personnel: _____	
		Engineer: Rhodri Jones	




**SICK (UK) LTD**  
Waldkirch House,  
39 Hedley Rd  
St Albans,  
Herts,  
AL1 5BN

# SICK

Tel : 0044 (0)1727 831121  
Fax : 0044 (0)1727 855332  
E – Mail : info@sick.co.uk

## Unit 7 - GM32

### BS EN 14181 – Functional Test & Maintenance Report

Customer	RWN nPower
Site	Aberthaw
Contact	James Beverage
Service Ref No.	SVON017322
Survey Dates	20 <sup>th</sup> February 2015
Device Type	GM32 Probe version
Serial Number	1406 8012
Device Location	Unit 7
MCERTS Cert No	MC100163/01
Certification Details	See Certificate
Measurement Principle	UV Spectroscopy
SICK Engineer	Rhodri Jones
Signature	

Actual installation data			Remarks
Location	Outside <input type="checkbox"/>	Under cover <input type="checkbox"/>	Inside <input checked="" type="checkbox"/>
Ambient temperature	8	°C	
Gas temperature	80	°C	
Length of the probe / Flange – Flange (CD)	1500	mm	
Active measurement distance	500	mm	
Zero path	N/A	mm	

<b>Plant operating status:</b>	Online
--------------------------------	--------

Measurement values before check					
	Unit	Range	Reading	Zero point	Span point
SO <sub>2</sub> Hi	ppm	0-1200	73	-0.014	838.148
NO	ppm	0-1000	448	-0.001	698.457
NO <sub>2</sub>	ppm	0-100	-0.0	-0.057	69.846
SO <sub>2</sub> Lo	ppm	0-250	73.1	-0.014	174.614
Temp.	°C	0-200	66	- X -	- X -
Press.	hPa	900-1200	1001	- X -	- X -

Lamp data			
Max. lamp intensity	18500	Exposure:	259 ms
		Lamp pulse:	100 mA

1. Alignment & Cleanliness		
1.1 Visual Inspection of the analyser	<input checked="" type="checkbox"/>	Brand new analyser
1.2 Internal check of the analyser	<input checked="" type="checkbox"/>	Checked. No problems detected.
1.3 Cleanliness of optical components	<input checked="" type="checkbox"/>	New analyser - Clean
1.4 Flushing of air supply	<input checked="" type="checkbox"/>	Air supply purged
1.5 Obstructions in the optical path	<input checked="" type="checkbox"/>	No obstructions found
1.6 Check optical alignment of the Analyser	<input checked="" type="checkbox"/>	Checked & adjusted alignment.



<b>2. Sampling System</b>		
2.1	System in good condition and free from faults	<input checked="" type="checkbox"/> Yes
2.2	Visual Inspection of sampling probe	<input checked="" type="checkbox"/> New
2.3	Check of the Gas Conditioning System	<input checked="" type="checkbox"/> N/A
2.4	Check of the sample pump	<input checked="" type="checkbox"/> N/A
2.5	Check of all hoses, gas lines and connections	<input checked="" type="checkbox"/> Checked analyser & air hoses.
2.6	Check of power supplies	<input checked="" type="checkbox"/> Checked. OK
2.7	Check of all filters	<input checked="" type="checkbox"/> Replaced purge air filter
2.8	Check of heated lines are operating correctly	<input checked="" type="checkbox"/> N/A
2.9	Efficiency of NOx converter > 95% (if applicable)	<input checked="" type="checkbox"/> N/A
2.10	Has the response time been checked by applying gas to probe	<input checked="" type="checkbox"/> N/A

<b>3. Serviceability</b>		
3.1	Safe & Clean working environment for servicing	<input checked="" type="checkbox"/> Working environment is ok
3.2	Adequate supply of reference gases	<input checked="" type="checkbox"/> N/A
3.3	Adequate supply of spare parts	<input checked="" type="checkbox"/> Spare analyser available
3.4	Ability to apply gas at probe	<input checked="" type="checkbox"/> N/A
3.5	Output signals checked	<input checked="" type="checkbox"/> Analogue outputs checked

<b>4. Zero Span &amp; Leak Checks</b>		
4.1	Leak check of whole system carried out	<input checked="" type="checkbox"/> N/A
4.2	Have span checks been carried out at suitable levels	<input checked="" type="checkbox"/> 70% of range
4.3	Have zero checks been carried out at suitable levels	<input checked="" type="checkbox"/> At Zero

<b>5. Linearity</b>		
5.1	Checks at 0, 20%, 40%, 60% & 80% of 2 X Daily ELV's	<input checked="" type="checkbox"/> Refer to linearity report
5.2	Different concentrations applied in a random sequence	<input checked="" type="checkbox"/> Refer to linearity report
5.3	Has the response time for each component been recorded (T90)	<input checked="" type="checkbox"/> Refer to linearity report
5.4	Have the first readings been taken at 3 x T90	<input checked="" type="checkbox"/> Refer to linearity report
5.5	Have three readings been taken at least 4 x T90 apart	<input checked="" type="checkbox"/> Refer to linearity report
5.6	Are the gases used traceable to ISO 17025 standards	<input checked="" type="checkbox"/> Refer to linearity report

<b>6. Visual inspection</b>		
6.1 Check measured values are plausible in control room	<input checked="" type="checkbox"/>	Checked. All ok
6.2 Check if control cycle is (zero/span) valid	<input checked="" type="checkbox"/>	Checked. Valid zero and span
6.3 Check if messages are active	<input checked="" type="checkbox"/>	Checked
Failure:	No	
Maintenance Req:	No	
Uncertain:	No	
6.4 Check for mechanical damages	<input checked="" type="checkbox"/>	No mechanical damage detected
6.5 Check attachments and connections	<input checked="" type="checkbox"/>	Checked. All ok

<b>7. Purge air unit (only GMP and Cross-Duct)</b>		
7.3 Check the hoses and clamps for hose	<input checked="" type="checkbox"/>	Checked hoses & clamps. All ok
7.4 Exchange purge air filter (order no 5306091)	<input checked="" type="checkbox"/>	Replaced purge air filter
7.5 Clean pre-filter housing	<input checked="" type="checkbox"/>	Checked and cleaned
7.6 Check noise of blower	<input checked="" type="checkbox"/>	Checked. OK
7.7 Check function of purge air heating and setting (option)	<input checked="" type="checkbox"/>	Checked. Operating correctly
7.8 Check function of differential pressure monitor (option)	<input checked="" type="checkbox"/>	Checked. Operating correctly

<b>8. Probe (GMP or GPP)</b>		
8.3 Clean probe	<input checked="" type="checkbox"/>	Brand new probe
8.4 Check Probe for corrosion	<input checked="" type="checkbox"/>	New probe
8.5 Check probe for damage and functionality	<input checked="" type="checkbox"/>	perfect

<b>9. Connection unit</b>		
9.3 Clean housing	<input checked="" type="checkbox"/>	Cleaned housing
9.4 Check power supply	<input checked="" type="checkbox"/>	Checked Power supply
26V±0,3V: 25.90      115V±1V: 114.0		
9.5 Check analogue input / output (option)	<input checked="" type="checkbox"/>	Checked. OK
9.6 Check digital input / output (option)	<input checked="" type="checkbox"/>	Checked. OK
9.7 Check parameterization of the I/O-Module	<input checked="" type="checkbox"/>	Checked. OK



<b>10. Sender / receiver unit (reflector)</b>		
10.3	Check drying agent cartridges, exchange if req.	<input checked="" type="checkbox"/> both dessicants replaced
10.4	Clean all optical surfaces.	<input checked="" type="checkbox"/> Cleaned all optical surfaces
10.5	Check stepper motor for proper function	<input checked="" type="checkbox"/> Checked Motor. OK
10.6	Check fans	<input checked="" type="checkbox"/> Fans are operating correctly.
10.7	Check spectrometer signals	<input checked="" type="checkbox"/> Checked spectrometer signal. All ok.
10.8	Check lamp	<input checked="" type="checkbox"/> New Lamp
10.9	Check emission free zero point	<input checked="" type="checkbox"/> ChReading zeroeeked.
10.10	Check the alignment adjust	<input checked="" type="checkbox"/> Checked & adjusted alignment
10.11	Check signal strength	<input checked="" type="checkbox"/> New Lamp - 100%
A: 270	B: 315	Good values
C: 312	D: 339	Good values
10.12	Check mirror steps	<input checked="" type="checkbox"/> OK
Dx: -0.053	Dy: -0.015	OK
10.13	Check parameterization	<input checked="" type="checkbox"/> As required

<b>11. Final check</b>		
11.3	Check function of complete measurement system	<input checked="" type="checkbox"/> Good
11.4	Check the availability of the system	<input checked="" type="checkbox"/> 100%
11.5	Save device data external	<input checked="" type="checkbox"/> Saved Data
11.6	Complete service report	<input checked="" type="checkbox"/> Service report complete


<b>12. Additional comments</b>	
Quality of maintenance work by plant personnel: Good	
Cleanliness of measuring station and devices: OK	
Maintenance work done (in addition to Annex 1): NA	
Problems found, solution applied: No problems	
Recommendation for repair, changes NA	
Availability of the unit:	100%
Proposal how to reach higher availability and better emission data quality:	NA
Training provided to plant personnel, proposal for settlement: TRaining Reccomended	
Training requirements detected: NA	

<b>13. Measurement values after check</b>					
	Unit	Range	Reading	Zero point	Span point
SO <sub>2</sub> Hi	ppm	0-1200	32	0.2	837.9
NO	ppm	0-100	412	-0.0	698.2
NO <sub>2</sub>	ppm	0-100	0.5	-0.1	69.8
SO <sub>2</sub> Lo	ppm	0-250	31.9	0.2	174.6
Temp.	°C	0-200	65	- X -	- X -
Press.	hPa	900-1200	1001	- X -	- X -

<b>Remarks</b>
Analyser in very good working order - No faults or warnings
Approved by customer Name:
Position:

# SICK

## Unit 7 - GM32 Linearity Report

Customer	RWE Npower
Site	Aberthaw Power Station
Permit Ref	A1234
Contact	James Beveridge
Service Ref No.	SVON017322
Survey Date	20.02.2015
Device Type	GM32
Device Location	Unit 7
Serial Number	1406 8012
MCERTS Cert No	MC100163/01
Certification Details	See Certificate
Measurement Principle	U.V DOAS
Gas Box Serial Number	0816 8035
SICK Engineer	Rhodri Jones
Signature	

Sick (UK) Ltd  
Waldkirch House  
39 Hedley Road  
St Albans  
Hertfordshire  
AL1 5BN



Tel: 0044 (0) 1727 831121  
Fax: 0044 (0) 1727 855332  
E-Mail: info@sick.co.uk

## GM32 SO<sub>2</sub> Linearity

Customer	RWE Npower
Site	Aberthaw Power Station
Product	GM32
Analysers Serial Number	1406 8012
MCERTS Certificate Number	MC10016301
Location	Unit 7
Active Path Length (mm)	500
Gas Box Serial Number	0816 8035
Calibration Date	01/12/2014
Calibration Gas Cylinder Ref:	VC3713C
Test Gas Conc. (ppm)	3752
Test Gas Conc. (mg.m <sup>-3</sup> )	10693
Gas Expiry Date	19.01.2018
Tolerance	2%


Site Permit Ref	A1234
SVON No	SVON017322
Order Number	AXC1235784133
QAL2/AST/Routine	QAL2
Date	16.02.2015
Engineer	Rhodri Jones
Response Time (T <sub>90</sub> ) Seconds	30
Permit Daily ELV for SO <sub>2</sub> (mg.m <sup>-3</sup> )	440
Permit Daily ELV x 2 (mg.m <sup>-3</sup> ) (C <sub>ELV</sub> )	880
Analysers Range	1200
Ambient Pressure (hPa)	1024
Results Visible on DCS?	Yes, when not in maintenance mode
Corrections Applied?	None - As measured (ppm)

Component 1		Linearity Results									
SO <sub>2</sub>		% 2 x ELV	Test Cell Length (mm)	Expected (ppm)	Reading 1 (ppm)	Reading 2 (ppm)	Reading 3 (ppm)	Average Reading	DCS	Time	
Set Pt Time (4xT <sub>90</sub> ) sec	120	244%	100	750.40	736.60	734.20	735.00	735.27			
Log Time (3xT <sub>90</sub> ) sec	90	183%	75	562.80	551.00	563.60	561.20	558.60			
Gasbox Temperature (°C)	80	305%	125	938.00	926.40	915.70	918.00	920.03			
Reference Temperature (K)	273	61%	25	187.60	193.10	192.20	193.50	192.97			
ppm to mg.m <sup>-3</sup> for SO <sub>2</sub>	2.85	122%	50	375.20	361.10	367.70	362.50	363.77			
		366%	150	1125.60	1090.00	1081.00	1090.00	1087.00			

Zero Results			
ZERO	Analysers Reading (ppm)	DCS (ppm)	Time
Zero Reading 1	0.00		09:38
Zero Reading 2	0.20		09:42
Zero Reading 3	1.00		09:46
Zero Reading 4	1.00		11:53
Zero Reading 5	0.00		11:55
Zero Reading 6	0.00		11:58
AVERAGE	0.37		

Test Cell Length (mm)	Actual Value (ppm)	% Range	% 2 x ELV	Average CEM Reading (ppm)	d <sub>C,rel</sub>	d <sub>C,rel</sub> <5% (EN 14181)
NA	0.00	0.0%	0%	0.37	-0.46%	PASS
25mm	187.60	15.6%	61%	192.97	0.74%	PASS
50mm	375.20	31.3%	122%	363.77	-0.54%	PASS
75mm	562.80	46.9%	183%	558.60	0.91%	PASS
100mm	750.40	62.5%	244%	735.27	0.30%	PASS
125mm	938.00	78.2%	305%	920.03	0.61%	PASS
150mm	1125.60	93.8%	366%	1087.00	-1.11%	PASS

Applied Concentration (ppm) v CEM Reading (ppm)



Engineer	Rhodri Jones
Signature	
Comments	Analysers exhibiting a linear response

Version 2.0, 07/08/2014 - Uncontrolled when printed



Sick (UK) Ltd  
Waldkirch House  
39 Hedley Road  
St Albans  
Hertfordshire  
AL1 5BN



Tel: 0044 (0) 1727 83111  
Fax: 0044 (0) 1727 85533  
E-Mail: info@sick.co.uk

### GM32 NO Linearity

Customer	RWE Npower
Site	Aberthaw Power Station
Product	GM32
Analysers Serial Number	1406 8012
MCERTS Certificate Number	MC100163/01
Location	Unit 7
Active Path Length (mm)	500

Site Permit Ref	A1234
SVON No	SVON017322
Order Number	AXC123784133
QAL2/AST/Routine	QAL2
Date	20.02.2015
Engineer	Rhodri Jones

Gas Box Serial Number	0816 8035
Calibration Date	01/12/2014
Calibration Gas Cylinder Ref:	VCSMG3285
Test Gas Conc. (ppm)	6817
Test Gas Conc. (mg.m <sup>-3</sup> )	9135
Gas Expiry Date	30.01.2018
Tolerance	2%

Response Time (T <sub>90</sub> ) Seconds	20
Permit Daily ELV for NO (mg.m <sup>-3</sup> )	1210
Permit Daily ELV x 2 (mg.m <sup>-3</sup> ) (C <sub>EL</sub> )	2420
Analysers Range	1000
Ambient Pressure (hPa)	1024
Results Visible on DCS?	Yes, when not in maintenance mode
Corrections Applied?	None - As measured (ppm)


Component 1	NO
Set Pt Time (4xT <sub>90</sub> ) sec	80
Log Time (3xT <sub>90</sub> ) sec	60
Gasbox Temperature (°C)	80
Reference Temperature (K)	273
ppm to mg.m <sup>-3</sup> for NO	1.34

Linearity Results								
% 2 x ELV	Test Cell Length (mm)	Expected (ppm)	Reading 1 (ppm)	Reading 2 (ppm)	Reading 3 (ppm)	Average Reading	DCS	DCS Time
75%	100	1363.40	1425.00	1425.00	1422.00	1424.00		15:01 - 15:04
57%	75	1022.55	1027.00	1026.00	1031.00	1028.00		15:17 - 15:20
19%	25	340.85	359.00	346.00	345.00	350.00		15:11 - 15:15
38%	50	681.70	679.00	679.00	680.00	679.33		15:07 - 15:10

Zero Results			
ZERO	Analysers Reading (ppm)	DCS (ppm)	Time
Zero Reading 1	1.00		11:50
Zero Reading 2	0.00		11:55
Zero Reading 3	1.00		12:00
Zero Reading 4	0.00		12:40
Zero Reading 5	0.00		12:45
Zero Reading 6	0.00		12:50
AVERAGE	0.33		

Test Cell Length (mm)	Actual Value (ppm)	% Range	% 2 x ELV	Average CEM Reading (ppm)	d <sub>c,rel</sub>	d <sub>c,rel</sub> <5% (EN 14181)
NA	0.00	0.0%	0%	0.33	0.23%	PASS
25mm	340.85	34.1%	14%	350.00	0.16%	PASS
50mm	681.70	68.2%	28%	679.33	-0.75%	PASS
75mm	1022.55	102.3%	42%	1028.00	-0.86%	PASS
100mm	1363.40	136.3%	56%	1424.00	0.98%	PASS

Applied Concentration (ppm) v CEM Reading (ppm)



Engineer	Rhodri Jones
Signature	
Comments	Analysers exhibiting a linear response

Version 2.0, 07/08/2014 - Uncontrolled when printed

Sick (UK) Ltd  
Waldkirch House  
39 Hedley Road  
St Albans  
Hertfordshire  
AL1 5BN



Tel: 0044 (0) 1727 831121  
Fax: 0044 (0) 1727 855332  
E-Mail: info@sick.co.uk

### GM32 NO<sub>2</sub> Linearity

Customer	RWE Npower
Site	Aberthaw Power Station
Product	GM32
Analysers Serial Number	1406 8012
MCERTS Certificate Number	MC100163/01
Location	Unit 7
Active Path Length (mm)	500

Site Permit Ref	A1234
SVON No	SVON017322
Order Number	AXC1235784133
QAL2/AST/Routine	AST
Date	20.02.2015
Engineer	Rhodri Jones

Gas Box Serial Number	0816 8035
Calibration Date	01/12/2014
Calibration Gas Cylinder Ref:	VCD116712
Test Gas Conc. (ppm)	986
Test Gas Conc. (mg.m <sup>-3</sup> )	2021
Gas Expiry Date	30.01.2018
Tolerance	2%

Response Time (T <sub>90</sub> ) Seconds	50
Permit Daily ELV for NO <sub>2</sub> (mg.m <sup>-3</sup> )	121.0
Permit Daily ELV x 2 (mg.m <sup>-3</sup> ) (C <sub>90</sub> )	242
Analysers Range	100
Ambient Pressure (hPa)	970
Results Visible on DCS?	Yes, when not in maintenance mode
Corrections Applied?	None - As measured (ppm)


Component 1	NO <sub>2</sub>
Set Pt Time (4xT <sub>90</sub> ) sec	200
Log Time (3xT <sub>90</sub> ) sec	150
Gasbox Temperature (°C)	80
Reference Temperature (K)	273
ppm to mg.m <sup>-3</sup> for NO <sub>2</sub>	2.05

Linearity Results								
% 2 x ELV	Test Cell Length (mm)	Expected (ppm)	Reading 1 (ppm)	Reading 2 (ppm)	Reading 3 (ppm)	Average Reading	DCS	Time
167%	100	197.20	192.60	207.90	212.30	204.27		15:26 - 15:29
125%	75	147.90	145.10	146.40	146.30	145.93		15:35 - 15:39
42%	25	49.30	45.70	45.30	46.70	45.90		15:30 - 15:34
84%	50	98.60	91.30	91.30	93.00	91.87		15:21 - 15:25

Zero Results			
ZERO	Analysers Reading (ppm)	DCS (ppm)	Time
Zero Reading 1	-0.10		09:35
Zero Reading 2	-0.70		09:38
Zero Reading 3	0.10		10:32
Zero Reading 4	-0.30		10:38
Zero Reading 5	2.00		
Zero Reading 6	-1.50		
AVERAGE	-0.08		

Test Cell Length (mm)	Actual Value (ppm)	% Range	% 2 x ELV	Average CEM Reading (ppm)	d <sub>C,rel</sub>	dc <sub>C,rel</sub> <5% (EN 14181)
NA	0.00	0.0%	0%	-0.08	1.06%	PASS
25mm	49.30	49.3%	20%	45.80	-0.75%	PASS
50mm	98.60	98.6%	41%	91.87	-2.57%	PASS
75mm	147.90	147.9%	61%	145.93	-1.04%	PASS
100mm	197.20	197.2%	81%	204.27	2.25%	PASS

Applied Concentration (ppm) v CEM Reading (ppm)



Engineer	Rhodri Jones
Signature	
Comments	

Version 2.0, 07/08/2014 - Uncontrolled when printed

**Table 5.2 - Functional Tests carried out by RPS**

Requirement	Compliance	Notes
<p>10 Documentation of Records</p> <p>The following documentation shall be controlled, readily accessible and up to date:</p> <ul style="list-style-type: none"> <li>- a plan of the CEMS;</li> </ul>	Yes	A plan of the AMS resides in the Electrical Control & Instrumentation section offices and at the analyser.
<ul style="list-style-type: none"> <li>- all manuals (maintenance, users, etc.);</li> </ul>	Yes	.
<ul style="list-style-type: none"> <li>- log books to document possible malfunctions and action taken;</li> </ul>	Yes	
<ul style="list-style-type: none"> <li>- service reports;</li> </ul>	Yes	All service reports are filed.
<ul style="list-style-type: none"> <li>- QAL 3 documentation including actions taken as a result of out of control situations</li> </ul>	Yes	There is a well established QAL3 system in place and analyser drift is routinely monitored using shewart charts. Evidence was provided to RPS.
<p>Management system procedures for maintenance, calibration and training;</p>	Yes	The Performance dept manage and review maintenance and calibration and training carried out by the EC&I team.
<p>Training records/certificates</p>	Yes	Evidence was provided to RPS



Requirement	Compliance	Notes
Maintenance schedules.	Yes	Maintenance scheduled by Performance.
Auditing Plans & Records – Evidence that the operator includes procedures for the management of the CEMS within the auditing cycle of the management system.	Yes	Carried out using dedicated Technical Team from Swindon Head Office. An auditor witnessed the parallel tests, which suggest a good level of organisation.

## **APPENDIX 1: SRM Calibration Data**

**Table A1.1– SRM On-Site Calibrations**

Sample Date	Equipment Name	Equipment ID Number	Span Gas Type	ID Number	Span Gas Concentration	Pre-Sampling Result*		Post-Sampling Result*	
						Zero	Span	Zero	Span
5 <sup>th</sup> – 6 <sup>th</sup> March 2015	Horiba PG 250	FYS403	NO (17025)	183469	911 ppm	1ppm	911ppm	1ppm	912ppm
			O <sub>2</sub> (V.U.)	221743	15.04 %	0.07 %	15.24 %	0.05%	15.02%
	Gasmet DX4000	1610	SO <sub>2</sub> (17025)	216990	54.98 ppm	-0.65ppm	22.16ppm	0.07ppm	58.98 <sup>#</sup> ppm
6 <sup>th</sup> March 2015	Horiba PG 250	FYS403	NO (17025)	183469	911 ppm	1ppm	912ppm	3ppm	914ppm
			O <sub>2</sub> (V.U.)	221743	15.04 %	0.05%	15.02%	0.04%	14.94%
	Gasmet DX4000	1610	SO <sub>2</sub> (17025)	216990	54.98 ppm	0.07ppm	58.98 <sup>#</sup> ppm	0.06ppm	55.3ppm
9 <sup>th</sup> – 10 <sup>th</sup> March 2015	Horiba PG 250	FYS403	NO (17025)	183469	911 ppm	0ppm	913ppm	2ppm	912ppm
			O <sub>2</sub> (V.U.)	221743	15.04 %	0.04 %	14.98 %	0.02 %	15.16 %
	Gasmet DX4000	1610	SO <sub>2</sub> (17025)	216990	54.98 ppm	-1.00 ppm	54.85 ppm	-0.85 ppm	55.2 ppm

**Notes**

- \*- Calibration values are those for the entire sample system.
- - Zero gas 99.999% N<sub>2</sub>

## **APPENDIX 2 – Accreditation Schedule**

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> TESTING 1709</p> <p>Accredited to ISO/IEC 17025:2005</p>	<b>The Environmental Consultancy Ltd</b> trading as <b>RPS Consultants</b>		
	<b>Issue No: 060    Issue date: 11 December 2014</b>		
	<table> <tr> <td data-bbox="491 577 794 667"> <b>14 Cornhill</b> <b>London</b> <b>EC3V 3ND</b> </td><td data-bbox="847 577 1118 725"> <b>Contact: Mr S Hurst</b> <b>Tel: +44 (0) 20 7280 3200</b> <b>Fax: +44 (0) 20 7283 9248</b> <b>E-Mail: hursts@rpsgroup.com</b> <b>Website: www.rpsgroup.com</b> </td></tr> </table>	<b>14 Cornhill</b> <b>London</b> <b>EC3V 3ND</b>	<b>Contact: Mr S Hurst</b> <b>Tel: +44 (0) 20 7280 3200</b> <b>Fax: +44 (0) 20 7283 9248</b> <b>E-Mail: hursts@rpsgroup.com</b> <b>Website: www.rpsgroup.com</b>
<b>14 Cornhill</b> <b>London</b> <b>EC3V 3ND</b>	<b>Contact: Mr S Hurst</b> <b>Tel: +44 (0) 20 7280 3200</b> <b>Fax: +44 (0) 20 7283 9248</b> <b>E-Mail: hursts@rpsgroup.com</b> <b>Website: www.rpsgroup.com</b>		
Testing performed by the Organisation at the locations specified below			

**Locations covered by the organisation and their relevant activities**

**Laboratory locations:**

Location details		Activity	Location code
<b>Address</b> Unit A1 Lowfields Business Park Elland HX 5 9DE	<b>Local contact</b> Mr B Durden  Tel: +44 (0)1422 324180 Fax: +44 (0)1422 324188 Email: rpslu@rpsgroup.com	Health and Hygiene	A
<b>Address</b> Tem Place Denmore Road Bridge of Don Aberdeen AB23 8JX	<b>Local contact</b> Mr I Edmans Tel: +44 (0)845 60 123 88 Fax: +44 (0)1224 706 998 Email: rpsasb@rpsgroup.com	Health and Hygiene	B
<b>Address</b> 14 Cornhill London EC3V 3ND	<b>Local contact</b> Mr D Blyton  Tel: +44 (0)20 7280 3200 Fax: +44 (0) 20 7283 9248 Email: rpslo@rpsgroup.com	Support Functions: Quality Management, including contract review, document control, auditing and quality control	D
<b>Address</b> Noble House Capital Drive Linford Wood Milton Keynes MK14 6QP	<b>Local contact</b> Mr B Durden  Tel: +44 (0)1235 437 100 Fax: +44 (0)1908 669899 Email: rpsmk@rpsgroup.com	Health and Hygiene	E
<b>Address</b> Suite 4C Rhodes Business Park Silburn Way Middleton Manchester M24 4NE	<b>Local contact</b> Mr B Durden  Tel: +44 (0) 161 6549069 Fax: +44 (0)161 6436495 Email: rpswn@rpsgroup.com	Health and Hygiene	F

 <p>1709 Accredited to ISO/IEC 17025:2005</p>	<p align="center"><b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK</p>
	<p align="center"><b>The Environmental Consultancy Ltd</b> trading as RPS Consultants <b>Issue No: 060    Issue date: 11 December 2014</b></p>
<p align="center">Testing performed by the Organisation at the locations specified</p>	

**Site activities performed away from the locations listed above:**


Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene Environmental Sampling and Testing	Elland - H Aberdeen - I London - K Milton Keynes - L Middleton - M
Customers sites requiring Stack Emissions Testing	Stack emissions Testing	Elland - H Milton Keynes - L
Mobile Laboratories	Health and Hygiene	London - J



 <p>1709 Accredited to ISO/IEC 17025:2005</p>	<p align="center"><b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK</p> <p align="center"><b>The Environmental Consultancy Ltd</b> trading as RPS Consultants</p> <p align="center"><b>Issue No: 060    Issue date: 11 December 2014</b></p>
Testing performed by the Organisation at the locations specified	


DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)	
	Sampling of air for fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	I, K, L, M
	Fibre counting	HSG 248:February 2005 (Documented In-House Procedure)	B, E, F, I, J, K, L, M
	4 Stage Clearance Process	HSG 248:February 2005 (Documented In-House Procedure)	I, K,L, M
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for asbestos identification	HSG 248:February 2005 (Documented In-House Procedure)	I, K, L, M
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	HSG 248:February 2005 (Documented In-House Procedure using stereo-microscopy, polarised light microscopy and dispersion staining)	B, F
Testing of Stack Emissions to Atmosphere	<u>Sampling with subsequent analysis by an ISO/IEC 17025 accredited laboratory</u>	National, International and other recognised standards using documented In-House work instructions to meet the requirements of DD CEN/TS 15675:2007/ BS EN 15259:2007	
	Gaseous Organic Compounds - sorbent tube method	USEPA Method 18 (RPSCE/1/19a)	L
	Total Particulate Matter (20 to 1000 mg/m <sup>3</sup> )	BS ISO 9096:2003 (RPSCE/1/7/d)	H, L


 1709 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants <b>Issue No: 060    Issue date: 11 December 2014</b>		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling with subsequent analysis by an ISO/IEC 17025 accredited laboratory (cont'd)</u>	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007	
	Total Particulate Matter	BS EN 13284-1:2002 (RPSCE/1/7c)	H, L
	Hydrogen Chloride	BS EN 1911:2010 (RPSCE/1/8b)	H, L
	Halides and Halogens: Hydrogen Bromide Chlorine Bromine	US EPA Method 26 and 26A (RPSCE/1/8a)	H, L
	Sulphur dioxide	BS EN 14791:2005 (RPSCE/1/23)	H, L
	Hydrogen Fluoride	BS ISO 15713:2006 (RPSCE/1/8c)	H, L
	Mercury	BS EN 13211:2002 (RPSCE/1/9b)	H, L
	Metals	BS EN 14385:2004 (RPSCE/1/9c)	H, L
	Dioxins and furans	BS EN 1948-1:2006 (RPSCE/1/10b)	H, L
	Dioxin-like Polychlorinated Biphenyls (PCBs)	BS EN 1948-4:2010 (RPSCE/1/10b)	H, L
	Polycyclic Aromatic hydrocarbons (PAH's)	BS ISO 11338-1:2003 (RPSCE/1/10c)	H, L

Assessment Manager: LS3

Page 4 of 8


 1709 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants <b>Issue No: 060    Issue date: 11 December 2014</b>		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling with subsequent analysis by an ISO/IEC 17025 accredited laboratory (cont'd)</u>	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007 (cont'd)	
	Formaldehyde	US EPA Method 316 (RPSCE/1/22)	H, L
	Speciated VOCs (carbon and other suitable tubes) (Dry stacks only): Aliphatic VOCs Aromatic VOCs Aliphatic amines Aromatic amines Cresols Phenols Acetic acid	BS EN 13649:2002 (RPSCE/1/19b)	H, L
	Amines (Total aromatic and aliphatic)	BS EN 13649:2002, NIOSH method 2010 + 2002 (RPSCE/1/19c)	H, L
	Isocyanates	USEPA Method 207-1 (documented in-house method RPSCE/1/18C)	H, L
	Isocyanates	USEPA CTM 036 (documented in-house method RPSCE/1/18D Rev A)	H, L
	Hydrogen cyanide	US EPA OTM 29 (RPSCE/1/16a)	L
	Hydrogen sulphide	US EPA Method 11 (RPSCE/1/17)	H, L
	Ammonia	BS EN 14791:2005 (RPSCE/1/14b)	H, L




 1709 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants <b>Issue No:</b> 060 <b>Issue date:</b> 11 December 2014		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling and On-Site Analysis</u>	National, European, International and Environment Agency specified standards including MID's and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007	
	Water Vapour	BS EN 14790:2005 (RPSCE/1/12b)	H, L
	<u>Sampling and On-Line Analysis</u>		
	Pressure, Temperature and Velocity (Point Velocity Method)	BS EN 16911-1:2013 (RPSCE 1/2 – Differential Pressure Device (Pitot Tube) Method	H, L
	Pressure, Temperature and Velocity	BS EN 13284-1:2002 BS ISO 9096:2003 (RPSCE/1/2)	H, L
	Water Vapour*	EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)	H, L
	Carbon Monoxide*	BS EN 15058:2006 (RPSCE/1/21h - NDIR analyser) EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)	H, L
	Carbon Dioxide*	ISO 12039:2001 (RPSCE/1/21e - NDIR analyser) EA TGN M22 (RPSCE/1/24 - FTIR analyser)	H, L
	Nitrogen Monoxide (NO)*	BS EN 14792:2005 (RPSCE/1/21f - Chemiluminescence analyser) EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)	H, L

Assessment Manager: LS3

Page 6 of 8

 1709 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants <b>Issue No: 060    Issue date: 11 December 2014</b>		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling and On-Line Analysis</u> (cont'd)	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007 (cont'd)	
	Nitrogen Dioxide (NO <sub>2</sub> )*	EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)	H, L
	Oxides of Nitrogen (NO <sub>x</sub> )*	BS EN 14792:2005 (RPSCE/1/21f - Chemiluminescence analyser) EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)	H, L
	Nitrous Oxide (N <sub>2</sub> O)*	EA TGN M22 (RPSCE/1/24 Rev B - Validated FTIR analyser)	H, L
	Sulphur dioxide*	EA TGN M22 (RPSCE/1/24 - FTIR analyser)	H, L
	Oxygen*	BS EN 14789:2005 (RPSCE/1/21g - validated Zirconium cell analyser)	H, L
	Total Gaseous Organic Carbon* (TOC/VOC) (0 to 1000 mg/m <sup>3</sup> )	BS EN 12619:2013 (RPSCE/1/4b - FID analyser)	H, L

 1709 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants <b>Issue No: 060    Issue date: 11 December 2014</b>		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Stack Emissions - Continuous Emissions Monitoring Systems (CEMS)	QAL 2 and the Annual Surveillance Test (AST) for CEMS	Documented in house procedure RPSCE/1/25 to meet the requirements of BS EN 14181:2004, Environment Agency MID 14181 and other requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007	H, L
END			

\* - The scale range of the analyser used for this test must be that detailed on its current MCERTS certificate or a range validated by the organisation to meet MCERTS requirements.