

**BS EN 14181 Report**

Permit Number: **RP3133LD**  
Operator: **RWE Generation UK plc.**  
Installation: **Aberthaw Power Station**  
Installation Type: **Coal-Fired Power Station**  
Emission Point: **Unit 8**  
Monitoring Dates: **26<sup>th</sup> October – 2<sup>nd</sup> November 2016**



Contract Reference: **FTBS 29312**

Operator: **RWE Generation UK plc.**

Address: **Aberthaw Power Plant  
The Leys  
Aberthaw, Nr Barry  
South Glamorgan  
CF62 4ZW**

Client Contact: **Richard Kadim**

Monitoring Organisation: **RPS Consultants**

Address: **Noble House, Capital Drive, Linford Wood,  
Milton Keynes. MK14 6QP.**

Report Date: **9<sup>th</sup> January 2017**

Report Author: **Glyn Harrison**

Report Approved By: **Richard Harvey**

Position: **Principal Consultant**

MCERTS Qualifications: **Level 2, Technical Endorsements 1, 2, 3 & 4**

MCERTS Registration No.: **MM 02 020**

Signature:

A handwritten signature in black ink, appearing to read 'Glyn Harrison'.

**RPS Consultants has produced this report within the term of the contract with the client and taking account of the resources devoted to it by agreement with the client.**

**We disclaim any responsibility to the client and others in respect of any matters outside the scope of the above. This report is confidential to the client and we accept no responsibility of whatsoever nature to third parties to whom this report, or any part thereof, is made known. Any such party relies on the report at their own risk.**

## CONTENTS

SECTION 1 – Executive Summary	3
SECTION 2 – Information about the regulated installation	6
SECTION 3 – Information about the monitoring campaign	12
SECTION 4A1 – Data & calculations – QAL2 Procal 1	15
SECTION 4A2 – Data & calculations – QAL2 Procal 2	49
SECTION 4B1 – Data & calculations – AST SICK OMD 41	84
SECTION 5 – Functional Test Results	92
APPENDIX 1 – SRM on site calibrations	127
APPENDIX 2 – Accreditation Schedule	129

## Section 1 – Executive Summary

### 1B.1 Result Summary – QAL2

EN 14181 Test Type		QAL2				
Stack designation		Unit 8				
Measurand	Correlation coefficient of parallel data ( $R^2$ )	Derived Calibration function ( $y_i$ )		Calibrated Range	Extrapolated Calibrated Range	Variability Test
		$y_i =$	$a + bx_i$			
Nitric Oxide as total NOx (Procal 1)	0.9652	2.835 <sup>1</sup>	0.94 <sup>1</sup>	0 – 869 mg/m <sup>3</sup> <sup>2</sup>	0 – 1869 mg/m <sup>3</sup> <sup>3</sup>	Pass
Sulphur Dioxide (Procal 1 – Low Range)	0.9562	-1.61 <sup>1</sup>	0.96 <sup>1</sup>	0 – 239 mg/m <sup>3</sup> <sup>2</sup>	0 – 639 mg/m <sup>3</sup> <sup>3</sup>	Pass
Sulphur Dioxide (Procal 1 – High Range)	0.9565	-1.69 <sup>1</sup>	1.01 <sup>1</sup>	0 – 240 mg/m <sup>3</sup> <sup>2</sup>	0 – 642 mg/m <sup>3</sup> <sup>3</sup>	Pass
Nitric Oxide as total NOx (Procal 2)	0.8555	-2.66 <sup>1</sup>	1.14 <sup>1</sup>	0 – 879 mg/m <sup>3</sup> <sup>2</sup>	0 – 1817 mg/m <sup>3</sup> <sup>3</sup>	Pass
Sulphur Dioxide (Procal 2 – Low Range)	0.9511	1.012 <sup>1</sup>	1.01 <sup>1</sup>	0 – 243 mg/m <sup>3</sup> <sup>2</sup>	0 – 643 mg/m <sup>3</sup> <sup>3</sup>	Pass
Sulphur Dioxide (Procal 2 – High Range)	0.9515	1.033 <sup>1</sup>	1.03 <sup>1</sup>	0 – 244 mg/m <sup>3</sup> <sup>2</sup>	0 – 643 mg/m <sup>3</sup> <sup>3</sup>	Pass

#### Notes

1 – Calibration function derived using Method B.

2 – Calibrated range derived using QAL2 parallel test data extended 10%.

3 – Extrapolated calibrated range derived using reference materials

**Note:** The calibration functions, once applied, only remain valid as long as the QAL 3 data remains within control limits, and there are no manual adjustments made to the CEMS other than those allowed to bring the settings back within the QAL 3 control limits.

## 1B.2 Result Summary – AST

EN 14181 Test Type		AST						
Stack designation		Unit 8						
Measurand	Correlation coefficient of parallel data ( $R^2$ )	Derived Calibration function ( $y_i$ )		Established Calibrated Range <sup>4</sup>	AST Calibrated Range	Extrapolated Calibrated Range	Variability Test	Test of Calibration Function
Particulate Matter (Erwin SICK)	0.9852	$y = 0.6^1 + 1.433^1$		0 – 55 mg/m <sup>3</sup>	-	Not applicable	Fail	Fail

### Notes

- 1 – Calibration function derived using Method A.
- 2 – Calibrated range derived using AST parallel test data extended 10%.
- 3 - Calibrated range derived using QAL2 parallel test data extended 100%.
- 4 – Established calibrated range from QAL 2 - derived using reference materials

**Note:** The calibration functions, once applied, only remain valid as long as the QAL 3 data remains within control limits, and there are no manual adjustments made to the CEMS other than those allowed to bring the settings back within the QAL 3 control limits.

## 1C Deviations

<b>SRM deviations</b>	None
<b>Reason for deviation</b>	None
<b>EN 14181 deviations</b>	None
<b>Reason for deviation</b>	N/A
<b>Impact on results</b>	N/A
<b>Further actions required</b>	None

## Section 2 - Information about the Regulated Installation

### 2.1 Regulatory Information

<b>Name of operator</b>	RWE Generation UK plc.
<b>Name of Installation</b>	Aberthaw Power Station
<b>Address of installation</b>	The Leys Aberthaw, Nr Barry South Glamorgan CF62 4ZW
<b>Sector</b>	LCPD
<b>Permit Number</b>	RP3133LD
<b>Date of last QAL 2/AST</b>	March 2015 AST – Procal 1 & 2 & SICK OMD 41

### Regulated Determinands

Determinand	Emission Point	Daily Mean	Calendar Monthly	Uncertainty Requirement
Oxides of Nitrogen	Unit 8	1080 mg/m <sup>3</sup> (NOx as NO <sub>2</sub> )	1050 mg/m <sup>3</sup> (NOx as NO <sub>2</sub> )	20% at the ELV
Sulphur Dioxide	Unit 8	440 mg/m <sup>3</sup>	350 mg/m <sup>3</sup>	20% at the ELV
Total Particulate Matter	Unit 8	35 mg/m <sup>3</sup>	20 mg/m <sup>3</sup>	30% at the ELV

Note: ELVs at reference conditions 273K, 101.3kPa, 6% oxygen, dry gas



### 2.2 Operational Information and site monitoring provisions

#### 2.2.1 Process type and emissions variations

<b>Process Type</b>	Continuous - Coal-Fired Power Station
<b>Process Variations</b>	Once operational at maximum load variation is minimal
<b>Expected emissions variations</b>	
<b>1 Oxides of nitrogen</b>	640 - 775 mg/m <sup>3</sup> as NO <sub>2</sub> – variable
<b>2 Sulphur Dioxide</b>	115 - 200 mg/m <sup>3</sup> – variable
<b>3 Oxygen</b>	6 – 8 % – constant
<b>Possible low level emissions</b>	Sulphur Dioxide
<b>Provision to deal with low level emissions</b>	Method C - Use zero and span value form functional checks, along with parallel test data to derive calibration function.
<b>Other factors affecting monitoring results</b>	None
<b>Fuel type</b>	Coal
<b>Abatement</b>	Flue Gas Desulphurisation & Combustion control

## 2.3 Monitoring Provisions at the installation

### 2.3.1 Stack & sampling ports



### Photograph of Stack

Photo of Monitoring Platform



Access was limited to ports due to the stack layout inside the windshield. The port shown is one of four.

### 2.3.2 Monitoring platform and site provisions

Requirement	Compliant	Notes
• A safe and clean working environment with sufficient space and weather protections.	CEMs cabinet – Monitoring probe -	Both cabinet and probe are located inside the windshield and are thus in a spacious, clean and weatherproof environment.
• Easy and safe access to the CEM.	CEMS cabinet – Monitoring probe -	Stairways to the relevant levels.
• Adequate supplies of reference materials, tools and spare parts.	Yes	
• Facilities to introduce the reference materials for gaseous-monitoring systems, both at the inlet of the sampling line (where present), and at the inlet of the CEM.	N/A	There is no facility to introduce reference materials to the sample line or CEM. The CEMs are cross-duct IR (NO, NO <sub>2</sub> , SO <sub>2</sub> ) and cross-duct forward scatter (particulate).
• Compliance with TGN M1	No	Unable to access sample port B
• Compliance with EN 15259 – flow stability criteria (if applicable), Stack gas homogeneity.	Yes Yes	Stack gas homogeneity carried out previously by Atkins.

## Temperature and Velocity Profile

Company Name: Aberthaw power station  
Site Name: Aberthaw  
Sampling Point Ref: unit 8  
Project Reference:

Date: 26/10/16  
Run: TPM

Mean Stack Temperature, °C	62.750
Traverse Stack Velocity, m/s	19.563
Stack Gas Volume Flow Rate, m³/s (acms)	553.123
Stack Gas Volume Flow Rate, m³/s. Wet, STP	447.213

Ap Measurement units (Pa or mmH2O)	mmH2O	Pitot Coefciant	0.831
Barometric	758 mmHg	Leak Test	
Static		Instrument range	250 mmH2O
Port A	-25 mmH2O	Δp for leak test	187.5 mmH2O
Port B	-25 mmH2O	Positive leakage rate	0 per 15secs
Mean	-25 mmH2O	Negative leakage rate	0 per 15secs
		Pass/Fail	Pass

Stagnation Test	
Static measurement	
Positive side	-25 mmH2O
Negative side	-25 mmH2O
Difference (Pa)	0
Pass/Fail	Pass

Stack Dimensions		
Rectangular A	(Width) m	
Rectangular B	(Length) m	
Circular diam A	6 m	
Circular diam B	6 m	
Circular Mean	6 m	
Area	28.27431 m²	

Traverse Point	Distance m	Port A						Port B					
		Δ p. mmH2O			Average	Swirl Degrees	Temp °C	Δ p. mmH2O			Average	Swirl Degrees	Temp °C
		Reading 1	Reading 2	Reading 3				Reading 1	Reading 2	Reading 3			
1	0.20	28	28	28	28	0	61	27	27	27	27	0	63
2	0.63	31	31	31	31	0	63	33	33	33	33	0	63
3	1.16	32	32	32	32	0	63	32	32	32	32	0	63
4	1.94												
5	4.06												
6	4.84	32	32	32	32	0	63	34	34	34	34	0	64
7	5.37	32	32	32	32	0	63	33	33	33	33	0	63
8	5.80	29	29	29	29	0	62	29	29	29	29	0	62
9													
10													

Gas Data	
Oxygen %	6.00
CO <sub>2</sub> %	13.00
CO %	

Oxygen Correction	
Required Correction Value	
Actual Oxygen Factor	1.00
<b>Leave BLANK if no O<sub>2</sub> correction is required</b>	

BS EN 13284-1 & M1 Sample Point Requirements			Requirement Met?
Duct gas Flow: angle with regard to duct access <15°?			Y
Duct Gas Flow: No Negative Velocity: Not Permitted			Y
Duct Gas Flow: Ratio of max to min velocity <3:1?			Y
Working Area > 5m <sup>2</sup> ?			
Handrails with removable chains / self closing gates across the top of the ladder?			
Handrails (approx 0.5 and 1.0 m high) and vertical baseboards (approx 0.25m high)?			
Scaffold Built to 'Heavy Duty' Scaffold Rating or at least 2.5kN/m <sup>2</sup> loading			
Handrails not restricting access to ports?			
Room opposite sampling port equal or greater than the length of the sampling probe plus 1 metre?			
Sufficient Power (Waterproof 110V BS4343 Standard) close or on the platform?			

## 2.4.1 Continuous Emissions Monitoring Systems at the installation

	Procal 1 & 2	SICK	ABB
Determinand	NOx & SO <sub>2</sub> & H <sub>2</sub> O	Particulate	Oxygen
Type	In Situ IR	Cross Duct Forward Scatter	Zirconia Cell
Make	Procal	Erwin Sick	ABB
Model	Pulsi 200 series	OMD41	AZ20
MCERTS Certificate	MC990006/13	MC040042/01	MC110191/01
QAL1 Compliance?	Yes	Yes	Yes
Certification ranges	NO = 0 – 1000 ppm SO <sub>2</sub> low = 0 – 250 ppm SO <sub>2</sub> high = 0 – 1000 ppm H <sub>2</sub> O = 0 – 30 %	0 – 0.3 extinction	0 to 25% 0 to 5%
Operational ranges	As Above	0 – 50 mg/m <sup>3</sup>	0 to 25%vol
Principle	Dual wavelength infra-red	Opacity	Zirconia Cell
Raw data units	SO <sub>2</sub> & NO = ppm H <sub>2</sub> O = %	Extinction	%
Reference condition of raw data	(SO <sub>2</sub> & NO only) wet gas, no oxygen, temp or pressure correction	wet gas, no oxygen, temp or pressure correction	wet gas
Signal output	4-20 mA	Fibre optics	4 – 20mA
Provision for logging of 14181 data	Data logged by DCS		
Location of sample	Measurement taken at monitoring platform. SRM sample obtained from monitoring platform.		
Moisture – Measurement or calculated	Measured	Measured	Measured

### Section 3 – Information about the Monitoring campaign

**Table 3.1 - Monitoring Organisation Staff Details**

Project Manager	Position	MCERTS Level	Technical Endorsements	Expiry Dates	MCERTS Registration Number
Glyn Harrison	Operations Manager (Stack Emissions)	2	1	10/19	MM 03 228
			2	06/17	
			3	12/21	
			4	07/20	

Site Team	Position	MCERTS Level	Technical Endorsements	Expiry Dates	MCERTS Registration Number
Edwin Powell	Consultant	2	1	12/17	MM 05 621
			2	12/17	
			3	12/17	
			4	12/17	
Daniel Lewis	Technician	1	--	-	MM 14 1291

Report Author	Position	MCERTS Level	Technical Endorsements	Expiry Dates	MCERTS Registration Number
Glyn Harrison	Operations Manager (Stack Emissions)	2	1	10/19	MM 03 228
			2	06/17	
			3	12/21	
			4	07/20	

Report Reviewer	Position	MCERTS Level	Technical Endorsements	Expiry Dates	MCERTS Registration Number
Richard Harvey	Principal Consultant	2	1	11/17	MM 02 020
			2	03/20	
			3	03/21	
			4	12/20	

### 3.2 - Monitoring Organisation Method Details

Emission Parameter	Standard Method	Monitoring Procedure No.	Monitoring Accreditation Status	Analysis Technique	Expected Uncertainty (%)	Analysis Procedure No.	Analytical Laboratory	Analysis Accreditation Status
Oxides of Nitrogen (as NO <sub>2</sub> )	BS EN 14792:2005	RPSCE/1/21f	MCERTS	Chemiluminescence	6	N/A	N/A	N/A
Oxygen	BS EN 14789:2005	RPSCE/1/21g	MCERTS	Zirconia Cell	5	N/A	N/A	N/A
Sulphur Dioxide	TGN M22	RPSCE/1/24	MCERTS	FTIR	10	N/A	N/A	N/A
Total particulate Matter	BS EN 13284:2002	RPSCE/1/7c	MCERTS	Gravimetric	10 - 30	D9	RPS	UKAS

## Equipment details

Emission Parameter	Analysis Technique	Analyser	Analyser Certification Status	Certified Ranges	Operational Ranges	Operating Principle
Oxygen	Zirconia Cell	Horiba PG 250	MCERTs certificate No MC 050056/04	0 – 25%	0-25%	Extractive, multicomponent dry gas analyser. Sample extracted through sample probe and 5metre heated sample line (with integral heated filter) – line temperature 180°C. Sample line connected directly to a gas conditioner (peltier cooler) set at 3°C. Cold dry sample then passes to analyser. Sample is drawn through system by integral pump built into analyser.
Oxides of Nitrogen	Chemiluminescence	Horiba PG 250	0 – 125 mg/m <sup>3</sup> NO/NOx	0 – 125 mg/m <sup>3</sup> NO/NOx	0 – 500ppm	
Sulphur Dioxide	FTIR	Gasmet DX4000	MCERTs certificate No MC30014/05	0 – 75mg/m <sup>3</sup>	0 – 500 ppm	Extractive wet gas analyser. Sample obtained non-isokinetically. Sample extracted through sample probe and filtered before passing through 5metre heated sample line (with integral heated filter) – line temperature 180°C. Sample line connected directly to a heated sample pump which in turn was connected to the FTIR. Hot, wet sample then passes to analyser.
Stack Gas Moisture	FTIR	Gasmet DX4000	MCERTs certificate No MC30014/05	0-40%	0-40%	
Total Particulate Matter	Multipoint isokinetic sampling with in stack filtration	N/A	N/A	0 – 50mg/m <sup>3</sup>	-	Extractive manual test. Sample obtained isokinetically through sharp edged nozzle. Sample gas passed through a pre weighed, pre blown filter. Filter holder mounted in-stack.

### **Section 4A1: Data & calculations – QAL2 – Unit 8, Procal 1**

**4A1.1 Table 4.1.1 – Raw monitoring Data – Oxides of Nitrogen**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (Wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw Value (dry)	SRM Oxygen (Dry)	SRM Moisture	SRM at CEMs Raw conditions
		hr:min		NO (ppm)	(%)	(%)	NOx (ppm)	(%)	(%)	NOx (ppm)
1	26-Oct-16	11:23	12:23	337.44	6.79	2.21	328.17	6.61	3.02	318.3
2	26-Oct-16	12:34	13:34	360.61	6.74	2.15	348.55	6.58	2.78	338.9
3	26-Oct-16	13:35	14:35	331.62	6.44	2.07	318.06	6.33	2.69	309.5
4	26-Oct-16	14:36	15:36	365.07	6.51	2.12	355.64	6.37	2.73	345.9
5	26-Oct-16	15:37	16:37	378.05	6.63	2.19	365.76	6.48	2.80	355.5
6	26-Oct-16	16:38	17:38	344.15	6.93	2.25	337.30	6.72	2.89	327.5
7	26-Oct-16	17:39	18:39	346.13	7.00	2.26	339.06	6.77	2.90	329.2
8	26-Oct-16	18:40	19:40	352.52	7.01	2.22	344.91	6.83	2.84	335.1
9	26-Oct-16	19:41	20:41	351.42	7.04	2.19	344.92	6.81	2.76	335.4
10	26-Oct-16	20:42	21:42	343.51	6.93	2.14	334.97	6.80	2.66	326.1
11	26-Oct-16	21:53	22:53	336.24	6.92	2.15	333.39	6.69	2.65	324.6
12	26-Oct-16	22:54	23:54	330.84	6.88	2.23	331.21	6.68	2.72	322.2
13	26-27/10/2016	23:55	0:55	330.27	7.01	2.19	331.05	6.78	2.66	322.2
14	27-Oct-16	0:56	1:56	332.76	7.15	2.12	331.55	6.93	2.59	323.0
15	27-Oct-16	1:57	2:57	332.48	7.01	2.13	326.22	6.83	2.62	317.7
16	27-Oct-16	2:58	3:58	331.60	7.08	2.21	326.35	6.85	2.73	317.5
17	27-Oct-16	3:59	4:59	333.34	7.01	2.27	329.32	6.79	2.80	320.1
18	27-Oct-16	5:00	6:00	340.56	7.01	2.24	335.27	6.79	2.74	326.1
19	27-Oct-16	6:01	7:01	342.75	6.99	2.26	337.56	6.77	2.73	328.4
20	27-Oct-16	7:12	8:12	334.69	6.93	2.26	329.93	6.71	2.73	320.9
21	27-Oct-16	8:13	9:13	336.57	7.00	2.21	330.14	6.78	2.67	321.3
22	27-Oct-16	9:14	10:14	322.59	7.16	2.22	316.48	6.93	2.68	308.0
23	27-Oct-16	10:50	11:50	317.85	6.79	2.06	313.70	6.68	2.58	305.6
24	27-Oct-16	11:51	12:51	330.36	6.78	2.11	324.34	6.58	2.84	315.1
25	27-Oct-16	13:02	14:02	313.56	6.96	1.99	302.95	6.75	2.63	295.0
26	27-Oct-16	14:03	15:03	318.95	7.00	1.95	307.08	6.78	2.63	299.0
27	27-Oct-16	15:04	16:04	320.81	6.99	1.99	312.45	6.78	2.62	304.3
28	27-Oct-16	16:05	17:05	319.73	7.18	2.18	314.18	6.95	2.80	305.4
29	27-Oct-16	17:06	18:06	322.33	7.26	2.17	317.90	7.03	2.84	308.9
30	27-Oct-16	18:07	19:07	318.47	7.05	2.18	313.87	6.84	2.83	305.0
31	27-Oct-16	19:08	20:08	316.93	7.02	2.17	312.16	6.80	2.85	303.3
32	27-Oct-16	20:09	21:09	326.56	6.95	2.16	322.07	6.74	2.82	313.0
33	27-Oct-16	21:10	22:10	330.99	6.91	2.19	326.00	6.72	2.86	316.7
34	27-Oct-16	22:21	23:21	315.23	6.84	2.24	309.61	6.63	2.93	300.5
35	27-28/10/2016	23:22	0:22	274.39	8.32	2.18	266.90	8.15	2.87	259.2
36	02-Nov-16	11:26	12:26	337.71	6.55	2.54	331.91	6.39	2.67	323.1
37	02-Nov-16	12:27	13:27	345.47	6.51	2.49	337.10	6.32	2.77	327.8

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4A1.2 Table 4.2.1 - Standardised monitoring Data – Oxides of Nitrogen**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(NOx as NO <sub>2</sub> mg/m <sup>3</sup> )	(NOx as NO <sub>2</sub> mg/m <sup>3</sup> )	(mg/m <sup>3</sup> )
1	26-Oct-16	11:23	12:23	748.1	702.6	23
2	26-Oct-16	12:34	13:34	796.5	744.7	24
3	26-Oct-16	13:35	14:35	716.7	667.9	22
4	26-Oct-16	14:36	15:36	793.2	749.0	25
5	26-Oct-16	15:37	16:37	829.0	776.3	25
6	26-Oct-16	16:38	17:38	771.0	727.6	24
7	26-Oct-16	17:39	18:39	779.3	734.1	24
8	26-Oct-16	18:40	19:40	794.3	750.2	25
9	26-Oct-16	19:41	20:41	793.0	749.1	25
10	26-Oct-16	20:42	21:42	768.7	727.0	24
11	26-Oct-16	21:53	22:53	752.1	718.0	24
12	26-Oct-16	22:54	23:54	738.4	712.7	23
13	26-27/10/2016	23:55	0:55	743.7	717.6	24
14	27-Oct-16	0:56	1:56	756.4	726.2	24
15	27-Oct-16	1:57	2:57	748.4	709.2	23
16	27-Oct-16	2:58	3:58	751.1	710.8	23
17	27-Oct-16	3:59	4:59	751.6	713.9	23
18	27-Oct-16	5:00	6:00	767.6	727.1	24
19	27-Oct-16	6:01	7:01	771.6	730.8	24
20	27-Oct-16	7:12	8:12	750.0	711.5	23
21	27-Oct-16	8:13	9:13	757.5	715.2	24
22	27-Oct-16	9:14	10:14	734.6	693.3	23
23	27-Oct-16	10:50	11:50	703.7	675.2	22
24	27-Oct-16	11:51	12:51	731.2	693.2	23
25	27-Oct-16	13:02	14:02	702.3	655.3	21
26	27-Oct-16	14:03	15:03	716.0	665.3	22
27	27-Oct-16	15:04	16:04	720.3	677.3	22
28	27-Oct-16	16:05	17:05	728.8	688.9	23
29	27-Oct-16	17:06	18:06	739.1	701.2	23
30	27-Oct-16	18:07	19:07	719.4	683.1	23
31	27-Oct-16	19:08	20:08	714.4	677.6	22
32	27-Oct-16	20:09	21:09	732.1	695.8	23
33	27-Oct-16	21:10	22:10	740.2	703.4	23
34	27-Oct-16	22:21	23:21	701.8	664.0	22
35	27-28/10/2016	23:22	0:22	682.4	640.4	21
36	02-Nov-16	11:26	12:26	738.9	699.7	23
37	02-Nov-16	12:27	13:27	753.5	707.3	23

**Note:**


Emission concentrations expresed at reference conditions 273K,

6 % Oxygen, dry gas

**4A1.3 – Plot 1.1 - Time Series of Standardised CEM versus Standardised SRM data – Oxides of Nitrogen, (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A1.4 – Elimination of Outliers – Oxides of Nitrogen



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.'

Additionally, any data points are not likely to be outliers unless they are more than three standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9652, an outlier test has not been undertaken.

#### 4A1.5.1 Determination of Method A, B or C - Oxides of Nitrogen

Test No	Test Start Time	Test End Time	SRM measured value (y)	SRM O2	SRM Standardised
	hr:min		(ppm)	(%)	(mg/m <sup>3</sup> )
1	11:23	12:23	328.2	6.6	702.6
2	12:34	13:34	348.6	6.6	744.7
3	13:35	14:35	318.1	6.3	667.9
4	14:36	15:36	355.6	6.4	749.0
5	15:37	16:37	365.8	6.5	776.3
6	16:38	17:38	337.3	6.7	727.6
7	17:39	18:39	339.1	6.8	734.1
8	18:40	19:40	344.9	6.8	750.2
9	19:41	20:41	344.9	6.8	749.1
10	20:42	21:42	335.0	6.8	727.0
11	21:53	22:53	333.4	6.7	718.0
12	22:54	23:54	331.2	6.7	712.7
13	23:55	0:55	331.0	6.8	717.6
14	0:56	1:56	331.6	6.9	726.2
15	1:57	2:57	326.2	6.8	709.2
16	2:58	3:58	326.4	6.9	710.8
17	3:59	4:59	329.3	6.8	713.9
18	5:00	6:00	335.3	6.8	727.1
19	6:01	7:01	337.6	6.8	730.8
20	7:12	8:12	329.9	6.7	711.5
21	8:13	9:13	330.1	6.8	715.2
22	9:14	10:14	316.5	6.9	693.3
23	10:50	11:50	313.7	6.7	675.2
24	11:51	12:51	324.3	6.6	693.2
25	13:02	14:02	303.0	6.8	655.3
26	14:03	15:03	307.1	6.8	665.3
27	15:04	16:04	312.5	6.8	677.3
28	16:05	17:05	314.2	6.9	688.9
29	17:06	18:06	317.9	7.0	701.2
30	18:07	19:07	313.9	6.8	683.1
31	19:08	20:08	312.2	6.8	677.6
32	20:09	21:09	322.1	6.7	695.8
33	21:10	22:10	326.0	6.7	703.4
34	22:21	23:21	309.6	6.6	664.0
35	23:22	0:22	266.9	8.1	640.4
36	11:26	12:26	331.9	6.4	699.7
37	12:27	13:27	337.1	6.3	707.3
Sum			12088.07		
Emission Limit Value (ELV) =	1080	mg/Nm <sup>3</sup>	$Y_{\max}$	776.29	
Maximum Permissible uncertainty =	20	%	$Y_{\min}$	640.38	
Maximum Permissible uncertainty (at 1	216	mg/Nm <sup>3</sup>			
15% of the ELV =	162	mg/Nm <sup>3</sup>			
Is $Y_{\max} - Y_{\min} > \text{MPU at ELV?}$	No		$Y_{\max} - Y_{\min}$	135.91	
Is $Y_{\min} > 15\%$ of ELV?	Yes				

Derivation of Calibration Function

Method B

**4A1.6.1 Table 4.3.1 - Data used to derive calibration function - Oxides of Nitrogen,**

Test No	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)
			(ppm)	(ppm)
1	Reference		0.0	-3.0
2	11:23	12:23	318.3	337.4
3	12:34	13:34	338.9	360.6
4	13:35	14:35	309.5	331.6
5	14:36	15:36	345.9	365.1
6	15:37	16:37	355.5	378.1
7	16:38	17:38	327.5	344.2
8	17:39	18:39	329.2	346.1
9	18:40	19:40	335.1	352.5
10	19:41	20:41	335.4	351.4
11	20:42	21:42	326.1	343.5
12	21:53	22:53	324.6	336.2
13	22:54	23:54	322.2	330.8
14	23:55	0:55	322.2	330.3
15	0:56	1:56	323.0	332.8
16	1:57	2:57	317.7	332.5
17	2:58	3:58	317.5	331.6
18	3:59	4:59	320.1	333.3
19	5:00	6:00	326.1	340.6
20	6:01	7:01	328.4	342.8
21	7:12	8:12	320.9	334.7
22	8:13	9:13	321.3	336.6
23	9:14	10:14	308.0	322.6
24	10:50	11:50	305.6	317.8
25	11:51	12:51	315.1	330.4
26	13:02	14:02	295.0	313.6
27	14:03	15:03	299.0	318.9
28	15:04	16:04	304.3	320.8
29	16:05	17:05	305.4	319.7
30	17:06	18:06	308.9	322.3
31	18:07	19:07	305.0	318.5
32	19:08	20:08	303.3	316.9
33	20:09	21:09	313.0	326.6
34	21:10	22:10	316.7	331.0
35	22:21	23:21	300.5	315.2
36	23:22	0:22	259.2	274.4
37	11:26	12:26	323.1	337.7
38	12:27	13:27	327.8	345.5
Sum			11755.10	12321.57

#### 4A1.7.1 Determination of Calibration Function - Oxides of Nitrogen

##### Method B

Formulae:- Number of tests conducted = N 38

If  $Y_{max} - Y_{min} < 15\%$  of the ELV, the following formulae are used:

$$b = (y/(x-Z)) \quad x = (1/\text{No. AMS Tests}) * \text{Total AMS}$$

$$x = (1/ 38) * 12321. \quad \text{or} - \quad x = 324.25$$

$$a = (-\beta \cdot Z) \quad y = (1/\text{No. SRM Tests}) * \text{Total SRM}$$

$$y = (1/ 38) * 11755 \quad \text{or} - \quad y = 309.34$$

The Slope is calculated by :

$$\beta = y/(x-Z) \quad 309.3 / ( 324.2 - 3) \quad \beta = 0.945$$

The offset is calculated by:

$$\alpha = -\beta \cdot Z \quad -0.94 * 3 \quad \alpha = 2.84$$

The calibration is function  $y_{im} = \alpha + \beta_{xi,m}$  or  $y_i = 2.835 + 0.94 * X_i$

**4A1.8.1 Table 4.4.1 - Calculation of calibrated CEMS values - Oxides of Nitrogen**


Test No	Test Start Time	Test End Time	CEMS Raw Value (x) (ppm)	CEMS Calibrated signal (ppm)	CEMS Dry Oxygen (%)	CEMS Moisture (%)	CEMS Standardised Value (dry) (mg/Nm <sup>3</sup> )	CEMS Calibrated Standardised Value (mg/Nm <sup>3</sup> )	SRM Standardised (mg/Nm <sup>3</sup> )
1	Reference		-3.0	0.0	N/A	N/A	-3.0	0.0	0.0
2	11:23	12:23	337.4	321.8	6.8	2.2	748.1	713.4	702.6
3	12:34	13:34	360.6	343.7	6.7	2.1	796.5	759.2	744.7
4	13:35	14:35	331.6	316.3	6.4	2.1	716.7	683.6	667.9
5	14:36	15:36	365.1	347.9	6.5	2.1	793.2	756.0	749.0
6	15:37	16:37	378.1	360.2	6.6	2.2	829.0	789.9	776.3
7	16:38	17:38	344.2	328.2	6.9	2.2	771.0	735.1	727.6
8	17:39	18:39	346.1	330.0	7.0	2.3	779.3	743.0	734.1
9	18:40	19:40	352.5	336.1	7.0	2.2	794.3	757.3	750.2
10	19:41	20:41	351.4	335.0	7.0	2.2	793.0	756.0	749.1
11	20:42	21:42	343.5	327.5	6.9	2.1	768.7	733.0	727.0
12	21:53	22:53	336.2	320.7	6.9	2.2	752.1	717.3	718.0
13	22:54	23:54	330.8	315.6	6.9	2.2	738.4	704.3	712.7
14	23:55	00:55	330.3	315.0	7.0	2.2	743.7	709.4	717.6
15	00:56	01:56	332.8	317.4	7.1	2.1	756.4	721.5	726.2
16	01:57	02:57	332.5	317.1	7.0	2.1	748.4	713.9	709.2
17	02:58	03:58	331.6	316.3	7.1	2.2	751.1	716.4	710.8
18	03:59	04:59	333.3	317.9	7.0	2.3	751.6	716.9	713.9
19	05:00	06:00	340.6	324.8	7.0	2.2	767.6	732.0	727.1
20	06:01	07:01	342.8	326.8	7.0	2.3	771.6	735.8	730.8
21	07:12	08:12	334.7	319.2	6.9	2.3	750.0	715.4	711.5
22	08:13	09:13	336.6	321.0	7.0	2.2	757.5	722.5	715.2
23	09:14	10:14	322.6	307.8	7.2	2.2	734.6	700.9	693.3
24	10:50	11:50	317.8	303.3	6.8	2.1	703.7	671.5	675.2
25	11:51	12:51	330.4	315.1	6.8	2.1	731.2	697.5	693.2
26	13:02	14:02	313.6	299.2	7.0	2.0	702.3	670.2	655.3
27	14:03	15:03	318.9	304.3	7.0	2.0	716.0	683.2	665.3
28	15:04	16:04	320.8	306.1	7.0	2.0	720.3	687.2	677.3
29	16:05	17:05	319.7	305.1	7.2	2.2	728.8	695.4	688.9
30	17:06	18:06	322.3	307.5	7.3	2.2	739.1	705.2	701.2
31	18:07	19:07	318.5	303.9	7.1	2.2	719.4	686.5	683.1
32	19:08	20:08	316.9	302.4	7.0	2.2	714.4	681.7	677.6
33	20:09	21:09	326.6	311.5	6.9	2.2	732.1	698.4	695.8
34	21:10	22:10	331.0	315.7	6.9	2.2	740.2	706.1	703.4
35	22:21	23:21	315.2	300.8	6.8	2.2	701.8	669.7	664.0
36	23:22	00:22	274.4	262.2	8.3	2.2	682.4	652.1	640.4
37	11:26	12:26	337.7	322.1	6.6	2.5	738.9	704.7	699.7
38	12:27	13:27	345.5	329.4	6.5	2.5	753.5	718.4	707.3
Sum							27634.16		
Emission Limit Value (ELV) = 1080 mg/Nm <sup>3</sup>									

Reference Oxygen

6

%

**4A1.9.1 Plot 2.1 CEM versus SRM Parallel Test Data at CEM measurement conditions –NOx ppm, wet gas.**




**4A1.10.1 Table 4.5.1 – Data used for the Variability Test – Oxides of Nitrogen**

Test No	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
	hr:min		mg/m3	mg/m3			
2	11:23	12:23	713.4	702.6	-10.82	-4.93	24.32
3	12:34	13:34	759.2	744.7	-14.52	-8.63	74.53
4	13:35	14:35	683.6	667.9	-15.68	-9.79	95.82
5	14:36	15:36	756.0	749.0	-6.99	-1.10	1.22
6	15:37	16:37	789.9	776.3	-13.59	-7.71	59.38
7	16:38	17:38	735.1	727.6	-7.49	-1.60	2.57
8	17:39	18:39	743.0	734.1	-8.92	-3.04	9.22
9	18:40	19:40	757.3	750.2	-7.03	-1.15	1.32
10	19:41	20:41	756.0	749.1	-6.90	-1.02	1.03
11	20:42	21:42	733.0	727.0	-5.98	-0.09	0.01
12	21:53	22:53	717.3	718.0	0.67	6.55	42.93
13	22:54	23:54	704.3	712.7	8.41	14.30	204.43
14	23:55	00:55	709.4	717.6	8.16	14.04	197.21
15	00:56	01:56	721.5	726.2	4.68	10.57	111.72
16	01:57	02:57	713.9	709.2	-4.64	1.24	1.54
17	02:58	03:58	716.4	710.8	-5.58	0.30	0.09
18	03:59	04:59	716.9	713.9	-2.99	2.90	8.42
19	05:00	06:00	732.0	727.1	-4.87	1.01	1.03
20	06:01	07:01	735.8	730.8	-5.02	0.86	0.74
21	07:12	08:12	715.4	711.5	-3.83	2.05	4.21
22	08:13	09:13	722.5	715.2	-7.21	-1.32	1.74
23	09:14	10:14	700.9	693.3	-7.60	-1.72	2.95
24	10:50	11:50	671.5	675.2	3.73	9.61	92.43
25	11:51	12:51	697.5	693.2	-4.25	1.64	2.68
26	13:02	14:02	670.2	655.3	-14.97	-9.08	82.43
27	14:03	15:03	683.2	665.3	-17.90	-12.01	144.29
28	15:04	16:04	687.2	677.3	-9.98	-4.09	16.72
29	16:05	17:05	695.4	688.9	-6.49	-0.61	0.37
30	17:06	18:06	705.2	701.2	-3.95	1.94	3.77
31	18:07	19:07	686.5	683.1	-3.43	2.46	6.05
32	19:08	20:08	681.7	677.6	-4.06	1.83	3.34
33	20:09	21:09	698.4	695.8	-2.57	3.31	10.97
34	21:10	22:10	706.1	703.4	-2.70	3.18	10.13
35	22:21	23:21	669.7	664.0	-5.70	0.19	0.04
36	23:22	00:22	652.1	640.4	-11.72	-5.83	34.04
37	11:26	12:26	704.7	699.7	-4.92	0.97	0.94
38	12:27	13:27	718.4	707.3	-11.15	-5.26	27.66
37 Tests		Mean			-5.89		
Sum							1282.27

#### 4A1.11.1 - Variability Test Calculation – Oxides of Nitrogen

SD=	$\text{Root}(1-\text{Number}).\text{Integral}(\text{D1}-\text{D})^2$	5.97	mg/m <sup>3</sup> (s,d), 6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2*1080 mg/m <sup>3</sup> (s,d,6%O <sub>2</sub> )/1.96	110.20	mg/m <sup>3</sup> (s,d), 6%O <sub>2</sub>
For 37 tests, kv =	0.9885		
Therefore variability=		5.97 <= 110.2 * 0.9885	
or	5.97	<=	108.94
Which is TRUE therefore the CEMS passes the test			

#### 4A1.12.1 Plot 3.1 – Standardised CEM data versus standardised SRM - Oxides of Nitrogen (as NO<sub>2</sub>) – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen



**4A1.1.2 Table 4.1.2 – Raw monitoring Data – Sulphur Dioxide, Low Range**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (wet)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMs Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	26-Oct-16	11:23	12:23	65.40	6.79	2.21	63.35	6.61	3.02	63.3
2	26-Oct-16	12:34	13:34	66.36	6.74	2.15	66.31	6.58	2.78	66.3
3	26-Oct-16	13:35	14:35	66.20	6.44	2.07	67.20	6.33	2.69	67.2
4	26-Oct-16	14:36	15:36	66.60	6.51	2.12	67.39	6.37	2.73	67.4
5	26-Oct-16	15:37	16:37	66.54	6.63	2.19	66.68	6.48	2.80	66.7
6	26-Oct-16	16:38	17:38	65.50	6.93	2.25	64.88	6.72	2.89	64.9
7	26-Oct-16	17:39	18:39	65.06	7.00	2.26	64.58	6.77	2.90	64.6
8	26-Oct-16	18:40	19:40	63.62	7.01	2.22	59.98	6.83	2.84	60.0
9	26-Oct-16	19:41	20:41	61.07	7.04	2.19	55.61	6.81	2.76	55.6
10	26-Oct-16	20:42	21:42	59.08	6.93	2.14	54.14	6.80	2.66	54.1
11	26-Oct-16	21:53	22:53	58.16	6.92	2.15	52.98	6.69	2.65	53.0
12	26-Oct-16	22:54	23:54	57.66	6.88	2.23	52.03	6.68	2.72	52.0
13	26-27/10/2016	23:55	0:55	55.09	7.01	2.19	49.45	6.78	2.66	49.4
14	27-Oct-16	0:56	1:56	52.96	7.15	2.12	47.16	6.93	2.59	47.2
15	27-Oct-16	1:57	2:57	53.68	7.01	2.13	48.38	6.83	2.62	48.4
16	27-Oct-16	2:58	3:58	54.80	7.08	2.21	49.52	6.85	2.73	49.5
17	27-Oct-16	3:59	4:59	55.21	7.01	2.27	50.03	6.79	2.80	50.0
18	27-Oct-16	5:00	6:00	52.96	7.01	2.24	47.23	6.79	2.74	47.2
19	27-Oct-16	6:01	7:01	51.65	6.99	2.26	45.39	6.77	2.73	45.4
20	27-Oct-16	7:12	8:12	52.71	6.93	2.26	46.08	6.71	2.73	46.1
21	27-Oct-16	8:13	9:13	52.21	7.00	2.21	45.45	6.78	2.67	45.5
22	27-Oct-16	9:14	10:14	52.38	7.16	2.22	44.75	6.93	2.68	44.8
23	27-Oct-16	10:50	11:50	56.08	6.79	2.06	53.11	6.68	2.58	53.1
24	27-Oct-16	11:51	12:51	59.48	6.78	2.11	56.52	6.58	2.84	56.5
25	27-Oct-16	13:02	14:02	62.32	6.96	1.99	58.67	6.75	2.63	58.7
26	27-Oct-16	14:03	15:03	63.59	7.00	1.95	60.69	6.78	2.63	60.7
27	27-Oct-16	15:04	16:04	64.84	6.99	1.99	63.57	6.78	2.62	63.6
28	27-Oct-16	16:05	17:05	62.66	7.18	2.18	59.02	6.95	2.80	59.0
29	27-Oct-16	17:06	18:06	60.21	7.26	2.17	56.66	7.03	2.84	56.7
30	27-Oct-16	18:07	19:07	57.10	7.05	2.18	53.37	6.84	2.83	53.4
31	27-Oct-16	19:08	20:08	52.29	7.02	2.17	49.14	6.80	2.85	49.1
32	27-Oct-16	20:09	21:09	51.76	6.95	2.16	48.63	6.74	2.82	48.6
33	27-Oct-16	21:10	22:10	49.63	6.91	2.19	46.83	6.72	2.86	46.8
34	27-Oct-16	22:21	23:21	47.46	6.84	2.24	44.76	6.63	2.93	44.8
35	27-28/10/2016	23:22	0:22	41.88	8.32	2.18	39.04	8.15	2.87	39.0
36	02-Nov-16	11:26	12:26	75.34	6.55	2.54	73.06	6.39	2.67	73.1
37	02-Nov-16	12:27	13:27	70.15	6.51	2.49	70.17	6.32	2.77	70.2

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4A1.2.2 Table 4.2.2 -Standardised monitoring Data – Sulphur Dioxide, Low Range**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(SO <sub>2</sub> mg/m <sup>3</sup> )	(SO <sub>2</sub> mg/m <sup>3</sup> )	(mg/m <sup>3</sup> )
1	26-Oct-16	11:23	12:23	201.7	194.6	8.7
2	26-Oct-16	12:34	13:34	203.9	202.7	9.1
3	26-Oct-16	13:35	14:35	199.0	201.8	9.1
4	26-Oct-16	14:36	15:36	201.3	203.0	9.1
5	26-Oct-16	15:37	16:37	203.0	202.6	9.1
6	26-Oct-16	16:38	17:38	204.1	200.5	9.0
7	26-Oct-16	17:39	18:39	203.8	200.4	9.0
8	26-Oct-16	18:40	19:40	199.5	186.8	8.4
9	26-Oct-16	19:41	20:41	191.7	172.8	7.8
10	26-Oct-16	20:42	21:42	183.9	168.0	7.6
11	26-Oct-16	21:53	22:53	181.0	163.1	7.3
12	26-Oct-16	22:54	23:54	179.0	160.1	7.2
13	26-27/10/2016	23:55	0:55	172.6	153.2	6.9
14	27-Oct-16	0:56	1:56	167.5	147.5	6.6
15	27-Oct-16	1:57	2:57	168.1	150.3	6.7
16	27-Oct-16	2:58	3:58	172.7	154.3	6.9
17	27-Oct-16	3:59	4:59	173.2	155.2	7.0
18	27-Oct-16	5:00	6:00	166.1	146.5	6.6
19	27-Oct-16	6:01	7:01	161.8	140.5	6.3
20	27-Oct-16	7:12	8:12	164.4	142.1	6.4
21	27-Oct-16	8:13	9:13	163.5	140.8	6.3
22	27-Oct-16	9:14	10:14	166.0	140.2	6.3
23	27-Oct-16	10:50	11:50	172.8	163.2	7.3
24	27-Oct-16	11:51	12:51	183.2	173.0	7.8
25	27-Oct-16	13:02	14:02	194.2	181.3	8.2
26	27-Oct-16	14:03	15:03	198.6	187.9	8.4
27	27-Oct-16	15:04	16:04	202.5	196.9	8.8
28	27-Oct-16	16:05	17:05	198.7	185.2	8.3
29	27-Oct-16	17:06	18:06	192.1	179.0	8.0
30	27-Oct-16	18:07	19:07	179.5	166.3	7.5
31	27-Oct-16	19:08	20:08	164.0	152.8	6.9
32	27-Oct-16	20:09	21:09	161.5	150.4	6.7
33	27-Oct-16	21:10	22:10	154.4	144.7	6.5
34	27-Oct-16	22:21	23:21	147.0	137.6	6.2
35	27-28/10/2016	23:22	0:22	144.9	134.2	6.0
36	02-Nov-16	11:26	12:26	229.3	220.2	9.9
37	02-Nov-16	12:27	13:27	212.9	210.7	9.4

Note:


Emission concentrations expresed at reference conditions 273K,

6 % Oxygen, dry gas

**4A1.3.2 – Plot 1.2 - Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide, Low Range (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A1.4.2 – Elimination of Outliers – Sulphur Dioxide, Low Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.'

Additionally, any data points are not likely to be outliers unless they are more than two standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9562, an outlier test has not been undertaken.

**4A1.5.2 Determination of Method A, B or C - Sulphur Dioxide, Low Range.**

Test No	Test Start Time	Test End Time	SRM measured value (y)	SRM O2	SRM Standardised
	hr:min		(ppm)	(%)	(mg/m <sup>3</sup> )
1	11:23	12:23	63.3	6.6	194.6
2	12:34	13:34	66.3	6.6	202.7
3	13:35	14:35	67.2	6.3	201.8
4	14:36	15:36	67.4	6.4	203.0
5	15:37	16:37	66.7	6.5	202.6
6	16:38	17:38	64.9	6.7	200.5
7	17:39	18:39	64.6	6.8	200.4
8	18:40	19:40	60.0	6.8	186.8
9	19:41	20:41	55.6	6.8	172.8
10	20:42	21:42	54.1	6.8	168.0
11	21:53	22:53	53.0	6.7	163.1
12	22:54	23:54	52.0	6.7	160.1
13	23:55	0:55	49.4	6.8	153.2
14	0:56	1:56	47.2	6.9	147.5
15	1:57	2:57	48.4	6.8	150.3
16	2:58	3:58	49.5	6.9	154.3
17	3:59	4:59	50.0	6.8	155.2
18	5:00	6:00	47.2	6.8	146.5
19	6:01	7:01	45.4	6.8	140.5
20	7:12	8:12	46.1	6.7	142.1
21	8:13	9:13	45.5	6.8	140.8
22	9:14	10:14	44.8	6.9	140.2
23	10:50	11:50	53.1	6.7	163.2
24	11:51	12:51	56.5	6.6	173.0
25	13:02	14:02	58.7	6.8	181.3
26	14:03	15:03	60.7	6.8	187.9
27	15:04	16:04	63.6	6.8	196.9
28	16:05	17:05	59.0	6.9	185.2
29	17:06	18:06	56.7	7.0	179.0
30	18:07	19:07	53.4	6.8	166.3
31	19:08	20:08	49.1	6.8	152.8
32	20:09	21:09	48.6	6.7	150.4
33	21:10	22:10	46.8	6.7	144.7
34	22:21	23:21	44.8	6.6	137.6
35	23:22	0:22	39.0	8.1	134.2
36	11:26	12:26	73.1	6.4	220.2
37	12:27	13:27	70.2	6.3	210.7
Sum			2041.79		
Emission Limit Value (ELV) =			440 mg/Nm <sup>3</sup>	Y <sub>max</sub>	220.18
Maximum Permissible uncertainty =			20 %	Y <sub>min</sub>	134.18
Maximum Permissible uncertainty (at 15% of ELV) =			88 mg/Nm <sup>3</sup>		
15% of the ELV =			66 mg/Nm <sup>3</sup>		
Is Y <sub>max</sub> - Y <sub>min</sub> > MPU at ELV?			No	Y <sub>max</sub> - Y <sub>min</sub>	86.00
Is Y <sub>min</sub> > 15% of ELV?			Yes		

Derivation of Calibration Function	Method B
------------------------------------	----------

**4A1.6.2 Table 4.3.2 - Data used to derive calibration function - Sulphur Dioxide, Low Range.**

Test No	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)
			(ppm)	(ppm)
1	Reference Zero		0.0	1.7
2	11:23	12:23	63.3	65.4
3	12:34	13:34	66.3	66.4
4	13:35	14:35	67.2	66.2
5	14:36	15:36	67.4	66.6
6	15:37	16:37	66.7	66.5
7	16:38	17:38	64.9	65.5
8	17:39	18:39	64.6	65.1
9	18:40	19:40	60.0	63.6
10	19:41	20:41	55.6	61.1
11	20:42	21:42	54.1	59.1
12	21:53	22:53	53.0	58.2
13	22:54	23:54	52.0	57.7
14	23:55	0:55	49.4	55.1
15	0:56	1:56	47.2	53.0
16	1:57	2:57	48.4	53.7
17	2:58	3:58	49.5	54.8
18	3:59	4:59	50.0	55.2
19	5:00	6:00	47.2	53.0
20	6:01	7:01	45.4	51.6
21	7:12	8:12	46.1	52.7
22	8:13	9:13	45.5	52.2
23	9:14	10:14	44.8	52.4
24	10:50	11:50	53.1	56.1
25	11:51	12:51	56.5	59.5
26	13:02	14:02	58.7	62.3
27	14:03	15:03	60.7	63.6
28	15:04	16:04	63.6	64.8
29	16:05	17:05	59.0	62.7
30	17:06	18:06	56.7	60.2
31	18:07	19:07	53.4	57.1
32	19:08	20:08	49.1	52.3
33	20:09	21:09	48.6	51.8
34	21:10	22:10	46.8	49.6
35	22:21	23:21	44.8	47.5
36	23:22	0:22	39.0	41.9
37	11:26	12:26	73.1	75.3
38	12:27	13:27	70.2	70.2
Sum			2041.79	2171.38

#### 4A1.7.2 Determination of Calibration Function - Sulphur Dioxide, Low Range.

## Method B

Formulae:- Number of tests conducted =	N	38
If $Y_{max} - Y_{min} < 15\%$ of the ELV, the following formulae are used:		
$b = (y/(x-Z))$ $x = (1/\text{No. AMS Tests}) * \text{Total AMS}$		
$x = (1/ 38) * 2171.3$ or -	$x =$	57.14
$a = (-\beta \cdot Z)$ $y = (1/\text{No. SRM Tests}) * \text{Total SRM}$		
$y = (1/ 38) * 2041.$ or -	$y =$	53.73
The Slope is calculated by :		
$\beta = y/(x-Z) = 53.73 / (57.14 - 1.67)$	$\beta =$	0.969
The offset is calculated by:		
$\alpha = -\beta \cdot Z = -0.96 * 1.67$	$\alpha =$	-1.62
The calibration is function $y_m = \alpha + \beta x_{1:m}$ or	$y_i =$	$-1.61 + 0.96 * X_i$

**4A1.8.2 Table A.4.4.2 - Calculation of calibrated CEMS values - Sulphur Dioxide, Low Range**

Test No	Test Start Time	Test End Time	CEMS Raw Value (x) (ppm)	CEMS Calibrated signal (ppm)	CEMS Dry Oxygen (%)	CEMS Moisture (%)	CEMS Standardised Value (dry) (mg/Nm <sup>3</sup> )	CEMS Calibrated Standardised Value (mg/Nm <sup>3</sup> )	SRM Standardised (mg/m <sup>3</sup> )
hr:min									
1	Reference Zero		1.7	0.0	N/A	N/A	1.7	0.0	0.0
2	11:23	12:23	65.4	61.7	6.8	2.2	201.7	190.4	194.6
3	12:34	13:34	66.4	62.7	6.7	2.1	203.9	192.6	202.7
4	13:35	14:35	66.2	62.5	6.4	2.1	199.0	187.9	201.8
5	14:36	15:36	66.6	62.9	6.5	2.1	201.3	190.1	203.0
6	15:37	16:37	66.5	62.8	6.6	2.2	203.0	191.7	202.6
7	16:38	17:38	65.5	61.8	6.9	2.2	204.1	192.7	200.5
8	17:39	18:39	65.1	61.4	7.0	2.3	203.8	192.3	200.4
9	18:40	19:40	63.6	60.0	7.0	2.2	199.5	188.1	186.8
10	19:41	20:41	61.1	57.5	7.0	2.2	191.7	180.6	172.8
11	20:42	21:42	59.1	55.6	6.9	2.1	183.9	173.1	168.0
12	21:53	22:53	58.2	54.7	6.9	2.2	181.0	170.3	163.1
13	22:54	23:54	57.7	54.2	6.9	2.2	179.0	168.4	160.1
14	23:55	00:55	55.1	51.7	7.0	2.2	172.6	162.1	153.2
15	00:56	01:56	53.0	49.7	7.1	2.1	167.5	157.1	147.5
16	01:57	02:57	53.7	50.4	7.0	2.1	168.1	157.8	150.3
17	02:58	03:58	54.8	51.5	7.1	2.2	172.7	162.2	154.3
18	03:59	04:59	55.2	51.9	7.0	2.3	173.2	162.7	155.2
19	05:00	06:00	53.0	49.7	7.0	2.2	166.1	155.8	146.5
20	06:01	07:01	51.6	48.4	7.0	2.3	161.8	151.6	140.5
21	07:12	08:12	52.7	49.4	6.9	2.3	164.4	154.2	142.1
22	08:13	09:13	52.2	49.0	7.0	2.2	163.5	153.3	140.8
23	09:14	10:14	52.4	49.1	7.2	2.2	166.0	155.6	140.2
24	10:50	11:50	56.1	52.7	6.8	2.1	172.8	162.4	163.2
25	11:51	12:51	59.5	56.0	6.8	2.1	183.2	172.4	173.0
26	13:02	14:02	62.3	58.7	7.0	2.0	194.2	183.1	181.3
27	14:03	15:03	63.6	60.0	7.0	2.0	198.6	187.3	187.9
28	15:04	16:04	64.8	61.2	7.0	2.0	202.5	191.1	196.9
29	16:05	17:05	62.7	59.1	7.2	2.2	198.7	187.4	185.2
30	17:06	18:06	60.2	56.7	7.3	2.2	192.1	180.9	179.0
31	18:07	19:07	57.1	53.7	7.1	2.2	179.5	168.8	166.3
32	19:08	20:08	52.3	49.0	7.0	2.2	164.0	153.8	152.8
33	20:09	21:09	51.8	48.5	6.9	2.2	161.5	151.3	150.4
34	21:10	22:10	49.6	46.5	6.9	2.2	154.4	144.5	144.7
35	22:21	23:21	47.5	44.4	6.8	2.2	147.0	137.4	137.6
36	23:22	00:22	41.9	39.0	8.3	2.2	144.9	134.8	134.2
37	11:26	12:26	75.3	71.4	6.6	2.5	229.3	217.2	220.2
38	12:27	13:27	70.2	66.3	6.5	2.5	212.9	201.3	210.7
Sum							6765.18		


Emission Limit Value (ELV) = 440 mg/Nm<sup>3</sup>

Reference Oxygen

6

%

**4A1.9.2 Plot 2.2 CEM versus SRM Parallel Test Data at CEM measurement conditions –Sulphur dioxide, wet gas**




**4A1.10.2 Table 4.5.2 – Data used for the Variability Test – Sulphur Dioxide, Low Range.**

Test No	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
	hr:min		mg/m3	mg/m3			
2	11:23	12:23	190.4	194.6	4.18	5.64	31.83
3	12:34	13:34	192.6	202.7	10.16	11.63	135.28
4	13:35	14:35	187.9	201.8	13.84	15.31	234.27
5	14:36	15:36	190.1	203.0	12.87	14.33	205.42
6	15:37	16:37	191.7	202.6	10.88	12.35	152.44
7	16:38	17:38	192.7	200.5	7.81	9.28	86.14
8	17:39	18:39	192.3	200.4	8.04	9.50	90.31
9	18:40	19:40	188.1	186.8	-1.30	0.17	0.03
10	19:41	20:41	180.6	172.8	-7.84	-6.37	40.59
11	20:42	21:42	173.1	168.0	-5.18	-3.72	13.81
12	21:53	22:53	170.3	163.1	-7.25	-5.78	33.45
13	22:54	23:54	168.4	160.1	-8.26	-6.79	46.11
14	23:55	00:55	162.1	153.2	-8.92	-7.45	55.56
15	00:56	01:56	157.1	147.5	-9.59	-8.12	65.98
16	01:57	02:57	157.8	150.3	-7.51	-6.04	36.49
17	02:58	03:58	162.2	154.3	-7.92	-6.46	41.71
18	03:59	04:59	162.7	155.2	-7.46	-5.99	35.91
19	05:00	06:00	155.8	146.5	-9.29	-7.82	61.20
20	06:01	07:01	151.6	140.5	-11.10	-9.63	92.75
21	07:12	08:12	154.2	142.1	-12.01	-10.54	111.18
22	08:13	09:13	153.3	140.8	-12.54	-11.08	122.71
23	09:14	10:14	155.6	140.2	-15.48	-14.01	196.23
24	10:50	11:50	162.4	163.2	0.90	2.36	5.58
25	11:51	12:51	172.4	173.0	0.54	2.01	4.03
26	13:02	14:02	183.1	181.3	-1.74	-0.27	0.08
27	14:03	15:03	187.3	187.9	0.54	2.00	4.02
28	15:04	16:04	191.1	196.9	5.74	7.20	51.88
29	16:05	17:05	187.4	185.2	-2.15	-0.68	0.46
30	17:06	18:06	180.9	179.0	-1.95	-0.48	0.23
31	18:07	19:07	168.8	166.3	-2.44	-0.97	0.95
32	19:08	20:08	153.8	152.8	-1.01	0.46	0.21
33	20:09	21:09	151.3	150.4	-0.92	0.55	0.30
34	21:10	22:10	144.5	144.7	0.16	1.63	2.66
35	22:21	23:21	137.4	137.6	0.22	1.68	2.84
36	23:22	00:22	134.8	134.2	-0.59	0.87	0.76
37	11:26	12:26	217.2	220.2	2.95	4.42	19.53
38	12:27	13:27	201.3	210.7	9.36	10.83	117.32
37 Tests		Mean			-1.47		
Sum							2100.25

#### 4A1.11.2 - Variability Test Calculation – Sulphur Dioxide, Low Range.

SD=	Root(1-Number).Integral(D1-D) <sup>2</sup>	7.64	mg/m <sub>3</sub> (s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2*440 mg/m <sub>3</sub> (s,d,6%O <sub>2</sub> )/1.96	44.90	mg/m <sub>3</sub> (s,d),6%O <sub>2</sub>
For 37 tests, kv =	0.9885		
Therefore variability=		7.64 <= 44.9 * 0.9885	
or	7.64	<=	44.38
Which is TRUE therefore the CEMS passes the test			

#### 4A1.12.2 Plot 3.2 – Standardised CEM data versus standardised SRM - Sulphur dioxide, Low Range – Reference conditions 273K, 101.3kPa 6% oxygen, dry gas.



#### 4A1.1.3 Table 4.1.3– Raw monitoring Data – Sulphur Dioxide, High Range

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (wet)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMs Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	26-Oct-16	11:23	12:23	62.50	6.79	2.21	63.35	6.61	3.02	63.3
2	26-Oct-16	12:34	13:34	63.45	6.74	2.15	66.31	6.58	2.78	66.3
3	26-Oct-16	13:35	14:35	63.28	6.44	2.07	67.20	6.33	2.69	67.2
4	26-Oct-16	14:36	15:36	63.67	6.51	2.12	67.39	6.37	2.73	67.4
5	26-Oct-16	15:37	16:37	63.61	6.63	2.19	66.68	6.48	2.80	66.7
6	26-Oct-16	16:38	17:38	62.61	6.93	2.25	64.88	6.72	2.89	64.9
7	26-Oct-16	17:39	18:39	62.17	7.00	2.26	64.58	6.77	2.90	64.6
8	26-Oct-16	18:40	19:40	60.79	7.01	2.22	59.98	6.83	2.84	60.0
9	26-Oct-16	19:41	20:41	58.32	7.04	2.19	55.61	6.81	2.76	55.6
10	26-Oct-16	20:42	21:42	56.40	6.93	2.14	54.14	6.80	2.66	54.1
11	26-Oct-16	21:53	22:53	55.51	6.92	2.15	52.98	6.69	2.65	53.0
12	26-Oct-16	22:54	23:54	55.02	6.88	2.23	52.03	6.68	2.72	52.0
13	26-27/10/2016	23:55	0:55	52.54	7.01	2.19	49.45	6.78	2.66	49.4
14	27-Oct-16	0:56	1:56	50.49	7.15	2.12	47.16	6.93	2.59	47.2
15	27-Oct-16	1:57	2:57	51.19	7.01	2.13	48.38	6.83	2.62	48.4
16	27-Oct-16	2:58	3:58	52.26	7.08	2.21	49.52	6.85	2.73	49.5
17	27-Oct-16	3:59	4:59	52.66	7.01	2.27	50.03	6.79	2.80	50.0
18	27-Oct-16	5:00	6:00	50.50	7.01	2.24	47.23	6.79	2.74	47.2
19	27-Oct-16	6:01	7:01	49.23	6.99	2.26	45.39	6.77	2.73	45.4
20	27-Oct-16	7:12	8:12	50.25	6.93	2.26	46.08	6.71	2.73	46.1
21	27-Oct-16	8:13	9:13	49.78	7.00	2.21	45.45	6.78	2.67	45.5
22	27-Oct-16	9:14	10:14	49.93	7.16	2.22	44.75	6.93	2.68	44.8
23	27-Oct-16	10:50	11:50	53.51	6.79	2.06	53.11	6.68	2.58	53.1
24	27-Oct-16	11:51	12:51	56.79	6.78	2.11	56.52	6.58	2.84	56.5
25	27-Oct-16	13:02	14:02	59.54	6.96	1.99	58.67	6.75	2.63	58.7
26	27-Oct-16	14:03	15:03	60.77	7.00	1.95	60.69	6.78	2.63	60.7
27	27-Oct-16	15:04	16:04	61.97	6.99	1.99	63.57	6.78	2.62	63.6
28	27-Oct-16	16:05	17:05	59.86	7.18	2.18	59.02	6.95	2.80	59.0
29	27-Oct-16	17:06	18:06	57.49	7.26	2.17	56.66	7.03	2.84	56.7
30	27-Oct-16	18:07	19:07	54.49	7.05	2.18	53.37	6.84	2.83	53.4
31	27-Oct-16	19:08	20:08	49.85	7.02	2.17	49.14	6.80	2.85	49.1
32	27-Oct-16	20:09	21:09	49.34	6.95	2.16	48.63	6.74	2.82	48.6
33	27-Oct-16	21:10	22:10	47.28	6.91	2.19	46.83	6.72	2.86	46.8
34	27-Oct-16	22:21	23:21	45.20	6.84	2.24	44.76	6.63	2.93	44.8
35	27-28/10/2016	23:22	0:22	39.84	8.32	2.18	39.04	8.15	2.87	39.0
36	02-Nov-16	11:26	12:26	72.15	6.55	2.54	73.06	6.39	2.67	73.1
37	02-Nov-16	12:27	13:27	67.11	6.51	2.49	70.17	6.32	2.77	70.2

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4A1.2.3 Table 4.2.3 -Standardised monitoring Data – Sulphur Dioxide, High Range**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(SO <sub>2</sub> mg/m <sup>3</sup> )	(SO <sub>2</sub> mg/m <sup>3</sup> )	(mg/m <sup>3</sup> )
1	26-Oct-16	11:23	12:23	192.8	194.6	8.7
2	26-Oct-16	12:34	13:34	195.0	202.7	9.1
3	26-Oct-16	13:35	14:35	190.3	201.8	9.1
4	26-Oct-16	14:36	15:36	192.5	203.0	9.1
5	26-Oct-16	15:37	16:37	194.1	202.6	9.1
6	26-Oct-16	16:38	17:38	195.1	200.5	9.0
7	26-Oct-16	17:39	18:39	194.8	200.4	9.0
8	26-Oct-16	18:40	19:40	190.6	186.8	8.4
9	26-Oct-16	19:41	20:41	183.1	172.8	7.8
10	26-Oct-16	20:42	21:42	175.6	168.0	7.6
11	26-Oct-16	21:53	22:53	172.8	163.1	7.3
12	26-Oct-16	22:54	23:54	170.8	160.1	7.2
13	26-27/10/2016	23:55	0:55	164.6	153.2	6.9
14	27-Oct-16	0:56	1:56	159.7	147.5	6.6
15	27-Oct-16	1:57	2:57	160.3	150.3	6.7
16	27-Oct-16	2:58	3:58	164.7	154.3	6.9
17	27-Oct-16	3:59	4:59	165.2	155.2	7.0
18	27-Oct-16	5:00	6:00	158.4	146.5	6.6
19	27-Oct-16	6:01	7:01	154.2	140.5	6.3
20	27-Oct-16	7:12	8:12	156.7	142.1	6.4
21	27-Oct-16	8:13	9:13	155.9	140.8	6.3
22	27-Oct-16	9:14	10:14	158.2	140.2	6.3
23	27-Oct-16	10:50	11:50	164.8	163.2	7.3
24	27-Oct-16	11:51	12:51	174.9	173.0	7.8
25	27-Oct-16	13:02	14:02	185.5	181.3	8.2
26	27-Oct-16	14:03	15:03	189.8	187.9	8.4
27	27-Oct-16	15:04	16:04	193.6	196.9	8.8
28	27-Oct-16	16:05	17:05	189.8	185.2	8.3
29	27-Oct-16	17:06	18:06	183.4	179.0	8.0
30	27-Oct-16	18:07	19:07	171.3	166.3	7.5
31	27-Oct-16	19:08	20:08	156.3	152.8	6.9
32	27-Oct-16	20:09	21:09	153.9	150.4	6.7
33	27-Oct-16	21:10	22:10	147.1	144.7	6.5
34	27-Oct-16	22:21	23:21	140.0	137.6	6.2
35	27-28/10/2016	23:22	0:22	137.8	134.2	6.0
36	02-Nov-16	11:26	12:26	219.6	220.2	9.9
37	02-Nov-16	12:27	13:27	203.6	210.7	9.4

**Note:**


Emission concentrations expresed at reference conditions 273K,

6 % Oxygen, dry gas

**4A1.3.3 – Plot 1.3 - Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide, High Range (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A1.4.3 – Elimination of Outliers – Sulphur Dioxide, High Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.

Additionally, any data points are not likely to be outliers unless they are more than two standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9565, an outlier test has not been undertaken.

#### 4A1.5.3 Determination of Method A, B or C - Sulphur Dioxide, High Range.

Test No	Test Start Time	Test End Time	SRM measured value (y)	SRM O2	SRM Standardised
	hr:min		(ppm)	(%)	(mg/m <sup>3</sup> )
1	11:23	12:23	63.3	6.6	194.6
2	12:34	13:34	66.3	6.6	202.7
3	13:35	14:35	67.2	6.3	201.8
4	14:36	15:36	67.4	6.4	203.0
5	15:37	16:37	66.7	6.5	202.6
6	16:38	17:38	64.9	6.7	200.5
7	17:39	18:39	64.6	6.8	200.4
8	18:40	19:40	60.0	6.8	186.8
9	19:41	20:41	55.6	6.8	172.8
10	20:42	21:42	54.1	6.8	168.0
11	21:53	22:53	53.0	6.7	163.1
12	22:54	23:54	52.0	6.7	160.1
13	23:55	0:55	49.4	6.8	153.2
14	0:56	1:56	47.2	6.9	147.5
15	1:57	2:57	48.4	6.8	150.3
16	2:58	3:58	49.5	6.9	154.3
17	3:59	4:59	50.0	6.8	155.2
18	5:00	6:00	47.2	6.8	146.5
19	6:01	7:01	45.4	6.8	140.5
20	7:12	8:12	46.1	6.7	142.1
21	8:13	9:13	45.5	6.8	140.8
22	9:14	10:14	44.8	6.9	140.2
23	10:50	11:50	53.1	6.7	163.2
24	11:51	12:51	56.5	6.6	173.0
25	13:02	14:02	58.7	6.8	181.3
26	14:03	15:03	60.7	6.8	187.9
27	15:04	16:04	63.6	6.8	196.9
28	16:05	17:05	59.0	6.9	185.2
29	17:06	18:06	56.7	7.0	179.0
30	18:07	19:07	53.4	6.8	166.3
31	19:08	20:08	49.1	6.8	152.8
32	20:09	21:09	48.6	6.7	150.4
33	21:10	22:10	46.8	6.7	144.7
34	22:21	23:21	44.8	6.6	137.6
35	23:22	0:22	39.0	8.1	134.2
36	11:26	12:26	73.1	6.4	220.2
37	12:27	13:27	70.2	6.3	210.7
Sum			2041.79		
Emission Limit Value (ELV) =	440	mg/Nm <sup>3</sup>	$Y_{\max}$	220.18	
Maximum Permissible uncertainty =	20	%	$Y_{\min}$	134.18	
Maximum Permissible uncertainty (at 15% of ELV) =	88	mg/Nm <sup>3</sup>			
15% of the ELV =	66	mg/Nm <sup>3</sup>			
Is $Y_{\max} - Y_{\min} > \text{MPU at ELV?}$	No		$Y_{\max} - Y_{\min}$	86.00	
Is $Y_{\min} > 15\%$ of ELV?	Yes				

Derivation of Calibration Function	Method B
------------------------------------	----------

**4A1.6.3 Table 4.3.3 - Data used to derive calibration function - Sulphur Dioxide, High Range.**

Test No	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)
	hr:min		(ppm)	(ppm)
1	Reference Zero		0.0	1.7
2	11:23	12:23	63.3	62.5
3	12:34	13:34	66.3	63.4
4	13:35	14:35	67.2	63.3
5	14:36	15:36	67.4	63.7
6	15:37	16:37	66.7	63.6
7	16:38	17:38	64.9	62.6
8	17:39	18:39	64.6	62.2
9	18:40	19:40	60.0	60.8
10	19:41	20:41	55.6	58.3
11	20:42	21:42	54.1	56.4
12	21:53	22:53	53.0	55.5
13	22:54	23:54	52.0	55.0
14	23:55	0:55	49.4	52.5
15	0:56	1:56	47.2	50.5
16	1:57	2:57	48.4	51.2
17	2:58	3:58	49.5	52.3
18	3:59	4:59	50.0	52.7
19	5:00	6:00	47.2	50.5
20	6:01	7:01	45.4	49.2
21	7:12	8:12	46.1	50.3
22	8:13	9:13	45.5	49.8
23	9:14	10:14	44.8	49.9
24	10:50	11:50	53.1	53.5
25	11:51	12:51	56.5	56.8
26	13:02	14:02	58.7	59.5
27	14:03	15:03	60.7	60.8
28	15:04	16:04	63.6	62.0
29	16:05	17:05	59.0	59.9
30	17:06	18:06	56.7	57.5
31	18:07	19:07	53.4	54.5
32	19:08	20:08	49.1	49.8
33	20:09	21:09	48.6	49.3
34	21:10	22:10	46.8	47.3
35	22:21	23:21	44.8	45.2
36	23:22	0:22	39.0	39.8
37	11:26	12:26	73.1	72.1
38	12:27	13:27	70.2	67.1
Sum			2041.79	2073.04

#### 4A1.7.3 Determination of Calibration Function - Sulphur Dioxide, High Range.

##### Method B

Formulae:- Number of tests conducted =	N	38
--	---	----

If Ymax-Ymin <15% of the ELV, the following formulae are used:

$$b=(y/(x-Z)) \quad x=(1/\text{No. AMS Tests}) * \text{Total AMS}$$

$$x=(1/ 38)*2073.0 \quad \text{or} - \quad x= 54.55$$

$$a=(-\beta.Z) \quad y=(1/\text{No. SRM Tests}) * \text{Total SRM}$$

$$y=(1/ 38)*2041. \quad \text{or} - \quad y= 53.73$$

The Slope is calculated by :

$$\beta=y/(x-Z) \quad 53.73 /( 54.55-1.67) \quad \beta= 1.016$$

The offset is calculated by:

$$\alpha=-\beta.Z \quad -1.01*1.67 \quad \alpha= -1.70$$

The calibration is function  $y_{im} = \alpha + \beta_{xi,m}$  or  $y_i = -1.69 + 1.01 * X_i$

**4A1.8.3 Table A.4.4.3 - Calculation of calibrated CEMS values - Sulphur Dioxide, High Range.**


Test No	Test Start Time	Test End Time	CEMS Raw Value (x) (ppm)	CEMS Calibrated signal (ppm)	CEMS Dry Oxygen (%)	CEMS Moisture (%)	CEMS Standardised Value (dry) (mg/Nm <sup>3</sup> )	CEMS Calibrated Standardised Value (mg/Nm <sup>3</sup> )	SRM Standardised (mg/m <sup>3</sup> )
1	Reference Zero		1.7	0.0	N/A	N/A	1.7	0.0	0.0
2	11:23	12:23	62.5	61.8	6.8	2.2	192.8	190.6	194.6
3	12:34	13:34	63.4	62.8	6.7	2.1	195.0	192.9	202.7
4	13:35	14:35	63.3	62.6	6.4	2.1	190.3	188.2	201.8
5	14:36	15:36	63.7	63.0	6.5	2.1	192.5	190.4	203.0
6	15:37	16:37	63.6	62.9	6.6	2.2	194.1	192.0	202.6
7	16:38	17:38	62.6	61.9	6.9	2.2	195.1	193.0	200.5
8	17:39	18:39	62.2	61.5	7.0	2.3	194.8	192.6	200.4
9	18:40	19:40	60.8	60.1	7.0	2.2	190.6	188.3	186.8
10	19:41	20:41	58.3	57.6	7.0	2.2	183.1	180.7	172.8
11	20:42	21:42	56.4	55.6	6.9	2.1	175.6	173.1	168.0
12	21:53	22:53	55.5	54.7	6.9	2.2	172.8	170.2	163.1
13	22:54	23:54	55.0	54.2	6.9	2.2	170.8	168.3	160.1
14	23:55	00:55	52.5	51.7	7.0	2.2	164.6	161.9	153.2
15	00:56	01:56	50.5	49.6	7.1	2.1	159.7	156.9	147.5
16	01:57	02:57	51.2	50.3	7.0	2.1	160.3	157.6	150.3
17	02:58	03:58	52.3	51.4	7.1	2.2	164.7	162.0	154.3
18	03:59	04:59	52.7	51.8	7.0	2.3	165.2	162.5	155.2
19	05:00	06:00	50.5	49.6	7.0	2.2	158.4	155.6	146.5
20	06:01	07:01	49.2	48.3	7.0	2.3	154.2	151.4	140.5
21	07:12	08:12	50.3	49.4	6.9	2.3	156.7	153.9	142.1
22	08:13	09:13	49.8	48.9	7.0	2.2	155.9	153.1	140.8
23	09:14	10:14	49.9	49.0	7.2	2.2	158.2	155.4	140.2
24	10:50	11:50	53.5	52.7	6.8	2.1	164.8	162.3	163.2
25	11:51	12:51	56.8	56.0	6.8	2.1	174.9	172.5	173.0
26	13:02	14:02	59.5	58.8	7.0	2.0	185.5	183.2	181.3
27	14:03	15:03	60.8	60.0	7.0	2.0	189.8	187.5	187.9
28	15:04	16:04	62.0	61.3	7.0	2.0	193.6	191.4	196.9
29	16:05	17:05	59.9	59.1	7.2	2.2	189.8	187.5	185.2
30	17:06	18:06	57.5	56.7	7.3	2.2	183.4	180.9	179.0
31	18:07	19:07	54.5	53.7	7.1	2.2	171.3	168.7	166.3
32	19:08	20:08	49.8	49.0	7.0	2.2	156.3	153.5	152.8
33	20:09	21:09	49.3	48.4	6.9	2.2	153.9	151.1	150.4
34	21:10	22:10	47.3	46.3	6.9	2.2	147.1	144.2	144.7
35	22:21	23:21	45.2	44.2	6.8	2.2	140.0	137.0	137.6
36	23:22	00:22	39.8	38.8	8.3	2.2	137.8	134.2	134.2
37	11:26	12:26	72.1	71.6	6.6	2.5	219.6	218.0	220.2
38	12:27	13:27	67.1	66.5	6.5	2.5	203.6	201.8	210.7
Sum							6458.51		
Emission Limit Value (ELV) = 440 mg/Nm <sup>3</sup>									

Reference Oxygen

6

%

**4A1.9.3 Plot 2.3 CEM versus SRM Parallel Test Data at CEM measurement conditions –Sulphur dioxide, wet gas.**




**4A1.10.3 Table 4.5.3 – Data used for the Variability Test – Sulphur Dioxide, High Range.**

Test No	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
	hr:min		mg/m3	mg/m3			
2	11:23	12:23	190.6	194.6	3.93	5.39	29.06
3	12:34	13:34	192.9	202.7	9.85	11.31	127.95
4	13:35	14:35	188.2	201.8	13.54	15.01	225.17
5	14:36	15:36	190.4	203.0	12.56	14.02	196.67
6	15:37	16:37	192.0	202.6	10.57	12.03	144.77
7	16:38	17:38	193.0	200.5	7.53	9.00	80.96
8	17:39	18:39	192.6	200.4	7.79	9.26	85.68
9	18:40	19:40	188.3	186.8	-1.49	-0.03	0.00
10	19:41	20:41	180.7	172.8	-7.91	-6.45	41.57
11	20:42	21:42	173.1	168.0	-5.17	-3.71	13.75
12	21:53	22:53	170.2	163.1	-7.19	-5.73	32.84
13	22:54	23:54	168.3	160.1	-8.18	-6.72	45.09
14	23:55	00:55	161.9	153.2	-8.75	-7.29	53.09
15	00:56	01:56	156.9	147.5	-9.36	-7.90	62.43
16	01:57	02:57	157.6	150.3	-7.30	-5.84	34.08
17	02:58	03:58	162.0	154.3	-7.74	-6.27	39.35
18	03:59	04:59	162.5	155.2	-7.30	-5.83	34.02
19	05:00	06:00	155.6	146.5	-9.05	-7.59	57.62
20	06:01	07:01	151.4	140.5	-10.82	-9.36	87.58
21	07:12	08:12	153.9	142.1	-11.76	-10.30	106.01
22	08:13	09:13	153.1	140.8	-12.30	-10.83	117.33
23	09:14	10:14	155.4	140.2	-15.22	-13.75	189.12
24	10:50	11:50	162.3	163.2	0.99	2.45	6.02
25	11:51	12:51	172.5	173.0	0.51	1.97	3.90
26	13:02	14:02	183.2	181.3	-1.90	-0.44	0.19
27	14:03	15:03	187.5	187.9	0.34	1.80	3.24
28	15:04	16:04	191.4	196.9	5.49	6.95	48.33
29	16:05	17:05	187.5	185.2	-2.29	-0.82	0.68
30	17:06	18:06	180.9	179.0	-1.97	-0.51	0.26
31	18:07	19:07	168.7	166.3	-2.37	-0.90	0.82
32	19:08	20:08	153.5	152.8	-0.75	0.72	0.51
33	20:09	21:09	151.1	150.4	-0.65	0.82	0.67
34	21:10	22:10	144.2	144.7	0.50	1.97	3.86
35	22:21	23:21	137.0	137.6	0.61	2.08	4.32
36	23:22	00:22	134.2	134.2	0.00	1.46	2.14
37	11:26	12:26	218.0	220.2	2.20	3.67	13.44
38	12:27	13:27	201.8	210.7	8.90	10.36	107.40
37 Tests		Mean			-1.46		
Sum							1999.90

#### 4A1.11.3 - Variability Test Calculation – Sulphur Dioxide, High Range.

SD=	$\text{Root}(1-\text{Number}).\text{Integral}(\text{D1}-\text{D})^2$	7.45	mg/m <sub>3</sub> (s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2*440 mg/m <sub>3</sub> (s,d,6%O <sub>2</sub> )/1.96	44.90	mg/m <sub>3</sub> (s,d),6%O <sub>2</sub>
For 37 tests, kv =	0.9885		
Therefore variability=		7.45 <= 44.9 * 0.9885	
or	7.45	<=	44.38
Which is TRUE therefore the CEMS passes the test			

#### 4A1.12.3 Plot 3.3 –Standardised CEM data versus standardised SRM - Sulphur dioxide, High Range – Reference conditions 273K, 101.3kPa., 6% oxygen, dry gas.



## **Section 4A2: Data & calculations – QAL2 – Unit 8, Procal 2**

**4A2.1 Table 4.1.1 – Raw monitoring Data – Oxides of Nitrogen**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (Wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw Value (dry)	SRM Oxygen (Dry)	SRM Moisture	SRM at CEMs Raw conditions
		hr:min		NO (ppm)	(%)	(%)	NOx (ppm)	(%)	(%)	NOx (ppm)
1	26-Oct-16	11:23	12:23	287.44	6.79	2.60	328.17	6.61	3.02	318.3
2	26-Oct-16	12:34	13:34	309.68	6.74	2.49	348.55	6.58	2.78	338.9
3	26-Oct-16	13:35	14:35	240.89	6.44	2.10	318.06	6.33	2.69	309.5
4	26-Oct-16	14:36	15:36	306.47	6.51	2.48	355.64	6.37	2.73	345.9
5	26-Oct-16	15:37	16:37	319.20	6.63	2.54	365.76	6.48	2.80	355.5
6	26-Oct-16	16:38	17:38	287.99	6.93	2.60	337.30	6.72	2.89	327.5
7	26-Oct-16	17:39	18:39	288.85	7.00	2.66	339.06	6.77	2.90	329.2
8	26-Oct-16	18:40	19:40	295.72	7.01	2.71	344.91	6.83	2.84	335.1
9	26-Oct-16	19:41	20:41	293.82	7.04	2.76	344.92	6.81	2.76	335.4
10	26-Oct-16	20:42	21:42	288.00	6.93	2.78	334.97	6.80	2.66	326.1
11	26-Oct-16	21:53	22:53	280.16	6.92	2.79	333.39	6.69	2.65	324.6
12	26-Oct-16	22:54	23:54	277.76	6.88	2.79	331.21	6.68	2.72	322.2
13	26-27/10/2016	23:55	0:55	273.58	7.01	2.73	331.05	6.78	2.66	322.2
14	27-Oct-16	0:56	1:56	280.66	7.15	2.66	331.55	6.93	2.59	323.0
15	27-Oct-16	1:57	2:57	280.74	7.01	2.44	326.22	6.83	2.62	317.7
16	27-Oct-16	2:58	3:58	278.98	7.08	2.35	326.35	6.85	2.73	317.5
17	27-Oct-16	3:59	4:59	282.66	7.01	2.40	329.32	6.79	2.80	320.1
18	27-Oct-16	5:00	6:00	286.19	7.01	2.45	335.27	6.79	2.74	326.1
19	27-Oct-16	6:01	7:01	292.48	6.99	2.50	337.56	6.77	2.73	328.4
20	27-Oct-16	7:12	8:12	280.62	6.93	2.54	329.93	6.71	2.73	320.9
21	27-Oct-16	8:13	9:13	284.77	7.00	2.55	330.14	6.78	2.67	321.3
22	27-Oct-16	9:14	10:14	271.92	7.16	2.56	316.48	6.93	2.68	308.0
23	27-Oct-16	10:50	11:50	273.91	6.79	2.57	313.70	6.68	2.58	305.6
24	27-Oct-16	11:51	12:51	279.28	6.78	2.53	324.34	6.58	2.84	315.1
25	27-Oct-16	13:02	14:02	265.67	6.96	2.43	302.95	6.75	2.63	295.0
26	27-Oct-16	14:03	15:03	266.19	7.00	2.15	307.08	6.78	2.63	299.0
27	27-Oct-16	15:04	16:04	267.87	6.99	2.12	312.45	6.78	2.62	304.3
28	27-Oct-16	16:05	17:05	265.10	7.18	2.19	314.18	6.95	2.80	305.4
29	27-Oct-16	17:06	18:06	269.37	7.26	2.27	317.90	7.03	2.84	308.9
30	27-Oct-16	18:07	19:07	264.62	7.05	2.34	313.87	6.84	2.83	305.0
31	27-Oct-16	19:08	20:08	261.75	7.02	2.38	312.16	6.80	2.85	303.3
32	27-Oct-16	20:09	21:09	272.19	6.95	2.40	322.07	6.74	2.82	313.0
33	27-Oct-16	21:10	22:10	274.67	6.91	2.42	326.00	6.72	2.86	316.7
34	27-Oct-16	22:21	23:21	259.82	6.84	2.45	309.61	6.63	2.93	300.5
35	27-28/10/2016	23:22	0:22	226.89	8.32	2.47	266.90	8.15	2.87	259.2
36	02-Nov-16	11:26	12:26	286.11	6.55	2.74	331.91	6.39	2.67	323.1
37	02-Nov-16	12:27	13:27	292.84	6.51	2.73	337.10	6.32	2.77	327.8

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4A2.2 Table 4.2.1 - Standardised monitoring Data – Oxides of Nitrogen**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(NOx as NO <sub>2</sub> mg/m <sup>3</sup> )	(NOx as NO <sub>2</sub> mg/m <sup>3</sup> )	(mg/m <sup>3</sup> )
1	26-Oct-16	11:23	12:23	639.7	702.6	23
2	26-Oct-16	12:34	13:34	686.4	744.7	24
3	26-Oct-16	13:35	14:35	520.8	667.9	22
4	26-Oct-16	14:36	15:36	668.4	749.0	25
5	26-Oct-16	15:37	16:37	702.5	776.3	25
6	26-Oct-16	16:38	17:38	647.5	727.6	24
7	26-Oct-16	17:39	18:39	653.0	734.1	24
8	26-Oct-16	18:40	19:40	669.7	750.2	25
9	26-Oct-16	19:41	20:41	666.9	749.1	25
10	26-Oct-16	20:42	21:42	648.7	727.0	24
11	26-Oct-16	21:53	22:53	630.8	718.0	24
12	26-Oct-16	22:54	23:54	623.4	712.7	23
13	26-27/10/2016	23:55	0:55	619.5	717.6	24
14	27-Oct-16	0:56	1:56	641.5	726.2	24
15	27-Oct-16	1:57	2:57	634.0	709.2	23
16	27-Oct-16	2:58	3:58	632.8	710.8	23
17	27-Oct-16	3:59	4:59	638.2	713.9	23
18	27-Oct-16	5:00	6:00	646.5	727.1	24
19	27-Oct-16	6:01	7:01	660.1	730.8	24
20	27-Oct-16	7:12	8:12	630.7	711.5	23
21	27-Oct-16	8:13	9:13	643.2	715.2	24
22	27-Oct-16	9:14	10:14	621.4	693.3	23
23	27-Oct-16	10:50	11:50	609.6	675.2	22
24	27-Oct-16	11:51	12:51	620.8	693.2	23
25	27-Oct-16	13:02	14:02	597.7	655.3	21
26	27-Oct-16	14:03	15:03	598.8	665.3	22
27	27-Oct-16	15:04	16:04	602.2	677.3	22
28	27-Oct-16	16:05	17:05	604.4	688.9	23
29	27-Oct-16	17:06	18:06	618.2	701.2	23
30	27-Oct-16	18:07	19:07	598.8	683.1	23
31	27-Oct-16	19:08	20:08	591.2	677.6	22
32	27-Oct-16	20:09	21:09	611.7	695.8	23
33	27-Oct-16	21:10	22:10	615.8	703.4	23
34	27-Oct-16	22:21	23:21	579.7	664.0	22
35	27-28/10/2016	23:22	0:22	565.9	640.4	21
36	02-Nov-16	11:26	12:26	627.3	699.7	23
37	02-Nov-16	12:27	13:27	640.3	707.3	23

Note:


Emission concentrations expresed at reference conditions 273K,

6 % Oxygen, dry gas

**4A2.3 – Plot 1.1 - Time Series of Standardised CEM versus Standardised SRM data – Oxides of Nitrogen, (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A2.4 – Elimination of Outliers – Oxides of Nitrogen



**4A2.4 (Cont...) – Elimination of Outliers – Oxides of Nitrogen**

Test No	Test Start Time	Test End Time	CEMS Raw Value	SRM Value at CEMS Raw conditions	Difference Di	Difference Di - $\bar{D}_i$	Is Result an Outlier - Di - $\bar{D}_i > 2SD$
	hr:min		(ppm)	(ppm)			
1	11:23	12:23	287.4	318.3	30.83	-8.10	No
2	12:34	13:34	309.7	338.9	29.18	-9.75	No
3	13:35	14:35	240.9	309.5	68.62	29.69	Yes
4	14:36	15:36	306.5	345.9	39.47	0.54	No
5	15:37	16:37	319.2	355.5	36.30	-2.63	No
6	16:38	17:38	288.0	327.5	39.56	0.63	No
7	17:39	18:39	288.8	329.2	40.39	1.46	No
8	18:40	19:40	295.7	335.1	39.39	0.47	No
9	19:41	20:41	293.8	335.4	41.58	2.65	No
10	20:42	21:42	288.0	326.1	38.06	-0.86	No
11	21:53	22:53	280.2	324.6	44.42	5.49	No
12	22:54	23:54	277.8	322.2	44.44	5.51	No
13	23:55	00:55	273.6	322.2	48.66	9.74	No
14	00:56	01:56	280.7	323.0	42.30	3.38	No
15	01:57	02:57	280.7	317.7	36.92	-2.01	No
16	02:58	03:58	279.0	317.5	38.47	-0.45	No
17	03:59	04:59	282.7	320.1	37.44	-1.48	No
18	05:00	06:00	286.2	326.1	39.91	0.98	No
19	06:01	07:01	292.5	328.4	35.88	-3.04	No
20	07:12	08:12	280.6	320.9	40.31	1.38	No
21	08:13	09:13	284.8	321.3	36.57	-2.36	No
22	09:14	10:14	271.9	308.0	36.08	-2.85	No
23	10:50	11:50	273.9	305.6	31.69	-7.23	No
24	11:51	12:51	279.3	315.1	35.87	-3.06	No
25	13:02	14:02	265.7	295.0	29.30	-9.62	No
26	14:03	15:03	266.2	299.0	32.81	-6.12	No
27	15:04	16:04	267.9	304.3	36.39	-2.53	No
28	16:05	17:05	265.1	305.4	40.29	1.37	No
29	17:06	18:06	269.4	308.9	39.50	0.57	No
30	18:07	19:07	264.6	305.0	40.35	1.43	No
31	19:08	20:08	261.8	303.3	41.51	2.59	No
32	20:09	21:09	272.2	313.0	40.80	1.87	No
33	21:10	22:10	274.7	316.7	42.02	3.09	No
34	22:21	23:21	259.8	300.5	40.73	1.80	No
35	23:22	00:22	226.9	259.2	32.34	-6.59	No
36	11:26	12:26	286.1	323.1	36.94	-1.98	No
37	12:27	13:27	292.8	327.8	34.93	-3.99	No
			Average $\bar{D}_i$	38.93			
			Standard Deviation	6.56			
			Standard Deviation x2	13.11			

#### 4A2.5.1 Determination of Method A, B or C - Oxides of Nitrogen

Test No	Test Start Time	Test End Time	SRM measured value (y) (ppm)	SRM O2 (%)	SRM Standardised (mg/m3)
	hr:min				
1	11:23	12:23	328.2	6.6	702.6
2	12:34	13:34	348.6	6.6	744.7
4	14:36	15:36	355.6	6.4	749.0
5	15:37	16:37	365.8	6.5	776.3
6	16:38	17:38	337.3	6.7	727.6
7	17:39	18:39	339.1	6.8	734.1
8	18:40	19:40	344.9	6.8	750.2
9	19:41	20:41	344.9	6.8	749.1
10	20:42	21:42	335.0	6.8	727.0
11	21:53	22:53	333.4	6.7	718.0
12	22:54	23:54	331.2	6.7	712.7
13	23:55	0:55	331.0	6.8	717.6
14	0:56	1:56	331.6	6.9	726.2
15	1:57	2:57	326.2	6.8	709.2
16	2:58	3:58	326.4	6.9	710.8
17	3:59	4:59	329.3	6.8	713.9
18	5:00	6:00	335.3	6.8	727.1
19	6:01	7:01	337.6	6.8	730.8
20	7:12	8:12	329.9	6.7	711.5
21	8:13	9:13	330.1	6.8	715.2
22	9:14	10:14	316.5	6.9	693.3
23	10:50	11:50	313.7	6.7	675.2
24	11:51	12:51	324.3	6.6	693.2
25	13:02	14:02	303.0	6.8	655.3
26	14:03	15:03	307.1	6.8	665.3
27	15:04	16:04	312.5	6.8	677.3
28	16:05	17:05	314.2	6.9	688.9
29	17:06	18:06	317.9	7.0	701.2
30	18:07	19:07	313.9	6.8	683.1
31	19:08	20:08	312.2	6.8	677.6
32	20:09	21:09	322.1	6.7	695.8
33	21:10	22:10	326.0	6.7	703.4
34	22:21	23:21	309.6	6.6	664.0
35	23:22	0:22	266.9	8.1	640.4
36	11:26	12:26	331.9	6.4	699.7
37	12:27	13:27	337.1	6.3	707.3
Sum			11770.01		
Emission Limit Value (ELV) =			1080	mg/Nm <sup>3</sup>	Y <sub>max</sub> 776.29
Maximum Permissible uncertainty =			20	%	Y <sub>min</sub> 640.38
Maximum Permissible uncertainty (at I)			216	mg/Nm <sup>3</sup>	
15% of the ELV =			162	mg/Nm <sup>3</sup>	
Is Y <sub>max</sub> - Y <sub>min</sub> > MPU at ELV?			No	Y <sub>max</sub> - Y <sub>min</sub>	135.91
Is Y <sub>min</sub> > 15% of ELV?			Yes		

Derivation of Calibration Function

Method B

**4A2.6.1 Table 4.3.1 - Data used to derive calibration function - Oxides of Nitrogen,**

Test No	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)
			(ppm)	(ppm)
1	Reference		0.0	2.3
2	11:23	12:23	318.3	287.4
3	12:34	13:34	338.9	309.7
5	14:36	15:36	345.9	306.5
6	15:37	16:37	355.5	319.2
7	16:38	17:38	327.5	288.0
8	17:39	18:39	329.2	288.8
9	18:40	19:40	335.1	295.7
10	19:41	20:41	335.4	293.8
11	20:42	21:42	326.1	288.0
12	21:53	22:53	324.6	280.2
13	22:54	23:54	322.2	277.8
14	23:55	0:55	322.2	273.6
15	0:56	1:56	323.0	280.7
16	1:57	2:57	317.7	280.7
17	2:58	3:58	317.5	279.0
18	3:59	4:59	320.1	282.7
19	5:00	6:00	326.1	286.2
20	6:01	7:01	328.4	292.5
21	7:12	8:12	320.9	280.6
22	8:13	9:13	321.3	284.8
23	9:14	10:14	308.0	271.9
24	10:50	11:50	305.6	273.9
25	11:51	12:51	315.1	279.3
26	13:02	14:02	295.0	265.7
27	14:03	15:03	299.0	266.2
28	15:04	16:04	304.3	267.9
29	16:05	17:05	305.4	265.1
30	17:06	18:06	308.9	269.4
31	18:07	19:07	305.0	264.6
32	19:08	20:08	303.3	261.8
33	20:09	21:09	313.0	272.2
34	21:10	22:10	316.7	274.7
35	22:21	23:21	300.5	259.8
36	23:22	0:22	259.2	226.9
37	11:26	12:26	323.1	286.1
38	12:27	13:27	327.8	292.8
Sum			11445.59	10076.30

#### 4A2.7.1 Determination of Calibration Function - Oxides of Nitrogen

##### Method B

Formulae:- Number of tests conducted =

N

37

If  $Y_{max} - Y_{min} < 15\%$  of the ELV, the following formulae are used:

$$b = (y/(x-Z)) \quad x = (1/\text{No. AMS Tests}) * \text{Total AMS}$$

$$x = (1/ 37) * 10076. \quad \text{or} - \quad x = 272.33$$

$$a = (-\beta \cdot Z) \quad y = (1/\text{No. SRM Tests}) * \text{Total SRM}$$

$$y = (1/ 37) * 11445 \quad \text{or} - \quad y = 309.34$$

The Slope is calculated by :

$$\beta = y/(x-Z) \quad 309.3 / ( 272.3 - 2.33 ) \quad \beta = 1.146$$

The offset is calculated by:

$$\alpha = -\beta \cdot Z \quad -1.14 * 2.33 \quad \alpha = -2.67$$

The calibration is function  $y_{im} = \alpha + \beta_{xi,m}$  or  $y_i = -2.66 + 1.14 * X_i$

**4A2.8.1 Table 4.4.1 - Calculation of calibrated CEMS values - Oxides of Nitrogen**

Test No	Test Start Time	Test End Time	CEMS Raw Value (x) (ppm)	CEMS Calibrated signal (ppm)	CEMS Dry Oxygen (%)	CEMS Moisture (%)	CEMS Standardised Value (dry) (mg/Nm <sup>3</sup> )	CEMS Calibrated Standardised Value (mg/Nm <sup>3</sup> )	SRM Standardised (mg/Nm <sup>3</sup> )
hr:min									
1	Reference		2.3	0.0	N/A	N/A	2.3	0.0	0.0
2	11:23	12:23	287.4	326.6	6.8	2.6	639.7	727.0	702.6
3	12:34	13:34	309.7	352.1	6.7	2.5	686.4	780.5	744.7
5	14:36	15:36	306.5	348.5	6.5	2.5	668.4	760.0	749.0
6	15:37	16:37	319.2	363.0	6.6	2.5	702.5	799.0	776.3
7	16:38	17:38	288.0	327.3	6.9	2.6	647.5	735.8	727.6
8	17:39	18:39	288.8	328.3	7.0	2.7	653.0	742.1	734.1
9	18:40	19:40	295.7	336.1	7.0	2.7	669.7	761.3	750.2
10	19:41	20:41	293.8	334.0	7.0	2.8	666.9	758.0	749.1
11	20:42	21:42	288.0	327.3	6.9	2.8	648.7	737.2	727.0
12	21:53	22:53	280.2	318.3	6.9	2.8	630.8	716.7	718.0
13	22:54	23:54	277.8	315.6	6.9	2.8	623.4	708.3	712.7
14	23:55	00:55	273.6	310.8	7.0	2.7	619.5	703.7	717.6
15	00:56	01:56	280.7	318.9	7.1	2.7	641.5	728.8	726.2
16	01:57	02:57	280.7	319.0	7.0	2.4	634.0	720.3	709.2
17	02:58	03:58	279.0	317.0	7.1	2.4	632.8	718.9	710.8
18	03:59	04:59	282.7	321.2	7.0	2.4	638.2	725.2	713.9
19	05:00	06:00	286.2	325.2	7.0	2.5	646.5	734.6	727.1
20	06:01	07:01	292.5	332.4	7.0	2.5	660.1	750.2	730.8
21	07:12	08:12	280.6	318.8	6.9	2.5	630.7	716.6	711.5
22	08:13	09:13	284.8	323.6	7.0	2.6	643.2	730.9	715.2
23	09:14	10:14	271.9	308.9	7.2	2.6	621.4	705.8	693.3
24	10:50	11:50	273.9	311.1	6.8	2.6	609.6	692.5	675.2
25	11:51	12:51	279.3	317.3	6.8	2.5	620.8	705.3	693.2
26	13:02	14:02	265.7	301.7	7.0	2.4	597.7	678.8	655.3
27	14:03	15:03	266.2	302.3	7.0	2.2	598.8	680.0	665.3
28	15:04	16:04	267.9	304.2	7.0	2.1	602.2	683.9	677.3
29	16:05	17:05	265.1	301.1	7.2	2.2	604.4	686.4	688.9
30	17:06	18:06	269.4	305.9	7.3	2.3	618.2	702.2	701.2
31	18:07	19:07	264.6	300.5	7.1	2.3	598.8	680.0	683.1
32	19:08	20:08	261.8	297.2	7.0	2.4	591.2	671.3	677.6
33	20:09	21:09	272.2	309.2	6.9	2.4	611.7	694.8	695.8
34	21:10	22:10	274.7	312.0	6.9	2.4	615.8	699.5	703.4
35	22:21	23:21	259.8	295.0	6.8	2.4	579.7	658.2	664.0
36	23:22	00:22	226.9	257.3	8.3	2.5	565.9	641.7	640.4
37	11:26	12:26	286.1	325.1	6.6	2.7	627.3	712.8	699.7
38	12:27	13:27	292.8	332.8	6.5	2.7	640.3	727.8	707.3
Sum							22689.78		


Emission Limit Value (ELV) = 1080 mg/Nm<sup>3</sup>

Reference Oxygen

6

%

**4A2.9.1 Plot 2.1 CEM versus SRM Parallel Test Data at CEM measurement conditions –NOx ppm, wet gas.**




**4A2.10.1 Table 4.5.1 – Data used for the Variability Test – Oxides of Nitrogen**

Test No	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
	hr:min		mg/m3	mg/m3			
2	11:23	12:23	727.0	702.6	-24.40	-16.02	256.55
3	12:34	13:34	780.5	744.7	-35.85	-27.48	754.91
5	14:36	15:36	760.0	749.0	-10.97	-2.59	6.72
6	15:37	16:37	799.0	776.3	-22.70	-14.32	205.09
7	16:38	17:38	735.8	727.6	-8.15	0.22	0.05
8	17:39	18:39	742.1	734.1	-7.99	0.39	0.15
9	18:40	19:40	761.3	750.2	-11.03	-2.66	7.05
10	19:41	20:41	758.0	749.1	-8.95	-0.57	0.33
11	20:42	21:42	737.2	727.0	-10.24	-1.86	3.46
12	21:53	22:53	716.7	718.0	1.33	9.71	94.31
13	22:54	23:54	708.3	712.7	4.43	12.81	163.97
14	23:55	00:55	703.7	717.6	13.89	22.27	496.04
15	00:56	01:56	728.8	726.2	-2.66	5.72	32.68
16	01:57	02:57	720.3	709.2	-11.10	-2.72	7.39
17	02:58	03:58	718.9	710.8	-8.12	0.26	0.07
18	03:59	04:59	725.2	713.9	-11.26	-2.88	8.31
19	05:00	06:00	734.6	727.1	-7.47	0.91	0.83
20	06:01	07:01	750.2	730.8	-19.47	-11.09	123.02
21	07:12	08:12	716.6	711.5	-5.05	3.33	11.07
22	08:13	09:13	730.9	715.2	-15.63	-7.25	52.60
23	09:14	10:14	705.8	693.3	-12.52	-4.14	17.15
24	10:50	11:50	692.5	675.2	-17.28	-8.90	79.22
25	11:51	12:51	705.3	693.2	-12.12	-3.74	14.00
26	13:02	14:02	678.8	655.3	-23.56	-15.19	230.61
27	14:03	15:03	680.0	665.3	-14.72	-6.34	40.18
28	15:04	16:04	683.9	677.3	-6.66	1.72	2.97
29	16:05	17:05	686.4	688.9	2.56	10.94	119.58
30	17:06	18:06	702.2	701.2	-0.98	7.40	54.70
31	18:07	19:07	680.0	683.1	3.06	11.44	130.76
32	19:08	20:08	671.3	677.6	6.26	14.64	214.21
33	20:09	21:09	694.8	695.8	0.96	9.34	87.24
34	21:10	22:10	699.5	703.4	3.87	12.25	150.14
35	22:21	23:21	658.2	664.0	5.82	14.20	201.60
36	23:22	00:22	641.7	640.4	-1.36	7.02	49.32
37	11:26	12:26	712.8	699.7	-13.08	-4.70	22.07
38	12:27	13:27	727.8	707.3	-20.49	-12.11	146.71
36 Tests		Mean			-8.38		
Sum							3785.05

#### 4A2.11.1 - Variability Test Calculation – Oxides of Nitrogen

SD=	$\text{Root}(1-\text{Number}).\text{Integral}(\text{D1}-\text{D})^2$	10.40	mg/m <sup>3</sup> (s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2*1080 mg/m <sup>3</sup> (s,d,6%O <sub>2</sub> )/1.96	110.20	mg/m <sup>3</sup> (s,d),6%O <sub>2</sub>
For 36 tests, kv =	0.9885		
Therefore variability=		10.4 <= 110.2 * 0.9885	
or	10.40	<=	108.94
Which is TRUE therefore the CEMS passes the test			

#### 4A2.12.1 Plot 3.1 –Standardised CEM data versus standardised SRM - Oxides of Nitrogen (as NO<sub>2</sub>) – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen



**4A2.1.2 Table 4.1.2 – Raw monitoring Data – Sulphur Dioxide, Low Range**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (wet)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMs Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	26-Oct-16	11:23	12:23	60.41	6.79	2.60	63.35	6.61	3.02	63.3
2	26-Oct-16	12:34	13:34	60.92	6.74	2.49	66.31	6.58	2.78	66.3
3	26-Oct-16	13:35	14:35	60.93	6.44	2.42	67.20	6.33	2.69	67.2
4	26-Oct-16	14:36	15:36	61.47	6.51	2.48	67.39	6.37	2.73	67.4
5	26-Oct-16	15:37	16:37	61.17	6.63	2.54	66.68	6.48	2.80	66.7
6	26-Oct-16	16:38	17:38	60.47	6.93	2.60	64.88	6.72	2.89	64.9
7	26-Oct-16	17:39	18:39	60.17	7.00	2.66	64.58	6.77	2.90	64.6
8	26-Oct-16	18:40	19:40	58.63	7.01	2.71	59.98	6.83	2.84	60.0
9	26-Oct-16	19:41	20:41	56.50	7.04	2.76	55.61	6.81	2.76	55.6
10	26-Oct-16	20:42	21:42	54.15	6.93	2.78	54.14	6.80	2.66	54.1
11	26-Oct-16	21:53	22:53	53.27	6.92	2.79	52.98	6.69	2.65	53.0
12	26-Oct-16	22:54	23:54	52.62	6.88	2.79	52.03	6.68	2.72	52.0
13	26-27/10/2016	23:55	0:55	50.63	7.01	2.73	49.45	6.78	2.66	49.4
14	27-Oct-16	0:56	1:56	49.60	7.15	2.66	47.16	6.93	2.59	47.2
15	27-Oct-16	1:57	2:57	48.29	7.01	2.44	48.38	6.83	2.62	48.4
16	27-Oct-16	2:58	3:58	49.78	7.08	2.35	49.52	6.85	2.73	49.5
17	27-Oct-16	3:59	4:59	48.93	7.01	2.40	50.03	6.79	2.80	50.0
18	27-Oct-16	5:00	6:00	47.63	7.01	2.45	47.23	6.79	2.74	47.2
19	27-Oct-16	6:01	7:01	45.47	6.99	2.50	45.39	6.77	2.73	45.4
20	27-Oct-16	7:12	8:12	46.44	6.93	2.54	46.08	6.71	2.73	46.1
21	27-Oct-16	8:13	9:13	46.93	7.00	2.55	45.45	6.78	2.67	45.5
22	27-Oct-16	9:14	10:14	46.00	7.16	2.56	44.75	6.93	2.68	44.8
23	27-Oct-16	10:50	11:50	50.19	6.79	2.57	53.11	6.68	2.58	53.1
24	27-Oct-16	11:51	12:51	53.35	6.78	2.53	56.52	6.58	2.84	56.5
25	27-Oct-16	13:02	14:02	55.44	6.96	2.43	58.67	6.75	2.63	58.7
26	27-Oct-16	14:03	15:03	57.72	7.00	2.15	60.69	6.78	2.63	60.7
27	27-Oct-16	15:04	16:04	60.30	6.99	2.12	63.57	6.78	2.62	63.6
28	27-Oct-16	16:05	17:05	58.02	7.18	2.19	59.02	6.95	2.80	59.0
29	27-Oct-16	17:06	18:06	56.06	7.26	2.27	56.66	7.03	2.84	56.7
30	27-Oct-16	18:07	19:07	52.75	7.05	2.34	53.37	6.84	2.83	53.4
31	27-Oct-16	19:08	20:08	47.71	7.02	2.38	49.14	6.80	2.85	49.1
32	27-Oct-16	20:09	21:09	46.91	6.95	2.40	48.63	6.74	2.82	48.6
33	27-Oct-16	21:10	22:10	44.58	6.91	2.42	46.83	6.72	2.86	46.8
34	27-Oct-16	22:21	23:21	42.31	6.84	2.45	44.76	6.63	2.93	44.8
35	27-28/10/2016	23:22	0:22	36.68	8.32	2.47	39.04	8.15	2.87	39.0
36	02-Nov-16	11:26	12:26	70.64	6.55	2.74	73.06	6.39	2.67	73.1
37	02-Nov-16	12:27	13:27	66.03	6.51	2.73	70.17	6.32	2.77	70.2

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4A2.2.2 Table 4.2.2 -Standardised monitoring Data – Sulphur Dioxide, Low Range**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(SO <sub>2</sub> mg/m <sup>3</sup> )	(SO <sub>2</sub> mg/m <sup>3</sup> )	(mg/m <sup>3</sup> )
1	26-Oct-16	11:23	12:23	187.1	194.6	8.7
2	26-Oct-16	12:34	13:34	187.9	202.7	9.1
3	26-Oct-16	13:35	14:35	183.9	201.8	9.1
4	26-Oct-16	14:36	15:36	186.5	203.0	9.1
5	26-Oct-16	15:37	16:37	187.3	202.6	9.1
6	26-Oct-16	16:38	17:38	189.2	200.5	9.0
7	26-Oct-16	17:39	18:39	189.2	200.4	9.0
8	26-Oct-16	18:40	19:40	184.8	186.8	8.4
9	26-Oct-16	19:41	20:41	178.4	172.8	7.8
10	26-Oct-16	20:42	21:42	169.7	168.0	7.6
11	26-Oct-16	21:53	22:53	166.9	163.1	7.3
12	26-Oct-16	22:54	23:54	164.3	160.1	7.2
13	26-27/10/2016	23:55	0:55	159.5	153.2	6.9
14	27-Oct-16	0:56	1:56	157.7	147.5	6.6
15	27-Oct-16	1:57	2:57	151.7	150.3	6.7
16	27-Oct-16	2:58	3:58	157.1	154.3	6.9
17	27-Oct-16	3:59	4:59	153.7	155.2	7.0
18	27-Oct-16	5:00	6:00	149.7	146.5	6.6
19	27-Oct-16	6:01	7:01	142.8	140.5	6.3
20	27-Oct-16	7:12	8:12	145.2	142.1	6.4
21	27-Oct-16	8:13	9:13	147.5	140.8	6.3
22	27-Oct-16	9:14	10:14	146.3	140.2	6.3
23	27-Oct-16	10:50	11:50	155.4	163.2	7.3
24	27-Oct-16	11:51	12:51	165.0	173.0	7.8
25	27-Oct-16	13:02	14:02	173.6	181.3	8.2
26	27-Oct-16	14:03	15:03	180.6	187.9	8.4
27	27-Oct-16	15:04	16:04	188.6	196.9	8.8
28	27-Oct-16	16:05	17:05	184.0	185.2	8.3
29	27-Oct-16	17:06	18:06	179.0	179.0	8.0
30	27-Oct-16	18:07	19:07	166.1	166.3	7.5
31	27-Oct-16	19:08	20:08	149.9	152.8	6.9
32	27-Oct-16	20:09	21:09	146.7	150.4	6.7
33	27-Oct-16	21:10	22:10	139.0	144.7	6.5
34	27-Oct-16	22:21	23:21	131.4	137.6	6.2
35	27-28/10/2016	23:22	0:22	127.3	134.2	6.0
36	02-Nov-16	11:26	12:26	215.5	220.2	9.9
37	02-Nov-16	12:27	13:27	200.9	210.7	9.4

*Note:*


Emission concentrations expresed at reference conditions 273K,

6 % Oxygen, dry gas

**4A2.3.2 – Plot 1.2 - Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide, Low Range (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A2.4.2 – Elimination of Outliers – Sulphur Dioxide, Low Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.'

Additionally, any data points are not likely to be outliers unless they are more than two standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9511, an outlier test has not been undertaken.

**4A2.5.2 Determination of Method A, B or C - Sulphur Dioxide, Low Range.**

Test No	Test Start Time	Test End Time	SRM measured value (y)	SRM O2	SRM Standardised
	hr:min		(ppm)	(%)	(mg/m <sup>3</sup> )
1	11:23	12:23	63.3	6.6	194.6
2	12:34	13:34	66.3	6.6	202.7
3	13:35	14:35	67.2	6.3	201.8
4	14:36	15:36	67.4	6.4	203.0
5	15:37	16:37	66.7	6.5	202.6
6	16:38	17:38	64.9	6.7	200.5
7	17:39	18:39	64.6	6.8	200.4
8	18:40	19:40	60.0	6.8	186.8
9	19:41	20:41	55.6	6.8	172.8
10	20:42	21:42	54.1	6.8	168.0
11	21:53	22:53	53.0	6.7	163.1
12	22:54	23:54	52.0	6.7	160.1
13	23:55	0:55	49.4	6.8	153.2
14	0:56	1:56	47.2	6.9	147.5
15	1:57	2:57	48.4	6.8	150.3
16	2:58	3:58	49.5	6.9	154.3
17	3:59	4:59	50.0	6.8	155.2
18	5:00	6:00	47.2	6.8	146.5
19	6:01	7:01	45.4	6.8	140.5
20	7:12	8:12	46.1	6.7	142.1
21	8:13	9:13	45.5	6.8	140.8
22	9:14	10:14	44.8	6.9	140.2
23	10:50	11:50	53.1	6.7	163.2
24	11:51	12:51	56.5	6.6	173.0
25	13:02	14:02	58.7	6.8	181.3
26	14:03	15:03	60.7	6.8	187.9
27	15:04	16:04	63.6	6.8	196.9
28	16:05	17:05	59.0	6.9	185.2
29	17:06	18:06	56.7	7.0	179.0
30	18:07	19:07	53.4	6.8	166.3
31	19:08	20:08	49.1	6.8	152.8
32	20:09	21:09	48.6	6.7	150.4
33	21:10	22:10	46.8	6.7	144.7
34	22:21	23:21	44.8	6.6	137.6
35	23:22	0:22	39.0	8.1	134.2
36	11:26	12:26	73.1	6.4	220.2
37	12:27	13:27	70.2	6.3	210.7
Sum			2041.79		
Emission Limit Value (ELV) =			440 mg/Nm <sup>3</sup>	Y <sub>max</sub>	220.18
Maximum Permissible uncertainty =			20 %	Y <sub>min</sub>	134.18
Maximum Permissible uncertainty (at 15% of ELV) =			88 mg/Nm <sup>3</sup>		
15% of the ELV =			66 mg/Nm <sup>3</sup>		
Is Y <sub>max</sub> - Y <sub>min</sub> > MPU at ELV?			No	Y <sub>max</sub> - Y <sub>min</sub>	86.00
Is Y <sub>min</sub> > 15% of ELV?			Yes		

Derivation of Calibration Function	Method B
------------------------------------	----------

**4A2.6.2 Table 4.3.2 - Data used to derive calibration function - Sulphur Dioxide, Low Range.**

Test No	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)
	hr:min		(ppm)	(ppm)
1	Reference Zero		0.0	-1.0
2	11:23	12:23	63.3	60.4
3	12:34	13:34	66.3	60.9
4	13:35	14:35	67.2	60.9
5	14:36	15:36	67.4	61.5
6	15:37	16:37	66.7	61.2
7	16:38	17:38	64.9	60.5
8	17:39	18:39	64.6	60.2
9	18:40	19:40	60.0	58.6
10	19:41	20:41	55.6	56.5
11	20:42	21:42	54.1	54.2
12	21:53	22:53	53.0	53.3
13	22:54	23:54	52.0	52.6
14	23:55	0:55	49.4	50.6
15	0:56	1:56	47.2	49.6
16	1:57	2:57	48.4	48.3
17	2:58	3:58	49.5	49.8
18	3:59	4:59	50.0	48.9
19	5:00	6:00	47.2	47.6
20	6:01	7:01	45.4	45.5
21	7:12	8:12	46.1	46.4
22	8:13	9:13	45.5	46.9
23	9:14	10:14	44.8	46.0
24	10:50	11:50	53.1	50.2
25	11:51	12:51	56.5	53.4
26	13:02	14:02	58.7	55.4
27	14:03	15:03	60.7	57.7
28	15:04	16:04	63.6	60.3
29	16:05	17:05	59.0	58.0
30	17:06	18:06	56.7	56.1
31	18:07	19:07	53.4	52.7
32	19:08	20:08	49.1	47.7
33	20:09	21:09	48.6	46.9
34	21:10	22:10	46.8	44.6
35	22:21	23:21	44.8	42.3
36	23:22	0:22	39.0	36.7
37	11:26	12:26	73.1	70.6
38	12:27	13:27	70.2	66.0
Sum			2041.79	1978.13

#### 4A2.7.2 Determination of Calibration Function - Sulphur Dioxide, Low Range.

##### Method B

Formulae:- Number of tests conducted =	N	38
If Ymax-Ymin <15% of the ELV, the following formulae are used: $b=(y/(x-Z))$ $x=(1/\text{No. AMS Tests}) * \text{Total AMS}$		
$x=(1/ 38)*1978.1$ or -	x=	52.06
$a=(-\beta.Z)$ $y=(1/\text{No. SRM Tests}) * \text{Total SRM}$ $y=(1/ 38)*2041.$ or -	y=	53.73
The Slope is calculated by: $\beta=y/(x-Z) = 53.73 / ( 52.05-1)$	$\beta=$	1.013
The offset is calculated by: $\alpha=-\beta.Z = -1.01*1$	$\alpha=$	1.01
The calibration is function $y_{im} = \alpha + \beta_{xi,m}$ or	$y_i =$	$1.012 + 1.01 * X_i$

**4A2.8.2 Table A.4.4.2 - Calculation of calibrated CEMS values - Sulphur Dioxide, Low Range**

Test No	Test Start Time	Test End Time	CEMS Raw Value (x) (ppm)	CEMS Calibrated signal (ppm)	CEMS Dry Oxygen (%)	CEMS Moisture (%)	CEMS Standardised Value (dry) (mg/Nm <sup>3</sup> )	CEMS Calibrated Standardised Value (mg/Nm <sup>3</sup> )	SRM Standardised (mg/m <sup>3</sup> )
1	Reference Zero		-1.0	0.0	N/A	N/A	-1.0	0.0	0.0
2	11:23	12:23	60.4	62.2	6.8	2.6	187.1	192.6	194.6
3	12:34	13:34	60.9	62.7	6.7	2.5	187.9	193.4	202.7
4	13:35	14:35	60.9	62.7	6.4	2.4	183.9	189.3	201.8
5	14:36	15:36	61.5	63.3	6.5	2.5	186.5	192.0	203.0
6	15:37	16:37	61.2	63.0	6.6	2.5	187.3	192.8	202.6
7	16:38	17:38	60.5	62.3	6.9	2.6	189.2	194.7	200.5
8	17:39	18:39	60.2	61.9	7.0	2.7	189.2	194.8	200.4
9	18:40	19:40	58.6	60.4	7.0	2.7	184.8	190.3	186.8
10	19:41	20:41	56.5	58.2	7.0	2.8	178.4	183.9	172.8
11	20:42	21:42	54.2	55.9	6.9	2.8	169.7	175.1	168.0
12	21:53	22:53	53.3	55.0	6.9	2.8	166.9	172.2	163.1
13	22:54	23:54	52.6	54.3	6.9	2.8	164.3	169.6	160.1
14	23:55	00:55	50.6	52.3	7.0	2.7	159.5	164.7	153.2
15	00:56	01:56	49.6	51.2	7.1	2.7	157.7	163.0	147.5
16	01:57	02:57	48.3	49.9	7.0	2.4	151.7	156.8	150.3
17	02:58	03:58	49.8	51.4	7.1	2.4	157.1	162.3	154.3
18	03:59	04:59	48.9	50.6	7.0	2.4	153.7	158.9	155.2
19	05:00	06:00	47.6	49.3	7.0	2.5	149.7	154.8	146.5
20	06:01	07:01	45.5	47.1	7.0	2.5	142.8	147.8	140.5
21	07:12	08:12	46.4	48.0	6.9	2.5	145.2	150.2	142.1
22	08:13	09:13	46.9	48.5	7.0	2.6	147.5	152.6	140.8
23	09:14	10:14	46.0	47.6	7.2	2.6	146.3	151.3	140.2
24	10:50	11:50	50.2	51.8	6.8	2.6	155.4	160.5	163.2
25	11:51	12:51	53.4	55.0	6.8	2.5	165.0	170.2	173.0
26	13:02	14:02	55.4	57.2	7.0	2.4	173.6	178.9	181.3
27	14:03	15:03	57.7	59.5	7.0	2.2	180.6	186.1	187.9
28	15:04	16:04	60.3	62.1	7.0	2.1	188.6	194.2	196.9
29	16:05	17:05	58.0	59.8	7.2	2.2	184.0	189.6	185.2
30	17:06	18:06	56.1	57.8	7.3	2.3	179.0	184.5	179.0
31	18:07	19:07	52.7	54.4	7.1	2.3	166.1	171.4	166.3
32	19:08	20:08	47.7	49.3	7.0	2.4	149.9	155.0	152.8
33	20:09	21:09	46.9	48.5	6.9	2.4	146.7	151.7	150.4
34	21:10	22:10	44.6	46.2	6.9	2.4	139.0	144.0	144.7
35	22:21	23:21	42.3	43.9	6.8	2.4	131.4	136.2	137.6
36	23:22	00:22	36.7	38.2	8.3	2.5	127.3	132.4	134.2
37	11:26	12:26	70.6	72.5	6.6	2.7	215.5	221.3	220.2
38	12:27	13:27	66.0	67.9	6.5	2.7	200.9	206.5	210.7
Sum							6188.32		


Emission Limit Value (ELV) = 440 mg/Nm<sup>3</sup>

Reference Oxygen

6

%

**4A2.9.2 Plot 2.2 CEM versus SRM Parallel Test Data at CEM measurement conditions –Sulphur dioxide, wet gas**




**4A2.10.2 Table 4.5.2 – Data used for the Variability Test – Sulphur Dioxide, Low Range.**

Test No	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
	hr:min		mg/m3	mg/m3			
2	11:23	12:23	192.6	194.6	2.00	4.03	16.28
3	12:34	13:34	193.4	202.7	9.34	11.37	129.33
4	13:35	14:35	189.3	201.8	12.51	14.54	211.49
5	14:36	15:36	192.0	203.0	11.01	13.05	170.22
6	15:37	16:37	192.8	202.6	9.80	11.84	140.07
7	16:38	17:38	194.7	200.5	5.79	7.83	61.24
8	17:39	18:39	194.8	200.4	5.53	7.56	57.16
9	18:40	19:40	190.3	186.8	-3.47	-1.43	2.05
10	19:41	20:41	183.9	172.8	-11.13	-9.09	82.65
11	20:42	21:42	175.1	168.0	-7.09	-5.06	25.60
12	21:53	22:53	172.2	163.1	-9.11	-7.08	50.10
13	22:54	23:54	169.6	160.1	-9.46	-7.42	55.10
14	23:55	00:55	164.7	153.2	-11.54	-9.50	90.31
15	00:56	01:56	163.0	147.5	-15.44	-13.40	179.59
16	01:57	02:57	156.8	150.3	-6.54	-4.50	20.27
17	02:58	03:58	162.3	154.3	-8.04	-6.00	36.00
18	03:59	04:59	158.9	155.2	-3.60	-1.57	2.46
19	05:00	06:00	154.8	146.5	-8.26	-6.22	38.73
20	06:01	07:01	147.8	140.5	-7.25	-5.21	27.15
21	07:12	08:12	150.2	142.1	-8.09	-6.06	36.69
22	08:13	09:13	152.6	140.8	-11.79	-9.75	95.10
23	09:14	10:14	151.3	140.2	-11.18	-9.14	83.57
24	10:50	11:50	160.5	163.2	2.73	4.77	22.72
25	11:51	12:51	170.2	173.0	2.73	4.76	22.70
26	13:02	14:02	178.9	181.3	2.37	4.41	19.45
27	14:03	15:03	186.1	187.9	1.77	3.81	14.49
28	15:04	16:04	194.2	196.9	2.70	4.73	22.41
29	16:05	17:05	189.6	185.2	-4.37	-2.33	5.45
30	17:06	18:06	184.5	179.0	-5.56	-3.52	12.42
31	18:07	19:07	171.4	166.3	-5.05	-3.01	9.09
32	19:08	20:08	155.0	152.8	-2.25	-0.22	0.05
33	20:09	21:09	151.7	150.4	-1.30	0.73	0.54
34	21:10	22:10	144.0	144.7	0.72	2.76	7.62
35	22:21	23:21	136.2	137.6	1.43	3.47	12.02
36	23:22	00:22	132.4	134.2	1.75	3.79	14.35
37	11:26	12:26	221.3	220.2	-1.12	0.92	0.85
38	12:27	13:27	206.5	210.7	4.13	6.17	38.07
37 Tests		Mean			-2.04		
	Sum						1813.38

#### 4A2.11.2 - Variability Test Calculation – Sulphur Dioxide, Low Range.

SD=	$\text{Root}(1-\text{Number}).\text{Integral}(\text{D1}-\text{D})^2$	7.10	mg/m <sup>3</sup> (s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2 * 440 mg/m <sup>3</sup> (s,d,6%O <sub>2</sub> ) / 1.96	44.90	mg/m <sup>3</sup> (s,d),6%O <sub>2</sub>
For 37 tests, kv =	0.9885		
Therefore variability=		7.1 <= 44.9 * 0.9885	
or	7.10	<=	44.38
Which is TRUE therefore the CEMS passes the test			

#### 4A2.12.2 Plot 3.2 –Standardised CEM data versus standardised SRM - Sulphur dioxide, Low Range – Reference conditions 273K, 101.3kPa 6% oxygen, dry gas.



**4A2.1.3 Table 4.1.3– Raw monitoring Data – Sulphur Dioxide, High Range**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (wet)	CEMS Oxygen (dry)	CEMS Moisture	SRM Raw value (wet)	SRM Oxygen (dry)	SRM Moisture	SRM at CEMs Raw conditions
		hr:min		(ppm)	(%)	(%)	(ppm)	(%)	(%)	(ppm)
1	26-Oct-16	11:23	12:23	59.23	6.79	2.60	63.35	6.61	3.02	63.3
2	26-Oct-16	12:34	13:34	59.68	6.74	2.49	66.31	6.58	2.78	66.3
3	26-Oct-16	13:35	14:35	59.78	6.44	2.42	67.20	6.33	2.69	67.2
4	26-Oct-16	14:36	15:36	60.25	6.51	2.48	67.39	6.37	2.73	67.4
5	26-Oct-16	15:37	16:37	59.94	6.63	2.54	66.68	6.48	2.80	66.7
6	26-Oct-16	16:38	17:38	59.37	6.93	2.60	64.88	6.72	2.89	64.9
7	26-Oct-16	17:39	18:39	59.06	7.00	2.66	64.58	6.77	2.90	64.6
8	26-Oct-16	18:40	19:40	57.45	7.01	2.71	59.98	6.83	2.84	60.0
9	26-Oct-16	19:41	20:41	55.36	7.04	2.76	55.61	6.81	2.76	55.6
10	26-Oct-16	20:42	21:42	52.98	6.93	2.78	54.14	6.80	2.66	54.1
11	26-Oct-16	21:53	22:53	52.15	6.92	2.79	52.98	6.69	2.65	53.0
12	26-Oct-16	22:54	23:54	51.47	6.88	2.79	52.03	6.68	2.72	52.0
13	26-27/10/2016	23:55	0:55	49.62	7.01	2.73	49.45	6.78	2.66	49.4
14	27-Oct-16	0:56	1:56	48.46	7.15	2.66	47.16	6.93	2.59	47.2
15	27-Oct-16	1:57	2:57	47.16	7.01	2.44	48.38	6.83	2.62	48.4
16	27-Oct-16	2:58	3:58	48.80	7.08	2.35	49.52	6.85	2.73	49.5
17	27-Oct-16	3:59	4:59	47.87	7.01	2.40	50.03	6.79	2.80	50.0
18	27-Oct-16	5:00	6:00	46.64	7.01	2.45	47.23	6.79	2.74	47.2
19	27-Oct-16	6:01	7:01	44.44	6.99	2.50	45.39	6.77	2.73	45.4
20	27-Oct-16	7:12	8:12	45.44	6.93	2.54	46.08	6.71	2.73	46.1
21	27-Oct-16	8:13	9:13	46.03	7.00	2.55	45.45	6.78	2.67	45.5
22	27-Oct-16	9:14	10:14	44.97	7.16	2.56	44.75	6.93	2.68	44.8
23	27-Oct-16	10:50	11:50	49.21	6.79	2.57	53.11	6.68	2.58	53.1
24	27-Oct-16	11:51	12:51	52.32	6.78	2.53	56.52	6.58	2.84	56.5
25	27-Oct-16	13:02	14:02	54.38	6.96	2.43	58.67	6.75	2.63	58.7
26	27-Oct-16	14:03	15:03	56.43	7.00	2.15	60.69	6.78	2.63	60.7
27	27-Oct-16	15:04	16:04	59.22	6.99	2.12	63.57	6.78	2.62	63.6
28	27-Oct-16	16:05	17:05	56.78	7.18	2.19	59.02	6.95	2.80	59.0
29	27-Oct-16	17:06	18:06	54.89	7.26	2.27	56.66	7.03	2.84	56.7
30	27-Oct-16	18:07	19:07	51.72	7.05	2.34	53.37	6.84	2.83	53.4
31	27-Oct-16	19:08	20:08	46.69	7.02	2.38	49.14	6.80	2.85	49.1
32	27-Oct-16	20:09	21:09	45.92	6.95	2.40	48.63	6.74	2.82	48.6
33	27-Oct-16	21:10	22:10	43.60	6.91	2.42	46.83	6.72	2.86	46.8
34	27-Oct-16	22:21	23:21	41.37	6.84	2.45	44.76	6.63	2.93	44.8
35	27-28/10/2016	23:22	0:22	35.75	8.32	2.47	39.04	8.15	2.87	39.0
36	02-Nov-16	11:26	12:26	69.40	6.55	2.74	73.06	6.39	2.67	73.1
37	02-Nov-16	12:27	13:27	64.70	6.51	2.73	70.17	6.32	2.77	70.2

Note:

Emission concentrations expressed at reference conditions 273K, 101.3kPa

**4A2.2.3 Table 4.2.3 -Standardised monitoring Data – Sulphur Dioxide, High Range**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised	SRM Standardised	SRM Uncertainty
		hr:min		(SO <sub>2</sub> mg/m <sup>3</sup> )	(SO <sub>2</sub> mg/m <sup>3</sup> )	(mg/m <sup>3</sup> )
1	26-Oct-16	11:23	12:23	183.4	194.6	8.7
2	26-Oct-16	12:34	13:34	184.0	202.7	9.1
3	26-Oct-16	13:35	14:35	180.4	201.8	9.1
4	26-Oct-16	14:36	15:36	182.8	203.0	9.1
5	26-Oct-16	15:37	16:37	183.5	202.6	9.1
6	26-Oct-16	16:38	17:38	185.7	200.5	9.0
7	26-Oct-16	17:39	18:39	185.8	200.4	9.0
8	26-Oct-16	18:40	19:40	181.0	186.8	8.4
9	26-Oct-16	19:41	20:41	174.8	172.8	7.8
10	26-Oct-16	20:42	21:42	166.0	168.0	7.6
11	26-Oct-16	21:53	22:53	163.4	163.1	7.3
12	26-Oct-16	22:54	23:54	160.7	160.1	7.2
13	26-27/10/2016	23:55	0:55	156.3	153.2	6.9
14	27-Oct-16	0:56	1:56	154.1	147.5	6.6
15	27-Oct-16	1:57	2:57	148.2	150.3	6.7
16	27-Oct-16	2:58	3:58	154.0	154.3	6.9
17	27-Oct-16	3:59	4:59	150.4	155.2	7.0
18	27-Oct-16	5:00	6:00	146.6	146.5	6.6
19	27-Oct-16	6:01	7:01	139.6	140.5	6.3
20	27-Oct-16	7:12	8:12	142.1	142.1	6.4
21	27-Oct-16	8:13	9:13	144.6	140.8	6.3
22	27-Oct-16	9:14	10:14	143.0	140.2	6.3
23	27-Oct-16	10:50	11:50	152.4	163.2	7.3
24	27-Oct-16	11:51	12:51	161.8	173.0	7.8
25	27-Oct-16	13:02	14:02	170.2	181.3	8.2
26	27-Oct-16	14:03	15:03	176.6	187.9	8.4
27	27-Oct-16	15:04	16:04	185.2	196.9	8.8
28	27-Oct-16	16:05	17:05	180.1	185.2	8.3
29	27-Oct-16	17:06	18:06	175.3	179.0	8.0
30	27-Oct-16	18:07	19:07	162.8	166.3	7.5
31	27-Oct-16	19:08	20:08	146.7	152.8	6.9
32	27-Oct-16	20:09	21:09	143.6	150.4	6.7
33	27-Oct-16	21:10	22:10	136.0	144.7	6.5
34	27-Oct-16	22:21	23:21	128.4	137.6	6.2
35	27-28/10/2016	23:22	0:22	124.1	134.2	6.0
36	02-Nov-16	11:26	12:26	211.7	220.2	9.9
37	02-Nov-16	12:27	13:27	196.8	210.7	9.4

*Note:*


Emission concentrations expresed at reference conditions 273K,

6 % Oxygen, dry gas

**4A2.3.3 – Plot 1.3 - Time Series of Standardised CEM versus Standardised SRM data – Sulphur Dioxide, High Range (Expressed at reference conditions 273K, 101.3kPa, dry gas, 6% oxygen)**



#### 4A2.4.3 – Elimination of Outliers – Sulphur Dioxide, High Range.



Guidance on performing tests for outliers in MID 14181 section 6.3, states the following:

'As a general guide, when plotting the raw SRM and raw CEM data, if the  $R^2$  value for the linear regression line is equal or more than 0.9, then it is not ordinarily necessary to perform an outlier test.'

'Additionally, any data points are not likely to be outliers unless they are more than two standard deviations from the regression line'

As the  $R^2$  value for this determinand was 0.9515, an outlier test has not been undertaken.

#### 4A2.5.3 Determination of Method A, B or C - Sulphur Dioxide, High Range.

Test No	Test Start Time	Test End Time	SRM measured value (y) (ppm)	SRM O2 (%)	SRM Standardised (mg/m3)
	hr:min				
1	11:23	12:23	63.3	6.6	194.6
2	12:34	13:34	66.3	6.6	202.7
3	13:35	14:35	67.2	6.3	201.8
4	14:36	15:36	67.4	6.4	203.0
5	15:37	16:37	66.7	6.5	202.6
6	16:38	17:38	64.9	6.7	200.5
7	17:39	18:39	64.6	6.8	200.4
8	18:40	19:40	60.0	6.8	186.8
9	19:41	20:41	55.6	6.8	172.8
10	20:42	21:42	54.1	6.8	168.0
11	21:53	22:53	53.0	6.7	163.1
12	22:54	23:54	52.0	6.7	160.1
13	23:55	0:55	49.4	6.8	153.2
14	0:56	1:56	47.2	6.9	147.5
15	1:57	2:57	48.4	6.8	150.3
16	2:58	3:58	49.5	6.9	154.3
17	3:59	4:59	50.0	6.8	155.2
18	5:00	6:00	47.2	6.8	146.5
19	6:01	7:01	45.4	6.8	140.5
20	7:12	8:12	46.1	6.7	142.1
21	8:13	9:13	45.5	6.8	140.8
22	9:14	10:14	44.8	6.9	140.2
23	10:50	11:50	53.1	6.7	163.2
24	11:51	12:51	56.5	6.6	173.0
25	13:02	14:02	58.7	6.8	181.3
26	14:03	15:03	60.7	6.8	187.9
27	15:04	16:04	63.6	6.8	196.9
28	16:05	17:05	59.0	6.9	185.2
29	17:06	18:06	56.7	7.0	179.0
30	18:07	19:07	53.4	6.8	166.3
31	19:08	20:08	49.1	6.8	152.8
32	20:09	21:09	48.6	6.7	150.4
33	21:10	22:10	46.8	6.7	144.7
34	22:21	23:21	44.8	6.6	137.6
35	23:22	0:22	39.0	8.1	134.2
36	11:26	12:26	73.1	6.4	220.2
37	12:27	13:27	70.2	6.3	210.7
Sum			2041.79		
Emission Limit Value (ELV) =	440	mg/Nm <sup>3</sup>	$Y_{\max}$	220.18	
Maximum Permissible uncertainty =	20	%	$Y_{\min}$	134.18	
Maximum Permissible uncertainty (at l)	88	mg/Nm <sup>3</sup>			
15% of the ELV =	66	mg/Nm <sup>3</sup>			
Is $Y_{\max} - Y_{\min} > \text{MPU at ELV?}$	No		$Y_{\max} - Y_{\min}$	86.00	
Is $Y_{\min} > 15\%$ of ELV?	Yes				

Derivation of Calibration Function	Method B
------------------------------------	----------

**4A2.6.3 Table 4.3.3 - Data used to derive calibration function - Sulphur Dioxide, High Range.**

Test No	Test Start Time	Test End Time	SRM measured value (y)	CEMS measured signal (x)
	hr:min		(ppm)	(ppm)
1	Reference Zero		0.0	-1.0
2	11:23	12:23	63.3	59.2
3	12:34	13:34	66.3	59.7
4	13:35	14:35	67.2	59.8
5	14:36	15:36	67.4	60.3
6	15:37	16:37	66.7	59.9
7	16:38	17:38	64.9	59.4
8	17:39	18:39	64.6	59.1
9	18:40	19:40	60.0	57.4
10	19:41	20:41	55.6	55.4
11	20:42	21:42	54.1	53.0
12	21:53	22:53	53.0	52.1
13	22:54	23:54	52.0	51.5
14	23:55	0:55	49.4	49.6
15	0:56	1:56	47.2	48.5
16	1:57	2:57	48.4	47.2
17	2:58	3:58	49.5	48.8
18	3:59	4:59	50.0	47.9
19	5:00	6:00	47.2	46.6
20	6:01	7:01	45.4	44.4
21	7:12	8:12	46.1	45.4
22	8:13	9:13	45.5	46.0
23	9:14	10:14	44.8	45.0
24	10:50	11:50	53.1	49.2
25	11:51	12:51	56.5	52.3
26	13:02	14:02	58.7	54.4
27	14:03	15:03	60.7	56.4
28	15:04	16:04	63.6	59.2
29	16:05	17:05	59.0	56.8
30	17:06	18:06	56.7	54.9
31	18:07	19:07	53.4	51.7
32	19:08	20:08	49.1	46.7
33	20:09	21:09	48.6	45.9
34	21:10	22:10	46.8	43.6
35	22:21	23:21	44.8	41.4
36	23:22	0:22	39.0	35.8
37	11:26	12:26	73.1	69.4
38	12:27	13:27	70.2	64.7
Sum			2041.79	1937.54

#### 4A2.7.3 Determination of Calibration Function - Sulphur Dioxide, High Range.

##### Method B

Formulae:- Number of tests conducted = N 38

If Ymax-Ymin <15% of the ELV, the following formulae are used:

$b=(y/(x-Z))$   $x=(1/\text{No. AMS Tests}) * \text{Total AMS}$

$$x=(1/ 38)*1937.5 \quad \text{or} - \quad x= 50.99$$

$a=(-\beta.Z)$   $y=(1/\text{No. SRM Tests}) * \text{Total SRM}$

$$y=(1/ 38)*2041. \quad \text{or} - \quad y= 53.73$$

The Slope is calculated by :

$$\beta=y/(x-Z) \quad 53.73 / ( 50.98--1) \quad \beta= 1.034$$

The offset is calculated by:

$$\alpha=-\beta.Z \quad -1.03*-1 \quad \alpha= 1.03$$

The calibration is function  $y_{im} = \alpha + \beta_{xi,m}$  or  $y_i = 1.033 + 1.03 * X_i$

**4A2.8.3 Table A.4.4.3 - Calculation of calibrated CEMS values - Sulphur Dioxide, High Range.**

Test No	Test Start Time	Test End Time	CEMS Raw Value (x) (ppm)	CEMS Calibrated signal (ppm)	CEMS Dry Oxygen (%)	CEMS Moisture (%)	CEMS Standardised Value (dry) (mg/Nm <sup>3</sup> )	CEMS Calibrated Standardised Value (mg/Nm <sup>3</sup> )	SRM Standardised (mg/m <sup>3</sup> )
1	Reference Zero		-1.0	0.0	N/A	N/A	-1.0	0.0	0.0
2	11:23	12:23	59.2	62.2	6.8	2.6	183.4	192.8	194.6
3	12:34	13:34	59.7	62.7	6.7	2.5	184.0	193.4	202.7
4	13:35	14:35	59.8	62.8	6.4	2.4	180.4	189.6	201.8
5	14:36	15:36	60.3	63.3	6.5	2.5	182.8	192.1	203.0
6	15:37	16:37	59.9	63.0	6.6	2.5	183.5	192.9	202.6
7	16:38	17:38	59.4	62.4	6.9	2.6	185.7	195.2	200.5
8	17:39	18:39	59.1	62.1	7.0	2.7	185.8	195.3	200.4
9	18:40	19:40	57.4	60.4	7.0	2.7	181.0	190.3	186.8
10	19:41	20:41	55.4	58.3	7.0	2.8	174.8	184.0	172.8
11	20:42	21:42	53.0	55.8	6.9	2.8	166.0	174.9	168.0
12	21:53	22:53	52.1	54.9	6.9	2.8	163.4	172.1	163.1
13	22:54	23:54	51.5	54.2	6.9	2.8	160.7	169.3	160.1
14	23:55	00:55	49.6	52.3	7.0	2.7	156.3	164.8	153.2
15	00:56	01:56	48.5	51.1	7.1	2.7	154.1	162.6	147.5
16	01:57	02:57	47.2	49.8	7.0	2.4	148.2	156.4	150.3
17	02:58	03:58	48.8	51.5	7.1	2.4	154.0	162.4	154.3
18	03:59	04:59	47.9	50.5	7.0	2.4	150.4	158.7	155.2
19	05:00	06:00	46.6	49.2	7.0	2.5	146.6	154.7	146.5
20	06:01	07:01	44.4	47.0	7.0	2.5	139.6	147.5	140.5
21	07:12	08:12	45.4	48.0	6.9	2.5	142.1	150.1	142.1
22	08:13	09:13	46.0	48.6	7.0	2.6	144.6	152.7	140.8
23	09:14	10:14	45.0	47.5	7.2	2.6	143.0	151.1	140.2
24	10:50	11:50	49.2	51.9	6.8	2.6	152.4	160.7	163.2
25	11:51	12:51	52.3	55.1	6.8	2.5	161.8	170.4	173.0
26	13:02	14:02	54.4	57.2	7.0	2.4	170.2	179.2	181.3
27	14:03	15:03	56.4	59.4	7.0	2.2	176.6	185.8	187.9
28	15:04	16:04	59.2	62.2	7.0	2.1	185.2	194.7	196.9
29	16:05	17:05	56.8	59.7	7.2	2.2	180.1	189.4	185.2
30	17:06	18:06	54.9	57.8	7.3	2.3	175.3	184.5	179.0
31	18:07	19:07	51.7	54.5	7.1	2.3	162.8	171.5	166.3
32	19:08	20:08	46.7	49.3	7.0	2.4	146.7	154.9	152.8
33	20:09	21:09	45.9	48.5	6.9	2.4	143.6	151.6	150.4
34	21:10	22:10	43.6	46.1	6.9	2.4	136.0	143.8	144.7
35	22:21	23:21	41.4	43.8	6.8	2.4	128.4	135.9	137.6
36	23:22	00:22	35.8	38.0	8.3	2.5	124.1	131.8	134.2
37	11:26	12:26	69.4	72.8	6.6	2.7	211.7	222.0	220.2
38	12:27	13:27	64.7	67.9	6.5	2.7	196.8	206.6	210.7
Sum							6061.29		


Emission Limit Value (ELV) = 440 mg/Nm<sup>3</sup>

Reference Oxygen

6

%

**4A2.9.3 Plot 2.3 CEM versus SRM Parallel Test Data at CEM measurement conditions –Sulphur dioxide, wet gas.**




**4A2.10.3 Table 4.5.3 – Data used for the Variability Test – Sulphur Dioxide, High Range.**

Test No	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
	hr:min		mg/m3	mg/m3			
2	11:23	12:23	192.8	194.6	1.82	3.85	14.82
3	12:34	13:34	193.4	202.7	9.34	11.37	129.26
4	13:35	14:35	189.6	201.8	12.20	14.23	202.50
5	14:36	15:36	192.1	203.0	10.90	12.93	167.25
6	15:37	16:37	192.9	202.6	9.72	11.75	138.04
7	16:38	17:38	195.2	200.5	5.34	7.37	54.29
8	17:39	18:39	195.3	200.4	5.11	7.14	50.95
9	18:40	19:40	190.3	186.8	-3.51	-1.48	2.19
10	19:41	20:41	184.0	172.8	-11.17	-9.14	83.50
11	20:42	21:42	174.9	168.0	-6.90	-4.87	23.67
12	21:53	22:53	172.1	163.1	-9.02	-6.98	48.76
13	22:54	23:54	169.3	160.1	-9.21	-7.18	51.51
14	23:55	00:55	164.8	153.2	-11.62	-9.59	91.92
15	00:56	01:56	162.6	147.5	-15.03	-13.00	169.02
16	01:57	02:57	156.4	150.3	-6.10	-4.07	16.57
17	02:58	03:58	162.4	154.3	-8.18	-6.15	37.77
18	03:59	04:59	158.7	155.2	-3.42	-1.39	1.94
19	05:00	06:00	154.7	146.5	-8.22	-6.18	38.25
20	06:01	07:01	147.5	140.5	-6.94	-4.91	24.09
21	07:12	08:12	150.1	142.1	-7.93	-5.90	34.78
22	08:13	09:13	152.7	140.8	-11.97	-9.94	98.81
23	09:14	10:14	151.1	140.2	-10.90	-8.86	78.57
24	10:50	11:50	160.7	163.2	2.57	4.60	21.16
25	11:51	12:51	170.4	173.0	2.54	4.57	20.91
26	13:02	14:02	179.2	181.3	2.13	4.16	17.33
27	14:03	15:03	185.8	187.9	2.10	4.13	17.06
28	15:04	16:04	194.7	196.9	2.19	4.22	17.83
29	16:05	17:05	189.4	185.2	-4.20	-2.17	4.72
30	17:06	18:06	184.5	179.0	-5.51	-3.48	12.10
31	18:07	19:07	171.5	166.3	-5.21	-3.18	10.10
32	19:08	20:08	154.9	152.8	-2.14	-0.11	0.01
33	20:09	21:09	151.6	150.4	-1.20	0.83	0.69
34	21:10	22:10	143.8	144.7	0.93	2.96	8.79
35	22:21	23:21	135.9	137.6	1.66	3.69	13.64
36	23:22	00:22	131.8	134.2	2.36	4.39	19.27
37	11:26	12:26	222.0	220.2	-1.77	0.26	0.07
38	12:27	13:27	206.6	210.7	4.07	6.10	37.22
37 Tests		Mean			-2.03		
Sum							1759.36

#### 4A2.11.3 - Variability Test Calculation – Sulphur Dioxide, High Range.

SD=	$\text{Root}(1-\text{Number}).\text{Integral}(D1-D)^2$	6.99	mg/m <sup>3</sup> (s,d),6%O <sub>2</sub>
The uncertainty laid down by the authorities is 20% ELV as a 95% confidence interval. O <sub>0</sub> is therefore calculated as:-			
O <sub>0</sub> =	0.2 * 440 mg/m <sup>3</sup> (s,d,6%O <sub>2</sub> ) / 1.96	44.90	mg/m <sup>3</sup> (s,d),6%O <sub>2</sub>
For 37 tests, kv =	0.9885		
Therefore variability=		6.99 <= 44.9 * 0.9885	
or	6.99	<=	44.38
Which is TRUE therefore the CEMS passes the test			

#### 4A2.12.3 Plot 3.3 –Standardised CEM data versus standardised SRM - Sulphur dioxide, High Range – Reference conditions 273K, 101.3kPa., 6% oxygen, dry gas.



## **Section 4B1: Data & Calculations – AST – Unit 8, SICK OMD 41**

**4B3 4.1 Table 4.1 – Raw monitoring Data – Total Particulate Matter**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (Extinction)	CEMS Oxygen (Wet)	CEMS Moisture (Wet)	CEM Stack Temp	CEM Stack Press	SRM Raw value (dry)	SRM Moisture (Wet FT)	SRM Oxygen (Dry)	SRM Stack Temp	SRM Stack Press	SRM at CEMs Raw conditions
		hr:min		%	(%)	(%)	C	kpa	(mg/m3)	(%)	(%)	C	kpa	(mg/m3)
1	26-Oct-16	10:43	11:43	18.2	6.5	2.60	65.1	101.7	117.9	2.9	6.6	63.0	100.8	92.6
2	26-Oct-16	14:13	15:13	22.6	6.3	2.46	64.5	101.8	106.4	2.7	6.3	64.0	100.8	83.5
3	27-Oct-16	10:47	11:47	12.0	6.9	2.56	64.6	102.1	118.4	2.6	6.7	65.0	100.8	92.8
4	27-Oct-16	11:55	12:55	12.6	6.6	2.52	64.3	102.1	56.1	2.8	6.6	64.0	100.8	44.0
5	27-Oct-16	13:03	14:03	13.4	6.8	2.43	63.8	102.2	73.5	2.6	6.8	63.0	100.8	57.8
6	27-Oct-16	14:10	15:10	14.8	6.8	2.11	63.3	102.2	87.9	2.6	6.8	64.0	100.8	69.0

Note:

Emission concentrations expresed at reference conditions 273K, 101.3kPa

**4B3 4.2 Table 4.2 – Standardised monitoring Data – Total Particulate Matter**


Test No	Date	Test Start Time	Test End Time	CEMS Standardised Value	SRM Standardised Value	SRM Uncertainty
		hr:min		(mg/m <sup>3</sup> )	mg/m <sup>3</sup> )	(mg/m <sup>3</sup> )
1	26-Oct-16	10:43	11:43	23.9	123.0	0.33
2	26-Oct-16	14:13	15:13	29.1	108.5	0.30
3	27-Oct-16	10:47	11:47	16.1	124.2	0.21
4	27-Oct-16	11:55	12:55	16.5	58.4	0.18
5	27-Oct-16	13:03	14:03	17.8	77.3	0.22
6	27-Oct-16	14:10	15:10	19.6	92.8	0.21

Note:


Emission concentrations expresed at reference conditions 273K, 101.3kPa

6 % Oxygen, dry gas

**4B3 4.3 Plot 1.1 –Time Series of Standardised CEM versus Standardised SRM data – Total Particulate Matter. Reference conditions: 273K, 101.3kPa, dry gas, 6% oxygen.**



#### 4B3 4.4 – Elimination of Outliers – Total Particulate Matter.




Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (Extinction)	SRM Value at CEMS Raw conditions	Difference Di	Difference Di - $\bar{D}_i$	Is Result an Outlier - Di - $\bar{D}_i > 2SD$
		hr:min		%	(mg/m³)			
1	26-Oct-16	10:43	11:43	18.2	92.6	74.45	16.78	No
2	26-Oct-16	14:13	15:13	22.6	83.5	60.90	3.23	No
3	27-Oct-16	10:47	11:47	12.0	92.8	80.72	23.06	No
4	27-Oct-16	11:55	12:55	12.6	44.0	31.39	-26.27	No
5	27-Oct-16	13:03	14:03	13.4	57.8	44.38	-13.28	No
6	27-Oct-16	14:10	15:10	14.8	69.0	54.15	-3.52	No
				Average $\bar{D}_i$		57.66		
				Standard Deviation		18.45		
				Standard Deviation x2		36.91		

**4B3 4.5 Table 4.3 - Data used to Calculate Calibrated Values – Total Particulate Matter.**

Test No	Date	Test Start Time	Test End Time	CEMS Raw Value (Extinction)	CEMS Calibrated signal	CEMS Moisture	CEMS Temp	CEMS Pressure	CEMS Dry Oxygen	CEMS Standardised Value (dry)	CEMS Calibrated Standardised Value	SRM Standardised
		hr:min		%	(mg/m <sup>3</sup> )	(%)	(°C)	(kPa)	(%)	(mg/Nm <sup>3</sup> )	(mg/Nm <sup>3</sup> )	(mg/m <sup>3</sup> )
1	17-Feb-16	Reference		0.0	-0.3	N/A	N/A	N/A	0.0	0.0	-0.2	0.0
2	26-Oct-16	10:43	11:43	18.2	21.7	2.6	65.1	101.7	6.5	23.9	28.4	123.0
3	26-Oct-16	14:13	15:13	22.6	27.0	2.5	64.5	101.8	6.3	29.1	34.8	108.5
4	27-Oct-16	10:47	11:47	12.0	14.2	2.6	64.6	102.1	6.9	16.1	19.0	124.2
5	27-Oct-16	11:55	12:55	12.6	14.9	2.5	64.3	102.1	6.6	16.5	19.5	58.4
6	27-Oct-16	13:03	14:03	13.4	15.9	2.4	63.8	102.2	6.8	17.8	21.2	77.3
7	27-Oct-16	14:10	15:10	14.8	17.6	2.1	63.3	102.2	6.8	19.6	23.3	92.8
Sum									123.04			
Emission Limit Value (ELV) =				35	mg/Nm <sup>3</sup>							
Reference Oxygen												
Established Calibration Function y <sub>i</sub> =												
-0.303 + 1.208xi												
6 %												

**Plot of CEM versus SRM Data at CEM measurement conditions– Total Particulate Matter, at stack temperature, pressure and moisture conditions.**



**4B3 4.6 Table 4.4 – Data used for the Variability Test – Total Particulate Matter.**

Test No	Test Start Time	Test End Time	CEMS Calibrated Standardised value	SRM Standardised value	Difference D1	Difference D1 - D	Squared Difference D1 - D
	hr:min		mg/m3	mg/m3			
1	Reference		-0.2	0.0	0.22	-62.38	3891.17
2	10:43	11:43	28.4	123.0	94.56	31.96	1021.45
3	14:13	15:13	34.8	108.5	73.74	11.14	124.17
4	10:47	11:47	19.0	124.2	105.15	42.56	1811.02
5	11:55	12:55	19.5	58.4	38.87	-23.73	562.94
6	13:03	14:03	21.2	77.3	56.19	-6.40	41.01
7	14:10	15:10	23.3	92.8	69.45	6.85	46.92
7 Tests		Mean			62.60		
Sum							7498.70

**4B3 4.7 - Variability Test and Test of Calibration Function Calculations – Total Particulate Matter.**

Variability Test

$$SD = \text{Root}(1-\text{Number}).\text{Integral}(D1-D)^2 \quad 35.35 \quad \text{mg/m3(s,d),6\%O}_2$$

The uncertainty laid down by the authorities is 30% ELV as a 95% confidence interval.  $O_0$  is therefore calculated as:-

$$O_0 = 0.3 * 35 \text{ mg/m3 (s,d,6\%O}_2) / 1.96 \quad 5.36 \quad \text{mg/m3(s,d),6\%O}_2$$

$$\text{For 8 tests, } k_v = 0.9521$$

$$\text{Therefore variability} = 35.35 \leq 5.36 * 0.9521 * 1.5 \\ \text{or} \quad 35.35 \leq 7.65$$

Which is FALSE therefore the CEMS fails the test

Test of Calibration Function

$$\text{The calibration of the AMS is accepted if:} \quad D1 \leq t_{0.95}(N-1)*(s,d/\text{root } N) + O_0$$

$$D1 = 62.60$$


$$\text{For 7 tests, } t_{0.95}(N-1) = 1.895$$

$$\text{Therefore test of calibration} = 62.6 \leq 1.895 * (35.35/\text{root } 7) + 5.36$$

$$\text{or} \quad 62.60 \leq 30.68$$

Which is FALSE therefore the calibration function is INVALID

**4B3 4.8 Plot 2 – Standardised CEM data versus standardised SRM - Total Particulate Matter – Reference conditions 273K, 101.3kPa, dry gas, 6% oxygen.**



## Section 5 – Results of Functional tests

**Table 5.1 - Audit of functional tests**

<b>Operator</b>	RWE Generation UK plc.	
<b>Site</b>	Aberthaw Power Plant	
<b>Stack</b>	Unit 8	
<b>Process Sector</b>	LCPD	
<b>Analyser A - Make Model</b> <b>MCERTs Certificate Number</b>	Procal 1 - Pulsi 2000, In Situ IR MC990006/07	
<b>Analyser B - Make Model</b> <b>MCERTs Certificate Number</b>	Procal 2 - Pulsi 2000, In Situ IR MC990006/07	
<b>Analyser C - Make Model</b> <b>MCERTs Certificate Number</b>	Erwin SICK OMD41, Cross Duct Forward Scatter MC040042/01	
<b>Parameters Tested</b>	<b>Daily ELV</b>	<b>Certified range</b>
<b>NO &amp; NO<sub>2</sub></b>	1080 mg/m <sup>3</sup> (NO <sub>x</sub> as NO <sub>2</sub> )	PROCAL: NO = 0 – 1000 ppm
<b>Total Particulate Matter</b>	35 mg/m <sup>3</sup>	Erwin SICK: TPM = 0 – 200 mg/m <sup>3</sup>
<b>SO<sub>2</sub></b>	440 mg/m <sup>3</sup>	PROCAL = SO2 low = 0 – 250 ppm SO2 high = 0 – 1000 ppm

<b>Analysers A &amp; B</b>	
<b>Organisation carrying out tests -</b>	Parker/Procal
<b>Status of organisation – CEMS manufacturer/operator/service contractor</b>	CEMS OEM
<b>Test engineer</b>	M Findley
<b>Date of tests</b>	1 <sup>st</sup> – 11 <sup>th</sup> October 2016

<b>Analyser C</b>	
<b>Organisation carrying out tests -</b>	SICK
<b>Status of organisation – CEMS manufacturer/operator/service contractor</b>	CEMS OEM
<b>Test engineer</b>	P Burgess
<b>Date of tests</b>	5 <sup>th</sup> - 10 <sup>th</sup> October 2016

## Functional Test compliance with EN 14181

Requirement	Compliance Y/N	Notes
1 – Alignment and cleanliness All checks specified in MID 14181 carried out ?  – Sampling System  A visual inspection of the sampling system shall be performed, noting the condition of the following components, when fitted: <ul style="list-style-type: none"><li>- sampling probe;</li><li>- gas conditioning systems;</li><li>- pumps;</li><li>- all connections;</li><li>- sample lines;</li><li>- power lines;</li><li>- filters.</li></ul> The sampling system shall be in good condition and free of any visible faults, which may decrease the quality of the testing.	Yes	<b>Procal</b> – Yes <b>SICK (OMD)</b> – Yes – Optics cleaned
2 - Leak Test  Leak testing shall be performed according to the AMS manuals. The test shall cover the entire sampling system.	N/A	<b>Procal</b> – N/A although differential pressure noted as OK
Results of leak check compliant with requirements of relevant standards	N/A	
3 - Zero and Span Check Analyser  Reference zero and span materials shall be used to verify the corresponding readings of the AMS.	Yes	<b>Procal</b> – Yes <b>SICK (OMD 41)</b> – Yes – although based on the use of filters
Results compliant with requirements of relevant standards  Parameter:		
NO SO <sub>2</sub> TPM	Yes (Procal) Yes (Procal) Yes	<b>SICK (OMD41)</b> - Span values recorded as mA signals

<b>Requirement</b>	<b>Compliance Y/N</b>	<b>Notes</b>
4 - Zero and Span Check Full System  Reference zero and span materials shall be used to verify the corresponding readings of the AMS.	N/A	Procal analysers are in situ type and therefore there is as such, no 'full system'.
5 – NOx converter efficiency check	N/A	
6 - Linearity  The linearity of the analysers shall be checked using five different reference materials, including zero concentration.	Yes – Procal, SICK OMD 41	<b>Procal</b> – Yes <b>SICK (OMD)</b> – Yes – although based on the use of filters
The reference material with zero concentration, as well as the reference materials with four different concentrations, shall have a verifiable quantity and quality.	Yes	<b>Procal</b> – Yes <b>SICK</b> – Yes – although based on the use of filters
The reference material concentrations shall be selected such that the measured values are approximately 20%, 40%, 60% and 80% of two times the emission limit.	See note	<b>Procal</b> – Yes <b>SICK</b> – Used 8 points using filters
The dry test reference material shall be applied to the inlet of the AMS.  Reference materials can be introduced directly into the analyser as long as the integrity of the sample system has been proved	N/A	N/A

Requirement	Compliance Y/N	Notes
6 – Linearity (continued) After each change in concentration, the first instrument readings shall be taken after a time period equal to at least three times the response time of the AMS. At each reference material concentration, at least three readings shall be made. The time period between the start of each of the three readings shall be separated by at least four times the response time.  A risk based approach may be adopted in order to reduce the time for the linearity tests	No  N/A	<b>Procal</b> – No times stated in report  <b>SICK (OMD)</b> – No times stated in report
Linearity Test Pass Parameter		<b>Procal</b> – Yes <b>SICK (OMD)</b> – Yes
NO, SO <sub>2</sub> & TPM	Yes	
6 – Interferences (only required in the event of a failure of the QAL 2/AST)	N/A	
7 – Zero and Span Drift (Audit)	Yes	
8- Response time  The response of the AMS shall be checked. This can be performed, if appropriate, by feeding the reference material at the end of the sampling probe. The response time shall not exceed the measured value as identified during QAL 1.	See Note	<b>Procal</b> – Yes <b>SICK (OMD 41)</b> – Yes

Requirement	Compliance Y/N	Notes
<b>9a – Service Report - PROCAL</b>		
• Document reference	Yes	
• Instrument manufacturer	Yes	Procal
• Instrument Type	Yes	In-situ
• Instrument model	Yes	PULSI 200
• Instrument Serial No's	Yes	
• Operating principal	Yes	
• Operating range	Yes	
• Certification details	Yes	
• Compliance with MCERTS	Yes	
• Location	Yes	
• Date and time of work	Yes	Date only
• Equipment used - Type serial no's etc	Yes	Gas divider
• Gases used – certificate numbers, expiry dates, type	No	Copy of gas certificates offered to customer
• NOx converter efficiency check	N/A	
• Calibration and linearity data	Yes	
• Logged data for period of calibration/linearity	No	
• Name & signature of test engineer	Yes	

Requirement	Compliance Y/N	Notes
9b – Service Report – <b>SICK OMD</b>		
• Document reference	Yes	
• Instrument manufacturer	Yes	
• Instrument Type	Yes	
• Instrument model	Yes	
• Instrument Serial No's	Yes	
• Operating principal	Yes	
• Operating range	Yes	
• Certification details	Yes	
• Compliance with MCERTS	Yes	
• Location	Yes	
• Date and time of work	Yes	Date Only
• Equipment used - Type serial no's etc	Yes	On Linearity Sheet
• Gases used – certificate numbers, expiry dates, type	N/A	Linearity carried out using filters. Filter set serial no stated.
• NOx converter efficiency check	N/A	
• Calibration and linearity data	Yes	
• Logged data for period of calibration/linearity	Yes	Data supplied separately by client
• Name & signature of test engineer	Yes	



OMD41- Particulate Linearity Report

Customer	ABERTHAW
Site	ABERTHAW
Permit Ref	
Contact	DAN PETERS
Service Ref No.	19128
Survey Date	10/10/2016
Device Type	OMD41
Serial Number	98078002
MCERTS Cert No	MC 040042/01
Certification Details	See Certificate
Measurement Principle	OPACITY
Gas Box Serial Number	91148002
SICK Engineer	P BURGESS
Signature	A handwritten signature in black ink, appearing to read "P Burgess".
	P BURGESS

Sick (UK) Ltd  
Waldekirch House  
39 Hedley Road  
St Albans  
Hertfordshire  
AL1 5BN



Tel: +44 (0) 1727 831121  
Fax: +44 (0) 1727 855332  
E-Mail: info@sick.co.uk

Particulate Linearity							
Customer		ABERTHAW					
Plant	ABERTHAW	Site Permit Ref	0				
S&Es Contact	DAN PETERS	SYNCH Ref	19128				
Product	OM01M1	Order Number	QAL2/AST/Routine				
MCERTS Certificate Number	MC 040042/01	Date	10/10/2016				
Location	SICK	Engineer	P BURGESS				
Filter Set SN	17103385	Response time ( $T_{\text{res}}$ ) seconds	2				
Calibration Date	2016-10-10	Plant ELV for $\text{mg.m}^{-3}$	250				
Results Visible on DCS?	Not possible disconnected from stack	Plant ELV $\times 2 \text{ mg.m}^{-3}$ ( $\text{c}_1$ )	400				
Corrections Applied?	NA	Analyser Range $\text{mg.m}^{-3}$	500				
Component 1		Particulates					
Extinction Time (Ex)	1	Expected Extinction	Reading 1	Reading 2	Reading 3	Average Reading	DCS
Log Time ( $\ln T_{\text{res}}$ )	0.693	0.6205	0.6192	0.6190	0.6190	0.6192	NA
		0.6496	0.6480	0.6480	0.6480	0.6480	NA
		0.6841	0.6820	0.6821	0.6820	0.6820	NA
		0.7155	0.7130	0.7131	0.7130	0.7130	NA
		0.7453	0.7430	0.7431	0.7430	0.7430	NA
		0.3818	0.3770	0.3770	0.3770	0.3770	NA
		0.8195	0.8000	0.8000	0.8000	0.8000	NA
		1.2124	1.1750	1.1730	1.1750	1.1743	NA
Linearity Results							
ZERO	Expected Transmission	Analyser Reading $\text{mg.m}^{-3}$	DCS $\text{mg.m}^{-3}$	DCS Time			
Zero Reading 1	0.00	0.00	NA	NA			
Zero Reading 2	0.00	0.00	NA	NA			
Zero Reading 3	0.00	0.00	NA	NA			
Zero Reading 4	0.00	0.00	NA	NA			
Zero Reading 5	0.00	0.00	NA	NA			
Zero Reading 6	0.00	0.00	NA	NA			
AVERAGE		0.00					
Actual Extinction	Average CEM Reading (Ex)	$d_{\text{CEM}}$	$d_{\text{CEM}} < 5\%$ (EN 14181)				
0.6205	0.6192	-0.03%	PASS				
0.6496	0.6480	-0.04%	PASS				
0.6841	0.6820	0.03%	PASS				
0.7155	0.7130	-0.04%	PASS				
0.7453	0.7430	-0.05%	PASS				
0.3818	0.3770	-0.04%	PASS				
0.8195	0.8000	0.03%	PASS				
1.2124	1.1743	-0.02%	PASS				
Actual Extinction v CEM Reading (Ex)							
Engineer Signature		P BURGESS					
Comments		Analyser exhibiting a linear response					
Version 2016 - Uncontrolled when printed							

Filter Extinction <i>i</i> %	Reference value <i>x<sub>i</sub></i> %	CEM value <i>y<sub>i</sub></i> mg.m <sup>-3</sup>	Difference <i>X<sub>i</sub>-X<sub>z</sub></i>	Product <i>Y<sub>i</sub>(X<sub>i</sub>-X<sub>z</sub>)</i>	Squared difference <i>(X<sub>i</sub>-X<sub>z</sub>)<sup>2</sup></i>
0%	0.00	0.00	-0.30	0.00	0.09
0%	0.00	0.00	-0.30	0.00	0.09
0%	0.00	0.00	-0.30	0.00	0.09
0.022	0.02	0.02	-0.28	-0.01	0.08
0.022	0.02	0.02	-0.28	-0.01	0.08
0.022	0.02	0.02	-0.28	-0.01	0.08
0.0496	0.05	0.05	-0.25	-0.01	0.06
0.0496	0.05	0.05	-0.25	-0.01	0.06
0.0496	0.05	0.05	-0.25	-0.01	0.06
0.0841	0.18	0.17	-0.12	-0.02	0.01
0.0841	0.18	0.17	-0.12	-0.02	0.01
0.1755	0.08	0.08	-0.21	-0.02	0.05
0.1755	0.08	0.08	-0.21	-0.02	0.05
0.1755	0.08	0.08	-0.21	-0.02	0.05
0.2334	0.23	0.23	-0.06	-0.01	0.00
0.2334	0.23	0.23	-0.06	-0.01	0.00
0.3818	0.38	0.38	0.08	0.03	0.01
0.3818	0.38	0.38	0.08	0.03	0.01
0.3818	0.38	0.38	0.08	0.03	0.01
0.8195	0.82	0.80	0.52	0.42	0.27
0.8195	0.82	0.80	0.52	0.42	0.27
0.8195	0.82	0.80	0.52	0.42	0.27
1.2124	1.21	1.18	0.91	1.07	0.84
1.2124	1.21	1.17	0.91	1.07	0.84
1.2124	1.21	1.18	0.91	1.07	0.84
0%	0.00	0.00	-0.30	0.00	0.09
0%	0.00	0.00	-0.30	0.00	0.09
0%	0.00	0.00	-0.30	0.00	0.09
Average	0.30	0.29			
Sum	8.93	8.72	0.00	4.36	4.49

a =	0.29
B =	0.97
A =	0.00

Residuals				
Averages	c %	Average Y <sub>c</sub>	d <sub>c</sub>	d <sub>c,rel</sub>
0	0.00	0.00	0.00	-0.03%
0.022	0.02	0.02	0.00	-0.09%
0.0496	0.05	0.05	0.00	-0.04%
0.0841	0.18	0.17	0.00	0.03%
0.1755	0.08	0.08	0.00	-0.02%
<b>0.2334</b>	<b>0.23</b>	<b>0.23</b>	<b>0.00</b>	<b>0.10%</b>

<b>SICK (UK)LTD</b> Waldkirch House, 39 Hedley Rd St Albans, Herts, AL1 5BN		Tel : 0044 (0)1727 831121 Fax : 0044 (0)1727 855332 E – Mail : info@sick.co.uk	
<b>Maintenance Report OMD41 Stack UNIT 8</b>			
<input checked="" type="checkbox"/> Maintenance <input type="checkbox"/> Repair			
Customer:	RWE Aberthaw	Customer no:	SVON 019128
Country:	UK	Location:	Aberthaw PS
Receiver:	1012184	Serial no.:	98078002
Reflector:	1012385	Serial no.:	98078002
Eval.Unit:	n/a	Serial no.:	n/a
Plant operating status: <b>Stack UNIT 8 unit not running</b>			

done    Remarks:

Check weekly/monthly/ quarterly log files: Log files ok?	<input type="checkbox"/>	Not available
Training recommended:	<input type="checkbox"/>	Not required

#### Dust reading before check

Opacity or Extinction	NA	Removed by RWE Staff.
Dust in Ex	Ex	0.065
Contamination	mA	
Dust reading, measured values	mA	4.7
Control cycle, Zero point values	mA	0
Control cycle, Span value	mA	70

#### 1. Visual Inspection

1.1	Check if measured values are plausible in the control room	<input checked="" type="checkbox"/>	OK
1.2	Check if control cycle is (zero/span) valid (control room or recorder or evaluation unit)	<input checked="" type="checkbox"/>	O.K.
1.3	Check if there is any error message	<input checked="" type="checkbox"/>	None
1.4	Check for mechanical or electrical connection damage	<input checked="" type="checkbox"/>	OK
1.5	Check mechanical mounting	<input checked="" type="checkbox"/>	O.K.

#### 2. Purge Air Unit

2.1	Check the hoses and ring nozzles	<input checked="" type="checkbox"/>	NA RWE STAFF
2.2	Clean the pre-filter	<input type="checkbox"/>	
2.3	Replace the filter insert, part no. 5306091 for OMD41 SN > 97xxxxxx, or part no.6700149 for OMD41 SN <96xxxxxx	<input type="checkbox"/>	
2.4	Check function of filter monitor purge switch.	<input type="checkbox"/>	

#### 3. Fail Safe Shutter (FSS Optional)

3.1	<input type="checkbox"/> Check function of fail-safe shutter	<input type="checkbox"/>	NA
-----	--	--------------------------	----

#### 4. Optical unit

4.1	Clean housing	<input checked="" type="checkbox"/>	O.K.
4.2	Check and, if necessary readjust optical axis	<input checked="" type="checkbox"/>	O.K.

4.3	Clean optical boundary surfaces for sender/receiver and reflector, 2 x part no. 4003353	<input checked="" type="checkbox"/>	O.K.
4.4	Check and replace desiccant cartridge, part no. 2008475	<input checked="" type="checkbox"/>	Not required Still blue.
4.5.	Check the transmitter diode	<input checked="" type="checkbox"/>	O.K.
4.6	Check the stepper motor	<input checked="" type="checkbox"/>	O.K.
4.7	Check zero point, re-calibrate at smoke-free measurement point if it deviates	<input checked="" type="checkbox"/>	O.K.
4.8	Check function of control cycle	<input checked="" type="checkbox"/>	O.K.
4.9	Linearity check	<input checked="" type="checkbox"/>	OK
Linearity Measurement	Set value	Actual Value	Deviation in %
Zero	0.0 mA	0.0 mA	0.00 0.0% OK
Span 1 (aprox. 30%)	14.06 mA	14.06 mA	0.02 0.0% OK
Span 2 (aprox. 70%)	16.80 mA	16.80 mA	0.00 0% OK
Span 3 (100%)	20.00 mA	20.00 mA	0.00 0.0% OK

##### 5. Connection unit

Check the analogue and digital input and output signals	<input type="checkbox"/>	n/a
---	--------------------------	-----

##### 6. Final check

6.1	Check function of complete measurement system	<input checked="" type="checkbox"/>	OK
6.2	Check the availability of the system	<input checked="" type="checkbox"/>	Not available
6.3	Write device data into the report file	<input checked="" type="checkbox"/>	OK
6.4	Complete Service Report and Measurement Log-file	<input checked="" type="checkbox"/>	OK

##### 7. Report in writing to the plant manager as follows:

Quality of maintenance work by plant personnel: OK
Cleanliness of measuring station and devices: OK
Maintenance work done(in addition to Annex 1): None
Problems found, solution applied: None
Recommendation for repair, changes: None
Availability of the unit: Off line
Proposal how to reach higher availability and better emission data quality: None
Training provided to plant personnel, proposal for settlement: None
Training requirements detected: None
Others:

##### Dust reading after check

Opacity or Extinction	EX	0.0047
Dust in mg / m <sup>3</sup>	mg/m <sup>3</sup>	0.0
Contamination	mA	0.0
Dust reading, measured values	mA	4.5
Control cycle, Zero point values	mA	4.00
Control cycle, Span value	mA	16.80

Remarks :

	Name
	Plant person: Dan Peters

Date : 10-10-16	Technician: Paul Burgess
-----------------	--------------------------



## CALIBRATION CERTIFICATE

Site	Aberthaw			
Date	11/10/2016			
Instrument	Unit 8 Primary			
Instrument Ref	8800539		8500916	
Engineer	Michael Findley			

### Results:

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
H2O	%	10	n/a				
CO2	mAU	400	n/a				
CO	PPM	500	336	site gas	316	319	-3.4
NO	PPM	1000	659		336	662	0.3
SO2-H	PPM	1000	674	site gas	700	678	0.4
SO2-L	PPM	250	207.9	936352001	211	213	2.0

### Cross sensitivity tests:

Test/	Response			% Error FSD		
	CO	NO	SO2-H	CO	NO	SO2-H
CO2 15%	1	24	4	0.2	2.4	0.4
H2O 4%	8	-2	4	1.6	-0.2	0.4
CO2 15% + H2O 3%	4	-3	7	0.8	-0.3	0.7

Test/	Response			% Error FSD		
	SO2-L			SO2-L		
CO2 15%	4			1.6		
H2O 4%	4			1.6		
CO2 15% + H2O 3%	7			2.8		

Signature: *M. Findley*

M.Findley

Procal Customer Support Engineer

Parker Procal, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK  
Tel: +44 (0) 17336 232495 Email: [procallsales@parker.com](mailto:procallsales@parker.com) Website: [www.procal.com](http://www.procal.com)

7-7853-02 cal cert site visit general use

Change note: 7007110

Date: 17/03/09




## LINEARISATION RESULTS

Site	Aberthaw		
Date	11/10/2016		
Instrument	Unit 8 Primary		
Instrument Ref	8800539		8500916M
Engineer	Michael Findley		


The above analyser has undergone a linearisation check and calibration, if required, in accordance with the company's quality system. The following results were obtained.

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
NO	PPM	1000	900	4000619281	908	908	0.8
					911	911	1.1
					911	911	1.1
					683	683	0.8
NO	PPM	1000	675	4000619281	684	684	0.9
					683	683	0.8
					439	439	-1.1
					452	452	0.2
NO	PPM	1000	450	4000619281	454	454	0.4
					221	221	-0.4
					230	230	0.5
					231	231	0.6

### NO




Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK  
Tel: +44 (0) 17336 232495 Email: [procalsales@parker.com](mailto:procalsales@parker.com) Website: [www.procal.com](http://www.procal.com)



7-7862-04 Linearisation Certificate

Change note: 7009251

Date: 30/07/12



Zero	H2O	CO2	CO	NO	SO2-H	SO2-L
#1	0	0	-1	-1	1	1
#2	0	0	-1	-4	0	0
#3	0	0	-1	-6	3	3
#4	0	0	0	-4	2	2
#5	0	0	0	-3	1	1
#6	0	0	-3	0	3	3

Signature: *M. Findley*

M.Findley

Procal Customer Support Engineer

Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK TEL:+44(0)1733 23495 FAX:+44(0)1733 235255  
E-mail: asampson@procal.com Website: www.procal.com



## Functionality Test Results

### DECLARATION OF CONFORMITY

We,

Kittiwake Procal Ltd  
5 Maxwell Road  
Woodston  
Peterborough  
Cambridgeshire  
PE2 7HU  
United Kingdom

Tel: (+44) (0) 1733 232495  
Fax: (+44) (0) 1733 235255  
e-mail: [post@procal.com](mailto:post@procal.com)  
Web site: [www.procal.com](http://www.procal.com)

declare that the product: Type PULSI 200 Analyser installed at

Customer	Aberthaw Power Station
Control Unit Serial Number	8800539
P200 Serial Number	8500916M
Site Identification	Unit 8 Primary

to which this declaration relates is in conformance with the MCERTS Performance Standard for Continuous Emission Monitoring Systems: Version 2. Revision 1 (April 2003).

#### MCERTS Product Conformance Certificates:

Licence	Certificate Number	Variation	Dated
ACU	SIRA MC 990006/	8	02/10/09
ACWn	SIRA MC 050060/	6	02/10/09

#### MCERTS Certification Body:

Sira Certification Service	
12 Acorn Industrial Park	Tel: (+44) (0) 1322 520500
Crayford Road	
Crayford	Fax: (+44) (0) 1322 520501
Dartford	
Kent	
DA1 4AL	e-mail: <a href="mailto:info@siraenvironmental.com">info@siraenvironmental.com</a>
United Kingdom	Web site: <a href="http://www.siracertification.com">www.siracertification.com</a>

#### Quality System:

ISO 9001:2008	Certificate Nr: 062043	Dated: 11/06/09
---------------	------------------------	-----------------

EN 15267-2	Dated: 2009	Technical Standards Used:
EN 15267-3	Dated: 2007	

All tests carried out to Kittiwake Procal's working procedures and those within Annex A of EN14181, Version 2.3 June 2010, relating to functionality testing by CEMs' manufacturers.

m. (Inaki)

Signed:..... Date:...11/10/2016.....



## 1.0 Alignment and Cleanliness.

### 1.1 Internal Check of Analyser

Analyser under test	Status	Notes
Optical path clear	Yes	
Optics clean	Yes	
Internal wiring ok	Yes	
Air supply on	Yes	
Purge flow >300 ltrs/hr	Yes	
Pressure differential <80mBar	Yes	36mbar

## 2.0 Zero and Span Checks.

### 2.1 Zero Check

Component:	H2O	CO2	CO	NO	SO2-H	SO2-L	Ta	Ts	Ps
Units:	%	mAU	ppm	ppm	ppm	ppm			
Process conc	3.5	64	49	65	115	103	23	132	1021
Zero conc	0	0	-2	3	3	3	23.2	129	1057
% error	<2	<2	<2	<2	<2	<2			
Pass/Fail	Pass	Pass	Pass	Pass	Pass	Pass			
T90 recovery									

### 2.2 Span Check

Component:	H2O	CO2	CO	NO	SO2-H	SO2-L	Ta	Ts	Ps
Units:	%	mAU	ppm	ppm	ppm	ppm			
Full range	10	-	500	1000	1000	250			
Zero Conc	0	0	-3	0	3	3	23.2	130	1057
Test Gas Conc			336	659	674	207.9			
Cert number				Site gas	Site gas	936352001			
Response Conc			336	662	678	207			
% Error			<2	<2	<2	<2			
Pass/Fail			Pass	Pass	Pass	Pass			
T90 time				80	80	85			

### 2.3 Filter Level Checks

All on zero air	F6	F2	F7	F5G	F3	F1G	F4	F8	
Factory	2822	4446	4654	2684	3487	2294	536	3053	
Last visit	2313	3028	3379	1981	2784	2036	457	2485	
This visit	1926	2643	2903	1678	2329	1696	401	2153	



## 2.4 Interference Checks

### 3.0 Data Acquisition.

Displayed \ Component						
Zero	Analyser					
	DCS					
	4-20mA					
Span	Analyser					
	DCS					
	4-20mA					

## 4.0 Linear coefficient record.

Coefficient \ Component	H2O	CO2	CO	NO	SO2-H	SO2-L
	Linear	2.200E-2	1	9.590E0	1.285E1	5.947E0
Last test	Squared	5.980E-5	0	9.496E-2	3.673E-2	-2.800E-3
	Cubed					
	Linear	2.200E-2	1	1.019E+01	1.285E+01	6.300E+00
This test	Squared	5.980E-5	0	1.009E-01	3.673E-02	-2.967E-03
	Cubed					

## 5.0 Test Information

5.1 Procal Engineer: Michael Findley

## 5.2 Test House:

### 5.3 Test House Engineer:



### Appendix 1. Pro-forma for assessing and reporting the results of the functional tests

Requirement	Notes			
<b>1. Alignment and Cleanliness</b>				
A visual inspection, with reference to the CEMs manuals, shall be carried out on the following when applicable:				
• Internal check of the CEM	✓	All functions ok		
• Cleanliness of the optical components	✓			
• Flushing air supply	✓	Ok		
• Obstructions in the optical path	✓	None		
• After re-assembly at the measurement location at least the following shall be checked	✓	No re-alignment required		
• Alignment of the measuring system	✓	No alignment required		
• Contamination control (internal check of optical surfaces)	✓			
<b>2. Sampling Systems</b>				
A visual inspection of the sampling system shall be performed, noting the condition of the following components, when fitted:				
• Sampling probe	n/a			
• Gas conditioning systems	n/a			
• Pumps	n/a			
• All connections	✓	Ok		
• Sample lines	n/a			
• Power supplies	✓	Ok		
• Filters	✓	Ok		
• NOx converters - if the sampling system contains a NOx converter, then the test laboratory shall record when the last efficiency-test was performed, and the result of this test.	n/a			
• The sampling system shall be in good condition and free of any visible faults, which may decrease the quality of data	✓	Ok		
<b>3. Leak testing</b>				
• Leak testing shall be performed according to the CEMs manuals. The tests shall cover the entire sampling system.	n/a			
<b>4. Zero and Span check</b>				
• Reference zero and span materials shall be used to verify the corresponding readings of the CEM.	✓	Ok		
• In case of non-extractive CEM, zero and span checks shall be performed using a reference-path free of flue gas before and after readjustment and after re-assembly of the CEM at the measurement location.	✓	Ok		
<b>5. Linearity</b>				
• During the calibration/linearity tests the applied concentrations should be logged onto the DCS to prove the complete system i.e. concentration applied to the instrument is represented by the instrument output and identical to the value logged on the DCS.	✓	Data collected on ACU MK3 customer DCS during testing		



Requirement		Notes
DCS logged values should be included in the instrument service report.	✓	Data collected on ACU MK3 & customer DCS during testing
<ul style="list-style-type: none"> <li>The linearity of the CEM's response shall be checked using five different reference materials, including a zero concentration.</li> </ul>	✓	See linearity Cert
<ul style="list-style-type: none"> <li>The reference material with zero concentration, as well as the reference materials with four different concentrations, shall have a verifiable quantity and quality.</li> </ul>	✓	See gas certs
<ul style="list-style-type: none"> <li>In case of gaseous reference materials, these four reference materials can be obtained from different gas cylinders or can be prepared by means of a calibrated dilution system from one single gas concentration.</li> </ul>	✓	Gas points derived from blending full gas range using a Procal P9000, a water generator/gas blender
<ul style="list-style-type: none"> <li>The reference material concentrations shall be selected such that the measured values are at approximately 20 %, 40 %, 60 % and 80 % of the range of two times the emission limit. It is necessary to know the values of the ratios of their concentrations precisely enough so that an incorrect failure of the linearity test does not occur. The dry test reference material shall be applied to the inlet of the CEM.</li> </ul>	✓	See linearity cert
The individual CEMs are tested using the following concentrations applied in a randomised sequence:		
<ul style="list-style-type: none"> <li>Reference material with zero concentration;</li> </ul>	✓	See linearity cert
Reference material concentration approximately 20 % of 2 times the emission limit;	✓	See linearity cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 40 % of 2 times the emission limit;</li> </ul>	✓	See linearity cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 60 % of 2 times the emission limit;</li> </ul>	✓	See linearity cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 80 % of 2 times the emission limit;</li> </ul>	✓	See linearity cert
<ul style="list-style-type: none"> <li>Reference material with zero concentration;</li> </ul>	✓	See linearity cert
After each change in concentration, the first instrument reading shall ordinarily be taken after a time period equal to at least three times the response time of the CEM. At each reference material concentration, at least three readings shall be made, six readings shall be taken at zero. The time period between the start of each of the three readings shall be separated by at least four times the response time.		
The test-laboratory may apply a risk-based approach to linearity testing in order to reduce the time for the tests. For example, the readings may be taken after less than 3x the response time; however, if the CEM fails the linearity test, then the test laboratory shall repeat the linearity test and wait at least 3x the response time as stated above. Alternatively, the number of repetitions of the test may be reduced if the CEM passes the required performance criteria by a factor of at least 2 (i.e. half the allowable residual). Increasing the waiting time to 5x the response time, for example, may be a means of meeting this requirement. Where no other method is possible, the linearity can also be performed with the aid of reference materials such as grating filters or gas filters.		



Requirement	Notes	
The linearity shall be calculated and tested using the procedure as given in EN 14181 annex B. If the CEM does not pass this test; then the problem shall be identified and rectified.		
6. Interferences		
<ul style="list-style-type: none"> <li>A test shall be undertaken if the process gases to be monitored contain components that are known interferences, as identified during QAL1 and there is a failure of the QAL2 or AST which could be due to interferences.</li> </ul>	✓	Interference checks made with Water vapour, CO <sub>2</sub> and each gas species
7. Zero and Span drift (Audit)		
<ul style="list-style-type: none"> <li>The test laboratory shall assess whether the operator has a QAL3 procedure in place, and whether the operator has applied this procedure. The evidence would comprise (i) a documented procedure, (ii) zero and span data, (iii) control charts.</li> </ul>	✓	Customer has procedure in place and has documented evidence
8. Response Time		
<ul style="list-style-type: none"> <li>The response time of the CEM shall be checked. This can be performed, if appropriate, by feeding of the reference material at the end of the sampling probe. The response time shall not exceed the performance requirement applied during the QAL1 tests.</li> </ul>	✓	Response time recorded on test sheet and this test sheet
9. Service Report		
As a minimum requirement the service report should include the following:		
<ul style="list-style-type: none"> <li>Document reference for work instruction for the type of work being undertaken</li> </ul>	✓	Procal Method Statement
<ul style="list-style-type: none"> <li>Instrument manufacturer</li> </ul>	✓	Procal
<ul style="list-style-type: none"> <li>Instrument type</li> </ul>	✓	Pulsi 2000
<ul style="list-style-type: none"> <li>Instrument model</li> </ul>	✓	P2000
<ul style="list-style-type: none"> <li>Instrument Serial No</li> </ul>	✓	8500916M
<ul style="list-style-type: none"> <li>Operating principle</li> </ul>	✓	GFCIR & NIR
<ul style="list-style-type: none"> <li>Operating range</li> </ul>	✓	Recorded on this report
<ul style="list-style-type: none"> <li>Certification details</li> </ul>	✓	Mccerted
<ul style="list-style-type: none"> <li>Compliance with MCERTS (including certificate no.)</li> </ul>	✓	SIRA MC 990006/08
<ul style="list-style-type: none"> <li>Location</li> </ul>	✓	Unit 8 Primary
<ul style="list-style-type: none"> <li>Date and time work was undertaken</li> </ul>	✓	11/06/2016
<ul style="list-style-type: none"> <li>Equipment used - type, serial no's, calibration dates</li> </ul>	✓	Procal Water Generator/Gas divider Pcal #6
<ul style="list-style-type: none"> <li>Gases used - certificate numbers, expiry dates, binary / mix</li> </ul>	✓	Copy of Gas Certs at customer request can be emailed
<ul style="list-style-type: none"> <li>NOx converter efficiency test, if applicable</li> </ul>	n/a	
<ul style="list-style-type: none"> <li>Calibration and linearity data as required by EN14181</li> </ul>	✓	See Linearity & (Calcrt Data)
<ul style="list-style-type: none"> <li>Logged data for period of calibration and linearity. Note there may be gaps in the data, in such cases, the test laboratory shall state why there are gaps in the data</li> </ul>	✓	See Linearity & (Calcrt Data)




## LINEARISATION RESULTS

Site	Matthew Gosling
Date	11/10/2016
Instrument	Unit 8 back up
Instrument Ref	8800624 8500726
Engineer	Matthew Gosling


The above analyser has undergone a linearisation check and calibration, if required, in accordance with the company's quality system. The following results were obtained.

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
NO	ppm	1000	675	4000619281	655	655	-2.0
					655	655	-2.0
					658	658	-1.7
NO	ppm	1000	225	4000619281	214	214	-1.1
					210	210	-1.5
					218	218	-0.7
NO	ppm	1000	900	4000619282	888	888	-1.2
					885	885	-1.5
					883	883	-1.7
NO	ppm	1000	450	4000619283	442	442	-0.8
					432	432	-1.8
					433	433	-1.7


### NO



Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK  
Tel: +44 (0) 17336 232495 Email: procalsales@parker.com Website: www.procal.com



Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK TEL:+44(0)1733 23495 FAX:+44(0)1733 235255  
E-mail: asampson@procal.com Website: www.procal.com



Zero	H2O	CO2	CO	NO	SO2 - H	SO2 - L
#1	0	0	3	6	1	1
#2	0	-1	1	0	-2	-2
#3	0	0	3	3	-3	-3
#4	0	0	3	2	0	0
#5	0	-1	1	2	-2	-2
#6	0	0	2	1	0	0

Signature:

Procal Customer Support Engineer

Kittiwake Procal Ltd, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK TEL:+44(0)1733 23495 FAX:+44(0)1733 235255  
E-mail: asampson@procal.com Website: www.procal.com

7-7862-04 Linearisation Certificate

Change note: 7009251

Date: 30/07/12



## CALIBRATION CERTIFICATE

Site	Aberthaw	
Date	01/10/2016	
Instrument	Unit 8 Back Up	
Instrument Ref	8800624	8500726
Engineer	Matthew Gosling	

### Results:

Component	Units	Range FSD	Gas conc	Cert ref	Results		% error FSD
					pre cal	post cal	
H2O	%	10	p9000				
CO2	mAU	400	14.9				
CO	PPM	500	336	site gas	327	327	-1.8
NO	PPM	1000	659		671	671	1.2
SO2-H	PPM	1000	674	site gas	607	679	0.5
SO2-L	PPM	250	207.9	6192686	191	209	0.4

### Cross sensitivity tests:

Test/	Response			% Error FSD		
	CO	NO	SO2-H	CO	NO	SO2-H
CO2 15%	0	14	-12	0.0	1.4	-1.2
H2O 3%	-5	6	5	-1.0	0.6	0.5
CO2 15% + H2O 3%	4	-3	7	0.8	-0.3	0.7

Test/	Response			% Error FSD		
	SO2-L			SO2-L		
CO2 15%	6			2.4		
H2O 3%	3			1.2		
CO2 15% + H2O 3%	7			2.8		

Signature:

M.Gosling  
Procal Customer Support Engineer

Parker Procal, 5 Maxwell Road, Woodston, Peterborough, PE2 7HU, UK  
Tel: +44 (0) 17336 232495 Email: procallsales@parker.com Website: www.procal.com

7-7853-02 cal cert site visit general use

Change note: 7007110

Date: 17/03/09



## Functionality Test Results

### DECLARATION OF CONFORMITY

We,

Kittiwake Procal Ltd  
5 Maxwell Road  
Woodston  
Peterborough  
Cambridgeshire  
PE2 7HU  
United Kingdom

Tel: (+44) (0) 1733 232495  
Fax: (+44) (0) 1733 235255  
e-mail: [post@procal.com](mailto:post@procal.com)  
Web site: [www.procal.com](http://www.procal.com)

declare that the product: Type PULSI 200 Analyser installed at

Customer	Aberthaw Power Station
Control Unit Serial Number	<b>8800624</b>
P200 Serial Number	<b>8500726M</b>
Site Identification	<b>Unit 8 back up</b>

to which this declaration relates is in conformance with the MCERTS Performance Standard for Continuous Emission Monitoring Systems: Version 2. Revision 1 (April 2003).

#### MCERTS Product Conformance Certificates:

Licence	Certificate Number	Variation	Dated
ACU	SIRA MC 990006/	8	02/10/09
ACWn	SIRA MC 050060/	6	02/10/09

#### MCERTS Certification Body:

Sira Certification Service	
12 Acorn Industrial Park	Tel: (+44) (0) 1322 520500
Crayford Road	
Crayford	Fax: (+44) (0) 1322 520501
Dartford	
Kent	
DA1 4AL	e-mail: <a href="mailto:info@siraenvironmental.com">info@siraenvironmental.com</a>
United Kingdom	Web site: <a href="http://www.siracertification.com">www.siracertification.com</a>

#### Quality System:

ISO 9001:2008	Certificate Nr: 062043	Dated: 11/06/09
EN 15267-2	Dated: 2009	
EN 15267-3	Dated: 2007	Technical Standards Used:

All tests carried out to Kittiwake Procal's working procedures and those within Annex A of EN14181, Version 2.3 June 2010, relating to functionality testing by CEMs' manufacturers.

Signed:..... Date:...11/09/2016.....



## 1.0 Alignment and Cleanliness.

### 1.1 Internal Check of Analyser

Analyser under test	Status	Notes
Optical path clear	Yes	
Optics clean	Yes	
Internal wiring ok	Yes	
Air supply on	Yes	
Purge flow >300 ltrs/hr	Yes	
Pressure differential <80mBar	Yes	24mbar

## 2.0 Zero and Span Checks.

### 2.1 Zero Check

Component:	H2O	CO2	CO	NO	SO2-H	SO2-L	Ta	Ts	Ps
Units:	%	mAU	ppm	ppm	ppm	ppm			
Process conc	4.4	71	39	7.1	120	127	25.2	130.5	1009
Zero conc	0	-1	1	0	0	0	25.3	130.1	1033
% error	<2	<2	<2	<2	<2	<2			
Pass/Fail	Pass	Pass	Pass	Pass	Pass	Pass			
T90 recovery									

### 2.2 Span Check

Component:	H2O	CO2	CO	NO	SO2-H	SO2-L	Ta	Ts	Ps
Units:	%	mAU	ppm	ppm	ppm	ppm			
Full range	10	-	500	1000	1000	250			
Zero Conc	0	-1	1	0	0	0	25.3	130.1	1033
Test Gas Conc			336	659	674	207.9			
Cert number				Site Gas	Site Gas	936352001			
Response Conc			327	671	672	209			
% Error			<2	<2	<2	<2			
Pass/Fail			Pass	Pass	Pass	Pass			
T90 time				82	60	106			

### 2.3 Filter Level Checks

All on zero air	F6	F2	F7	F5G	F3	F1G	F4	F8	
Factory	3254	4455	3769	2280	3242	1962	556	3716	
Last visit	2007	2717	3007	1767	2444	1773	395	2144	
This visit	2203	2964	2534	1544	2205	1535	417	2866	



## 2.4 Interference Checks

### 3.0 Data Acquisition.

Displayed \ Component						
Zero	Analyser					
	DCS					
	4-20mA					
Span	Analyser					
	DCS					
	4-20mA					

#### 4.0 Linear coefficient record.

Coefficient \ Component	H2O	CO2	CO	NO	SO2-H	SO2-L
Last test	Linear	1.860E-02	1	9.603E0	1.285E1	5.947E0
	Squared	4.760E-05	0	9.509E-2	3.673E-2	-2.800E-3
	Cubed					
This test	Linear	1.860E-02	1	6.858E+00	1.064E+01	4.714E+00
	Squared	4.760E-05	0	9.144E-02	4.890E-02	4.000E-04
	Cubed					

## 5.0 Test Information

5.1 Procal Engineer: M Gosling

## 5.2 Test House:

### 5.3 Test House Engineer:

## Appendix 1. Pro-forma for assessing and reporting the results of the functional tests



Requirement	Notes			
<b>1. Alignment and Cleanliness</b>				
A visual inspection, with reference to the CEMs manuals, shall be carried out on the following when applicable:				
• Internal check of the CEM	✓	All functions ok		
• Cleanliness of the optical components	✓			
• Flushing air supply	✓	Ok		
• Obstructions in the optical path	✓	None		
• After re-assembly at the measurement location at least the following shall be checked	✓	No re-alignment required		
• Alignment of the measuring system	✓	No alignment required		
• Contamination control (internal check of optical surfaces)	✓			
<b>2. Sampling Systems</b>				
A visual inspection of the sampling system shall be performed, noting the condition of the following components, when fitted:				
• Sampling probe	n/a			
• Gas conditioning systems	n/a			
• Pumps	n/a			
• All connections	✓	Ok		
• Sample lines	n/a			
• Power supplies	✓	Ok		
• Filters	✓	Ok		
• NOx converters - if the sampling system contains a NOx converter, then the test laboratory shall record when the last efficiency-test was performed, and the result of this test.	n/a			
• The sampling system shall be in good condition and free of any visible faults, which may decrease the quality of data	✓	Ok		
<b>3. Leak testing</b>				
• Leak testing shall be performed according to the CEMs manuals. The tests shall cover the entire sampling system.	n/a			
<b>4. Zero and Span check</b>				
• Reference zero and span materials shall be used to verify the corresponding readings of the CEM.	✓	Ok		
• In case of non-extractive CEM, zero and span checks shall be performed using a reference-path free of flue gas before and after readjustment and after re-assembly of the CEM at the measurement location.	✓	Ok		
<b>5. Linearity</b>				
• During the calibration/linearity tests the applied concentrations should be logged onto the DCS to prove the complete system i.e. concentration applied to the instrument is represented by the instrument output and identical to the value logged on the DCS.	✓	Data collected on ACU MK4		

Requirement	Notes



DCS logged values should be included in the instrument service report.	<input checked="" type="checkbox"/>	Data collected on ACU MK4
<ul style="list-style-type: none"> <li>The linearity of the CEM's response shall be checked using five different reference materials, including a zero concentration.</li> </ul>	<input checked="" type="checkbox"/>	See linearity Cert
<ul style="list-style-type: none"> <li>The reference material with zero concentration, as well as the reference materials with four different concentrations, shall have a verifiable quantity and quality.</li> </ul>	<input checked="" type="checkbox"/>	See gas certs
<ul style="list-style-type: none"> <li>In case of gaseous reference materials, these four reference materials can be obtained from different gas cylinders or can be prepared by means of a calibrated dilution system from one single gas concentration.</li> </ul>	<input checked="" type="checkbox"/>	Gas points derived from blending full gas range using a Procal P9000, a water generator/gas blender
<ul style="list-style-type: none"> <li>The reference material concentrations shall be selected such that the measured values are at approximately 20 %, 40 %, 60 % and 80 % of the range of two times the emission limit. It is necessary to know the values of the ratios of their concentrations precisely enough so that an incorrect failure of the linearity test does not occur. The dry test reference material shall be applied to the inlet of the CEM.</li> </ul>		
The individual CEMs are tested using the following concentrations applied in a randomised sequence:		
<ul style="list-style-type: none"> <li>Reference material with zero concentration;</li> </ul>	<input checked="" type="checkbox"/>	See linearity cert
Reference material concentration approximately 20 % of 2 times the emission limit;	<input checked="" type="checkbox"/>	See linearity cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 40 % of 2 times the emission limit;</li> </ul>	<input checked="" type="checkbox"/>	See linearity cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 60 % of 2 times the emission limit;</li> </ul>	<input checked="" type="checkbox"/>	See linearity cert
<ul style="list-style-type: none"> <li>Reference material concentration approximately 80 % of 2 times the emission limit;</li> </ul>	<input checked="" type="checkbox"/>	See linearity cert
<ul style="list-style-type: none"> <li>Reference material with zero concentration;</li> </ul>	<input checked="" type="checkbox"/>	See linearity cert
After each change in concentration, the first instrument reading shall ordinarily be taken after a time period equal to at least three times the response time of the CEM. At each reference material concentration, at least three readings shall be made, six readings shall be taken at zero. The time period between the start of each of the three readings shall be separated by at least four times the response time.		
The test-laboratory may apply a risk-based approach to linearity testing in order to reduce the time for the tests. For example, the readings may be taken after less than 3x the response time; however, if the CEM fails the linearity test, then the test laboratory shall repeat the linearity test and wait at least 3x the response time as stated above. Alternatively, the number of repetitions of the test may be reduced if the CEM passes the required performance criteria by a factor of at least 2 (i.e. half the allowable residual). Increasing the waiting time to 5x the response time, for example, may be a means of meeting this requirement. Where no other method is possible, the linearity can also be performed with the aid of reference materials such as grating filters or gas filters.		

Requirement	Notes
-------------	-------



The linearity shall be calculated and tested using the procedure as given in EN 14181 annex B. If the CEM does not pass this test; then the problem shall be identified and rectified.		
<b>6. Interferences</b>		
<ul style="list-style-type: none"> <li>A test shall be undertaken if the process gases to be monitored contain components that are known interferences, as identified during QAL1 and there is a failure of the QAL2 or AST which could be due to interferences.</li> </ul>	✓	Interference checks made with Water vapour, CO2 and each gas species
<b>7. Zero and Span drift (Audit)</b>		
<ul style="list-style-type: none"> <li>The test laboratory shall assess whether the operator has a QAL3 procedure in place, and whether the operator has applied this procedure. The evidence would comprise (i) a documented procedure, (ii) zero and span data, (iii) control charts.</li> </ul>	✓	Customer has procedure in place and has documented evidence
<b>8. Response Time</b>		
<ul style="list-style-type: none"> <li>The response time of the CEM shall be checked. This can be performed, if appropriate, by feeding of the reference material at the end of the sampling probe. The response time shall not exceed the performance requirement applied during the QAL1 tests.</li> </ul>	✓	Response time recorded on test sheet and this test sheet
<b>9. Service Report</b>		
As a minimum requirement the service report should include the following:		
<ul style="list-style-type: none"> <li>Document reference for work instruction for the type of work being undertaken</li> </ul>	✓	Procal Method Statement
<ul style="list-style-type: none"> <li>Instrument manufacturer</li> </ul>	✓	Procal
<ul style="list-style-type: none"> <li>Instrument type</li> </ul>	✓	Pulsi 2000
<ul style="list-style-type: none"> <li>Instrument model</li> </ul>	✓	P2000
<ul style="list-style-type: none"> <li>Instrument Serial No</li> </ul>	✓	8500726M
<ul style="list-style-type: none"> <li>Operating principle</li> </ul>	✓	GFCIR & NIR
<ul style="list-style-type: none"> <li>Operating range</li> </ul>	✓	Recorded on this report
<ul style="list-style-type: none"> <li>Certification details</li> </ul>	✓	Mcerted
<ul style="list-style-type: none"> <li>Compliance with MCERTS (including certificate)</li> </ul>	✓	SIRA MC 990006/08
<ul style="list-style-type: none"> <li>Location</li> </ul>	✓	Unit 8 Back up
<ul style="list-style-type: none"> <li>Date and time work was undertaken</li> </ul>	✓	11/09/2016
<ul style="list-style-type: none"> <li>Equipment used - type, serial no's, calibration dates</li> </ul>	✓	Procal Water Generator/Gas divider Pcal #6
<ul style="list-style-type: none"> <li>Gases used - certificate numbers, expiry dates, binary / mix</li> </ul>	✓	Copy of Gas Certs at customer request can be emailed
<ul style="list-style-type: none"> <li>NOx converter efficiency test, if applicable</li> </ul>	n/a	
<ul style="list-style-type: none"> <li>Calibration and linearity data as required by EN14181</li> </ul>	✓	See Linearity & (Calcert Data)
<ul style="list-style-type: none"> <li>Logged data for period of calibration and linearity. Note there may be gaps in the data, in such cases, the test laboratory shall state why there are gaps in the data</li> </ul>	✓	See Linearity & (Calcert Data)

# CERTIFICATE OF CALIBRATION

Unit No	U 7
Application	Stack Flue Gas Oxygen Analyser
Probe Serial No	3K220000139813
WOC No	3-16-513044-00

TESTED AGAINST		
Calibration Meter	Type	Beamex
	Serial No	25514526
	Calibration Date	05-02-16
2 % Oxygen	Certified Input Valve	1.98 %
	Cylinder No	S1358263
	Expiry Date	04/10/2017
	Theoretical Output	7.168mA
8 % Oxygen	Certified Input Valve	8.04 %
	Cylinder No	S1144548
	Expiry Date	19/06/2019
	Theoretical Output	16.864mA

## Results

		Local Display	Current	Procal ACU Display
1.98 % Oxygen (7.168 mA)	Before	1.97%	7.133mA	2%
	After			
8.04 % Oxygen (16.864 mA)	Before	7.94%	16.7mA	8%
	After			

Signed – Ian Jauncey

Date – 16/02 /16

RWE Npower  
C & I Department  
Aberthaw Power Station  
West Aberthaw  
The Leys  
Barry  
CF62 4ZW



**Table 5.2 - Functional Tests carried out by RPS**

Requirement	Compliance	Notes
<p>10 Documentation of Records</p> <p>The following documentation shall be controlled, readily accessible and up to date:</p> <ul style="list-style-type: none"> <li>- a plan of the CEMS;</li> </ul>	Yes	A plan of the AMS resides in the Electrical Control & Instrumentation section offices and at the analyser.
<ul style="list-style-type: none"> <li>- all manuals (maintenance, users, etc.);</li> </ul>	Yes	.
<ul style="list-style-type: none"> <li>- log books to document possible malfunctions and action taken;</li> <li>- service reports;</li> </ul>	Yes	All service reports are filed.
<ul style="list-style-type: none"> <li>- QAL 3 documentation including actions taken as a result of out of control situations</li> </ul>	Yes	There is a well established QAL3 system in place and analyser drift is routinely monitored using shewart charts. Evidence was provided to RPS.
Management system procedures for maintenance, calibration and training;	Yes	The Performance dept manage and review maintenance and calibration and training carried out by the EC&I team.
Training records/certificates	Yes	Evidence was provided to RPS

Requirement	Compliance	Notes
Maintenance schedules.	Yes	Maintenance scheduled by Performance.
Auditing Plans & Records – Evidence that the operator includes procedures for the management of the CEMS within the auditing cycle of the management system.	Yes	Carried out using dedicated Technical Team from Swindon Head Office. An auditor witnessed the parallel tests, which suggest a good level of organisation.

## **APPENDIX 1: SRM Calibration Data**

**Table A1.1– SRM On-Site Calibrations**

Sample Date	Equipment Name	Equipment ID Number	Span Gas Type	ID Number	Span Gas Concentration	Pre-Sampling Result*		Post-Sampling Result*	
						Zero	Span	Zero	Span
26- 27/10/16	Horiba PG 250	955	NO (V.U.)	222094	405.7 ppm	2.3 ppm	405.6 ppm	1.6 ppm	404.1 ppm
			O <sub>2</sub> (V.U.)	216463	14.63 %	0.10 %	14.55 %	0.06 %	14.52 %
	Gasmet DX4000	1610	SO <sub>2</sub> (V.U.)	237324	201 ppm	0.1 ppm	197.8 ppm	0.0 ppm	197.6 ppm
27- 28/10/16	Horiba PG 250	955	NO (17025)	83649	405.7 ppm	1.6 ppm	404.1 ppm	0.8 ppm	405.1 ppm
			O <sub>2</sub> (V.U.)	216463	14.63 %	0.06 %	14.52 %	0.05 %	14.53 %
	Gasmet DX4000	1610	SO <sub>2</sub> (17025)	237324	201 ppm	0.0 ppm	197.6 ppm	0.0 ppm	196.7 ppm
2/11/16	Horiba PG 250	955	NO (17025)	83649	405.7 ppm	0.0 ppm	407.8 ppm	1.3 ppm	404.4 ppm
			O <sub>2</sub> (V.U.)	216463	14.63 %	0.19 %	14.60 %	0.20 %	14.50 %
	Gasmet DX4000	1610	SO <sub>2</sub> (17025)	237324	201 ppm	0.0 ppm	196.9 ppm	0.0 ppm	192.9 ppm

*Notes*

- \*- Calibration values are those for the entire sample system.
- - Zero gas 99.999% N<sub>2</sub>

**APPENDIX 2 – Accreditation Schedule**

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 1709 Accredited to ISO/IEC 17025:2005	<p style="text-align: center;"><b>The Environmental Consultancy Ltd</b> trading as RPS Consultants</p> <p style="text-align: center;">Issue No: 066 Issue date: 15 April 2016</p> <p>35 New Bridge Street, London EC4V 6BW</p> <p>Contact: Mr S Hurst Tel: +44 (0) 20 7280 3200 Fax: +44 (0) 20 7283 9248 E-Mail: <a href="mailto:hursts@rpsgroup.com">hursts@rpsgroup.com</a> Website: <a href="http://www.rpsgroup.com">www.rpsgroup.com</a></p>	
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities  
Laboratory locations:

Location details	Activity	Location code	
Address 35 New Bridge Street London EC4V 6BW United Kingdom	Local contact Mr D Blyton  Tel: +44 (0)20 7280 3200 Fax: +44 (0) 20 7283 9248 Email: <a href="mailto:rpslo@rpsgroup.com">rpslo@rpsgroup.com</a>	Health and Hygiene Support Functions: Quality Management, including contract review, document control, auditing and quality control	D
Address Noble House Capital Drive Linford Wood Milton Keynes MK14 6QP	Local contact Mr M Bates  Tel: +44 (0)1235 437 100 Fax: +44 (0)1908 669899 Email: <a href="mailto:rpsmk@rpsgroup.com">rpsmk@rpsgroup.com</a>	Health and Hygiene	E
Address Suite 4C Rhodes Business Park Silburn Way Middleton Manchester M24 4NE	Local contact Mr S Pepper  Tel: +44 (0) 161 6549069 Fax: +44 (0)161 6436495 Email: <a href="mailto:rpswn@rpsgroup.com">rpswn@rpsgroup.com</a>	Health and Hygiene	F

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Premises including domestic, commercial and industrial	Health and Hygiene Environmental Sampling and Testing	London - K Milton Keynes - L Middleton - M
Customers sites requiring Stack Emissions Testing	Stack emissions Testing	Milton Keynes - L
Mobile Laboratories	Health and Hygiene	London - J

 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants Issue No: 066 Issue date: 15 April 2016
Testing performed by the Organisation at the locations specified	

**DETAIL OF ACCREDITATION**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR  ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<u>Health and Hygiene</u>  Sampling of air for fibre counting  Fibre counting  4 Stage Clearance Process  Sampling of bulk materials for asbestos identification  Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Health and Safety Executive Asbestos: The analysts' guide for sampling, analysis and clearance procedures (HSG 248)  HSG 248:February 2005 (Documented In-House Procedure)  HSG 248:February 2005 (Documented In-House Procedure)  HSG 248:February 2005 (Documented In-House Procedure)  HSG 248:February 2005 (Documented In-House Procedure)	K, L, M  E, F, J, K, L, M  K,L, M  K, L, M  F

 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants Issue No: 066 Issue date: 15 April 2016		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere	<u>Sampling with subsequent analysis by an ISO/IEC 17025 accredited laboratory</u>  Gaseous Organic Compounds - sorbent tube method  Total Particulate Matter (20 to 1000 mg/m <sup>3</sup> )	National, International and other recognised standards using documented in-house work instructions to meet the requirements of DD CEN/TS 15675:2007/ BS EN 15259:2007  USEPA Method 18 (RPSCE/1/19a)  BS ISO 9096:2003 (RPSCE/1/7/d)	L
Testing of Stack Emissions to Atmosphere	<u>Sampling and On-Line Analysis</u>  Pressure, Temperature and Velocity	BS EN 13284-1:2002 BS ISO 9096:2003 (RPSCE/1/2)	L
Testing of Stack Emissions to Atmosphere	<u>Sampling with subsequent analysis by an ISO/IEC 17025 accredited laboratory</u>  Total Particulate Matter  Hydrogen Chloride  Halides and Halogens: Hydrogen Bromide Chlorine Bromine	National, European, International and Environment Agency specified standards including MIDs and documented in-house work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007  BS EN 13284-1:2002 (RPSCE/1/7/c)  BS EN 1911:2010 (RPSCE/1/8/b)  US EPA Method 26 and 26A (RPSCE/1/8/a)	L

 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants Issue No: 066 Issue date: 15 April 2016		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling with subsequent analysis by an ISO/IEC 17025 accredited laboratory</u> (cont'd)  Sulphur dioxide Hydrogen Fluoride Mercury Metals Dioxins and furans Dioxin-like Polychlorinated Biphenyls (PCBs) Polycyclic Aromatic hydrocarbons (PAH's) Formaldehyde Formaldehyde – sorbent tube method	National, European, International and Environment Agency specified standards including MIDs and documented in-house work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007  BS EN 14791:2005 (RPSCE/1/23) BS ISO 15713:2006 (RPSCE/1/8c) BS EN 13211:2002 (RPSCE/1/9b) BS EN 14385:2004 (RPSCE/1/9c) BS EN 1948-1:2006 (RPSCE/1/10b) BS EN 1948-4:2010 (RPSCE/1/10b) BS ISO 11338-1:2003 (RPSCE/1/10c) US EPA Method 316 (RPSCE/1/22) PD CEN/TS 13649:2014 RPSCE/1/19b Rev D	L

 <b>1709</b> Accredited to ISO/IEC 17025:2005	<p style="text-align: center;"><b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK</p> <p style="text-align: center;"><b>The Environmental Consultancy Ltd</b> trading as <b>RPS Consultants</b></p> <p style="text-align: center;">Issue No: 066 Issue date: 15 April 2016</p>		
	Testing performed by the Organisation at the locations specified		
	Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling with subsequent analysis by an ISO/IEC 17025 accredited laboratory (cont'd)</u>  Speciated VOCs (carbon and other suitable tubes) (Dry stacks only): Aliphatic VOCs Aromatic VOCs Aliphatic amines Aromatic amines Cresols Phenols Acetic acid  Amines (Total aromatic and aliphatic)  Isocyanates  Isocyanates  Hydrogen cyanide  Hydrogen sulphide  Ammonia	National, European, International and Environment Agency specified standards including MIDs and documented in-house work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007  PD CEN/TS 13649:2014 (RPSCE/1/19b)  PD CEN/TS 13649:2014, NIOSH method 2010 + 2002 (RPSCE/1/19c)  USEPA Method 207-1 (documented in-house method RPSCE/1/18C)  USEPA CTM 036 (documented in-house method RPSCE/1/18D Rev A)  US EPA OTM 29 (RPSCE/1/16a)  US EPA Method 11 (RPSCE/1/17)  BS EN 14791:2005 (RPSCE/1/14b)	L  L  L  L  L  L  L

 <b>Accredited to</b> <b>ISO/IEC 17025:2005</b>	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants Issue No: 066 Issue date: 15 April 2016		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling with subsequent analysis by an ISO/IEC 17025 accredited laboratory (cont'd)</u>  PM <sub>10</sub> /PM <sub>2.5</sub> at low concentrations (0-40 mg/m <sup>3</sup> ) Chrome VI	National, European, International and Environment Agency specified standards including MIDs and documented in-house work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007  BS EN ISO 23210 (RPSCE/1/26 Rev A)  USEPA Method 0061 (RPSCE/1/9d Rev A)	L
Testing of Stack Emissions to Atmosphere	<u>Sampling and On-Site Analysis</u>  Water Vapour	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007  BS EN 14790:2005 (RPSCE/1/12b)	L

 Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants Issue No: 066 Issue date: 15 April 2016		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling and On-Line Analysis</u>  Pressure, Temperature and Velocity (Point Velocity Method)  Water Vapour*  Carbon Monoxide*  Carbon Dioxide*  Nitrogen Monoxide (NO)*  Nitrogen Dioxide (NO <sub>2</sub> )*	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007  BS EN 16911-1:2013 (RPSCE 1/2 - Differential Pressure Device (Pitot Tube) Method)  EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)  BS EN 15058:2006 (RPSCE/1/21h - NDIR analyser) EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)  ISO 12039:2001 (RPSCE/1/21e - NDIR analyser) EA TGN M22 (RPSCE/1/24 - FTIR analyser)  BS EN 14792:2005 (RPSCE/1/21f - Chemiluminescence analyser) EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)  EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)	L L L L L L

 <b>1709</b> Accredited to ISO/IEC 17025:2005	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants Issue No: 066 Issue date: 15 April 2016		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere (cont'd)	<u>Sampling and On-Line Analysis</u> (cont'd)  Oxides of Nitrogen (NOx)*  Nitrous Oxide (N <sub>2</sub> O)  Sulphur dioxide*  Oxygen*  Total Gaseous Organic Carbon* (TOC/VOC) (0 to 1000 mg/m <sup>3</sup> )	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007  BS EN 14792:2005 (RPSCE/1/21f - Chemiluminescence analyser) EA TGN M22 (RPSCE/1/24 - Validated FTIR analyser)  EA TGN M22 (RPSCE/1/24 Rev B - Validated FTIR analyser)  EA TGN M22 (RPSCE/1/24 - FTIR analyser)  BS EN 14789:2005 (RPSCE/1/21g - validated Zirconium cell analyser)  BS EN 12619:2013 (RPSCE/1/4b - FID analyser)	L  L  L  L

 <b>Accredited to</b> <b>ISO/IEC 17025:2005</b>	<b>Schedule of Accreditation</b> issued by <b>United Kingdom Accreditation Service</b> 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK		
	<b>The Environmental Consultancy Ltd</b> trading as RPS Consultants Issue No: 066 Issue date: 15 April 2016		
Testing performed by the Organisation at the locations specified			
Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used	Location Code
Stack Emissions - Continuous Emissions Monitoring Systems (CEMS)	QAL 2 and the Annual Surveillance Test (AST) for CEMS	Documented in house procedure RPSCE/1/25 to meet the requirements of BS EN 14181:2014, Environment Agency MID 14181 (TGN M20 Annex A) and other requirements of the Environment Agency (MCERTS) Performance Standard and DD CEN/TS 15675:2007/ BS EN 15259:2007	L
END			

\* - The scale range of the analyser used for this test must be that detailed on its current MCERTS certificate or a range validated by the organisation to meet MCERTS requirements.