


Schedule 6 - Notification

Part B - to be submitted as soon as practicable

Permit Number	BR9383ID	Notification Reference	EP_EX_215
Name of operator	Knauf Insulation Ltd		
Location of Facility	Chemistry Lane, Queensferry, Deeside, Flintshire, CH5 2DA		
Time and date of the detection	0900 6 th November 2018		

Any more accurate information on the matters for notification under Part A.	Reinstatement of primary Opsis analyser on a managed basis resulted in 33 minutes loss of CEMs data during a period of stable operation (missing data is 0819 to 0852 in CEMS database)
Measures taken, or intended to be taken, to prevent a recurrence of the incident	None – planned equipment change over to return primary analyser to service after repair
Measures taken, or intended to be taken, to rectify, limit or prevent any pollution of the environment which has been or may be caused by the emission	None required, emissions are highly unlikely to have been significantly different during the 30 minute period to those either side that were well within ELVs on CEMS validated data records
The dates of any unauthorised emissions from the facility in the preceding 24 months.	25/04/2017, 23/05/17, 05/07/17, 14-15/08/17, 06/09/17, 14/09/17, 16-20/09/17, 26/09/17, 1-3/10/17, 9-11/10/17, 21/10/17, 23/10/17, 31/10/17, 08/12/17, 29/12/17, 09-11/01/18, 24/01/18, 26/03/18, 30/03/18, 17/04/18, 21/04/18, 26-27/04/18, 16/05/18, 11/7/18, 5/9/18, 7/9/18, 9/9/18, 10/9/18, 16/9/18, 15/10/18, 17/10/18, 19/10/18, 22/10/18, 3/11/18, 4/11/18

Name*	Graham Jones
Post	Process and Energy Manager
Signature	
Date	25/4/2019

* authorised to sign on behalf of Knauf Insulation Ltd