

## Appendix F – Extract from CAR\_NRW0038426 Outlining current permit is missing a legacy condition for monitoring A55 Morfa Main Stack, and will be updated and submitted as part of permit condition.



### Compliance Assessment Report for: CAR\_NRW0038426

Permit number	BL7108IM	Operator name	Tata Steel UK Ltd.
Site name	Port Talbot Steelworks		
Site address	Port Talbot Steelworks, Port Talbot, SA13 2NG		
Type of assessment	Report Review		
Date of assessment	16 <sup>th</sup> July 2021		

#### A55 (Morfa Main Stack) – CEMS for measurement of visible smoke (Ringelmann)

- Results have been submitted for Q1.
- The reported results for the quarter were below the limits specified in the permit.

Note: there is an acknowledged issue with incomplete permit condition(s) applying to this monitoring requirement. Current permit is missing legacy condition 2.10.4<sup>8</sup> which confirms its calibration according to the Ringelmann scale. Also, a legacy regulatory position in which we agreed to assess compliance against a weekly mean limit has not carried forwards to the current permit. This will be corrected as part of the forthcoming variation. The application has been submitted by TATA and is now awaiting assessment. The current reporting form accurately reflects the previously agreed positions in respect of calibration (Ringelmann 1 = CEMS is calibrated to 25% opacity) and weekly mean % opacity reporting. Operator to continue using this form.

Title:	[title]
Author	[author]

